



Alcohol Testing Program

INSTRUMENT PROCESSING SHEET

Agency South Daytona PD S/N 80 001246

Date In 10/19/16 Date Out 11/3/16 Ship P/U H/D CMI EE

Intake Performed By <u>PO</u>		Quality Checks Performed By <u>PO</u>		Flow Calibration Performed By <u>PO</u>																
<input type="checkbox"/> Registration <input checked="" type="checkbox"/> Annual <input type="checkbox"/> Return from CMI <input type="checkbox"/> Return from Enforcement <input type="checkbox"/> Electronics <input type="checkbox"/> Other _____		<input checked="" type="checkbox"/> Breath Tube Screen <input checked="" type="checkbox"/> Replace O-Rings <input checked="" type="checkbox"/> Instrument Set Up Verified <input checked="" type="checkbox"/> R-Value <u>121</u> <input checked="" type="checkbox"/> Flow Verification (L/s) Flow Column # <u>ATP102</u> 32mm <u>.121</u> (.139 - .169) 36mm <u>.140</u> (.156 - .190) 53mm <u>.207</u> (.228 - .278) 103mm <u>.464</u> (.447 - .547)		<input type="checkbox"/> Flow Calibration N/A <input checked="" type="checkbox"/> Flow Calibration Complete Flow Column # <u>ATP105</u> <input checked="" type="checkbox"/> 5L/min - 17mm <input checked="" type="checkbox"/> 15L/min - 53mm <input checked="" type="checkbox"/> 30L/min - 103mm <input checked="" type="checkbox"/> R-Value <u>121</u> <input checked="" type="checkbox"/> Post Calibration Verification (L/s) Flow Column # <u>ATP102</u> 32mm <u>.144</u> (.139 - .169) 36mm <u>.160</u> (.156 - .190) 53mm <u>.230</u> (.228 - .278) 103mm <u>.451</u> (.447 - .547)																
Visual Inspection: <input checked="" type="checkbox"/> Case <input checked="" type="checkbox"/> Handle <input checked="" type="checkbox"/> Dry Gas Holder <input checked="" type="checkbox"/> Feet <input checked="" type="checkbox"/> Keyboard/Plug <input checked="" type="checkbox"/> Back/Plugs <input checked="" type="checkbox"/> Screws tight <input checked="" type="checkbox"/> Breath Hose		<input checked="" type="checkbox"/> Barometric Pressure Check Gauge ID # <u>28427</u> <input checked="" type="checkbox"/> Stability Checks		<table border="1"> <thead> <tr> <th>Simulator</th> <th>Serial #</th> <th>Lot #/Exp</th> </tr> </thead> <tbody> <tr> <td>0.05</td> <td>SD1018</td> <td>201507A 7/14/17</td> </tr> <tr> <td>0.08</td> <td>SD1011</td> <td>201601F 1/26/18</td> </tr> <tr> <td>0.20</td> <td>SD1025</td> <td>201604C 4/5/18</td> </tr> <tr> <td>0.08 DGS</td> <td>N/A</td> <td>AG619605 7/14/18</td> </tr> </tbody> </table>		Simulator	Serial #	Lot #/Exp	0.05	SD1018	201507A 7/14/17	0.08	SD1011	201601F 1/26/18	0.20	SD1025	201604C 4/5/18	0.08 DGS	N/A	AG619605 7/14/18
Simulator	Serial #	Lot #/Exp																		
0.05	SD1018	201507A 7/14/17																		
0.08	SD1011	201601F 1/26/18																		
0.20	SD1025	201604C 4/5/18																		
0.08 DGS	N/A	AG619605 7/14/18																		
Other Equipment: <input type="checkbox"/> Power cord <input type="checkbox"/> Printer Cable <input type="checkbox"/> Other _____		<table border="1"> <thead> <tr> <th>Simulator</th> <th>Serial #</th> <th>Lot #/Exp</th> </tr> </thead> <tbody> <tr> <td>0.05</td> <td>SD1018</td> <td>201507A 7/14/17</td> </tr> <tr> <td>0.08</td> <td>SD1011</td> <td>201601F 1/26/18</td> </tr> <tr> <td>0.20</td> <td>SD1025</td> <td>201604C 4/5/18</td> </tr> <tr> <td>0.08 DGS</td> <td>N/A</td> <td>AG619605 7/14/18</td> </tr> </tbody> </table>		Simulator	Serial #	Lot #/Exp	0.05	SD1018	201507A 7/14/17	0.08	SD1011	201601F 1/26/18	0.20	SD1025	201604C 4/5/18	0.08 DGS	N/A	AG619605 7/14/18	Maintenance Performed By _____ <input type="checkbox"/> Battery Replacement <input type="checkbox"/> Dry Gas Regulator Replacement <input type="checkbox"/> Breath Tube Replacement <input type="checkbox"/> Other _____	
Simulator	Serial #	Lot #/Exp																		
0.05	SD1018	201507A 7/14/17																		
0.08	SD1011	201601F 1/26/18																		
0.20	SD1025	201604C 4/5/18																		
0.08 DGS	N/A	AG619605 7/14/18																		
Notes: _____ _____ _____		Suggested Service _____ _____		<div style="text-align: right; font-size: 2em; opacity: 0.5;">RECEIVED NOV 04 2016 FDLE Alcohol Testing Program</div>																

Optical Bench Calibration Performed By <u>PO</u>		Department Inspection Performed By <u>PO</u>																																									
<input type="checkbox"/> Optical Bench Calibration N/A <input checked="" type="checkbox"/> Optical Bench Calibration Complete Barometric Pressure Gauge <u>21082</u> ID # <u>26932</u>		<input checked="" type="checkbox"/> Barometric Pressure <u>1024</u> Gauge ID# <u>28427</u> <u>1024</u> Instrument Mouth Alcohol Solution Lot # <u>2016-A</u> Acetone Stock Solution Lot # <u>2016-B</u>																																									
<table border="1"> <thead> <tr> <th>Simulator</th> <th>Serial Number</th> <th>Lot Number</th> <th>Expiration</th> </tr> </thead> <tbody> <tr> <td>0.000</td> <td>64444</td> <td>N/A</td> <td>N/A</td> </tr> <tr> <td>0.040</td> <td>SD1024</td> <td>16101</td> <td>2/2/18</td> </tr> <tr> <td>0.100</td> <td>SD1013</td> <td>16001</td> <td>5/8/18</td> </tr> <tr> <td>0.200</td> <td>SD1012</td> <td>16103</td> <td>6/14/18</td> </tr> <tr> <td>0.400</td> <td>66621</td> <td>16102</td> <td>3/22/18</td> </tr> <tr> <td>0.080 DGS</td> <td>N/A</td> <td>09415080A1</td> <td>3/5/17</td> </tr> </tbody> </table>		Simulator	Serial Number	Lot Number	Expiration	0.000	64444	N/A	N/A	0.040	SD1024	16101	2/2/18	0.100	SD1013	16001	5/8/18	0.200	SD1012	16103	6/14/18	0.400	66621	16102	3/22/18	0.080 DGS	N/A	09415080A1	3/5/17	<table border="1"> <thead> <tr> <th>Simulator</th> <th>Serial Number</th> </tr> </thead> <tbody> <tr> <td>0.00</td> <td>SD1019</td> </tr> <tr> <td>Interferent</td> <td>SD1021</td> </tr> <tr> <td>0.05</td> <td>SD1018</td> </tr> <tr> <td>0.08</td> <td>SD1011</td> </tr> <tr> <td>0.20</td> <td>SD1025</td> </tr> </tbody> </table>		Simulator	Serial Number	0.00	SD1019	Interferent	SD1021	0.05	SD1018	0.08	SD1011	0.20	SD1025
Simulator	Serial Number	Lot Number	Expiration																																								
0.000	64444	N/A	N/A																																								
0.040	SD1024	16101	2/2/18																																								
0.100	SD1013	16001	5/8/18																																								
0.200	SD1012	16103	6/14/18																																								
0.400	66621	16102	3/22/18																																								
0.080 DGS	N/A	09415080A1	3/5/17																																								
Simulator	Serial Number																																										
0.00	SD1019																																										
Interferent	SD1021																																										
0.05	SD1018																																										
0.08	SD1011																																										
0.20	SD1025																																										
<input checked="" type="checkbox"/> Post Calibration Stability Checks		Attachments																																									
<table border="1"> <thead> <tr> <th>Simulator</th> <th>Serial Number</th> <th>Lot Number</th> <th>Expiration</th> </tr> </thead> <tbody> <tr> <td>0.05</td> <td>SD1018</td> <td>201507A</td> <td>7/14/17</td> </tr> <tr> <td>0.08</td> <td>SD1011</td> <td>201601F</td> <td>1/26/18</td> </tr> <tr> <td>0.20</td> <td>SD1025</td> <td>201604C</td> <td>4/5/18</td> </tr> <tr> <td>0.08 DGS</td> <td>N/A</td> <td>AG619605</td> <td>7/14/18</td> </tr> </tbody> </table>		Simulator	Serial Number	Lot Number	Expiration	0.05	SD1018	201507A	7/14/17	0.08	SD1011	201601F	1/26/18	0.20	SD1025	201604C	4/5/18	0.08 DGS	N/A	AG619605	7/14/18	<input checked="" type="checkbox"/> Form 41 <input checked="" type="checkbox"/> Pre-Stability Tests <input checked="" type="checkbox"/> Flow Calibration <input checked="" type="checkbox"/> Optical Bench Cal <input checked="" type="checkbox"/> Post-Stability Tests <input type="checkbox"/> Other _____																					
Simulator	Serial Number	Lot Number	Expiration																																								
0.05	SD1018	201507A	7/14/17																																								
0.08	SD1011	201601F	1/26/18																																								
0.20	SD1025	201604C	4/5/18																																								
0.08 DGS	N/A	AG619605	7/14/18																																								

Notes: QC: SP

Instrument Complies with Chapter 11D-8, FAC
 Instrument Does Not Comply with Chapter 11D-8, FAC
 Return to/Place into Evidentiary Use
 Remain Out of Evidentiary Use
 Conduct an Agency Inspection Before Evidentiary Use

T. Smith Kirkland
Quality Control Review

11/4/16
Date

PRE - CALIBRATION

SOUTH DAYTONA P.D.
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-001246
10/28/2016
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	09:12
Control Test	0.052	09:12
Air Blank	0.000	09:13
Control Test	0.052	09:14
Air Blank	0.000	09:14
Control Test	0.052	09:15
Air Blank	0.000	09:15
Control Test Stats		
Average	0.0520	
Std Dev	0.0000	
Rel Std Dev(%)	0.0000	

SOUTH DAYTONA P.D.
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-001246
10/28/2016
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	09:17
Control Test	0.202	09:18
Air Blank	0.000	09:18
Control Test	0.201	09:19
Air Blank	0.000	09:20
Control Test	0.201	09:20
Air Blank	0.000	09:21
Control Test Stats		
Average	0.2013	
Std Dev	0.0006	
Rel Std Dev(%)	0.2868	

BK

P Murphy

Operator's Signature

P Murphy

Operator's Signature

SOUTH DAYTONA P.D.
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-001246
10/28/2016
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	09:23
Control Test	0.083	09:23
Air Blank	0.000	09:24
Control Test	0.082	09:25
Air Blank	0.000	09:25
Control Test	0.082	09:26
Air Blank	0.000	09:26
Control Test Stats		
Average	0.0823	
Std Dev	0.0006	
Rel Std Dev(%)	0.7012	

SOUTH DAYTONA P.D.
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-001246
10/28/2016
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	09:29
Control Test	0.078	09:30
Air Blank	0.000	09:30
Control Test	0.078	09:31
Air Blank	0.000	09:31
Control Test	0.078	09:31
Air Blank	0.000	09:32
Control Test Stats		
Average	0.0780	
Std Dev	0.0000	
Rel Std Dev(%)	0.0000	

SP

OGS

P Murphy

Operator's Signature

P Murphy

Operator's Signature

SOUTH DAYTONA P.D.
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-001246
11/03/2016 09:59:23

Auto Calibration

Max Power Res Value = 3i
Auto Range Res Value = 15
Sol Value = 0.000 g/210L ***
Fit value = 0.0000 mg/l %%%
Samples Taken = 4, Discarded = 1
Sum Io = 12654, Sum Io = 14072

<<<< CHANNEL 1 >>>>
Sample % Abs (% Abs Ref)
Sample #1 = 0.1320 (-0.0280)
Sample #2 = 0.1230 (0.0190)
Sample #3 = 0.1430 (0.0540)
Sample #4 = 0.1420 (0.0630)
Aug % Abs = 0.1360 (0.0453)
STD DEV = 0.0113 (0.0232)
REL STD DEV = 8.286 (51.276)

<<<< CHANNEL 2 >>>>
Sample % Abs (% Abs Ref)
Sample #1 = 0.1630 (-0.0140)
Sample #2 = 0.1020 (0.0270)
Sample #3 = 0.1360 (0.0340)
Sample #4 = 0.1300 (0.0350)
Aug % Abs = 0.1227 (0.0320)
STD DEV = 0.0181 (0.0044)
REL STD DEV = 14.794 (13.622)

Sol Value = 0.040 g/210L ***
Fit value = 0.1905 mg/l %%%
Samples Taken = 4, Discarded = 1
Sum Io = 12640, Sum Io = 14062

<<<< CHANNEL 1 >>>>
Sample % Abs (% Abs Ref)
Sample #1 = 0.8380 (-0.0110)
Sample #2 = 0.8340 (0.0170)
Sample #3 = 0.8210 (0.0270)
Sample #4 = 0.8230 (0.0390)
Aug % Abs = 0.8260 (0.0277)
STD DEV = 0.0070 (0.0110)
REL STD DEV = 0.847 (39.814)

<<<< CHANNEL 2 >>>>
Sample % Abs (% Abs Ref)
Sample #1 = 1.5660 (-0.0300)
Sample #2 = 1.5470 (0.0000)
Sample #3 = 1.5160 (0.0070)
Sample #4 = 1.5410 (0.0250)
Aug % Abs = 1.5347 (0.0107)
STD DEV = 0.0164 (0.0129)
REL STD DEV = 1.071 (120.910)

Sol Value = 0.100 g/210L ***
Fit value = 0.4762 mg/l %%%
Samples Taken = 4, Discarded = 1
Sum Io = 12633, Sum Io = 14059

<<<< CHANNEL 1 >>>>
Sample % Abs (% Abs Ref)
Sample #1 = 1.9010 (-0.0250)
Sample #2 = 1.8960 (-0.0200)
Sample #3 = 1.8860 (-0.0100)
Sample #4 = 1.8670 (0.0060)
Aug % Abs = 1.8830 (-0.0080)
STD DEV = 0.0147 (0.0131)
REL STD DEV = 0.782 (163.936)

<<<< CHANNEL 2 >>>>
Sample % Abs (% Abs Ref)
Sample #1 = 3.6720 (-0.0090)
Sample #2 = 3.6220 (0.0340)
Sample #3 = 3.6200 (0.0350)
Sample #4 = 3.6390 (0.0280)
Aug % Abs = 3.6270 (0.0323)
STD DEV = 0.0104 (0.0038)
REL STD DEV = 0.288 (11.709)

Sol Value = 0.200 g/210L ***
Fit value = 0.9524 mg/l %%%
Samples Taken = 4, Discarded = 1
Sum Io = 12635, Sum Io = 14061

<<<< CHANNEL 1 >>>>
Sample % Abs (% Abs Ref)
Sample #1 = 3.6430 (-0.0190)
Sample #2 = 3.6160 (0.0160)
Sample #3 = 3.6050 (0.0430)
Sample #4 = 3.6020 (0.0380)
Aug % Abs = 3.6077 (0.0323)
STD DEV = 0.0074 (0.0144)
REL STD DEV = 0.204 (44.426)

<<<< CHANNEL 2 >>>>
Sample % Abs (% Abs Ref)
Sample #1 = 7.0700 (0.0060)
Sample #2 = 7.0230 (0.0740)
Sample #3 = 7.0070 (0.1060)
Sample #4 = 6.9850 (0.1190)
Aug % Abs = 7.0050 (0.0997)
STD DEV = 0.0191 (0.0232)
REL STD DEV = 0.272 (23.236)

Sol Value = 0.400 g/210L ***
Fit value = 1.9048 mg/l %%%
Samples Taken = 4, Discarded = 1
Sum Io = 12633, Sum Io = 14053

<<<< CHANNEL 1 >>>>
Sample % Abs (% Abs Ref)
Sample #1 = 6.8050 (-0.0200)
Sample #2 = 6.7760 (0.0330)
Sample #3 = 6.8060 (0.0380)
Sample #4 = 6.7750 (0.0480)
Aug % Abs = 6.7857 (0.0397)
STD DEV = 0.0176 (0.0076)
REL STD DEV = 0.260 (19.255)

<<<< CHANNEL 2 >>>>
Sample % Abs (% Abs Ref)
Sample #1 = 13.0440 (-0.0010)
Sample #2 = 12.9530 (0.1060)
Sample #3 = 12.9730 (0.1130)
Sample #4 = 12.9840 (0.1160)
Aug % Abs = 12.9700 (0.1117)
STD DEV = 0.0157 (0.0051)
REL STD DEV = 0.121 (4.595)

***** AUTO CAL DATA *****
<<<< CHANNEL 1 >>>>
Sol Val = 0.0000 mg/l or 0.000 g/210L
% Abs = 0.136
Std Dev = 0.01 Rel Std Dev = 8.29
Sol Val = 0.1905 mg/l or 0.040 g/210L
% Abs = 0.826
Std Dev = 0.01 Rel Std Dev = 0.85
Sol Val = 0.4762 mg/l or 0.100 g/210L
% Abs = 1.883
Std Dev = 0.01 Rel Std Dev = 0.78
Sol Val = 0.9524 mg/l or 0.200 g/210L
% Abs = 3.608
Std Dev = 0.01 Rel Std Dev = 0.20
Sol Val = 1.9048 mg/l or 0.400 g/210L
% Abs = 6.786
Std Dev = 0.02 Rel Std Dev = 0.26
Zero Order Coef = -314.57
First Order Coef = 2609.98
Second Order Coef = 35.68
Standard Deviation = 39.704227

BK

SP

```

<<<< CHANNEL 2 >>>>
Sol Val = 0.0000 mg/l or 0.000 g/210L
% Abs = 0.123
Std Dev = 0.02 Rel Std Dev = 14.79
Sol Val = 0.1905 mg/l or 0.040 g/210L
% Abs = 1.535
Std Dev = 0.02 Rel Std Dev = 1.07
Sol Val = 0.4762 mg/l or 0.100 g/210L
% Abs = 3.627
Std Dev = 0.01 Rel Std Dev = 0.29
Sol Val = 0.9524 mg/l or 0.200 g/210L
% Abs = 7.005
Std Dev = 0.02 Rel Std Dev = 0.27
Sol Val = 1.9048 mg/l or 0.400 g/210L
% Abs = 12.970
Std Dev = 0.02 Rel Std Dev = 0.12
Zero Order Coef = -117.50
First Order Coef = 1274.64
Second Order Coef = 15.58
Standard Deviation = 44.396057

```

```

-----
: Solution Stats Quadratic Fit Chan 1 :
: Act      Fit      Residual :
: g/210L   g/210L   g/210L   :
: 0.000    0.001    -0.0009   :
: 0.040    0.039    0.0008    :
: 0.100    0.099    0.0007    :
: 0.200    0.201    -0.0009   :
: 0.400    0.400    0.0002    :
-----

```

```

-----
: Solution Stats Quadratic Fit Chan 2 :
: Act      Fit      Residual :
: g/210L   g/210L   g/210L   :
: 0.000    0.001    -0.0008   :
: 0.040    0.039    0.0006    :
: 0.100    0.099    0.0011    :
: 0.200    0.201    -0.0011   :
: 0.400    0.400    0.0003    :
-----

```

```

SOUTH DAYTONA P.D.
Intoxilyzer - Alcohol Analyzer
Model 8000          SN 80-001246
10/28/2016
Software: 8100.27

```

```

Flow Rate Calibration*****
1: Rate (Liters/min) = 5
  SQRT(Diff) ) = 6.082
2: Rate (Liters/min) = 15
  SQRT(Diff) ) = 11.355
3: Rate (Liters/min) = 30
  SQRT(Diff) ) = 21.883
Dependent Data Scale Factor = 100000 L/min
Independent Data Scale Factor = 256
Rounded Slope = 609
Rounded Intercept = -377712
Correlation = 0.99721

```

```

Sol Value = 0.080 g/210L ***
Fit value = 0.3810 mg/l %%%
Samples Taken = 4, Discarded = 1
**** CHANNEL 1
Sample #1 = 2805.00
Sample #2 = 2733.00
Sample #3 = 2737.00
Sample #4 = 2829.00
Average Result = 2766.3333
STD DEV = 54.3078
REL STD DEV = 1.963
*****
**** CHANNEL 2
Sample #1 = 3299.00
Sample #2 = 3247.00
Sample #3 = 3334.00
Sample #4 = 3339.00
Average Result = 3306.6667
STD DEV = 51.7333
REL STD DEV = 1.565
*****

```

```

Dry Gas H2O Adjust Results *****
Barometric Pressure = 1024
3 um H2O Adjust (mg/l*10,000) = 1043
9 um H2O Adjust (mg/l*10,000) = 503
**** AUTO CAL PASS

```

```

SOUTH DAYTONA P.D.
Intoxilyzer - Alcohol Analyzer
Model 8000          SN 80-001246
10/28/2016
Software: 8100.27

```

BLK

```

Flow Rate Calibration*****
1: Rate (Liters/min) = 5
  SQRT(Diff) ) = 6.164
2: Rate (Liters/min) = 15
  SQRT(Diff) ) = 12.000
3: Rate (Liters/min) = 30
  SQRT(Diff) ) = 22.203
Dependent Data Scale Factor = 100000 L/min
Independent Data Scale Factor = 256
Rounded Slope = 605
Rounded Intercept = -417242
Correlation = 0.99916

```

SP

POST - CALIBRATION

SOUTH DAYTONA P.D.
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-001246
11/03/2016
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	10:40
Control Test	0.050	10:41
Air Blank	0.000	10:41
Control Test	0.050	10:42
Air Blank	0.000	10:42
Control Test	0.050	10:43
Air Blank	0.000	10:44
Control Test Stats		
Average	0.0500	
Std Dev	0.0000	
Rel Std Dev(%)	0.0000	

SOUTH DAYTONA P.D.
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-001246
11/03/2016
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	10:50
Control Test	0.197	10:51
Air Blank	0.000	10:52
Control Test	0.198	10:52
Air Blank	0.000	10:53
Control Test	0.197	10:54
Air Blank	0.000	10:54
Control Test Stats		
Average	0.1973	
Std Dev	0.0006	
Rel Std Dev(%)	0.2926	

P Murphy

Operator's Signature

P Murphy

Operator's Signature

SOUTH DAYTONA P.D.
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-001246
11/03/2016
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	10:45
Control Test	0.080	10:46
Air Blank	0.000	10:47
Control Test	0.080	10:47
Air Blank	0.000	10:48
Control Test	0.080	10:49
Air Blank	0.000	10:49
Control Test Stats		
Average	0.0800	
Std Dev	0.0000	
Rel Std Dev(%)	0.0000	

SOUTH DAYTONA P.D.
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-001246
11/03/2016
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	10:56
Control Test	0.081	10:57
Air Blank	0.000	10:57
Control Test	0.081	10:57
Air Blank	0.000	10:58
Control Test	0.082	10:58
Air Blank	0.000	10:59
Control Test Stats		
Average	0.0813	
Std Dev	0.0006	
Rel Std Dev(%)	0.7099	

ASK

DGS

P Murphy

Operator's Signature

P Murphy

Operator's Signature

SP