

## INSTRUMENT PROCESSING SHEET

Agency Broward County Sheriff's Office S/N 80-001247

Date In 6/23/2016 Date Out 6/23/2016  Ship  P/U  H/D  CMI  EE

| <b>Intake</b> Performed By <u>DELL</u><br><input type="checkbox"/> Registration<br><input checked="" type="checkbox"/> Annual<br><input checked="" type="checkbox"/> Return from CMI<br><input type="checkbox"/> Return from Enforcement<br>Electronics<br><input type="checkbox"/> Other _____<br><b>Visual Inspection:</b><br><u>OK</u> Case <u>OK</u> Handle<br><u>OK</u> Dry Gas Holder <u>OK</u> Feet<br><u>OK</u> Keyboard/Plug <u>OK</u> Back/Plugs<br><u>OK</u> Screws tight <u>OK</u> Breath Hose<br><b>Other Equipment:</b><br><input type="checkbox"/> Power cord<br><input type="checkbox"/> Printer Cable<br><input type="checkbox"/> Other _____<br>Notes: _____<br>_____<br>_____ | <b>Quality Checks</b> Performed By <u>DELL</u><br><input checked="" type="checkbox"/> Breath Tube Screen<br><input checked="" type="checkbox"/> Replace O-Rings<br><input checked="" type="checkbox"/> Instrument Set Up Verified<br><input checked="" type="checkbox"/> R-Value <u>190</u><br><input checked="" type="checkbox"/> Flow Verification (L/s)<br>Flow Column # <u>ATP101</u><br>32mm <u>167</u> (.139 - .169)<br>36mm <u>187</u> (.156 - .190)<br>53mm <u>261</u> (.228 - .278)<br>103mm <u>520</u> (.447 - .547)<br><input checked="" type="checkbox"/> Barometric Pressure Check<br>Gauge ID # <u>28163</u><br><input checked="" type="checkbox"/> Stability Checks<br><table border="1" style="width:100%; border-collapse: collapse;"> <thead> <tr> <th>Simulator</th> <th>Serial #</th> <th>Lot #/Exp</th> </tr> </thead> <tbody> <tr> <td>0.05</td> <td><u>SD3967</u></td> <td><u>201507A</u><br/><u>07/14/2017</u></td> </tr> <tr> <td>0.08</td> <td><u>SD3968</u></td> <td><u>201601F</u><br/><u>01/26/2018</u></td> </tr> <tr> <td>0.20</td> <td><u>SD3969</u></td> <td><u>201505A</u><br/><u>05/12/2017</u></td> </tr> <tr> <td>0.08 DGS</td> <td>N/A</td> <td><u>A6600504</u><br/><u>01/05/2018</u></td> </tr> </tbody> </table> | Simulator                            | Serial # | Lot #/Exp | 0.05 | <u>SD3967</u> | <u>201507A</u><br><u>07/14/2017</u> | 0.08 | <u>SD3968</u> | <u>201601F</u><br><u>01/26/2018</u> | 0.20 | <u>SD3969</u> | <u>201505A</u><br><u>05/12/2017</u> | 0.08 DGS | N/A | <u>A6600504</u><br><u>01/05/2018</u> | <b>Flow Calibration</b> Performed By _____<br><input checked="" type="checkbox"/> Flow Calibration N/A<br><input type="checkbox"/> Flow Calibration Complete<br>Flow Column # _____<br><input type="checkbox"/> 5L/min - 17mm<br><input type="checkbox"/> 15L/min - 53mm<br><input type="checkbox"/> 30L/min<br><input type="checkbox"/> R-Value _____<br><input type="checkbox"/> Post Calibration Verification (L/s)<br>Flow Column # _____<br>32mm _____ (.139 - .169)<br>36mm _____ (.156 - .190)<br>53mm _____ (.228 - .278)<br>103mm _____ (.447 - .547)<br><b>Maintenance</b> Performed By _____<br><input type="checkbox"/> Battery Replacement<br><input type="checkbox"/> Dry Gas Regulator Replacement<br><input type="checkbox"/> Breath Tube Replacement<br><input type="checkbox"/> Other _____<br><b>Suggested Service</b><br>_____<br>_____ |
|--|--|--------------------------------------|----------|-----------|------|---------------|-------------------------------------|------|---------------|-------------------------------------|------|---------------|-------------------------------------|----------|-----|--------------------------------------|---|
| Simulator  | Serial #   | Lot #/Exp                            |          |           |      |               |                                     |      |               |                                     |      |               |                                     |          |     |                                      |   |
| 0.05   | <u>SD3967</u>  | <u>201507A</u><br><u>07/14/2017</u>  |          |           |      |               |                                     |      |               |                                     |      |               |                                     |          |     |                                      |   |
| 0.08   | <u>SD3968</u>  | <u>201601F</u><br><u>01/26/2018</u>  |          |           |      |               |                                     |      |               |                                     |      |               |                                     |          |     |                                      |   |
| 0.20   | <u>SD3969</u>  | <u>201505A</u><br><u>05/12/2017</u>  |          |           |      |               |                                     |      |               |                                     |      |               |                                     |          |     |                                      |   |
| 0.08 DGS   | N/A  | <u>A6600504</u><br><u>01/05/2018</u> |          |           |      |               |                                     |      |               |                                     |      |               |                                     |          |     |                                      |   |

**RECEIVED**  
 JUN 30 2016  
 FDLE  
 Alcohol Testing Program

|   |               |            |            |
|---|---------------|------------|------------|
| <b>Optical Bench Calibration</b> Performed By _____               |               |            |            |
| <input checked="" type="checkbox"/> Optical Bench Calibration N/A |               |            |            |
| <input type="checkbox"/> Optical Bench Calibration Complete       |               |            |            |
| Barometric Pressure Gauge   |               | ID #       |            |
| Simulator   | Serial Number | Lot Number | Expiration |
| 0.000   |               | N/A        | N/A        |
| 0.040   |               |            |            |
| 0.100   |               |            |            |
| 0.200   |               |            |            |
| 0.400   |               |            |            |
| 0.080 DGS   | N/A           |            |            |
| <b>Post Calibration Stability Checks</b>                          |               |            |            |
| Simulator   | Serial Number | Lot Number | Expiration |
| 0.05  |               |            |            |
| 0.08  |               |            |            |
| 0.20  |               |            |            |
| 0.08 DGS  | N/A           |            |            |

|   |   |
|---|---|
| <b>Department Inspection</b> Performed By <u>DELL</u>   |   |
| <input checked="" type="checkbox"/> Barometric Pressure | <u>1019</u> Gauge                             |
| ID# <u>28199</u>  | <u>1019</u> Instrument                        |
| Mouth Alcohol Solution Lot #                            | <u>2016-A</u>                                 |
| Acetone Stock Solution Lot #                            | <u>2016-A</u>                                 |
| Simulator   | Serial Number                                 |
| 0.00  | <u>SD 3965</u>                                |
| Interferent   | <u>SD 3966</u>                                |
| 0.05  | <u>SD 3967</u>                                |
| 0.08  | <u>SD 3968</u>                                |
| 0.20  | <u>SD 3969</u>                                |
| <b>Attachments</b>                                      |   |
| <input checked="" type="checkbox"/> Form 41             | <input type="checkbox"/> Optical Bench Cal    |
| <input checked="" type="checkbox"/> Pre-Stability Tests | <input type="checkbox"/> Post-Stability Tests |
| <input type="checkbox"/> Flow Calibration               | <input type="checkbox"/> Other _____          |

Notes: **E-MAILED**  **APPROVED**  
6/23/2016  
QA/QC OK [Signature]

|   |
|---|
| <input checked="" type="checkbox"/> Instrument Complies with Chapter 11D-8, FAC         |
| <input type="checkbox"/> Instrument Does Not Comply with Chapter 11D-8, FAC             |
| <input checked="" type="checkbox"/> Return to/Place into Evidentiary Use                |
| <input type="checkbox"/> Remain Out of Evidentiary Use                                  |
| <input checked="" type="checkbox"/> Conduct an Agency Inspection Before Evidentiary Use |

Brett Highland  
 Quality Control Review

6/30/16  
 Date

| TYPE OF TEST | SERIAL NUMBER | AGENCY                          | DATE       | PERFORMED BY |
|--------------|---------------|---------------------------------|------------|--------------|
| Stabilities  | 80-001247     | Broward County Sheriff's Office | 06/23/2016 | DeLL         |

AK

| 0.05g/210L   | 0.08g/210L   | 0.20g/210L   | DGS 0.08g/210L                                     |
|--|--|--|--|
| 0.047 to 0.053 <input checked="" type="checkbox"/> | 0.077 to 0.083 <input checked="" type="checkbox"/> | 0.194 to 0.206 <input checked="" type="checkbox"/> | 0.077 to 0.083 <input checked="" type="checkbox"/> |

BROWARD COUNTY SO  
Intoxilizer - Alcotest Analyzer  
Model 8000 SN 80-001247  
06/23/2016  
Software: 8100.27

Test 9/21/0L Time

Air Blank 0.000 08:46  
Control Test 0.051 08:47  
Air Blank 0.030 08:48  
Control Test 0.051 08:48  
Air Blank 0.000 08:49  
Control Test 0.050 08:50  
Air Blank 0.000 08:50  
Control Test Status  
Average 0.0507  
Std Dev 0.0035  
Rel Std Dev(%) 1.1395

DELL  
Operator's Signature

BROWARD COUNTY SO  
Intoxilizer - Alcotest Analyzer  
Model 8000 SN 80-001247  
06/23/2016  
Software: 8100.27

Test 9/21/0L Time

Air Blank 0.000 09:52  
Control Test 0.081 09:53  
Air Blank 0.000 09:54  
Control Test 0.082 09:54  
Air Blank 0.000 09:55  
Control Test 0.083 09:55  
Air Blank 0.000 09:56  
Control Test Status  
Average 0.0883  
Std Dev 0.0035  
Rel Std Dev(%) 0.2187

DELL  
Operator's Signature

BROWARD COUNTY SO  
Intoxilizer - Alcotest Analyzer  
Model 8000 SN 80-001247  
06/23/2016  
Software: 8100.27

Test 9/21/0L Time

Air Blank 0.000 09:58  
Control Test 0.202 09:59  
Air Blank 0.000 09:59  
Control Test 0.202 09:59  
Air Blank 0.000 09:00  
Control Test 0.201 09:00  
Air Blank 0.000 09:01  
Control Test Status  
Average 0.2017  
Std Dev 0.0036  
Rel Std Dev(%) 0.2863

DELL  
Operator's Signature

BROWARD COUNTY SO  
Intoxilizer - Alcotest Analyzer  
Model 8000 SN 80-001247  
06/23/2016  
Software: 8100.27

Test 9/21/0L Time

Air Blank 0.000 09:04  
Control Test 0.083 09:04  
Air Blank 0.001 09:05  
Control Test 0.083 09:05  
Air Blank 0.000 09:05  
Control Test 0.083 09:06  
Air Blank 0.000 09:06  
Control Test Status  
Average 0.0833  
Std Dev 0.0033  
Rel Std Dev(%) 0.0000

DELL  
Operator's Signature

OSR



Alcohol Testing Program

INSTRUMENT PROCESSING SHEET

Agency Broward County Sheriff's Office S/N 80-001247  
Date In 4/19/2016 Date Out 4/20/2016  Ship  P/U  H/D  CMI  EE

**Intake** Performed By DEAR

Registration  
 Annual  
 Return from CMI  
 Return from Enforcement Electronics  
 Other \_\_\_\_\_

Visual Inspection:  
OK Case OK Handle  
OK Dry Gas Holder OK Feet  
OK Keyboard/Plug OK Back/Plugs  
OK Screws tight OK Breath Hose

Other Equipment:  
 Power cord  
 Printer Cable  
 Other ADD STATIC BAG

Notes: \_\_\_\_\_

**Quality Checks** Performed By JELL

Breath Tube Screen  
 Replace O-Rings  
 Instrument Set Up Verified  
 R-Value 189  
 Flow Verification (L/s)  
 Flow Column # ATP 101  
 32mm 160 (.139 - .169)  
 36mm 179 (.156 - .190)  
 53mm 253 (.228 - .278)  
 103mm 519 (.447 - .547)

Barometric Pressure Check  
 Gauge ID # 68639

Stability Checks

| Simulator | Serial #       | Lot #/Exp                            |
|-----------|----------------|--------------------------------------|
| 0.05      | <u>SD 3967</u> | <u>201507A</u><br><u>07/14/2017</u>  |
| 0.08      | <u>SD 3968</u> | <u>201502G</u><br><u>02/24/2017</u>  |
| 0.20      | <u>SD 3969</u> | <u>201505A</u><br><u>05/12/2017</u>  |
| 0.08 DGS  | <u>N/A</u>     | <u>AG519701</u><br><u>07/16/2017</u> |

**Flow Calibration** Performed By \_\_\_\_\_

Flow Calibration N/A  
 Flow Calibration Complete  
 Flow Column # \_\_\_\_\_  
 5L/min - 17mm  
 15L/min - 54mm  
 30L/min - 103mm

R-Value \_\_\_\_\_  
 Post Calibration Verification (L/s)  
 Flow Column # \_\_\_\_\_  
 32mm \_\_\_\_\_ (.139 - .169)  
 36mm \_\_\_\_\_ (.156 - .190)  
 53mm \_\_\_\_\_ (.228 - .278)  
 103mm \_\_\_\_\_ (.447 - .547)

**Maintenance** Performed By \_\_\_\_\_

Battery Replacement  
 Dry Gas Regulator Replacement  
 Breath Tube Replacement  
 Other \_\_\_\_\_

**Suggested Service** \_\_\_\_\_

RECEIVED  
MAY 02 2016  
FDLE  
Alcohol Testing Program

**Optical Bench Calibration** Performed By \_\_\_\_\_

Optical Bench Calibration N/A  
 Optical Bench Calibration Complete

**Barometric Pressure Gauge** ID # \_\_\_\_\_

| Simulator | Serial Number | Lot Number | Expiration |
|-----------|---------------|------------|------------|
| 0.000     |               | N/A        | N/A        |
| 0.040     |               |            |            |
| 0.100     |               |            |            |
| 0.200     |               |            |            |
| 0.400     |               |            |            |
| 0.080 DGS | <u>N/A</u>    |            |            |

Post Calibration Stability Checks

| Simulator | Serial Number | Lot Number | Expiration |
|-----------|---------------|------------|------------|
| 0.05      |               |            |            |
| 0.08      |               |            |            |
| 0.20      |               |            |            |
| 0.08 DGS  | <u>N/A</u>    |            |            |

**Department Inspection** Performed By DEAR

Barometric Pressure 1019 Gauge  
 ID# 28663 1019 Instrument

Mouth Alcohol Solution Lot # 2015-A  
 Acetone Stock Solution Lot # 2015-B

| Simulator   | Serial Number  |
|-------------|----------------|
| 0.00        | <u>SD 3965</u> |
| Interferent | <u>SD 3966</u> |
| 0.05        | <u>SD 3967</u> |
| 0.08        | <u>SD 3968</u> |
| 0.20        | <u>SD 3969</u> |

**Attachments**

Form 41  
 Pre-Stability Tests  
 Flow Calibration

Optical Bench Cal  
 Post-Stability Tests  
 Other \_\_\_\_\_

Notes: E-MAILED  APPROVED  
4/20/2016  
QA/QC OK GPM

Instrument Complies with Chapter 11D-8, FAC  
 Instrument Does Not Comply with Chapter 11D-8, FAC  
 Return to/Place into Evidentiary Use  
 Remain Out of Evidentiary Use  
 Conduct an Agency Inspection Before Evidentiary Use

Bruce Whitland  
Quality Control Review

5/2/16  
Date

68911

| TYPE OF TEST | SERIAL NUMBER | AGENCY                          | DATE       | PERFORMED BY       |
|--------------|---------------|---------------------------------|------------|--------------------|
| Stabilities  | 80-001247     | Broward County Sheriff's Office | 04/20/2016 | <i>[Signature]</i> |

AK

|  |  |  |  |
|--|--|--|--|
| 0.05g/210L <input checked="" type="checkbox"/> | 0.077 to 0.083 <input checked="" type="checkbox"/> | 0.20g/210L <input checked="" type="checkbox"/> | DGS 0.08g/210L <input checked="" type="checkbox"/> |
|--|--|--|--|

BROWARD COUNTY SO  
 InOXiLizer - AlControl Analyzer  
 Model 8000 SN 80-001247  
 04/20/2016  
 Software: 8100.27

| Test           | g/210L | Time  |
|----------------|--------|-------|
| Air Blank      | 0.000  | 08:21 |
| Control Test   | 0.049  | 08:22 |
| Air Blank      | 0.000  | 08:22 |
| Control Test   | 0.051  | 08:23 |
| Air Blank      | 0.000  | 08:24 |
| Control Test   | 0.051  | 08:24 |
| Air Blank      | 0.000  | 08:25 |
| Control Test   | 0.051  | 08:25 |
| Average        | 0.0503 |       |
| Std Dev        | 0.0012 |       |
| Rel Std Dev(%) | 2.2941 |       |

*[Signature]*  
Operator's Signature

BROWARD COUNTY SO  
 InOXiLizer - AlControl Analyzer  
 Model 8000 SN 80-001247  
 04/20/2016  
 Software: 8100.27

| Test           | g/210L | Time  |
|----------------|--------|-------|
| Air Blank      | 0.000  | 08:26 |
| Control Test   | 0.050  | 08:27 |
| Air Blank      | 0.000  | 08:28 |
| Control Test   | 0.051  | 08:28 |
| Air Blank      | 0.000  | 08:29 |
| Control Test   | 0.050  | 08:29 |
| Air Blank      | 0.000  | 08:30 |
| Control Test   | 0.050  | 08:30 |
| Average        | 0.0503 |       |
| Std Dev        | 0.0006 |       |
| Rel Std Dev(%) | 1.1787 |       |

*[Signature]*  
Operator's Signature

BROWARD COUNTY SO  
 InOXiLizer - AlControl Analyzer  
 Model 8000 SN 80-001247  
 04/20/2016  
 Software: 8100.27

| Test           | g/210L | Time  |
|----------------|--------|-------|
| Air Blank      | 0.000  | 08:34 |
| Control Test   | 0.200  | 08:34 |
| Air Blank      | 0.000  | 08:35 |
| Control Test   | 0.201  | 08:35 |
| Air Blank      | 0.000  | 08:36 |
| Control Test   | 0.202  | 08:36 |
| Air Blank      | 0.000  | 08:37 |
| Control Test   | 0.200  | 08:37 |
| Average        | 0.2010 |       |
| Std Dev        | 0.0010 |       |
| Rel Std Dev(%) | 0.4975 |       |

*[Signature]*  
Operator's Signature

BROWARD COUNTY SO  
 InOXiLizer - AlControl Analyzer  
 Model 8000 SN 80-001247  
 04/20/2016  
 Software: 8100.27

| Test           | g/210L | Time  |
|----------------|--------|-------|
| Air Blank      | 0.000  | 08:43 |
| Control Test   | 0.078  | 08:43 |
| Air Blank      | 0.000  | 08:44 |
| Control Test   | 0.079  | 08:44 |
| Air Blank      | 0.000  | 08:44 |
| Control Test   | 0.079  | 08:44 |
| Air Blank      | 0.000  | 08:45 |
| Control Test   | 0.079  | 08:45 |
| Average        | 0.0787 |       |
| Std Dev        | 0.0006 |       |
| Rel Std Dev(%) | 0.7335 |       |

*[Signature]*  
Operator's Signature

## INSTRUMENT PROCESSING SHEET

Agency Broward County Sheriff's Office S/N 80-001247

Date In 3/3/2016 Date Out 3/8/2016  Ship  P/U  H/D  CMI  EE

**Intake** Performed By DELL

Registration  
 Annual  
 Return from CMI  
 Return from Enforcement  
 Electronics  
 Other Checking Control tests

Visual Inspection:  
OK Case OK Handle  
OK Dry Gas Holder OK Feet  
OK Keyboard/Plug OK Back/Plugs  
OK Screws tight OK Breath Hose

Other Equipment:  
 Power cord  
 Printer Cable  
 Other ANT STATIC BAG

Notes: \_\_\_\_\_  
 \_\_\_\_\_  
 \_\_\_\_\_

**Quality Checks** Performed By \_\_\_\_\_

Breath Tube Screen  
 Replace O-Rings  
 Instrument Set Up Verified  
 R-Value \_\_\_\_\_  
 Flow Verification (L/s)  
 Flow Column # \_\_\_\_\_  
 32mm \_\_\_\_\_ (.139 - .169)  
 36mm \_\_\_\_\_ (.156 - .190)  
 53mm \_\_\_\_\_ (.228 - .278)  
 103mm \_\_\_\_\_ (.447 - .547)

Barometric Pressure Check  
 Gauge ID # \_\_\_\_\_

Stability Checks

| Simulator | Serial # | Lot #/Exp |
|-----------|----------|-----------|
| 0.05      |          |           |
| 0.08      |          |           |
| 0.20      |          |           |
| 0.08 DGS  | N/A      |           |

**Flow Calibration** Performed By \_\_\_\_\_

Flow Calibration N/A  
 Flow Calibration Complete  
 Flow Column # \_\_\_\_\_  
 5L/min - 47mm  
 15L/min - 53mm  
 30L/min - 103mm

R-Value \_\_\_\_\_  
 Post Calibration Verification (L/s)  
 Flow Column # \_\_\_\_\_  
 32mm \_\_\_\_\_ (.139 - .169)  
 36mm \_\_\_\_\_ (.156 - .190)  
 53mm \_\_\_\_\_ (.228 - .278)  
 103mm \_\_\_\_\_ (.447 - .547)

**Maintenance** Performed By \_\_\_\_\_

Battery Replacement  
 Dry Gas Regulator Replacement  
 Breath Tube Replacement  
 Other \_\_\_\_\_

**Suggested Service**  
Sent to repair facility

**RECEIVED**  
 MAR 15 2016  
 FDLE  
 Alcohol Testing Program

**Optical Bench Calibration** Performed By \_\_\_\_\_

Optical Bench Calibration N/A  
 Optical Bench Calibration Complete

Barometric Pressure Gauge ID # \_\_\_\_\_

| Simulator | Serial Number | Lot Number | Expiration |
|-----------|---------------|------------|------------|
| 0.000     |               | N/A        | N/A        |
| 0.040     |               |            |            |
| 0.100     |               |            |            |
| 0.200     |               |            |            |
| 0.400     |               |            |            |
| 0.080 DGS | N/A           |            |            |

Post Calibration Stability Checks

| Simulator | Serial Number | Lot Number | Expiration |
|-----------|---------------|------------|------------|
| 0.05      |               |            |            |
| 0.08      |               |            |            |
| 0.20      |               |            |            |
| 0.08 DGS  | N/A           |            |            |

**Department Inspection** Performed By \_\_\_\_\_

Barometric Pressure \_\_\_\_\_ Gauge  
 ID# \_\_\_\_\_ Instrument

Mouth Alcohol Solution Lot # \_\_\_\_\_  
 Acetone Stock Solution Lot # \_\_\_\_\_

| Simulator   | Serial Number |
|-------------|---------------|
| 0.00        |               |
| Interferent |               |
| 0.05        |               |
| 0.08        |               |
| 0.20        |               |

**Attachments**

Form 41  
 Pre-Stability Tests  
 Flow Calibration

Optical Bench Cal  
 Post-Stability Tests  
 Other FORM 40

Notes: \_\_\_\_\_  
E-MAILED QC-13K  
3/8/2016

[Signature]  
 Quality Control Review

Instrument Complies with Chapter 11D-8, FAC  
 Instrument Does Not Comply with Chapter 11D-8, FAC

Return to/Place into Evidentiary Use  
 Remain Out of Evidentiary Use

Conduct an Agency Inspection Before Evidentiary Use

3/15/16  
 Date

|              |               |                                 |            |              |
|--------------|---------------|---------------------------------|------------|--------------|
| TYPE OF TEST | SERIAL NUMBER | AGENCY                          | DATE       | PERFORMED BY |
| Stabilities  | 80-001247     | Broward County Sheriff's Office | 03/07/2016 | <i>PPM</i>   |

|   |   |   |   |
|---|---|---|---|
| 0.05g/210L                              | 0.08g/210L                              | 0.20g/210L                              | DGS 0.08g/210L                          |
| 0.047 to 0.053 <input type="checkbox"/> | 0.077 to 0.083 <input type="checkbox"/> | 0.194 to 0.206 <input type="checkbox"/> | 0.077 to 0.083 <input type="checkbox"/> |

BROWARD COUNTY SO  
 Intoxilyzer - Alcohol Analyzer  
 Model: 8100 SN 80-001247  
 03/07/2016  
 Software: 8100.27

| Test         | g/210L | Time  |
|--------------|--------|-------|
| Air Blank    | 0.000  | 08:11 |
| Control Test | 0.081  | 08:11 |
| Air Blank    | 0.000  | 08:12 |
| Control Test | 0.083  | 08:13 |
| Air Blank    | 0.000  | 08:13 |
| Control Test | 0.082  | 08:14 |
| Air Blank    | 0.000  | 08:14 |
| Control Test | 0.083  | 08:15 |
| Air Blank    | 0.000  | 08:15 |
| Control Test | 0.083  | 08:16 |
| Air Blank    | 0.000  | 08:17 |

Control Test Stats  
 Average 0.0824  
 Std Dev 0.0009  
 Rel Std Dev(%) 1.0855

BROWARD COUNTY SO  
 Intoxilyzer - Alcohol Analyzer  
 Model: 8100 SN 80-001247  
 03/07/2016  
 Software: 8100.27

| Test         | g/210L | Time  |
|--------------|--------|-------|
| Air Blank    | 0.000  | 08:18 |
| Control Test | 0.080  | 08:19 |
| Air Blank    | 0.000  | 08:19 |
| Control Test | 0.080  | 08:19 |
| Air Blank    | 0.000  | 08:20 |
| Control Test | 0.080  | 08:20 |
| Air Blank    | 0.000  | 08:21 |
| Control Test | 0.079  | 08:21 |
| Air Blank    | 0.000  | 08:22 |
| Control Test | 0.081  | 08:22 |
| Air Blank    | 0.000  | 08:22 |

Control Test Stats  
 Average 0.0800  
 Std Dev 0.0007  
 Rel Std Dev(%) 0.8839

Operator's Signature

*PPM*

Operator's Signature

*PPM*