

TYPE OF TEST	SERIAL NUMBER	AGENCY	DATE	PERFORMED BY
Stabilities	80-001250	Broward County Sheriff's Office	12/26/2016	<i>WLL</i>

0.05g/210L <input checked="" type="checkbox"/>	0.08g/210L <input checked="" type="checkbox"/>	0.20g/210L <input checked="" type="checkbox"/>	DGS 0.08g/210L <input checked="" type="checkbox"/>
<p>BROWARD COUNTY SO Intoxilizer - Alcotest Analyzer Model 8000 SN 80-001250 12/26/2016 Software: 8100.27</p> <p>Test g/210L Time</p> <p>Air Blank 0.000 08:10 Control Test 0.048 08:11 Air Blank 0.000 08:11 Control Test 0.048 08:12 Air Blank 0.000 08:12 Control Test 0.048 08:13 Air Blank 0.000 08:14</p> <p>Control Test Stats Average 0.0480 Std Dev 0.0000 Rel Std Dev(%) 0.0000</p> <p><i>WLL</i> Operator's Signature</p>	<p>BROWARD COUNTY SO Intoxilizer - Alcotest Analyzer Model 8000 SN 80-001250 12/26/2016 Software: 8100.27</p> <p>Test g/210L Time</p> <p>Air Blank 0.000 08:15 Control Test 0.078 08:15 Air Blank 0.000 08:16 Control Test 0.078 08:17 Air Blank 0.000 08:17 Control Test 0.079 08:18 Air Blank 0.000 08:18</p> <p>Control Test Stats Average 0.0783 Std Dev 0.0006 Rel Std Dev(%) 0.7370</p> <p><i>WLL</i> Operator's Signature</p>	<p>BROWARD COUNTY SO Intoxilizer - Alcotest Analyzer Model 8000 SN 80-001250 12/26/2016 Software: 8100.27</p> <p>Test g/210L Time</p> <p>Air Blank 0.000 08:20 Control Test 0.200 08:20 Air Blank 0.000 08:21 Control Test 0.200 08:22 Air Blank 0.000 08:22 Control Test 0.199 08:23 Air Blank 0.000 08:23</p> <p>Control Test Stats Average 0.1997 Std Dev 0.0006 Rel Std Dev(%) 0.2892</p> <p><i>WLL</i> Operator's Signature</p>	<p>BROWARD COUNTY SO Intoxilizer - Alcotest Analyzer Model 8000 SN 80-001250 12/26/2016 Software: 8100.27</p> <p>Test g/210L Time</p> <p>Air Blank 0.000 08:24 Control Test 0.079 08:25 Air Blank 0.000 08:25 Control Test 0.079 08:25 Air Blank 0.000 08:26 Control Test 0.080 08:26 Air Blank 0.000 08:27</p> <p>Control Test Stats Average 0.0793 Std Dev 0.0006 Rel Std Dev(%) 0.7277</p> <p><i>WLL</i> Operator's Signature</p>

WLL

WLL

INSTRUMENT PROCESSING SHEET

Agency Broward County Sheriff's Office S/N 80-001250
 Date In 6/3/2016 Date Out 6/7/2016 Ship P/U H/D CMI EE

Intake Performed By <u>DELL</u> <input type="checkbox"/> Registration <input checked="" type="checkbox"/> Annual <input type="checkbox"/> Return from CMI <input type="checkbox"/> Return from Enforcement Electronics <input type="checkbox"/> Other _____ Visual Inspection: <u>OK</u> Case <u>OK</u> Handle <u>OK</u> Dry Gas Holder <u>OK</u> Feet <u>OK</u> Keyboard/Plug <u>OK</u> Back/Plugs <u>OK</u> Screws tight <u>OK</u> Breath Hose Other Equipment: <input type="checkbox"/> Power cord <input type="checkbox"/> Printer Cable <input type="checkbox"/> Other _____ Notes: _____ _____ _____	Quality Checks Performed By <u>DELL</u> <input checked="" type="checkbox"/> Breath Tube Screen <input checked="" type="checkbox"/> Replace O-Rings <input checked="" type="checkbox"/> Instrument Set Up Verified <input checked="" type="checkbox"/> R-Value <u>131</u> <input checked="" type="checkbox"/> Flow Verification (L/s) Flow Column # <u>ATP101</u> 32mm <u>148</u> (.139 - .169) 36mm <u>171</u> (.156 - .190) 53mm <u>242</u> (.228 - .278) 103mm <u>507</u> (.447 - .547) <input checked="" type="checkbox"/> Barometric Pressure Check Gauge ID # <u>28663</u> <input checked="" type="checkbox"/> Stability Checks <table border="1" style="width:100%; border-collapse: collapse;"> <thead> <tr> <th>Simulator</th> <th>Serial #</th> <th>Lot #/Exp</th> </tr> </thead> <tbody> <tr> <td>0.05</td> <td><u>503967</u></td> <td><u>201507A</u> <u>07/14/2017</u></td> </tr> <tr> <td>0.08</td> <td><u>503968</u></td> <td><u>201601F</u> <u>01/26/2018</u></td> </tr> <tr> <td>0.20</td> <td><u>503969</u></td> <td><u>201505A</u> <u>05/12/2017</u></td> </tr> <tr> <td>0.08 DGS</td> <td>N/A</td> <td><u>AG 600504</u> <u>01/05/2018</u></td> </tr> </tbody> </table>	Simulator	Serial #	Lot #/Exp	0.05	<u>503967</u>	<u>201507A</u> <u>07/14/2017</u>	0.08	<u>503968</u>	<u>201601F</u> <u>01/26/2018</u>	0.20	<u>503969</u>	<u>201505A</u> <u>05/12/2017</u>	0.08 DGS	N/A	<u>AG 600504</u> <u>01/05/2018</u>	Flow Calibration Performed By _____ <input checked="" type="checkbox"/> Flow Calibration N/A <input type="checkbox"/> Flow Calibration Complete Flow Column # _____ <input type="checkbox"/> 5L/min - 3mm <input type="checkbox"/> 15L/min - 53mm <input type="checkbox"/> 30L/min - 103mm <input type="checkbox"/> R-Value _____ <input type="checkbox"/> Post Calibration Verification (L/s) Flow Column # _____ 32mm _____ (.139 - .169) 36mm _____ (.156 - .190) 53mm _____ (.228 - .278) 103mm _____ (.447 - .547)
Simulator	Serial #	Lot #/Exp															
0.05	<u>503967</u>	<u>201507A</u> <u>07/14/2017</u>															
0.08	<u>503968</u>	<u>201601F</u> <u>01/26/2018</u>															
0.20	<u>503969</u>	<u>201505A</u> <u>05/12/2017</u>															
0.08 DGS	N/A	<u>AG 600504</u> <u>01/05/2018</u>															
		Maintenance Performed By _____ <input type="checkbox"/> Battery Replacement <input type="checkbox"/> Dry Gas Regulator Replacement <input type="checkbox"/> Breath Tube Replacement <input type="checkbox"/> Other _____ Suggested Service _____ _____															

RECEIVED
JUN 15 2016
FDLE
Alcohol Testing Program

Optical Bench Calibration Performed By _____			
<input checked="" type="checkbox"/> Optical Bench Calibration N/A <input type="checkbox"/> Optical Bench Calibration Complete Barometric Pressure Gauge ID # _____			
Simulator	Serial Number	Lot Number	Expiration
0.000		N/A	N/A
0.040			
0.100			
0.200			
0.400			
0.080 DGS	N/A		
<input type="checkbox"/> Post Calibration Stability Checks			
Simulator	Serial Number	Lot Number	Expiration
0.05			
0.08			
0.20			
0.08 DGS	N/A		

Department Inspection Performed By <u>DELL</u>	
<input checked="" type="checkbox"/> Barometric Pressure ID# <u>28199</u>	Gauge <u>1004</u> Instrument <u>1007</u>
Mouth Alcohol Solution Lot # <u>2016-A</u>	
Acetone Stock Solution Lot # <u>2016-B</u>	
Simulator	Serial Number
0.00	<u>SD 3965</u>
Interferent	<u>SD 3966</u>
0.05	<u>SD 3967</u>
0.08	<u>SD 3968</u>
0.20	<u>SD 3969</u>

Attachments	
<input checked="" type="checkbox"/> Form 41 <input checked="" type="checkbox"/> Pre-Stability Tests <input type="checkbox"/> Flow Calibration	<input type="checkbox"/> Optical Bench Cal <input type="checkbox"/> Post-Stability Tests <input type="checkbox"/> Other _____

Notes: **E-MAILED** **APPROVED**
6/7/2016
QA/PC OK BJM

<input checked="" type="checkbox"/> Instrument Complies with Chapter 11D-8, FAC <input type="checkbox"/> Instrument Does Not Comply with Chapter 11D-8, FAC <input checked="" type="checkbox"/> Return to/Place into Evidentiary Use <input type="checkbox"/> Remain Out of Evidentiary Use <input checked="" type="checkbox"/> Conduct an Agency Inspection Before Evidentiary Use

Britt Kirkland
 Quality Control Review

6/15/16
 Date

Agency Broward County Sheriff's Office S/N 80-001250
 Date In 12/28/2015 Date Out 2/9/2016 Ship P/U H/D CMI EE

Intake Performed By <u>DELL</u> <input type="checkbox"/> Registration <input checked="" type="checkbox"/> Annual <input type="checkbox"/> Return from CMI <input type="checkbox"/> Return from Enforcement Electronics <input checked="" type="checkbox"/> Other <u>Replace Dry Gas Regulator</u> Visual Inspection: <u>OK</u> Case <u>OK</u> Handle <u>OK</u> Dry Gas Holder <u>OK</u> Feet <u>OK</u> Keyboard/Plug <u>OK</u> Back/Plugs <u>OK</u> Screws tight <u>OK</u> Breath Hose Other Equipment: <input checked="" type="checkbox"/> Power cord <input type="checkbox"/> Printer Cable <input type="checkbox"/> Other <u>ANTI STATIC BAG</u> Notes: <u>Waiting on Dry Gas regulator from CMI</u>	Quality Checks Performed By <u>DELL</u> <input checked="" type="checkbox"/> Breath Tube Screen <input checked="" type="checkbox"/> Replace O-Rings <input checked="" type="checkbox"/> Instrument Set Up Verified <input checked="" type="checkbox"/> R-Value <u>132</u> <input checked="" type="checkbox"/> Flow Verification (L/s) Flow Column # <u>ATP 101</u> 32mm <u>144</u> (.139 - .169) 36mm <u>167</u> (.156 - .190) 53mm <u>238</u> (.228 - .278) 103mm <u>520</u> (.447 - .547) <input checked="" type="checkbox"/> Barometric Pressure Check Gauge ID # <u>68639</u> <input checked="" type="checkbox"/> Stability Checks <table border="1" style="width:100%; border-collapse: collapse;"> <thead> <tr> <th>Simulator</th> <th>Serial #</th> <th>Lot #/Exp</th> </tr> </thead> <tbody> <tr> <td>0.05</td> <td>SD3947</td> <td>201507A 07/14/2017</td> </tr> <tr> <td>0.08</td> <td>SD3968</td> <td>2015036 03/24/2017</td> </tr> <tr> <td>0.20</td> <td>SD3969</td> <td>201505A 05/12/2017</td> </tr> <tr> <td>0.08 DGS</td> <td>N/A</td> <td>A6507523 03/16/2017</td> </tr> </tbody> </table>	Simulator	Serial #	Lot #/Exp	0.05	SD3947	201507A 07/14/2017	0.08	SD3968	2015036 03/24/2017	0.20	SD3969	201505A 05/12/2017	0.08 DGS	N/A	A6507523 03/16/2017	Flow Calibration Performed By <u>DELL</u> <input checked="" type="checkbox"/> Flow Calibration N/A <input type="checkbox"/> Flow Calibration Complete Flow Column # <u>ATP 101</u> <input type="checkbox"/> 5L/min - 17mm <input type="checkbox"/> 15L/min - 53mm <input type="checkbox"/> 30L/min - 103mm <input type="checkbox"/> R-Value _____ <input type="checkbox"/> Post Calibration Verification (L/s) Flow Column # _____ 32mm _____ (.139 - .169) 36mm _____ (.156 - .190) 53mm _____ (.228 - .278) 103mm _____ (.447 - .547)
Simulator	Serial #	Lot #/Exp															
0.05	SD3947	201507A 07/14/2017															
0.08	SD3968	2015036 03/24/2017															
0.20	SD3969	201505A 05/12/2017															
0.08 DGS	N/A	A6507523 03/16/2017															
Maintenance Performed By <u>DELL</u> <input type="checkbox"/> Battery Replacement <input checked="" type="checkbox"/> Dry Gas Regulator Replacement <input type="checkbox"/> Breath Tube Replacement <input checked="" type="checkbox"/> Other <u>Tighten handle</u> Suggested Service																	

RECEIVED
 FEB 26 2016
 FDLE
 Alcohol Testing Program

Optical Bench Calibration Performed By _____ <input checked="" type="checkbox"/> Optical Bench Calibration N/A <input type="checkbox"/> Optical Bench Calibration Complete Barometric Pressure Gauge ID # _____ <table border="1" style="width:100%; border-collapse: collapse;"> <thead> <tr> <th>Simulator</th> <th>Serial Number</th> <th>Lot Number</th> <th>Expiration</th> </tr> </thead> <tbody> <tr> <td>0.000</td> <td></td> <td>N/A</td> <td>N/A</td> </tr> <tr> <td>0.040</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.100</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.200</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.400</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.080 DGS</td> <td>N/A</td> <td></td> <td></td> </tr> </tbody> </table> <input type="checkbox"/> Post Calibration Stability Checks <table border="1" style="width:100%; border-collapse: collapse;"> <thead> <tr> <th>Simulator</th> <th>Serial Number</th> <th>Lot Number</th> <th>Expiration</th> </tr> </thead> <tbody> <tr> <td>0.05</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.08</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.20</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.08 DGS</td> <td>N/A</td> <td></td> <td></td> </tr> </tbody> </table>	Simulator	Serial Number	Lot Number	Expiration	0.000		N/A	N/A	0.040				0.100				0.200				0.400				0.080 DGS	N/A			Simulator	Serial Number	Lot Number	Expiration	0.05				0.08				0.20				0.08 DGS	N/A			Department Inspection Performed By <u>DELL</u> <input checked="" type="checkbox"/> Barometric Pressure <u>1018</u> Gauge ID# <u>28663</u> <u>1015</u> Instrument Mouth Alcohol Solution Lot # <u>2015-A</u> Acetone Stock Solution Lot # <u>2015-B</u> <table border="1" style="width:100%; border-collapse: collapse;"> <thead> <tr> <th>Simulator</th> <th>Serial Number</th> </tr> </thead> <tbody> <tr> <td>0.00</td> <td>SD 3965</td> </tr> <tr> <td>Interferent</td> <td>SD 3966</td> </tr> <tr> <td>0.05</td> <td>SD 3967</td> </tr> <tr> <td>0.08</td> <td>SD 3968</td> </tr> <tr> <td>0.20</td> <td>SD 3969</td> </tr> </tbody> </table>	Simulator	Serial Number	0.00	SD 3965	Interferent	SD 3966	0.05	SD 3967	0.08	SD 3968	0.20	SD 3969
Simulator	Serial Number	Lot Number	Expiration																																																										
0.000		N/A	N/A																																																										
0.040																																																													
0.100																																																													
0.200																																																													
0.400																																																													
0.080 DGS	N/A																																																												
Simulator	Serial Number	Lot Number	Expiration																																																										
0.05																																																													
0.08																																																													
0.20																																																													
0.08 DGS	N/A																																																												
Simulator	Serial Number																																																												
0.00	SD 3965																																																												
Interferent	SD 3966																																																												
0.05	SD 3967																																																												
0.08	SD 3968																																																												
0.20	SD 3969																																																												
Attachments <input checked="" type="checkbox"/> Form 41 <input checked="" type="checkbox"/> Pre-Stability Tests <input type="checkbox"/> Flow Calibration <input type="checkbox"/> Optical Bench Cal <input type="checkbox"/> Post-Stability Tests <input type="checkbox"/> Other _____																																																													

Notes: 2/9/2016 **APPROVED**
QC-TBK
[Signature]
 Quality Control Review

Instrument Complies with Chapter 11D-8, FAC
 Instrument Does Not Comply with Chapter 11D-8, FAC
 Return to/Place into Evidentiary Use
 Remain Out of Evidentiary Use
 Conduct an Agency Inspection Before Evidentiary Use
2/26/16
 Date

TYPE OF TEST	SERIAL NUMBER	AGENCY	DATE	PERFORMED BY
Stabilities	80-001250	Broward County Sheriff's Office	02/09/2016	<i>WLL</i>

WLL

0.05g/210L	0.08g/210L	0.20g/210L	DG5 0.08g/210L
0.047 to 0.053 <input checked="" type="checkbox"/>	0.077 to 0.083 <input checked="" type="checkbox"/>	0.194 to 0.206 <input checked="" type="checkbox"/>	0.077 to 0.083 <input checked="" type="checkbox"/>

BROWARD COUNTY SC
Intoxilizer - Alcohol Analyzer
Model 8000 SN 80-001250
02/09/2016
Software: 8100.27

Test 9/210L Time

Air Blank 0.009 09:06
Control Test 0.059 09:06
Air Blank 0.003 09:07
Control Test 0.050 09:08
Air Blank 0.006 09:08
Control Test 0.050 09:09
Air Blank 0.005 09:09
Control Test Stats
Average 0.053
Std Dev 0.000
Rel Std Dev(%) 0.000

WLL
Operator's Signature

BROWARD COUNTY SC
Intoxilizer - Alcohol Analyzer
Model 8000 SN 90-001250
02/09/2016
Software: 8100.27

Test 9/210L Time

Air Blank 0.006 09:11
Control Test 0.079 09:12
Air Blank 0.000 09:12
Control Test 0.079 09:13
Air Blank 0.000 09:13
Control Test 0.079 09:14
Air Blank 0.002 09:14
Control Test Stats
Average 0.079
Std Dev 0.000
Rel Std Dev(%) 0.000

WLL
Operator's Signature

BROWARD COUNTY SC
Intoxilizer - Alcohol Analyzer
Model 8000 SN 80-001250
02/09/2016
Software: 9100.27

Test 9/210L Time

Air Blank 0.000 09:16
Control Test 0.198 09:17
Air Blank 0.000 09:17
Control Test 0.199 09:18
Air Blank 0.000 09:18
Control Test 0.200 09:19
Air Blank 0.000 09:20
Control Test Stats
Average 0.199
Std Dev 0.000
Rel Std Dev(%) 0.000

WLL
Operator's Signature

BROWARD COUNTY SC
Intoxilizer - Alcohol Analyzer
Model 8000 SN 80-001250
02/09/2016
Software: 8100.27

Test 9/210L Time

Air Blank 0.000 09:22
Control Test 0.081 09:22
Air Blank 0.000 09:22
Control Test 0.081 09:23
Air Blank 0.000 09:23
Control Test 0.081 09:24
Air Blank 0.000 09:24
Control Test Stats
Average 0.081
Std Dev 0.000
Rel Std Dev(%) 0.000

WLL
Operator's Signature

WLL