

INSTRUMENT PROCESSING SHEET

Agency Brevard County S.O S/N 80-001253
 Date In 2015 Date Out 6/2/16 Ship P/U H/D CMI EE

Intake Performed By _____

Registration
 Annual *See 2015*
 Return from CMI *IPS*
 Return from Enforcement
 Electronics
 Other _____

Visual Inspection:
 ___ Case ___ Handle
 ___ Dry Gas Holder ___ Feet
 ___ Keyboard/Plug ___ Back/Plugs
 ___ Screws tight ___ Breath Hose

Other Equipment:
 Power cord
 Printer Cable
 Other _____

Notes: _____

Quality Checks Performed By FR

Breath Tube Screen
 Replace O-Rings
 Instrument Set Up Verified
 R-Value 179
 Flow Verification (L/s)
 Flow Column # ATP103
 32mm 148 (.139 - .169)
 36mm 170 (.156 - .190)
 53mm 251 (.228 - .278)
 103mm 500 (.447 - .547)

Barometric Pressure Check
 Gauge ID # 28427

Stability Checks

Simulator	Serial #	Lot #/Exp
0.05	SD1018	201507A 7/14/17
0.08	SD1011	201601F 1/26/16
0.20	G4444	201505A 5/12/17
0.08 DGS	N/A	AT-605301 2/2/18

Flow Calibration Performed By _____

Flow Calibration N/A
 Flow Calibration Complete
 Flow Column # _____
 5L/min - 17mm
 15L/min - 53mm
 30L/min - 103mm

R-Value _____
 Post Calibration Verification (L/s)
 Flow Column # _____
 32mm _____ (.139 - .169)
 36mm _____ (.156 - .190)
 53mm _____ (.228 - .278)
 103mm _____ (.447 - .547)

Maintenance Performed By _____

Battery Replacement
 Dry Gas Regulator Replacement
 Breath Tube Replacement
 Other _____

Suggested Service

RECEIVED
 JUN 03 2016
 FDLE
 Alcohol Testing Program

Optical Bench Calibration Performed By _____

Optical Bench Calibration N/A
 Optical Bench Calibration Complete

Barometric Pressure Gauge ID # _____

Simulator	Serial Number	Lot Number	Expiration
0.000		N/A	N/A
0.040			
0.100			
0.200			
0.400			
0.080 DGS	N/A		

Post Calibration Stability Checks

Simulator	Serial Number	Lot Number	Expiration
0.05			
0.08			
0.20			
0.08 DGS	N/A		

Department Inspection Performed By FR

Barometric Pressure 1015 Gauge
 ID# 28427 1016 Instrument

Mouth Alcohol Solution Lot # 2016A
 Acetone Stock Solution Lot # 2016R

Simulator	Serial Number
0.00	SD1022 62880
Interferent	SD1027 62834
0.05	SD1016 62403
0.08	SD1011 503964
0.20	G4444 64444

Attachments

Form 41
 Pre-Stability Tests
 Flow Calibration

Optical Bench Cal
 Post-Stability Tests
 Other Com UC

Notes: Power loss mid Department inspection
4/21/16. Dam Spect.

QC - DUS

Brett Huckland

Instrument Complies with Chapter 11D-8, FAC
 Instrument Does Not Comply with Chapter 11D-8, FAC

Return to/Place into Evidentiary Use
 Remain Out of Evidentiary Use
 Conduct an Agency Inspection Before Evidentiary Use

Quality Control Review

6/3/16
 Date

80-001253
 Stability Checks
 4/21/16
 (P)

BREVARD COUNTY S.O.
 Intoxilyzer - Alcohol Analyzer
 Model 8000 SN 80-001253
 04/21/2016
 Software: 8100.27

INTOXILYZER 8000
 Instrument Initialization
 08:07 04/21/2016

Test	g/210L	Time
Air Blank	0.000	08:36
Control Test	0.077	08:36
Air Blank	0.000	08:37
Control Test	0.077	08:37
Air Blank	0.000	08:38
Control Test	0.077	08:38
Air Blank	0.000	08:39

Control Test Stats
 Average 0.0770
 Std Dev 0.0000
 Rel Std Dev(%) 0.0000

DGS



 Operator's Signature

BREVARD COUNTY S.O.
 Intoxilyzer - Alcohol Analyzer
 Model 8000 SN 80-001253
 04/21/2016
 Software: 8100.27


Test	g/210L	Time
Air Blank	0.000	08:40
Control Test	0.197	08:41
Air Blank	0.000	08:41
Control Test	0.197	08:42
Air Blank	0.000	08:42
Control Test	0.198	08:43
Air Blank	0.000	08:44

Control Test Stats
 Average 0.1973
 Std Dev 0.0006
 Rel Std Dev(%) 0.2926


BREVARD COUNTY S.O.
 Intoxilyzer - Alcohol Analyzer
 Model 8000 SN 80-001253
 04/21/2016
 Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	08:45
Control Test	0.050	08:45
Air Blank	0.000	08:46
Control Test	0.049	08:47
Air Blank	0.000	08:47
Control Test	0.050	08:48
Air Blank	0.000	08:48

Control Test Stats
 Average 0.0497
 Std Dev 0.0006
 Rel Std Dev(%) 1.1625



 Operator's Signature



 Operator's Signature

DGS

BREVARD COUNTY
 Intoxilyzer - Alcohol Analyzer
 Model 8000 SN 80-001253
 04/21/2016
 Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	08:51
Control Test	0.079	08:51
Air Blank	0.000	08:52
Control Test	0.079	08:53
Air Blank	0.000	08:53
Control Test	0.079	08:54
Air Blank	0.000	08:54

Control Test Stats
 Average 0.0790
 Std Dev 0.0000
 Rel Std Dev(%) 0.0000



 Operator's Signature