

INSTRUMENT PROCESSING SHEET

Agency Orange County SO S/N 80-001256
 Date In 6/29/16 Date Out 7/28/16 Ship P/U H/D CMI EE

RECEIVED
 JUL 29 2016
 FDLE
 Alcohol Testing Program

Intake Performed By [Signature]

Registration
 Annual
 Return from CMI
 Return from Enforcement
 Electronics
 Other _____

Visual Inspection:

OK Case OK Handle
OK Dry Gas Holder OK Feet
OK Keyboard/Plug OK Back/Plugs
OK Screws tight OK Breath Hose

Other Equipment:

Power cord
 Printer Cable
 Other Static Bag

Notes: _____

Quality Checks Performed By [Signature]

Breath Tube Screen
 Replace O-Rings
 Instrument Set Up Verified
 R-Value 208
 Flow Verification (L/s)
 Flow Column # ATR102
 32mm 150 (.139 - .169)
 36mm 174 (.156 - .190)
 53mm 259 (.228 - .278)
 103mm 512 (.447 - .547)

Barometric Pressure Check
 Gauge ID # 28427

Stability Checks

Simulator	Serial #	Lot #/Exp
0.05	G4444	201507A 7/14/17
0.08	G6621	201601F 1/26/18
0.20	SD1012	201505A 5/12/17
0.08 DGS	N/A	19G612-105 5/31/18

Flow Calibration Performed By _____

Flow Calibration N/A
 Flow Calibration Complete
 Flow Column # _____
 5L/min - 17mm
 15L/min - 53mm
 30L/min - 103mm

R-Value _____
 Post Calibration Verification (L/s)
 Flow Column # _____
 32mm _____ (.139 - .169)
 36mm _____ (.156 - .190)
 53mm _____ (.228 - .278)
 103mm _____ (.447 - .547)

Maintenance Performed By _____

Battery Replacement
 Dry Gas Regulator Replacement
 Breath Tube Replacement
 Other _____

Suggested Service

Optical Bench Calibration Performed By _____

Optical Bench Calibration N/A
 Optical Bench Calibration Complete

Barometric Pressure Gauge ID # _____

Simulator	Serial Number	Lot Number	Expiration
0.000		N/A	N/A
0.040			
0.100			
0.200			
0.400			
0.080 DGS	N/A		

Post Calibration Stability Checks

Simulator	Serial Number	Lot Number	Expiration
0.05			
0.08			
0.20			
0.08 DGS	N/A		

Department Inspection Performed By [Signature]

Barometric Pressure 1016 Gauge
 ID# 28427 1016 Instrument

Mouth Alcohol Solution Lot # 2016-A
 Acetone Stock Solution Lot # 2016-B

Simulator	Serial Number
0.00	G2835
Interferent	G2840
0.05	G4444
0.08	G6621
0.20	SD1012

Attachments

Form 41 Optical Bench Cal
 Pre-Stability Tests Post-Stability Tests
 Flow Calibration Other _____

Notes: PA/OC OK [Signature] 7/28/16

Brett Herbrand
 Quality Control Review

Instrument Complies with Chapter 11D-8, FAC
 Instrument Does Not Comply with Chapter 11D-8, FAC

Return to/Place Into Evidentiary Use
 Remain Out of Evidentiary Use
 Conduct an Agency Inspection Before Evidentiary Use

7/29/16
 Date

80-001256
 Stability Checks
 7/28/16

AK

INTOXILYZER 8000
 Instrument Initialization
 07:58 07/28/2016

ORANGE COUNTY S.O.
 Intoxilyzer - Alcohol Analyzer
 Model 8000 SN 80-001256
 07/28/2016
 Software: 8100.27

Test	9/21/0L	Time
Air Blank	0.000	08:48
Control Test	0.201	08:49
Air Blank	0.000	08:50
Control Test	0.202	08:50
Air Blank	0.000	08:51
Control Test	0.202	08:51
Air Blank	0.000	08:52
Control Test Stats		
Average	0.2017	
Std Dev	0.0006	
Rel Std Dev(%)	0.2863	

[Signature]
 Operator's Signature

ORANGE COUNTY S.O.
 Intoxilyzer - Alcohol Analyzer
 Model 8000 SN 80-001256
 07/28/2016
 Software: 8100.27

Test	9/21/0L	Time
Air Blank	0.000	08:53
Control Test	0.078	08:54
Air Blank	0.000	08:54
Control Test	0.078	08:55
Air Blank	0.000	08:56
Control Test	0.079	08:56
Air Blank	0.000	08:57
Control Test Stats		
Average	0.0783	
Std Dev	0.0006	
Rel Std Dev(%)	0.7370	

[Signature]
 Operator's Signature

ORANGE COUNTY S.O.
 Intoxilyzer - Alcohol Analyzer
 Model 8000 SN 80-001256
 07/28/2016
 Software: 8100.27

Test	9/21/0L	Time
Air Blank	0.000	08:58
Control Test	0.048	08:59
Air Blank	0.000	08:59
Control Test	0.048	09:00
Air Blank	0.000	09:01
Control Test	0.048	09:02
Air Blank	0.000	
Control Test Stats		
Average	0.0480	
Std Dev	0.0000	
Rel Std Dev(%)	0.0000	

[Signature]
 Operator's Signature

ORANGE COUNTY S.O.
 Intoxilyzer - Alcohol Analyzer
 Model 8000 SN 80-001256
 07/28/2016
 Software: 8100.27

Test	9/21/0L	Time
Air Blank	0.000	08:34
Control Test	0.081	08:35
Air Blank	0.000	08:35
Control Test	0.080	08:35
Air Blank	0.000	08:36
Control Test	0.080	08:36
Air Blank	0.000	08:37
Control Test Stats		
Average	0.0803	
Std Dev	0.0006	
Rel Std Dev(%)	0.7187	

DSS

[Signature]

[Signature]
 Operator's Signature