



Alcohol Testing Program

INSTRUMENT PROCESSING SHEET

Agency Cocoa PD S/N 80-001260
 Date In 3/5/16 Date Out 4/7/16 Ship P/U H/D CMI EE

Intake Performed By [Signature]

Registration
 Annual
 Return from CMI
 Return from Enforcement Electronics
 Other _____

Visual Inspection:
OK Case OK Handle
OK Dry Gas Holder OK Feet
OK Keyboard/Plug OK Back/Plugs
OK Screws tight OK Breath Hose

Other Equipment:
 Power cord
 Printer Cable
 Other _____

Notes: _____

Quality Checks Performed By [Signature]

Breath Tube Screen
 Replace O-Rings
 Instrument Set Up Verified
 R-Value 163
 Flow Verification (L/s)
 Flow Column # AP102
 32mm 147 (.139 - .169)
 36mm 166 (.156 - .190)
 53mm 241 (.228 - .278)
 103mm 497 (.447 - .547)

Barometric Pressure Check
 Gauge ID # 26427

Stability Checks

Simulator	Serial #	Lot #/Exp
0.05	SD106	201507A 7/14/17
0.08	SD1011	201601F 1/26/18
0.20	G4444	201505A 5/12/17
0.08 DGS	N/A	AG605301 2/22/18

Flow Calibration Performed By _____

Flow Calibration N/A
 Flow Calibration Complete
 Flow Column # _____
 5L/min - 17mm
 15L/min - 53mm
 30L/min - 103mm
 R-Value _____
 Post Calibration Verification (L/s)
 Flow Column # _____
 32mm _____ (.139 - .169)
 36mm _____ (.156 - .190)
 53mm _____ (.228 - .278)
 103mm _____ (.447 - .547)

Maintenance Performed By _____

Battery Replacement
 Dry Gas Regulator Replacement
 Breath Tube Replacement
 Other _____

Suggested Service

RECEIVED
 APR 08 2016
 FDLE
 Alcohol Testing Program

Optical Bench Calibration Performed By _____

Optical Bench Calibration N/A
 Optical Bench Calibration Complete

Barometric Pressure Gauge ID # _____

Simulator	Serial Number	Lot Number	Expiration
0.000		N/A	N/A
0.040			
0.100			
0.200			
0.400			
0.080 DGS	N/A		

Post Calibration Stability Checks

Simulator	Serial Number	Lot Number	Expiration
0.05			
0.08			
0.20			
0.08 DGS	N/A		

Department Inspection Performed By [Signature]

Barometric Pressure 1012 Gauge
 ID# 26427 1007 Instrument

Mouth Alcohol Solution Lot # 2016 A
 Acetone Stock Solution Lot # 2016-B

Simulator	Serial Number
0.00	SD1027
Interferent	SD1021
0.05	SD106
0.08	SD104
0.20	G4444

Attachments

Form 41 Optical Bench Cal
 Pre-Stability Tests Post-Stability Tests
 Flow Calibration Other _____

Notes: QC - DUB

Instrument Complies with Chapter 11D-8, FAC
 Instrument Does Not Comply with Chapter 11D-8, FAC

Return to/Place into Evidentiary Use
 Remain Out of Evidentiary Use
 Conduct an Agency Inspection Before Evidentiary Use

Smith Kirkland

Quality Control Review

4/8/16

Date

80-001260
 Stability Checks
 4/7/16

INTOXILYZER 8000
 Instrument Initialization
 08:11 04/07/2016

COCOA P.D.
 Intoxilyzer - Alcohol Analyzer
 Model 8000 SN 80-001260
 04/07/2016
 Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	09:13
Control Test	0.079	09:14
Air Blank	0.000	09:15
Control Test	0.078	09:15
Air Blank	0.000	09:16
Control Test	0.078	09:17
Air Blank	0.000	09:17
Control Test Stats		
Average	0.0783	
Std Dev	0.0006	
Rel Std Dev(%)	0.7370	

Operator's Signature

COCOA P.D.
 Intoxilyzer - Alcohol Analyzer
 Model 8000 SN 80-001260
 04/07/2016
 Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	09:18
Control Test	0.049	09:19
Air Blank	0.000	09:19
Control Test	0.050	09:20
Air Blank	0.000	09:21
Control Test	0.049	09:21
Air Blank	0.000	09:22
Control Test Stats		
Average	0.0493	
Std Dev	0.0006	
Rel Std Dev(%)	1.1703	

Operator's Signature

COCOA P.D.
 Intoxilyzer - Alcohol Analyzer
 Model 8000 SN 80-001260
 04/07/2016
 Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	09:24
Control Test	0.201	09:24
Air Blank	0.000	09:25
Control Test	0.201	09:26
Air Blank	0.000	09:26
Control Test	0.200	09:27
Air Blank	0.000	09:28
Control Test Stats		
Average	0.2007	
Std Dev	0.0006	
Rel Std Dev(%)	0.2877	

DKS

Operator's Signature

Operator's Signature

DKS