

INSTRUMENT PROCESSING SHEET

Agency Rockledge PD S/N 80-001261
 Date In 2/1/16 Date Out 2/7/16 Ship P/U H/D CMI EE

Intake Performed By JP

Registration
 Annual
 Return from CMI
 Return from Enforcement Electronics
 Other _____

Visual Inspection:
OK Case OK Handle
OK Dry Gas Holder OK Feet
OK Keyboard/Plug OK Back/Plugs
OK Screws tight OK Breath Hose

Other Equipment:
 Power cord
 Printer Cable
 Other Static Bag

Notes: _____

Quality Checks Performed By JP

Breath Tube Screen
 Replace O-Rings
 Instrument Set Up Verified
 R-Value 225
 Flow Verification (L/s)
 Flow Column # A7P105
 32mm 151 (.139 - .169)
 36mm 179 (.156 - .190)
 53mm 248 (.228 - .278)
 103mm 508 (.447 - .547)

Barometric Pressure Check
 Gauge ID # 28427

Stability Checks

Simulator	Serial #	Lot #/Exp
0.05	SD1018	2015026 AG 7/14/17
0.08	SD1011	2015026 2/24/17
0.20	SD1025	201505A 5/12/17
0.08 DGS	N/A	AG 5/17/09 4/27/17

Flow Calibration Performed By JP

Flow Calibration N/A
 Flow Calibration Complete
 Flow Column # _____
 5L/min - 17mm
 15L/min - 53mm
 30L/min - 103mm
 R-Value _____
 Post Calibration Verification (L/s)
 Flow Column # _____
 32mm _____ (.139 - .169)
 36mm _____ (.156 - .190)
 53mm _____ (.228 - .278)
 103mm _____ (.447 - .547)

Maintenance Performed By _____

Battery Replacement
 Dry Gas Regulator Replacement
 Breath Tube Replacement
 Other _____

Suggested Service

RECEIVED
 FEB 09 2016
 FDLE
 Alcohol Testing Program

Optical Bench Calibration Performed By _____

Optical Bench Calibration N/A
 Optical Bench Calibration Complete

Barometric Pressure Gauge ID # _____

Simulator	Serial Number	Lot Number	Expiration
0.000		N/A	N/A
0.040			
0.100			
0.200			
0.400			
0.080 DGS	N/A		

Post Calibration Stability Checks

Simulator	Serial Number	Lot Number	Expiration
0.05			
0.08			
0.20			
0.08 DGS	N/A		

Department Inspection Performed By JP

Barometric Pressure
 ID# 28427 1019 1019 Gauge
1021 Instrument

Mouth Alcohol Solution Lot # 2015-A
 Acetone Stock Solution Lot # 2015-B

Simulator	Serial Number
0.00	SD1022
Interferent	SD1021
0.05	SD1018
0.08	SD1011
0.20	SD1025

Attachments

Form 41
 Pre-Stability Tests
 Flow Calibration
 Optical Bench Cal
 Post-Stability Tests
 Other _____

Notes: QC - BK

Instrument Complies with Chapter 11D-8, FAC
 Instrument Does Not Comply with Chapter 11D-8, FAC

Return to/Place into Evidentiary Use
 Remain Out of Evidentiary Use

Conduct an Agency Inspection Before Evidentiary Use

Patrick Murphy
 Quality Control Review

2/9/16
 Date

80-001261
 Stability checks
 2/7/16

RAM

ROCKLEDGE P.D.
 Intoxilyzer - Alcohol Analyzer
 Model 8000 SN 80-001261
 02/07/2016
 Software: 8100.27

ROCKLEDGE P.D.
 Intoxilyzer - Alcohol Analyzer
 Model 8000 SN 80-001261
 02/07/2016
 Software: 8100.27

ROCKLEDGE P.D.
 Intoxilyzer - Alcohol Analyzer
 Model 8000 SN 80-001261
 02/07/2016
 Software: 8100.27

ROCKLEDGE P.D.
 Intoxilyzer - Alcohol Analyzer
 Model 8000 SN 80-001261
 02/07/2016
 Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	10:08
Control Test	0.203	10:09
Air Blank	0.000	10:09
Control Test	0.203	10:10
Air Blank	0.000	10:11
Control Test	0.203	10:11
Air Blank	0.000	10:12
Control Test	0.203	10:12
Average	0.2030	
Std Dev	0.0000	
Rel Std Dev(%)	0.0000	

Test	g/210L	Time
Air Blank	0.000	10:02
Control Test	0.080	10:02
Air Blank	0.000	10:03
Control Test	0.080	10:03
Air Blank	0.000	10:04
Control Test	0.080	10:04
Air Blank	0.000	10:04
Control Test	0.081	10:04
Air Blank	0.000	10:04
Control Test	0.083	10:04
Average	0.0803	
Std Dev	0.0006	
Rel Std Dev(%)	0.7187	

Test	g/210L	Time
Air Blank	0.000	09:57
Control Test	0.051	09:57
Air Blank	0.000	09:58
Control Test	0.050	09:58
Air Blank	0.000	09:59
Control Test	0.050	09:59
Air Blank	0.000	10:00
Control Test	0.050	10:00
Air Blank	0.000	10:00
Control Test	0.050	10:00
Average	0.0503	
Std Dev	0.0006	
Rel Std Dev(%)	1.1471	

Test	g/210L	Time
Air Blank	0.000	09:51
Control Test	0.081	09:52
Air Blank	0.000	09:53
Control Test	0.080	09:53
Air Blank	0.000	09:54
Control Test	0.080	09:55
Air Blank	0.000	09:55
Control Test	0.080	09:55
Average	0.0803	
Std Dev	0.0006	
Rel Std Dev(%)	0.7187	

SK
 DOS

Operator's Signature

Operator's Signature

Operator's Signature

Operator's Signature