

RECEIVED
JUL 06 2016

INSTRUMENT PROCESSING SHEET

Agency Palm Bay PD S/N 80-001265

Date In 6/30/16 Date Out 7/6/16

Ship P/U H/D CMI EE

Alcohol Testing Program



Intake Performed By [Signature]

- Registration
- Annual
- Return from CMI
- Return from Enforcement Electronics
- Other _____

Visual Inspection:

OK Case di Handle

OK Dry Gas Holder OK Feet

OK Keyboard/Plug OK Back/Plugs

OK Screws tight OK Breath Hose

Other Equipment:

- Power cord
- Printer Cable
- Other Static Bag

Notes: _____

Quality Checks Performed By [Signature]

- Breath Tube Screen
- Replace O-Rings
- Instrument Set Up Verified
- R-Value 173
- Flow Verification (L/s)

Flow Column # ATP102

32mm 147 (.139 - .169)

36mm 168 (.156 - .190)

53mm 240 (.228 - .278)

103mm 498 (.447 - .547)

- Barometric Pressure Check
- Stability Checks

Simulator	Serial #	Lot #/Exp
0.05	SD1018	201507A 7/14/17
0.08	SD1011	201601F 1/26/18
0.20	SD1025	201505A 5/12/17
0.08 DGS	N/A	16605301 2/22/18

Flow Calibration Performed By [Signature]

- Flow Calibration N/A
- Flow Calibration Complete
- Flow Column # _____
- 5L/min - 17mm
- 15L/min - 53mm
- 30L/min - 103mm
- R-Value _____
- Post Calibration Verification (L/s)

Flow Column # _____

32mm _____ (.139 - .169)

36mm _____ (.156 - .190)

53mm _____ (.228 - .278)

103mm _____ (.447 - .547)

Department Inspection Performed By [Signature]

- Barometric Pressure 1019 Gauge
- ID# 28427 1019 Instrument

Mouth Alcohol Solution Lot # 2016-A

Acetone Stock Solution Lot # 2016-B

Simulator	Serial Number
0.00	SD1019
Interferent	SD1021
0.05	SD1018
0.08	SD1011
0.20	SD1025

Attachments

- Form 41
- Pre-Stability Tests
- Flow Calibration
- Optical Bench Cal
- Post-Stability Tests
- Other AT Software

Optical Bench Calibration Performed By _____

- Optical Bench Calibration N/A
- Optical Bench Calibration Complete

Barometric Pressure Gauge ID # _____

Simulator	Serial Number	Lot Number	Expiration
0.000		N/A	N/A
0.040			
0.100			
0.200			
0.400			
0.080 DGS	N/A		

Post Calibration Stability Checks

Simulator	Serial Number	Lot Number	Expiration
0.05			
0.08			
0.20			
0.08 DGS	N/A		

Notes: AT received intermittent data messages and low readings on .05 solution. Sent to ATP.

-Conducted stability tests and never received an issue. .05 readings as well as all others were within range. -Conducted Department inspection, in compliance. Advise AT to check simulators are properly cleaned and test for air leaks.

Instrument Complies with Chapter 11D-8, FAC

Instrument Does Not Comply with Chapter 11D-8, FAC

Return to/Place into Evidentiary Use

Remain Out of Evidentiary Use

Conduct an Agency Inspection Before Evidentiary Use

Quality Control Review Brett Kunkland

Date 7/6/16

QC-pws

80-001265
 Stability Checks
 7/6/16

INTOXILYZER 8000
 Instrument Initialization
 07:11 07/06/2016

PALM BAY P.D.
 Intoxilyzer - Alcohol Analyzer
 Model 8000 SN 80-001265
 07/06/2016
 Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	08:05
Control Test	0.047	08:06
Air Blank	0.000	08:06
Control Test	0.048	08:07
Air Blank	0.000	08:07
Control Test	0.048	08:08
Air Blank	0.000	08:08
Control Test Stats	0.0477	
Average	0.0006	
Std Dev	1.2112	
Rel Std Dev(%)		

[Signature]
 Operator's Signature

PALM BAY P.D.
 Intoxilyzer - Alcohol Analyzer
 Model 8000 SN 80-001265
 07/06/2016
 Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	08:15
Control Test	0.079	08:16
Air Blank	0.000	08:16
Control Test	0.079	08:17
Air Blank	0.000	08:17
Control Test	0.079	08:18
Air Blank	0.000	08:18
Control Test Stats		
Average	0.0790	
Std Dev	0.0000	
Rel Std Dev(%)	0.0000	

JDWS
BK

[Signature]
 Operator's Signature

PALM BAY P.D.
 Intoxilyzer - Alcohol Analyzer
 Model 8000 SN 80-001265
 07/06/2016
 Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	08:10
Control Test	0.199	08:11
Air Blank	0.000	08:11
Control Test	0.200	08:12
Air Blank	0.000	08:12
Control Test	0.199	08:13
Air Blank	0.000	08:13
Control Test Stats		
Average	0.1993	
Std Dev	0.0006	
Rel Std Dev(%)	0.2896	

[Signature]
 Operator's Signature

PALM BAY P.D.
 Intoxilyzer - Alcohol Analyzer
 Model 8000 SN 80-001265
 07/06/2016
 Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	07:52
Control Test	0.080	07:52
Air Blank	0.000	07:52
Control Test	0.080	07:53
Air Blank	0.000	07:53
Control Test	0.080	07:53
Air Blank	0.000	07:54
Control Test Stats		
Average	0.0800	
Std Dev	0.0000	
Rel Std Dev(%)	0.0000	

DGS

[Signature]
 Operator's Signature