



Alcohol Testing Program

INSTRUMENT PROCESSING SHEET

Agency Palm Bay P.D.

S/N 80-001266

Date In 1/10/15

Date Out 1/25/16

Ship P/U H/D CMI EE

Intake Performed By _____

Registration
 Annual
 Return from CMI
 Return from Enforcement Electronics
 Other _____

Visual Inspection:
 _____ Case _____ Handle
 _____ Dry Gas Holder _____ Feet
 _____ Keyboard/Plug _____ Back/Plugs
 _____ Screws tight _____ Breath Hose

Other Equipment:
 Power cord
 Printer Cable
 Other _____

Notes: _____

Quality Checks Performed By [Signature]

Breath Tube Screen
 Replace O-Rings
 Instrument Set Up Verified
 R-Value 162
 Flow Verification (L/s)
 Flow Column # A7P102
 32mm 146 (.139 - .169)
 36mm 165 (.156 - .190)
 53mm 240 (.228 - .278)
 103mm 500 (.447 - .547)

Barometric Pressure Check
 Gauge ID # 26932

Stability Checks

Simulator	Serial #	Lot #/Exp
0.05	SD1024	201507A 7/14/17
0.08	DR2035	201502G 2/24/17
0.20	DR3856	201505A 5/12/17
0.08 DGS	N/A	AG511701 4/27/17

Flow Calibration Performed By _____

Flow Calibration N/A
 Flow Calibration Complete
 Flow Column # _____
 5L/min - 17mm
 15L/min - 53mm
 30L/min - 103mm

R-Value _____
 Post Calibration Verification (L/s)
 Flow Column # _____
 32mm _____ (.139 - .169)
 36mm _____ (.156 - .190)
 53mm _____ (.228 - .278)
 103mm _____ (.447 - .547)

Maintenance Performed By _____

Battery Replacement
 Dry Gas Regulator Replacement
 Breath Tube Replacement
 Other _____

Suggested Service

RECEIVED
JAN 25 2016
FDLE
Alcohol Testing Program

Optical Bench Calibration Performed By _____

Optical Bench Calibration N/A
 Optical Bench Calibration Complete

Barometric Pressure Gauge ID # _____

Simulator	Serial Number	Lot Number	Expiration
0.000		N/A	N/A
0.040			
0.100			
0.200			
0.400			
0.080 DGS	N/A		

Post Calibration Stability Checks

Simulator	Serial Number	Lot Number	Expiration
0.05			
0.08			
0.20			
0.08 DGS	N/A		

Department Inspection Performed By [Signature]

Barometric Pressure 1024 Gauge
 ID# 26932 1022 Instrument

Mouth Alcohol Solution Lot # 2015-A
 Acetone Stock Solution Lot # 2015-B

Simulator	Serial Number
0.00	SD1016
Interferent	G-2840
0.05	SD1024
0.08	DR2035
0.20	DR3856

Attachments

Form 41
 Pre-Stability Tests
 Flow Calibration
 Optical Bench Cal
 Post-Stability Tests
 Other 2015 Paperwork

Notes: QC-13K

Instrument Complies with Chapter 11D-8, FAC
 Instrument Does Not Comply with Chapter 11D-8, FAC

Return to/Place into Evidentiary Use
 Remain Out of Evidentiary Use

Conduct an Agency Inspection Before Evidentiary Use

Potuck Murphy
Quality Control Review

1/25/16
Date

80-001266
 STAB, HYC, LOCKS
 1/25/16

INTOXILYZER 8000
 Instrument Initialization
 08:01 01/25/2016

RAM

PALM BRY P. D.
 Intoxilyzer - Alcohol Analyzer
 Model 8000 SN 80-001266
 01/25/2016
 Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	08:47
Control Test	0.203	08:48
Air Blank	0.000	08:48
Control Test	0.204	08:49
Air Blank	0.000	08:49
Control Test	0.204	08:50
Air Blank	0.000	08:51
Control Test Stats		
Average	0.2037	
Std Dev	0.0006	
Rel Std Dev(%)	0.2835	

Operator's Signature

PALM BRY P. D.
 Intoxilyzer - Alcohol Analyzer
 Model 8000 SN 80-001266
 01/25/2016
 Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	08:52
Control Test	0.079	08:53
Air Blank	0.000	08:53
Control Test	0.080	08:54
Air Blank	0.000	08:54
Control Test	0.080	08:55
Air Blank	0.000	08:55
Control Test Stats		
Average	0.0797	
Std Dev	0.0006	
Rel Std Dev(%)	0.7247	

SK

Operator's Signature

PALM BRY P. D.
 Intoxilyzer - Alcohol Analyzer
 Model 8000 SN 80-001266
 01/25/2016
 Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	08:57
Control Test	0.050	08:58
Air Blank	0.000	08:58
Control Test	0.050	08:59
Air Blank	0.000	09:00
Control Test	0.051	09:00
Air Blank	0.000	09:01
Control Test Stats		
Average	0.0503	
Std Dev	0.0006	
Rel Std Dev(%)	1.1471	

Operator's Signature

PALM BRY P. D.
 Intoxilyzer - Alcohol Analyzer
 Model 8000 SN 80-001266
 01/25/2016
 Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	08:41
Control Test	0.080	08:42
Air Blank	0.000	08:42
Control Test	0.081	08:43
Air Blank	0.000	08:43
Control Test	0.080	08:43
Air Blank	0.000	08:44
Control Test Stats		
Average	0.0803	
Std Dev	0.0006	
Rel Std Dev(%)	0.7187	

DGS

Operator's Signature