

INSTRUMENT PROCESSING SHEET

Agency Seminole County S/N 80-001272
 Date In 10/7/16 Date Out 10/25/16 Ship P/U H/D CMI EE

Intake Performed By [Signature]

Registration
 Annual
 Return from CMI
 Return from Enforcement Electronics
 Other _____

Visual Inspection:
OK Case OK Handle
OK Dry Gas Holder OK Feet
OK Keyboard/Plug OK Back/Plugs
OK Screws tight OK Breath Hose

Other Equipment:
 Power cord
 Printer Cable
 Other Static Bag

Notes: _____

Quality Checks Performed By [Signature]

Breath Tube Screen
 Replace O-Rings
 Instrument Set Up Verified
 R-Value 121
 Flow Verification (L/s)
 Flow Column # AP102
 32mm 143 (.139 - .169)
 36mm 164 (.156 - .190)
 53mm 231 (.228 - .278)
 103mm 496 (.447 - .547)

Barometric Pressure Check
 Gauge ID # 28427

Stability Checks

Simulator	Serial #	Lot #/Exp
0.05	SD1018	201507A 7/14/17
0.08	SD1011	201607F 1/20/18
0.20	SD1025	201604C 4/5/18
0.08 DGS	N/A	AC-1019605 7/14/18

Flow Calibration Performed By [Signature]

Flow Calibration N/A
 Flow Calibration Complete
 Flow Column # _____
 5L/min - 47mm
 15L/min - 53mm
 30L/min - 103mm

R-Value _____
 Post Calibration Verification (L/s)
 Flow Column # _____
 32mm _____ (.139 - .169)
 36mm _____ (.156 - .190)
 53mm _____ (.228 - .278)
 103mm _____ (.447 - .547)

Maintenance Performed By [Signature]

Battery Replacement
 Dry Gas Regulator Replacement
 Breath Tube Replacement
 Other _____

Suggested Service

RECEIVED
 OCT 25 2016
 FDLE
 Alcohol Testing Program

Optical Bench Calibration Performed By _____

Optical Bench Calibration N/A
 Optical Bench Calibration Complete

Barometric Pressure Gauge ID # _____

Simulator	Serial Number	Lot Number	Expiration
0.000		N/A	N/A
0.040			
0.100			
0.200			
0.400			
0.080 DGS	N/A		

Post Calibration Stability Checks

Simulator	Serial Number	Lot Number	Expiration
0.05			
0.08			
0.20			
0.08 DGS	N/A		

Department Inspection Performed By [Signature]

Barometric Pressure 1019 Gauge
 ID# 28427 1018 Instrument

Mouth Alcohol Solution Lot # 2016-A
 Acetone Stock Solution Lot # 2016-B

Simulator	Serial Number
0.00	SD1019
Interferent	SD1021
0.05	SD1018
0.08	SD1011
0.20	SD1025

Attachments

Form 41
 Pre-Stability Tests
 Flow Calibration
 Optical Bench Cal
 Post-Stability Tests
 Other _____

Notes: QC by [Signature]

Instrument Complies with Chapter 11D-8, FAC
 Instrument Does Not Comply with Chapter 11D-8, FAC

Return to/Place into Evidentiary Use
 Remain Out of Evidentiary Use

Conduct an Agency Inspection Before Evidentiary Use

Brett Kirkland
 Quality Control Review

10/25/16
 Date

80-001272
 Stability checks
 10/25/16

INTOXILYZER 8000
 Instrument Initialization
 08:46 10/25/2016

SENINOLE COUNTY S.O.
 Intoxilyzer - Alcohol Analyzer
 Model 8000 SN 80-001272
 10/25/2016
 Software: 8100.27

Test	9/210L	Time
Air Blank	0.000	09:04
Control Test	0.051	09:05
Air Blank	0.000	09:05
Control Test	0.050	09:06
Air Blank	0.000	09:06
Control Test	0.051	09:07
Air Blank	0.000	09:08
Control Test Stats		
Average	0.0507	
Std Dev	0.0006	
Rel Std Dev(%)	1.1395	

Operator's Signature

SENINOLE COUNTY S.O.
 Intoxilyzer - Alcohol Analyzer
 Model 8000 SN 80-001272
 10/25/2016
 Software: 8100.27

Test	9/210L	Time
Air Blank	0.000	09:20
Control Test	0.078	09:20
Air Blank	0.000	09:21
Control Test	0.080	09:21
Air Blank	0.000	09:22
Control Test	0.079	09:22
Air Blank	0.000	09:22
Control Test Stats		
Average	0.0790	
Std Dev	0.0010	
Rel Std Dev(%)	1.2658	

DES

Operator's Signature

SENINOLE COUNTY S.O.
 Intoxilyzer - Alcohol Analyzer
 Model 8000 SN 80-001272
 10/25/2016
 Software: 8100.27

Test	9/210L	Time
Air Blank	0.000	09:15
Control Test	0.199	09:16
Air Blank	0.000	09:17
Control Test	0.199	09:17
Air Blank	0.000	09:18
Control Test	0.200	09:18
Air Blank	0.000	09:19
Control Test Stats		
Average	0.1993	
Std Dev	0.0006	
Rel Std Dev(%)	0.2896	

Operator's Signature

SENINOLE COUNTY S.O.
 Intoxilyzer - Alcohol Analyzer
 Model 8000 SN 80-001272
 10/25/2016
 Software: 8100.27

Test	9/210L	Time
Air Blank	0.000	09:23
Control Test	0.079	09:23
Air Blank	0.000	09:24
Control Test	0.080	09:24
Air Blank	0.000	09:25
Control Test	0.080	09:25
Air Blank	0.000	09:26
Control Test Stats		
Average	0.0797	
Std Dev	0.0006	
Rel Std Dev(%)	0.7247	

Operator's Signature

DES
 TSK