



Alcohol Testing Program

INSTRUMENT PROCESSING SHEET

Agency Pinellas County SO

S/N 80-001273

Date In 1/7/16

Date Out 1/20/16

Ship P/U H/D CMI EE

Intake Performed By TP

Registration
 Annual
 Return from CMI
 Return from Enforcement Electronics
 Other _____

Visual Inspection:
OK Case OK Handle
OK Dry Gas Holder OK Feet
OK Keyboard/Plug OK Back/Plugs
OK Screws tight OK Breath Hose

Other Equipment:
 Power cord
 Printer Cable
 Other _____

Notes: _____

Quality Checks Performed By FWB

Breath Tube Screen
 Replace O-Rings
 Instrument Set Up Verified
 R-Value 194
 Flow Verification (L/s)
 Flow Column # ATP 102
 32mm 0.152 (.139 - .169)
 36mm 0.167 (.156 - .190)
 53mm 0.238 (.228 - .278)
 103mm 0.511 (.447 - .547)

Barometric Pressure Check
 Gauge ID # 26932

Stability Checks

| Simulator | Serial # | Lot #/Exp |
|-----------|----------|---------------------|
| 0.05 | SD1024 | 201507A 7/14/17 |
| 0.08 | DR2035 | 201502G 2/24/17 |
| 0.20 | DR3856 | 201505A 5/12/17 |
| 0.08 DGS | N/A | AG511701 4/27/17 |

Flow Calibration Performed By _____

Flow Calibration N/A
 Flow Calibration Complete
 Flow Column # _____
 5L/min - 17mm
 15L/min - 53mm
 30L/min - 103mm
 R-Value _____
 Post Calibration Verification (L/s)
 Flow Column # _____
 32mm _____ (.139 - .169)
 36mm _____ (.156 - .190)
 53mm _____ (.228 - .278)
 103mm _____ (.447 - .547)

Maintenance Performed By _____

Battery Replacement
 Dry Gas Regulator Replacement
 Breath Tube Replacement
 Other _____

Suggested Service

RECEIVED
JAN 20 2016
FDLE
Alcohol Testing Program

Optical Bench Calibration Performed By _____

Optical Bench Calibration N/A
 Optical Bench Calibration Complete

Barometric Pressure Gauge ID # _____

| Simulator | Serial Number | Lot Number | Expiration |
|-----------|---------------|------------|------------|
| 0.000 | | N/A | N/A |
| 0.040 | | | |
| 0.100 | | | |
| 0.200 | | | |
| 0.400 | | | |
| 0.080 DGS | N/A | | |

Post Calibration Stability Checks

| Simulator | Serial Number | Lot Number | Expiration |
|-----------|---------------|------------|------------|
| 0.05 | | | |
| 0.08 | | | |
| 0.20 | | | |
| 0.08 DGS | N/A | | |

Department Inspection Performed By FWB

Barometric Pressure 1034 Gauge
 ID# 26932 1017 Instrument

Mouth Alcohol Solution Lot # 2015-A
 Acetone Stock Solution Lot # 2015-B

| Simulator | Serial Number |
|-------------|---------------|
| 0.00 | SD1016 |
| Interferent | G2840 |
| 0.05 | SD1024 |
| 0.08 | DR2035 |
| 0.20 | DR3856 |

Attachments

Form 41
 Pre-Stability Tests
 Flow Calibration
 Optical Bench Cal
 Post-Stability Tests
 Other _____

Notes: QC # 1/20/16

Instrument Complies with Chapter 11D-8, FAC
 Instrument Does Not Comply with Chapter 11D-8, FAC
 Return to/Place into Evidentiary Use
 Remain Out of Evidentiary Use
 Conduct an Agency Inspection Before Evidentiary Use

T. Brett Kirkland

1/20/16

Quality Control Review

Date

Stability Checks 8D-001273 Pinellas County S.O. 1/20/16 *DMB*

DS

PINELLAS COUNTY SO
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-001273
01/20/2016
Software: 8100.27

| Test | 9/21/0L | Time |
|--------------------|---------|-------|
| Air Blank | 0.000 | 10:57 |
| Control Test | 0.052 | 10:58 |
| Air Blank | 0.000 | 10:58 |
| Control Test | 0.051 | 10:59 |
| Air Blank | 0.000 | 10:59 |
| Control Test | 0.052 | 11:00 |
| Air Blank | 0.000 | 11:01 |
| Control Test Stats | | |
| Average | 0.0517 | |
| Std Dev | 0.0006 | |
| Rel Std Dev(%) | 1.1175 | |

DMB
Operator's Signature

PINELLAS COUNTY SO
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-001273
01/20/2016
Software: 8100.27

| Test | 9/21/0L | Time |
|--------------------|---------|-------|
| Air Blank | 0.000 | 10:46 |
| Control Test | 0.201 | 10:47 |
| Air Blank | 0.000 | 10:47 |
| Control Test | 0.202 | 10:48 |
| Air Blank | 0.000 | 10:48 |
| Control Test | 0.202 | 10:49 |
| Air Blank | 0.000 | 10:50 |
| Control Test Stats | | |
| Average | 0.2017 | |
| Std Dev | 0.0006 | |
| Rel Std Dev(%) | 0.2863 | |

DMB
Operator's Signature

PINELLAS COUNTY SO
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-001273
01/20/2016
Software: 8100.27

| Test | 9/21/0L | Time |
|--------------------|---------|-------|
| Air Blank | 0.000 | 10:52 |
| Control Test | 0.082 | 10:52 |
| Air Blank | 0.000 | 10:53 |
| Control Test | 0.082 | 10:53 |
| Air Blank | 0.000 | 10:54 |
| Control Test | 0.082 | 10:54 |
| Air Blank | 0.000 | 10:55 |
| Control Test Stats | | |
| Average | 0.0820 | |
| Std Dev | 0.0000 | |
| Rel Std Dev(%) | 0.0000 | |

DMB
Operator's Signature

AK