



Alcohol Testing Program

INSTRUMENT PROCESSING SHEET

Agency Jacksonville SO

S/N 80-001278

Date In 8/31/16

Date Out 9/13/16

Ship P/U H/D CMI EE

Intake Performed By [Signature]

Registration
 Annual
 Return from CMI
 Return from Enforcement
 Electronics
 Other _____

Visual Inspection:
 Case OK Handle OK
 Dry Gas Holder OK Feet OK
 Keyboard/Plug OK Back/Plugs OK
 Screws tight OK Breath Hose OK

Other Equipment:
 Power cord
 Printer Cable
 Other Static Bag

Notes: _____

Quality Checks Performed By [Signature]

Breath Tube Screen
 Replace O-Rings
 Instrument Set Up Verified
 R-Value 134
 Flow Verification (L/s)
 Flow Column # ATP102
 32mm .156 (.139 - .169)
 36mm .171 (.156 - .190)
 53mm .242 (.228 - .278)
 103mm .527 (.447 - .547)

Barometric Pressure Check
 Gauge ID # 28427

Stability Checks

Simulator	Serial #	Lot #/Exp
0.05	SD1018	201507A 7/14/17
0.08	SD1011	201601F 1/26/18
0.20	SD1025	201604C 4/5/18
0.08 DGS	N/A	AG612405 5/3/18

Flow Calibration Performed By [Signature]

Flow Calibration N/A
 Flow Calibration Complete
 Flow Column # _____
 5L/min - 4.7mm
 15L/min - 53mm
 30L/min - 103mm

R-Value _____
 Post Calibration Verification (L/s)
 Flow Column # _____
 32mm _____ (.139 - .169)
 36mm _____ (.156 - .190)
 53mm _____ (.228 - .278)
 103mm _____ (.447 - .547)

Maintenance Performed By _____

Battery Replacement
 Dry Gas Regulator Replacement
 Breath Tube Replacement
 Other _____

Suggested Service

RECEIVED
SEP 14 2016
FDLE
Alcohol Testing Program

Optical Bench Calibration Performed By _____

Optical Bench Calibration N/A
 Optical Bench Calibration Complete

Barometric Pressure Gauge ID # _____

Simulator	Serial Number	Lot Number	Expiration
0.000		N/A	N/A
0.040			
0.100			
0.200			
0.400			
0.080 DGS	N/A		

Post Calibration Stability Checks

Simulator	Serial Number	Lot Number	Expiration
0.05			
0.08			
0.20			
0.08 DGS	N/A		

Department Inspection Performed By [Signature]

Barometric Pressure 1015 Gauge
 ID# 28427 1015 Instrument

Mouth Alcohol Solution Lot # 2016-A
 Acetone Stock Solution Lot # 2016-B

Simulator	Serial Number
0.00	SD1019
Interferent	SD1021
0.05	SD1018
0.08	SD1011
0.20	SD1025

Attachments

Form 41
 Pre-Stability Tests
 Flow Calibration
 Optical Bench Cal
 Post-Stability Tests
 Other _____

Notes: [Signature]

Instrument Complies with Chapter 11D-8, FAC
 Instrument Does Not Comply with Chapter 11D-8, FAC

Return to/Place into Evidentiary Use
 Remain Out of Evidentiary Use
 Conduct an Agency Inspection Before Evidentiary Use

Brett Whitland
Quality Control Review

9/14/16
Date

JACKSONVILLE SO
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-001278
09/13/2016
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	09:48
Control Test	0.049	09:49
Air Blank	0.000	09:49
Control Test	0.048	09:50
Air Blank	0.000	09:51
Control Test	0.048	09:51
Air Blank	0.000	09:52
Control Test Stats		
Average	0.0483	
Std Dev	0.0006	
Rel Std Dev(%)	1.1945	


Operator's Signature

JACKSONVILLE SO
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-001278
09/13/2016
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	09:42
Control Test	0.078	09:43
Air Blank	0.000	09:43
Control Test	0.079	09:44
Air Blank	0.000	09:45
Control Test	0.079	09:45
Air Blank	0.000	09:46
Control Test Stats		
Average	0.0787	
Std Dev	0.0006	
Rel Std Dev(%)	0.7339	


Operator's Signature

JACKSONVILLE SO
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-001278
09/13/2016
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	09:54
Control Test	0.198	09:55
Air Blank	0.000	09:55
Control Test	0.198	09:56
Air Blank	0.000	09:57
Control Test	0.198	09:57
Air Blank	0.000	09:58
Control Test Stats		
Average	0.1980	
Std Dev	0.0000	
Rel Std Dev(%)	0.0000	


Operator's Signature

JACKSONVILLE SO
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-001278
09/13/2016
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	10:05
Control Test	0.079	10:05
Air Blank	0.000	10:05
Control Test	0.079	10:06
Air Blank	0.000	10:06
Control Test	0.079	10:07
Air Blank	0.000	10:07
Control Test Stats		
Average	0.0790	
Std Dev	0.0000	
Rel Std Dev(%)	0.0000	

DBS


Operator's Signature

AK
DWS