

INSTRUMENT PROCESSING SHEET

Agency Jacksonville SD S/N 80-001279
 Date In 6/20/16 Date Out 6/27/16 Ship P/U H/D CMI EE

Intake Performed By (Signature)

Registration
 Annual
 Return from CMI
 Return from Enforcement Electronics
 Other _____

Visual Inspection:
 Case ok Handle ok
 Dry Gas Holder ok Feet ok
 Keyboard/Plug ok Back/Plugs ok
 Screws tight ok Breath Hose ok

Other Equipment:
 Power cord
 Printer Cable
 Other Static Bag

Notes: _____

Quality Checks Performed By (Signature)

Breath Tube Screen
 Replace O-Rings
 Instrument Set Up Verified
 R-Value 140
 Flow Verification (L/s)
 Flow Column # ATP102
 32mm .160 (.139 - .169)
 36mm .171 (.156 - .190)
 53mm .234 (.228 - .278)
 103mm .519 (.447 - .547)

Barometric Pressure Check
 Gauge ID # 28427

Stability Checks

| Simulator | Serial # | Lot #/Exp |
|-----------|----------|---------------------|
| 0.05 | SD 3962 | 201507A 7/14/17 |
| 0.08 | SD 3964 | 201601F 01/26/18 |
| 0.20 | G4444 | 201505A 5/12/17 |
| 0.08 DGS | N/A | AG605301 2/22/18 |

Flow Calibration Performed By _____

Flow Calibration N/A
 Flow Calibration Complete
 Flow Column # _____
 5L/min - 19mm
 15L/min - 53mm
 30L/min - 103mm

R-Value _____
 Post Calibration Verification (L/s)
 Flow Column # _____
 32mm _____ (.139 - .169)
 36mm _____ (.156 - .190)
 53mm _____ (.228 - .278)
 103mm _____ (.447 - .547)

Maintenance Performed By _____

Battery Replacement
 Dry Gas Regulator Replacement
 Breath Tube Replacement
 Other _____

Suggested Service

RECEIVED
 JUN 28 2016
 FDLE
 Alcohol Testing Program

Optical Bench Calibration Performed By _____

Optical Bench Calibration N/A
 Optical Bench Calibration Complete

Barometric Pressure Gauge ID # _____

| Simulator | Serial Number | Lot Number | Expiration |
|-----------|---------------|------------|------------|
| 0.000 | | N/A | N/A |
| 0.040 | | | |
| 0.100 | | | |
| 0.200 | | | |
| 0.400 | | | |
| 0.080 DGS | N/A | | |

Post Calibration Stability Checks

| Simulator | Serial Number | Lot Number | Expiration |
|-----------|---------------|------------|------------|
| 0.05 | | | |
| 0.08 | | | |
| 0.20 | | | |
| 0.08 DGS | N/A | | |

Department Inspection Performed By (Signature)

Barometric Pressure 1020 Gauge
 ID# 28427 1017 Instrument

Mouth Alcohol Solution Lot # 2016-A
 Acetone Stock Solution Lot # 2016-B

| Simulator | Serial Number |
|-------------|---------------|
| 0.00 | G2880 |
| Interferent | G2834 |
| 0.05 | SD3962 |
| 0.08 | SD3964 |
| 0.20 | G4444 |

Attachments

| | |
|---|---|
| <input checked="" type="checkbox"/> Form 41 | <input type="checkbox"/> Optical Bench Cal |
| <input checked="" type="checkbox"/> Pre-Stability Tests | <input type="checkbox"/> Post-Stability Tests |
| <input type="checkbox"/> Flow Calibration | <input type="checkbox"/> Other _____ |

Notes: QC (Signature) 6/27/16

Instrument Complies with Chapter 11D-8, FAC
 Instrument Does Not Comply with Chapter 11D-8, FAC

Return to/Place Into Evidentiary Use
 Remain Out of Evidentiary Use

Conduct an Agency Inspection Before Evidentiary Use

Brett Kirkland
 Quality Control Review

6/28/16
 Date

JACKSONVILLE SO
 Intoxilyzer - Alcohol Analyzer
 Model 8000 SN 80-001279
 06/24/2016
 Software: 8100.27

| Test | g/210L | Time |
|--------------------|--------|-------|
| Air Blank | 0.000 | 10:45 |
| Control Test | 0.047 | 10:45 |
| Air Blank | 0.000 | 10:46 |
| Control Test | 0.048 | 10:47 |
| Air Blank | 0.000 | 10:47 |
| Control Test | 0.049 | 10:48 |
| Air Blank | 0.000 | 10:48 |
| Control Test Stats | | |
| Average | 0.0480 | |
| Std Dev | 0.0010 | |
| Rel Std Dev(%) | 2.0833 | |

P Murphy

 Operator's Signature

JACKSONVILLE SO
 Intoxilyzer - Alcohol Analyzer
 Model 8000 SN 80-001279
 06/24/2016
 Software: 8100.27

| Test | g/210L | Time |
|--------------------|--------|-------|
| Air Blank | 0.000 | 10:52 |
| Control Test | 0.077 | 10:53 |
| Air Blank | 0.000 | 10:53 |
| Control Test | 0.078 | 10:54 |
| Air Blank | 0.000 | 10:54 |
| Control Test | 0.078 | 10:55 |
| Air Blank | 0.000 | 10:55 |
| Control Test Stats | | |
| Average | 0.0777 | |
| Std Dev | 0.0006 | |
| Rel Std Dev(%) | 0.7434 | |

P Murphy

 Operator's Signature

JACKSONVILLE SO
 Intoxilyzer - Alcohol Analyzer
 Model 8000 SN 80-001279
 06/24/2016
 Software: 8100.27

| Test | g/210L | Time |
|--------------------|--------|-------|
| Air Blank | 0.000 | 10:57 |
| Control Test | 0.200 | 10:58 |
| Air Blank | 0.000 | 10:59 |
| Control Test | 0.200 | 10:59 |
| Air Blank | 0.000 | 11:00 |
| Control Test | 0.200 | 11:01 |
| Air Blank | 0.000 | 11:01 |
| Control Test Stats | | |
| Average | 0.2000 | |
| Std Dev | 0.0000 | |
| Rel Std Dev(%) | 0.0000 | |

P Murphy

 Operator's Signature

JACKSONVILLE SO
 Intoxilyzer - Alcohol Analyzer
 Model 8000 SN 80-001279
 06/24/2016
 Software: 8100.27

| Test | g/210L | Time |
|--------------------|--------|-------|
| Air Blank | 0.000 | 11:05 |
| Control Test | 0.080 | 11:05 |
| Air Blank | 0.000 | 11:06 |
| Control Test | 0.081 | 11:06 |
| Air Blank | 0.000 | 11:07 |
| Control Test | 0.081 | 11:07 |
| Air Blank | 0.000 | 11:07 |
| Control Test Stats | | |
| Average | 0.0807 | |
| Std Dev | 0.0006 | |
| Rel Std Dev(%) | 0.7157 | |

BK

D65

P Murphy

 Operator's Signature