



Alcohol Testing Program

INSTRUMENT PROCESSING SHEET

Agency Jacksonville SO S/N 80-001280
 Date In 4/27/16 Date Out 5/28/16 Ship P/U H/D CMI EE

Intake Performed By TR

Registration
 Annual
 Return from CMI
 Return from Enforcement
 Electronics
 Other _____

Visual Inspection:
OK Case OK Handle
OK Dry Gas Holder OK Feet
OK Keyboard/Plug OK Back/Plugs
OK Screws tight OK Breath Hose

Other Equipment:
 Power cord
 Printer Cable
 Other Static Bag

Notes: _____

Quality Checks Performed By YDGM

Breath Tube Screen
 Replace O-Rings
 Instrument Set Up Verified
 R-Value 121
 Flow Verification (L/s)
 Flow Column # ATP103
 32mm 144 (.139 - .169)
 36mm 164 (.156 - .190)
 53mm 234 (.228 - .278)
 103mm 500 (.447 - .547)

Barometric Pressure Check
 Gauge ID # 28427

Stability Checks

Simulator	Serial #	Lot #/Exp
0.05	SD1018	201507A 7/14/17
0.08	SD1011	201601F 1/26/18
0.20	G4444	201505A 5/12/17
0.08 DGS	N/A	AG605301 2/22/18

Flow Calibration Performed By _____

Flow Calibration N/A
 Flow Calibration Complete
 Flow Column # _____
 5L/min - 17mm
 15L/min - 150mm
 30L/min - 103mm

R-Value _____
 Post Calibration Verification (L/s)
 Flow Column # _____
 32mm _____ (.139 - .169)
 36mm _____ (.156 - .190)
 53mm _____ (.228 - .278)
 103mm _____ (.447 - .547)

Maintenance Performed By _____

Battery Replacement
 Dry Gas Regulator Replacement
 Breath Tube Replacement
 Other _____

Suggested Service

RECEIVED
MAY 02 2016
FDLE
Alcohol Testing Program

Optical Bench Calibration Performed By _____

Optical Bench Calibration N/A
 Optical Bench Calibration Complete
 Barometric Pressure Gauge ID # _____

Simulator	Serial Number	Lot Number	Expiration
0.000		N/A	N/A
0.040			
0.100			
0.200			
0.400			
0.080 DGS	N/A		

Post Calibration Stability Checks

Simulator	Serial Number	Lot Number	Expiration
0.05			
0.08			
0.20			
0.08 DGS	N/A		

Department Inspection Performed By YDGM

Barometric Pressure 1012 Gauge
 ID# 28427 1010 Instrument

Mouth Alcohol Solution Lot # 2016-A
 Acetone Stock Solution Lot # 2016-B

Simulator	Serial Number
0.00	SD1022
Interferent	SD1021
0.05	SD1018
0.08	SD1011
0.20	G4444

Attachments

Form 41 Optical Bench Cal
 Pre-Stability Tests Post-Stability Tests
 Flow Calibration Other _____

Notes: QC-PWS

Instrument Complies with Chapter 11D-8, FAC
 Instrument Does Not Comply with Chapter 11D-8, FAC

Return to/Place into Evidentiary Use
 Remain Out of Evidentiary Use
 Conduct an Agency Inspection Before Evidentiary Use

Brett Kirkland
Quality Control Review

5/2/16
Date

JACKSONVILLE SO
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-001280
04/27/2016
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	10:58
Control Test	0.049	10:59
Air Blank	0.000	10:59
Control Test	0.049	11:00
Air Blank	0.000	11:01
Control Test	0.050	11:01
Air Blank	0.000	11:02
Control Test Stats		
Average	0.0493	
Std Dev	0.0006	
Rel Std Dev(%)	1.1703	

JACKSONVILLE SO
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-001280
04/27/2016
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	11:17
Control Test	0.200	11:17
Air Blank	0.000	11:18
Control Test	0.201	11:19
Air Blank	0.000	11:19
Control Test	0.201	11:20
Air Blank	0.000	11:20
Control Test Stats		
Average	0.2007	
Std Dev	0.0006	
Rel Std Dev(%)	0.2877	

JACKSONVILLE SO
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-001280
04/27/2016
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	11:06
Control Test	0.078	11:06
Air Blank	0.000	11:07
Control Test	0.079	11:08
Air Blank	0.000	11:08
Control Test	0.079	11:09
Air Blank	0.000	11:09
Control Test Stats		
Average	0.0787	
Std Dev	0.0006	
Rel Std Dev(%)	0.7339	

P. Murphy
Operator's Signature

P. Murphy
Operator's Signature

P. Murphy
Operator's Signature

BSK
PWS

JACKSONVILLE SO
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-001280
04/27/2016
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	11:23
Control Test	0.080	11:23
Air Blank	0.000	11:24
Control Test	0.081	11:24
Air Blank	0.000	11:25
Control Test	0.081	11:25
Air Blank	0.000	11:25
Control Test Stats		
Average	0.0807	
Std Dev	0.0006	
Rel Std Dev(%)	0.7157	

DGS

P. Murphy
Operator's Signature