



Alcohol Testing Program

INSTRUMENT PROCESSING SHEET

Agency UF PD

S/N 80-001282

Date In 11/9/16

Date Out 11/16/16

Ship  P/U  H/D  CMI  EE

**Intake** Performed By [Signature]

Registration  
 Annual  
 Return from CMI  
 Return from Enforcement  
 Electronics  
 Other \_\_\_\_\_

Visual Inspection:  
 Case  Handle  
 Dry Gas Holder  Feet  
 Keyboard/Plug  Back/Plugs  
 Screws tight  Breath Hose

Other Equipment:  
 Power cord  
 Printer Cable  
 Other Static Bag x2

Notes: \_\_\_\_\_

**Quality Checks** Performed By [Signature]

Breath Tube Screen  
 Replace O-Rings  
 Instrument Set Up Verified  
 R-Value 155  
 Flow Verification (L/s)  
 Flow Column # AP102  
 32mm 149 (.139 - .169)  
 36mm 170 (.156 - .190)  
 53mm 239 (.228 - .278)  
 103mm 499 (.447 - .547)

Barometric Pressure Check  
 Gauge ID # 28427

Stability Checks

Simulator	Serial #	Lot #/Exp
0.05	SD1016	201507A 7/14/18
0.08	SD1011	201601F 1/26/18
0.20	SD1025	201604C 4/15/18
0.08 DGS	N/A	AG626605 9/22/18

**Flow Calibration** Performed By [Signature]

Flow Calibration N/A  
 Flow Calibration Complete  
 Flow Column # \_\_\_\_\_  
 5L/min - 17mm  
 15L/min - 59mm  
 30L/min - 103mm  
 R-Value \_\_\_\_\_  
 Post Calibration Verification (L/s)  
 Flow Column # \_\_\_\_\_  
 32mm \_\_\_\_\_ (.139 - .169)  
 36mm \_\_\_\_\_ (.156 - .190)  
 53mm \_\_\_\_\_ (.228 - .278)  
 103mm \_\_\_\_\_ (.447 - .547)

**Maintenance** Performed By \_\_\_\_\_

Battery Replacement  
 Dry Gas Regulator Replacement  
 Breath Tube Replacement  
 Other \_\_\_\_\_

**Suggested Service**

RECEIVED  
NOV 16 2016  
FDLE  
Alcohol Testing Program

**Optical Bench Calibration** Performed By \_\_\_\_\_

Optical Bench Calibration N/A  
 Optical Bench Calibration Complete  
 Barometric Pressure Gauge ID # \_\_\_\_\_

Simulator	Serial Number	Lot Number	Expiration
0.000		N/A	N/A
0.040			
0.100			
0.200			
0.400			
0.080 DGS	N/A		

Post Calibration Stability Checks

Simulator	Serial Number	Lot Number	Expiration
0.05			
0.08			
0.20			
0.08 DGS	N/A		

**Department Inspection** Performed By [Signature]

Barometric Pressure Gauge ID# 28427 Gauge # 1015 Instrument # 1015

Mouth Alcohol Solution Lot # 2016-A  
 Acetone Stock Solution Lot # 2016-B

Simulator	Serial Number
0.00	SD1019
Interferent	SD1021
0.05	SD1018
0.08	SD1011
0.20	SD1025

**Attachments**

Form 41  
 Pre-Stability Tests  
 Flow Calibration  
 Optical Bench Cal  
 Post-Stability Tests  
 Other \_\_\_\_\_

Notes: QA/OK OK [Signature] 11/15/2016

Instrument Complies with Chapter 11D-8, FAC  
 Instrument Does Not Comply with Chapter 11D-8, FAC  
 Return to/Place into Evidentiary Use  
 Remain Out of Evidentiary Use  
 Conduct an Agency Inspection Before Evidentiary Use

Brett Hankland  
Quality Control Review

11/16/16  
Date

80-001282  
 Stability checks  
 11/15/16

INTOXILYZER 8000  
 Instrument Initialization  
 10:54 11/15/2016

UNIVERSITY OF FL PD  
 Intoxilyzer - Alcohol Analyzer  
 Model 8000 SN 80-001282  
 11/15/2016  
 Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	13:08
Control Test	0.201	13:09
Air Blank	0.000	13:09
Control Test	0.204	13:10
Air Blank	0.000	13:11
Control Test	0.204	13:11
Air Blank	0.000	13:12
Control Test Stats		
Average	0.2030	
Std Dev	0.0017	
Rel Std Dev(%)	0.8532	

Operator's Signature

UNIVERSITY OF FL PD  
 Intoxilyzer - Alcohol Analyzer  
 Model 8000 SN 80-001282  
 11/15/2016  
 Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	12:57
Control Test	0.049	12:58
Air Blank	0.000	12:59
Control Test	0.051	12:59
Air Blank	0.000	13:00
Control Test	0.050	13:01
Air Blank	0.000	13:01
Control Test Stats		
Average	0.0500	
Std Dev	0.0010	
Rel Std Dev(%)	2.0000	

Operator's Signature

UNIVERSITY OF FL PD  
 Intoxilyzer - Alcohol Analyzer  
 Model 8000 SN 80-001282  
 11/15/2016  
 Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	13:13
Control Test	0.078	13:13
Air Blank	0.000	13:14
Control Test	0.079	13:14
Air Blank	0.000	13:15
Control Test	0.078	13:15
Air Blank	0.000	13:15
Control Test Stats		
Average	0.0783	
Std Dev	0.0006	
Rel Std Dev(%)	0.7370	

Operator's Signature

BSK

UNIVERSITY OF FL PD  
 Intoxilyzer - Alcohol Analyzer  
 Model 8000 SN 80-001282  
 11/15/2016  
 Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	12:51
Control Test	0.078	12:52
Air Blank	0.000	12:52
Control Test	0.079	12:52
Air Blank	0.000	12:53
Control Test	0.079	12:53
Air Blank	0.000	12:54
Control Test Stats		
Average	0.0787	
Std Dev	0.0006	
Rel Std Dev(%)	0.7339	

BSK

Operator's Signature