

## INSTRUMENT PROCESSING SHEET

Agency Baker County SO S/N 80-001286  
 Date In 6/16/16 Date Out 10/14/16  Ship  P/U  H/D  CMI  EE

**Intake** Performed By [Signature]

Registration  
 Annual  
 Return from CMI  
 Return from Enforcement  
 Electronics  
 Other \_\_\_\_\_

Visual Inspection:  
ok Case ok Handle  
ok Dry Gas Holder ok Feet  
ok Keyboard/Plug ok Back/Plugs  
ok Screws tight ok Breath Hose

Other Equipment:  
 Power cord  
 Printer Cable  
 Other Static Bag

Notes: \_\_\_\_\_

**Quality Checks** Performed By [Signature]

Breath Tube Screen  
 Replace O-Rings  
 Instrument Set Up Verified  
 R-Value 189  
 Flow Verification (L/s)  
 Flow Column # ATP102  
 32mm 148 (.139 - .169)  
 36mm 164 (.156 - .190)  
 53mm 238 (.228 - .278)  
 103mm 488 (.447 - .547)

Barometric Pressure Check  
 Gauge ID # 28427

Stability Checks

Simulator	Serial #	Lot #/Exp
0.05	SD1018	201507A 7/14/17
0.08	SD1011	201601F 1/26/18
0.20	SD1025	201604C 4/5/18
0.08 DGS	N/A	RG619605 7/14/18

**Flow Calibration** Performed By \_\_\_\_\_

Flow Calibration N/A  
 Flow Calibration Complete  
 Flow Column # \_\_\_\_\_  
 5L/min - 17mm  
 15L/min - 9mm  
 30L/min - 103mm

R-Value \_\_\_\_\_  
 Post Calibration Verification (Program)  
 Flow Column # \_\_\_\_\_  
 32mm \_\_\_\_\_ (.139 - .169)  
 36mm \_\_\_\_\_ (.156 - .190)  
 53mm \_\_\_\_\_ (.228 - .278)  
 103mm \_\_\_\_\_ (.447 - .547)

**Maintenance** Performed By \_\_\_\_\_

Battery Replacement  
 Dry Gas Regulator Replacement  
 Breath Tube Replacement  
 Other \_\_\_\_\_

**Suggested Service**

RECEIVED  
 OCT 14 2016  
 FDLE  
 Alcohol Testing Program

**Optical Bench Calibration** Performed By \_\_\_\_\_

Optical Bench Calibration N/A  
 Optical Bench Calibration Complete

Barometric Pressure Gauge ID # \_\_\_\_\_

Simulator	Serial Number	Lot Number	Expiration
0.000		N/A	N/A
0.040			
0.100			
0.200			
0.400			
0.080 DGS	N/A		

Post Calibration Stability Checks

Simulator	Serial Number	Lot Number	Expiration
0.05			
0.08			
0.20			
0.08 DGS	N/A		

**Department Inspection** Performed By [Signature]

Barometric Pressure 1022 Gauge  
 ID# 28427 1017 Instrument

Mouth Alcohol Solution Lot # 2016-A  
 Acetone Stock Solution Lot # 2016-B

Simulator	Serial Number
0.00	SD1019
Interferent	SD1021
0.05	SD1018
0.08	SD1011
0.20	SD1025

**Attachments**

Form 41  
 Pre-Stability Tests  
 Flow Calibration

Optical Bench Cal  
 Post-Stability Tests  
 Other Form 40

Notes: QC: SP

\_\_\_\_\_

\_\_\_\_\_

\_\_\_\_\_

\_\_\_\_\_

Instrument Complies with Chapter 11D-8, FAC  
 Instrument Does Not Comply with Chapter 11D-8, FAC

Return to/Place into Evidentiary Use  
 Remain Out of Evidentiary Use  
 Conduct an Agency Inspection Before Evidentiary Use

Date 10/14/16

[Signature]  
 Quality Control Review


BAKER COUNTY SO  
Intoxilyzer - Alcohol Analyzer  
Model 8000 SN 80-001286  
10/14/2016  
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	10:16
Control Test	0.050	10:17
Air Blank	0.000	10:17
Control Test	0.050	10:18
Air Blank	0.000	10:18
Control Test	0.049	10:19
Air Blank	0.000	10:20
Control Test Stats		
Average	0.0497	
Std Dev	0.0006	
Rel Std Dev(%)	1.1625	

  
-----  
Operator's Signature

BAKER COUNTY SO  
Intoxilyzer - Alcohol Analyzer  
Model 8000 SN 80-001286  
10/14/2016  
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	10:28
Control Test	0.079	10:29
Air Blank	0.000	10:29
Control Test	0.080	10:30
Air Blank	0.000	10:31
Control Test	0.081	10:31
Air Blank	0.000	10:32
Control Test Stats		
Average	0.0800	
Std Dev	0.0010	
Rel Std Dev(%)	1.2500	

  
-----  
Operator's Signature

BAKER COUNTY SO  
Intoxilyzer - Alcohol Analyzer  
Model 8000 SN 80-001286  
10/14/2016  
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	10:22
Control Test	0.199	10:23
Air Blank	0.000	10:23
Control Test	0.198	10:24
Air Blank	0.000	10:25
Control Test	0.199	10:25
Air Blank	0.000	10:26
Control Test Stats		
Average	0.1987	
Std Dev	0.0006	
Rel Std Dev(%)	0.2906	

  
-----  
Operator's Signature

BAKER COUNTY SO  
Intoxilyzer - Alcohol Analyzer  
Model 8000 SN 80-001286  
10/14/2016  
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	10:34
Control Test	0.079	10:34
Air Blank	0.000	10:35
Control Test	0.079	10:35
Air Blank	0.000	10:36
Control Test	0.080	10:36
Air Blank	0.000	10:37
Control Test Stats		
Average	0.0793	
Std Dev	0.0006	
Rel Std Dev(%)	0.7277	

SP

BK

D65

  
-----  
Operator's Signature



Alcohol Testing Program

INSTRUMENT PROCESSING SHEET

Agency Baker County SO S/N 80-001286

Date In 1/7/16 Date Out 4/28/16 Ship P/U H/D CMI EEE

Intake Performed By [Signature]
Registration
Annual
Return from CMI
Return from Enforcement Electronics
Other
Visual Inspection:
Case Handle
Dry Gas Holder Feet
Keyboard/Plug Back/Plugs
Screws tight Breath Hose
Other Equipment:
Power cord
Printer Cable
Other Static Bag

Quality Checks Performed By [Signature]
Breath Tube Screen
Replace O-Rings
Instrument Set Up Verified
R-Value 184
Flow Verification (L/s)
Flow Column # ATP103
32mm .148 (.139 -.169)
36mm .164 (.156 -.190)
53mm .238 (.228 -.278)
103mm .496 (.447 -.547)
Barometric Pressure Check
Gauge ID # 28662
Stability Checks
Table with Simulator, Serial #, Lot #/Exp

Flow Calibration Performed By [Signature]
Flow Calibration N/A
Flow Calibration Complete
Flow Column #
5L/min - 17mm
15L/min - 53mm
30L/min - 103mm
R-Value
Post Calibration Verification (L/s)
Flow Column #
32mm (.139 -.169)
36mm (.156 -.190)
53mm (.228 -.278)
103mm (.447 -.547)
Maintenance Performed By
Battery Replacement
Dry Gas Regulator Replacement
Breath Tube Replacement
Other
Suggested Service

RECEIVED MAY 02 2016 FDLE Alcohol Testing Program

Optical Bench Calibration Performed By
Optical Bench Calibration N/A
Optical Bench Calibration Complete
Barometric Pressure Gauge ID #
Table with Simulator, Serial Number, Lot Number, Expiration
Post Calibration Stability Checks
Table with Simulator, Serial Number, Lot Number, Expiration

Department Inspection Performed By [Signature]
Barometric Pressure 1024/1012 Gauge
ID# 28662 1020/1007 Instrument
Mouth Alcohol Solution Lot # 2015-A/2016-B
Acetone Stock Solution Lot # 2015-B/2016-B
Table with Simulator, Serial Number
Attachments
Form 41
Pre-Stability Tests
Flow Calibration
Optical Bench Cal
Post-Stability Tests
Other Form 40

Notes: Replaced BREATH TUBE, REINSPECTED
[Signature]

Instrument Complies with Chapter 11D-8, FAC
Instrument Does Not Comply with Chapter 11D-8, FAC
Return to/Place into Evidentiary Use
Remain Out of Evidentiary Use
Conduct an Agency Inspection Before Evidentiary Use

[Signature]
Quality Control Review

5/2/16
Date

# Stability Checks - Instrument # 80-001286 Baker County SO 1/12/2016 80

ASK

BAKER COUNTY SO  
Intoxilyzer - Alcohol Analyzer  
Model 8000 SN 80-001286  
01/12/2016  
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	09:52
Control Test	0.050	09:53
Air Blank	0.000	09:53
Control Test	0.049	09:54
Air Blank	0.000	09:55
Control Test	0.049	09:55
Air Blank	0.000	09:56
Control Test Stats		
Average	0.0493	
Std Dev	0.0006	
Rel Std Dev(%)	1.1703	

Operator's Signature \_\_\_\_\_

BAKER COUNTY SO  
Intoxilyzer - Alcohol Analyzer  
Model 8000 SN 80-001286  
01/12/2016  
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	09:57
Control Test	0.079	09:58
Air Blank	0.000	09:59
Control Test	0.079	09:59
Air Blank	0.000	10:00
Control Test	0.000	10:00
Air Blank	0.000	10:01
Control Test Stats		
Average	0.0793	
Std Dev	0.0006	
Rel Std Dev(%)	0.7277	

Operator's Signature \_\_\_\_\_

BAKER COUNTY SO  
Intoxilyzer - Alcohol Analyzer  
Model 8000 SN 80-001286  
01/12/2016  
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	10:03
Control Test	0.205	10:03
Air Blank	0.000	10:04
Control Test	0.203	10:04
Air Blank	0.000	10:05
Control Test	0.203	10:06
Air Blank	0.000	10:06
Control Test Stats		
Average	0.2037	
Std Dev	0.0012	
Rel Std Dev(%)	0.5670	

Operator's Signature \_\_\_\_\_

BAKER COUNTY SO  
Intoxilyzer - Alcohol Analyzer  
Model 8000 SN 80-001286  
01/12/2016  
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	09:48
Control Test	0.080	09:48
Air Blank	0.000	09:49
Control Test	0.080	09:49
Air Blank	0.000	09:50
Control Test	0.081	09:50
Air Blank	0.000	09:50
Control Test Stats		
Average	0.0803	
Std Dev	0.0006	
Rel Std Dev(%)	0.7187	

DGS

Operator's Signature \_\_\_\_\_

PWS