



Alcohol Testing Program

INSTRUMENT PROCESSING SHEET

Agency Bradford County SO S/N 80-001290
Date In 3/1/16 Date Out 6/27/16 Ship P/U H/D CMI EE

Intake Performed By JP

Registration
 Annual
 Return from CMI
 Return from Enforcement Electronics
 Other _____

Visual Inspection:
OK Case OK Handle
OK Dry Gas Holder OK Feet
OK Keyboard/Plug OK Back/Plugs
OK Screws tight OK Breath Hose

Other Equipment:
 Power cord
 Printer Cable
 Other Static Bag

Notes: _____

Quality Checks Performed By JSM

Breath Tube Screen
 Replace O-Rings
 Instrument Set Up Verified
 R-Value 140
 Flow Verification (L/s)
 Flow Column # ATP102
 32mm .169 (.139 - .169)
 36mm .171 (.156 - .190)
 53mm .246 (.228 - .278)
 103mm .520 (.447 - .547)

Barometric Pressure Check
 Gauge ID # 28427

Stability Checks

Simulator	Serial #	Lot #/Exp
0.05	5D3962	201507A 7/14/17
0.08	5D3964	201601A 1/25/18
0.20	64444	201505A 5/12/17
0.08 DGS	N/A	AG605301 2/22/18

Flow Calibration Performed By JSM

Flow Calibration N/A
 Flow Calibration Complete
 Flow Column # _____
 5L/min - 17mm
 15L/min - 53mm
 30L/min - 103mm
 R-Value _____
 Post Calibration Verification (L/s)
 Flow Column # _____
 32mm _____ (.139 - .169)
 36mm _____ (.156 - .190)
 53mm _____ (.228 - .278)
 103mm _____ (.447 - .547)

Maintenance Performed By JSM

Battery Replacement
 Dry Gas Regulator Replacement
 Breath Tube Replacement
 Other _____

Suggested Service

RECEIVED
JUN 28 2016
FDLE
Alcohol Testing Program

Optical Bench Calibration Performed By _____

Optical Bench Calibration N/A
 Optical Bench Calibration Complete

Barometric Pressure Gauge ID # _____

Simulator	Serial Number	Lot Number	Expiration
0.000		N/A	N/A
0.040			
0.100			
0.200			
0.400			
0.080 DGS	N/A		

Post Calibration Stability Checks

Simulator	Serial Number	Lot Number	Expiration
0.05			
0.08			
0.20			
0.08 DGS	N/A		

Department Inspection Performed By Dem

Barometric Pressure 1020 Gauge
 ID# 28427 1012 Instrument

Mouth Alcohol Solution Lot # 2016-A
 Acetone Stock Solution Lot # 2016-B

Simulator	Serial Number
0.00	62880
Interferent	62834
0.05	5D3962
0.08	5D3964
0.20	64444

Attachments

Form 41 Optical Bench Cal
 Pre-Stability Tests Post-Stability Tests
 Flow Calibration Other FORM 40

Notes: Came in with temp res failure
Needs new breath tube
Replaced BREATH TUBE

Instrument Complies with Chapter 11D-8, FAC
 Instrument Does Not Comply with Chapter 11D-8, FAC

Return to/Place Into Evidentiary Use
 Remain Out of Evidentiary Use

Conduct an Agency Inspection Before Evidentiary Use

JC JP 6/27/16 Britt Kirkland
Quality Control Review

6/28/16
Date

BRADFORD COUNTY SO
 Intoxilyzer - Alcohol Analyzer
 Model 8000 SN 80-001290
 06/24/2016
 Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	10:59
Control Test	0.050	11:00
Air Blank	0.000	11:00
Control Test	0.050	11:01
Air Blank	0.000	11:01
Control Test	0.049	11:02
Air Blank	0.000	11:03
Control Test Stats		
Average	0.0497	
Std Dev	0.0006	
Rel Std Dev(%)	1.1625	

✓

P Murphy
 Operator's Signature

BRADFORD COUNTY SO
 Intoxilyzer - Alcohol Analyzer
 Model 8000 SN 80-001290
 06/24/2016
 Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	11:04
Control Test	0.079	11:05
Air Blank	0.000	11:05
Control Test	0.080	11:06
Air Blank	0.000	11:06
Control Test	0.080	11:07
Air Blank	0.000	11:08
Control Test Stats		
Average	0.0797	
Std Dev	0.0006	
Rel Std Dev(%)	0.7247	

✓

P Murphy
 Operator's Signature

BRADFORD COUNTY SO
 Intoxilyzer - Alcohol Analyzer
 Model 8000 SN 80-001290
 06/24/2016
 Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	10:52
Control Test	0.199	10:53
Air Blank	0.000	10:54
Control Test	0.200	10:54
Air Blank	0.000	10:55
Control Test	0.199	10:56
Air Blank	0.000	10:56
Control Test Stats		
Average	0.1993	
Std Dev	0.0006	
Rel Std Dev(%)	0.2896	

P Murphy
 Operator's Signature

BRADFORD COUNTY SO
 Intoxilyzer - Alcohol Analyzer
 Model 8000 SN 80-001290
 06/24/2016
 Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	11:20
Control Test	0.079	11:21
Air Blank	0.000	11:21
Control Test	0.079	11:21
Air Blank	0.000	11:22
Control Test	0.079	11:22
Air Blank	0.000	11:23
Control Test Stats		
Average	0.0790	
Std Dev	0.0000	
Rel Std Dev(%)	0.0000	

PK

DGS

P Murphy
 Operator's Signature



Alcohol Testing Program

INSTRUMENT PROCESSING SHEET

Agency Bradford County SO S/N 90-00290
Date In 2/19/16 Date Out 2/22/2016 Ship P/U H/D CMI EE

Intake Performed By <u>CTV</u> <input type="checkbox"/> Registration <input checked="" type="checkbox"/> Annual <input type="checkbox"/> Return from CMI <input type="checkbox"/> Return from Enforcement Electronics <input type="checkbox"/> Other _____ Visual Inspection: <u>OK</u> Case <u>OK</u> Handle <u>OK</u> Dry Gas Holder <u>OK</u> Feet <u>OK</u> Keyboard/Plug <u>OK</u> Back/Plugs <u>OK</u> Screws tight <u>OK</u> Breath Hose Other Equipment: <input checked="" type="checkbox"/> Power cord <input type="checkbox"/> Printer Cable <input checked="" type="checkbox"/> Other <u>Static Bag</u> Notes: _____ _____ _____	Quality Checks Performed By <u>JS</u> <input checked="" type="checkbox"/> Breath Tube Screen <input checked="" type="checkbox"/> Replace O-Rings <input checked="" type="checkbox"/> Instrument Set Up Verified <input checked="" type="checkbox"/> R-Value <u>145</u> <input checked="" type="checkbox"/> Flow Verification (L/s) Flow Column # <u>AP102</u> 32mm <u>.144</u> (.139 - .169) 36mm <u>.160</u> (.156 - .190) 53mm <u>.230</u> (.228 - .278) 103mm <u>.500</u> (.447 - .547) <input checked="" type="checkbox"/> Barometric Pressure Check Gauge ID # <u>28427</u> <input checked="" type="checkbox"/> Stability Checks <table border="1"> <thead> <tr> <th>Simulator</th> <th>Serial #</th> <th>Lot #/Exp</th> </tr> </thead> <tbody> <tr> <td>0.05</td> <td>SD1018</td> <td>201501A 07/14/2017</td> </tr> <tr> <td>0.08</td> <td>SD1011</td> <td>201502G 02/24/2017</td> </tr> <tr> <td>0.20</td> <td>SD3933</td> <td>201505A 05/12/2017</td> </tr> <tr> <td>0.08 DGS</td> <td>N/A</td> <td>A6511701 01/27/2017</td> </tr> </tbody> </table>	Simulator	Serial #	Lot #/Exp	0.05	SD1018	201501A 07/14/2017	0.08	SD1011	201502G 02/24/2017	0.20	SD3933	201505A 05/12/2017	0.08 DGS	N/A	A6511701 01/27/2017	Flow Calibration Performed By _____ <input checked="" type="checkbox"/> Flow Calibration N/A <input type="checkbox"/> Flow Calibration Complete Flow Column # _____ <input type="checkbox"/> 5L/min - 17mm <input type="checkbox"/> 15L/min - 53mm <input type="checkbox"/> 30L/min - 103mm <input type="checkbox"/> R-Value _____ <input type="checkbox"/> Post Calibration Verification (L/s) Flow Column # _____ 32mm _____ (.139 - .169) 36mm _____ (.156 - .190) 53mm _____ (.228 - .278) 103mm _____ (.447 - .547) Maintenance Performed By _____ <input type="checkbox"/> Battery Replacement <input type="checkbox"/> Dry Gas Regulator Replacement <input type="checkbox"/> Breath Tube Replacement <input type="checkbox"/> Other _____ Suggested Service _____ _____
Simulator	Serial #	Lot #/Exp															
0.05	SD1018	201501A 07/14/2017															
0.08	SD1011	201502G 02/24/2017															
0.20	SD3933	201505A 05/12/2017															
0.08 DGS	N/A	A6511701 01/27/2017															

RECEIVED
FEB 23 2016
FDLE
Alcohol Testing Program

Optical Bench Calibration Performed By _____ <input checked="" type="checkbox"/> Optical Bench Calibration N/A <input type="checkbox"/> Optical Bench Calibration Complete Barometric Pressure Gauge _____ ID # _____ <table border="1"> <thead> <tr> <th>Simulator</th> <th>Serial Number</th> <th>Lot Number</th> <th>Expiration</th> </tr> </thead> <tbody> <tr> <td>0.000</td> <td></td> <td>N/A</td> <td>N/A</td> </tr> <tr> <td>0.040</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.100</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.200</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.400</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.080 DGS</td> <td>N/A</td> <td></td> <td></td> </tr> </tbody> </table> <input type="checkbox"/> Post Calibration Stability Checks <table border="1"> <thead> <tr> <th>Simulator</th> <th>Serial Number</th> <th>Lot Number</th> <th>Expiration</th> </tr> </thead> <tbody> <tr> <td>0.05</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.08</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.20</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.08 DGS</td> <td>N/A</td> <td></td> <td></td> </tr> </tbody> </table>	Simulator	Serial Number	Lot Number	Expiration	0.000		N/A	N/A	0.040				0.100				0.200				0.400				0.080 DGS	N/A			Simulator	Serial Number	Lot Number	Expiration	0.05				0.08				0.20				0.08 DGS	N/A			Department Inspection Performed By <u>JS</u> <input checked="" type="checkbox"/> Barometric Pressure <u>1014</u> Gauge ID# <u>28427</u> <u>1014</u> Instrument Mouth Alcohol Solution Lot # <u>2015-A</u> Acetone Stock Solution Lot # <u>2015-B</u> <table border="1"> <thead> <tr> <th>Simulator</th> <th>Serial Number</th> </tr> </thead> <tbody> <tr> <td>0.00</td> <td>SD1022</td> </tr> <tr> <td>Interferent</td> <td>SD1021</td> </tr> <tr> <td>0.05</td> <td>SD1018</td> </tr> <tr> <td>0.08</td> <td>SD1011</td> </tr> <tr> <td>0.20</td> <td>SD3933</td> </tr> </tbody> </table>	Simulator	Serial Number	0.00	SD1022	Interferent	SD1021	0.05	SD1018	0.08	SD1011	0.20	SD3933
Simulator	Serial Number	Lot Number	Expiration																																																										
0.000		N/A	N/A																																																										
0.040																																																													
0.100																																																													
0.200																																																													
0.400																																																													
0.080 DGS	N/A																																																												
Simulator	Serial Number	Lot Number	Expiration																																																										
0.05																																																													
0.08																																																													
0.20																																																													
0.08 DGS	N/A																																																												
Simulator	Serial Number																																																												
0.00	SD1022																																																												
Interferent	SD1021																																																												
0.05	SD1018																																																												
0.08	SD1011																																																												
0.20	SD3933																																																												

Attachments <input checked="" type="checkbox"/> Form 41 <input checked="" type="checkbox"/> Pre-Stability Tests <input type="checkbox"/> Flow Calibration	<input type="checkbox"/> Optical Bench Cal <input type="checkbox"/> Post-Stability Tests <input type="checkbox"/> Other _____
---	---

Notes: OK - PWS

 Instrument Complies with Chapter 11D-8, FAC
 Instrument Does Not Comply with Chapter 11D-8, FAC
 Return to/Place into Evidentiary Use
 Remain Out of Evidentiary Use
 Conduct an Agency Inspection Before Evidentiary Use

Patrick Murphy
Quality Control Review

2/23/16
Date

Stability checks - Instrument # 80-001290 Bradford County SO 2/22/2016

OPM

Plus

BRADFORD COUNTY SO
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-001290
02/22/2016
Software: 8100.27

BRADFORD COUNTY SO
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-001290
02/22/2016
Software: 8100.27

BRADFORD COUNTY SO
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-001290
02/22/2016
Software: 8100.27

BRADFORD COUNTY SO
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-001290
02/22/2016
Software: 8100.27

Test	9/21/0L	Time
Air Blank	0.000	12:27
Control Test	0.051	12:28
Air Blank	0.000	12:29
Control Test	0.051	12:29
Air Blank	0.000	12:30
Control Test	0.051	12:31
Air Blank	0.000	12:31
Control Test Stats		
Average	0.0510	
Std Dev	0.0000	
Rel Std Dev(%)	0.0000	

Test	9/21/0L	Time
Air Blank	0.000	12:32
Control Test	0.081	12:33
Air Blank	0.000	12:33
Control Test	0.081	12:34
Air Blank	0.000	12:35
Control Test	0.081	12:35
Air Blank	0.000	12:36
Control Test Stats		
Average	0.0810	
Std Dev	0.0000	
Rel Std Dev(%)	0.0000	

Test	9/21/0L	Time
Air Blank	0.000	12:37
Control Test	0.203	12:38
Air Blank	0.000	12:38
Control Test	0.202	12:39
Air Blank	0.000	12:40
Control Test	0.202	12:40
Air Blank	0.000	12:41
Control Test Stats		
Average	0.2023	
Std Dev	0.0006	
Rel Std Dev(%)	0.2853	

Test	9/21/0L	Time
Air Blank	0.000	12:24
Control Test	0.079	12:25
Air Blank	0.000	12:25
Control Test	0.079	12:25
Air Blank	0.000	12:26
Control Test	0.079	12:26
Air Blank	0.000	12:27
Control Test Stats		
Average	0.0790	
Std Dev	0.0000	
Rel Std Dev(%)	0.0000	

DES

Operator's Signature

Operator's Signature

Operator's Signature

Operator's Signature