

INSTRUMENT PROCESSING SHEET

Agency Bradford County SO

S/N 80-00291

Date In 12/28/15

Date Out 1/6/2016

Ship P/U H/D CMI EE

Intake Performed By TR

Registration
 Annual
 Return from CMI
 Return from Enforcement Electronics
 Other _____

Visual Inspection:
OK Case OK Handle
OK Dry Gas Holder OK Feet
OK Keyboard/Plug OK Back/Plugs
OK Screws tight OK Breath Hose

Other Equipment:
 Power cord
 Printer Cable
 Other Static Bag

Notes: _____

Quality Checks Performed By TR

Breath Tube Screen
 Replace O-Rings
 Instrument Set Up Verified
 R-Value 138
 Flow Verification (L/s)
 Flow Column # ATP103
 32mm .144 (.139 - .169)
 36mm .160 (.156 - .190)
 53mm .234 (.228 - .278)
 103mm .503 (.447 - .547)

Barometric Pressure Check
 Gauge ID # 28662

Stability Checks

Simulator	Serial #	Lot #/Exp
0.05	SD1018	201507A 07/14/2017
0.08	SD1011	201502G 02/27/2017
0.20	SD1025	201505A 05/12/2017
0.08 DGS	N/A	AG511 701 04/27/2017

Flow Calibration Performed By _____

Flow Calibration N/A
 Flow Calibration Complete
 Flow Column # _____
 5L/min - 17mm
 15L/min - 53mm
 30L/min - 103mm

R-Value _____
 Post Calibration Verification (L/s)
 Flow Column # _____
 32mm _____ (.139 - .169)
 36mm _____ (.156 - .190)
 53mm _____ (.228 - .278)
 103mm _____ (.447 - .547)

Maintenance Performed By _____

Battery Replacement
 Dry Gas Regulator Replacement
 Breath Tube Replacement
 Other _____

Suggested Service

RECEIVED
 JAN 06 2016
 FDLE
 Alcohol Testing Program

Optical Bench Calibration Performed By _____

Optical Bench Calibration N/A
 Optical Bench Calibration Complete

Barometric Pressure Gauge ID # _____

Simulator	Serial Number	Lot Number	Expiration
0.000		N/A	N/A
0.040			
0.100			
0.200			
0.400			
0.080 DGS	N/A		

Post Calibration Stability Checks

Simulator	Serial Number	Lot Number	Expiration
0.05			
0.08			
0.20			
0.08 DGS	N/A		

Department Inspection Performed By TR

Barometric Pressure 1027 Gauge
 ID# 28662 1023 Instrument

Mouth Alcohol Solution Lot # 2015-A
 Acetone Stock Solution Lot # 2015-B

Simulator	Serial Number
0.00	SD1022
Interferent	SD1021
0.05	SD1018
0.08	SD1011
0.20	SD1025

Attachments

Form 41
 Pre-Stability Tests
 Flow Calibration

Optical Bench Cal
 Post-Stability Tests
 Other _____

Notes: 2016 Annual
QC-TR

Instrument Complies with Chapter 11D-8, FAC
 Instrument Does Not Comply with Chapter 11D-8, FAC

Return to/Place into Evidentiary Use
 Remain Out of Evidentiary Use
 Conduct an Agency Inspection Before Evidentiary Use

Patrick Murphy
 Quality Control Review

1/6/16
 Date

Stability Checks - Instrument #80-001291 Bradford County So 1/6/2016 89

PGM

BRADFORD COUNTY SO
Intoxilyzer - Alcohol Analyzer
Model 8000
01/06/2016
Software: 8100.27

BRADFORD COUNTY SO
Intoxilyzer - Alcohol Analyzer
Model 8000
01/06/2016
Software: 8100.27

BRADFORD COUNTY SO
Intoxilyzer - Alcohol Analyzer
Model 8000
01/06/2016
Software: 8100.27

BRADFORD COUNTY SO
Intoxilyzer - Alcohol Analyzer
Model 8000
01/06/2016
Software: 8100.27

BRADFORD COUNTY SO
Intoxilyzer - Alcohol Analyzer
Model 8000
01/06/2016
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	08:46
Control Test	0.049	08:46
Air Blank	0.000	08:47
Control Test	0.050	08:48
Air Blank	0.000	08:48
Control Test	0.049	08:49
Air Blank	0.000	08:49
Control Test Stats		
Average	0.0493	
Std Dev	0.0006	
Rel Std Dev(%)	1.1703	

Test	g/210L	Time
Air Blank	0.000	08:51
Control Test	0.078	08:51
Air Blank	0.000	08:52
Control Test	0.079	08:53
Air Blank	0.000	08:53
Control Test	0.079	08:54
Air Blank	0.000	08:54
Control Test Stats		
Average	0.0787	
Std Dev	0.0006	
Rel Std Dev(%)	0.7339	

Test	g/210L	Time
Air Blank	0.000	08:56
Control Test	0.200	08:57
Air Blank	0.000	08:57
Control Test	0.202	08:58
Air Blank	0.000	08:58
Control Test	0.202	08:59
Air Blank	0.000	09:00
Control Test Stats		
Average	0.2013	
Std Dev	0.0012	
Rel Std Dev(%)	0.5735	

Test	g/210L	Time
Air Blank	0.000	08:42
Control Test	0.081	08:42
Air Blank	0.000	08:43
Control Test	0.081	08:43
Air Blank	0.000	08:43
Control Test	0.080	08:44
Air Blank	0.000	08:44
Control Test Stats		
Average	0.0807	
Std Dev	0.0006	
Rel Std Dev(%)	0.7157	

Test	g/210L	Time
Air Blank	0.000	08:42
Control Test	0.081	08:42
Air Blank	0.000	08:43
Control Test	0.081	08:43
Air Blank	0.000	08:43
Control Test	0.080	08:44
Air Blank	0.000	08:44
Control Test Stats		
Average	0.0807	
Std Dev	0.0006	
Rel Std Dev(%)	0.7157	

BK

D65

92 Operator's Signature

92 Operator's Signature

92 Operator's Signature

92 Operator's Signature