



Alcohol Testing Program

INSTRUMENT PROCESSING SHEET

Agency Levy County SO S/N 80-001292

Date In 8/4/16 Date Out 8/12/16 Ship P/U H/D CMI EE

Intake Performed By DK

Registration
 Annual
 Return from CMI
 Return from Enforcement
 Electronics
 Other _____

Visual Inspection:
OK Case OK Handle
OK Dry Gas Holder OK Feet
OK Keyboard/Plug OK Back/Plugs
OK Screws tight OK Breath Hose

Other Equipment:
 Power cord
 Printer Cable
 Other Static Bag

Notes: _____

Quality Checks Performed By PGM

Breath Tube Screen
 Replace O-Rings
 Instrument Set Up Verified
 R-Value 198
 Flow Verification (L/s)
 Flow Column # ATP102
 32mm .140 (.139 - .169)
 36mm .156 (.156 - .190)
 53mm .238 (.228 - .278)
 103mm .515 (.447 - .547)

Barometric Pressure Check
 Gauge ID # 28427

Stability Checks

| Simulator | Serial # | Lot #/Exp |
|-----------|----------|--------------------|
| 0.05 | SD1018 | 201507A 7/14/17 |
| 0.08 | SD1011 | 201601F 1/26/18 |
| 0.20 | SD1025 | 201604C 4/5/18 |
| 0.08 DGS | N/A | AG612405 5/3/18 |

Flow Calibration Performed By PGM

Flow Calibration N/A
 Flow Calibration Complete
 Flow Column # _____
 5L/min - 17mm
 15L/min - 30mm
 30L/min - 103mm

R-Value _____
 Post Calibration Verification (L/s)
 Flow Column # _____
 32mm _____ (.139 - .169)
 36mm _____ (.156 - .190)
 53mm _____ (.228 - .278)
 103mm _____ (.447 - .547)

Maintenance Performed By _____

Battery Replacement
 Dry Gas Regulator Replacement
 Breath Tube Replacement
 Other _____

Suggested Service

RECEIVED
 AUG 15 2016
 FDLE
 Alcohol Testing Program

Optical Bench Calibration Performed By _____

Optical Bench Calibration N/A
 Optical Bench Calibration Complete

Barometric Pressure Gauge ID # _____

| Simulator | Serial Number | Lot Number | Expiration |
|-----------|---------------|------------|------------|
| 0.000 | | N/A | N/A |
| 0.040 | | | |
| 0.100 | | | |
| 0.200 | | | |
| 0.400 | | | |
| 0.080 DGS | N/A | | |

Post Calibration Stability Checks

| Simulator | Serial Number | Lot Number | Expiration |
|-----------|---------------|------------|------------|
| 0.05 | | | |
| 0.08 | | | |
| 0.20 | | | |
| 0.08 DGS | N/A | | |

Department Inspection Performed By PGM

Barometric Pressure 1016 Gauge
 ID# 28427 1014 Instrument

Mouth Alcohol Solution Lot # 2016-A
 Acetone Stock Solution Lot # 2016-B

| Simulator | Serial Number |
|-------------|---------------|
| 0.00 | SD1019 |
| Interferent | SD1021 |
| 0.05 | SD1018 |
| 0.08 | SD1011 |
| 0.20 | SD1025 |

Attachments

Form 41
 Pre-Stability Tests
 Flow Calibration

Optical Bench Cal
 Post-Stability Tests
 Other _____

Notes: OK PWS

Instrument Complies with Chapter 11D-8, FAC
 Instrument Does Not Comply with Chapter 11D-8, FAC

Return to/Place into Evidentiary Use
 Remain Out of Evidentiary Use
 Conduct an Agency Inspection Before Evidentiary Use

Scott Hankland
 Quality Control Review

8/15/16
 Date

LEUY COUNTY SO
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-001292
08/12/2016
Software: 8100.27

| Test | g/210L | Time |
|--------------------|--------|-------|
| Air Blank | 0.000 | 11:08 |
| Control Test | 0.049 | 11:08 |
| Air Blank | 0.000 | 11:09 |
| Control Test | 0.049 | 11:10 |
| Air Blank | 0.000 | 11:10 |
| Control Test | 0.050 | 11:11 |
| Air Blank | 0.000 | 11:11 |
| Control Test Stats | | |
| Average | 0.0493 | |
| Std Dev | 0.0006 | |
| Rel Std Dev(%) | 1.1703 | |



Operator's Signature

LEUY COUNTY SO
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-001292
08/12/2016
Software: 8100.27

| Test | g/210L | Time |
|--------------------|--------|-------|
| Air Blank | 0.000 | 11:02 |
| Control Test | 0.080 | 11:02 |
| Air Blank | 0.000 | 11:03 |
| Control Test | 0.080 | 11:03 |
| Air Blank | 0.000 | 11:04 |
| Control Test | 0.080 | 11:05 |
| Air Blank | 0.000 | 11:05 |
| Control Test Stats | | |
| Average | 0.0800 | |
| Std Dev | 0.0000 | |
| Rel Std Dev(%) | 0.0000 | |



Operator's Signature

LEUY COUNTY SO
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-001292
08/12/2016
Software: 8100.27

| Test | g/210L | Time |
|--------------------|--------|-------|
| Air Blank | 0.000 | 10:17 |
| Control Test | 0.197 | 10:18 |
| Air Blank | 0.000 | 10:18 |
| Control Test | 0.197 | 10:19 |
| Air Blank | 0.000 | 10:20 |
| Control Test | 0.197 | 10:20 |
| Air Blank | 0.000 | 10:21 |
| Control Test Stats | | |
| Average | 0.1970 | |
| Std Dev | 0.0000 | |
| Rel Std Dev(%) | 0.0000 | |



Operator's Signature

AS
BK

LEUY COUNTY SO
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-001292
08/12/2016
Software: 8100.27

| Test | g/210L | Time |
|--------------------|--------|-------|
| Air Blank | 0.000 | 11:36 |
| Control Test | 0.080 | 11:36 |
| Air Blank | 0.000 | 11:36 |
| Control Test | 0.081 | 11:37 |
| Air Blank | 0.000 | 11:37 |
| Control Test | 0.080 | 11:38 |
| Air Blank | 0.000 | 11:38 |
| Control Test Stats | | |
| Average | 0.0803 | |
| Std Dev | 0.0006 | |
| Rel Std Dev(%) | 0.7187 | |

DGS



Operator's Signature