



Alcohol Testing Program

INSTRUMENT PROCESSING SHEET

Agency Williston P.D S/N 80-001293

Date In 12/21/16 Date Out 12/21/16 Ship P/U H/D CMI EE

Intake Performed By <u>[Signature]</u> <input type="checkbox"/> Registration <input checked="" type="checkbox"/> Annual <input type="checkbox"/> Return from CMI <input type="checkbox"/> Return from Enforcement Electronics <input type="checkbox"/> Other _____ Visual Inspection: <u>ok</u> Case <u>ok</u> Handle <u>ok</u> Dry Gas Holder <u>ok</u> Feet <u>ok</u> Keyboard/Plug <u>ok</u> Back/Plugs <u>ok</u> Screws tight <u>ok</u> Breath Hose Other Equipment: <input checked="" type="checkbox"/> Power cord <input type="checkbox"/> Printer Cable <input checked="" type="checkbox"/> Other <u>static Bag</u> Notes: _____ _____ _____	Quality Checks Performed By <u>[Signature]</u> <input checked="" type="checkbox"/> Breath Tube Screen <input checked="" type="checkbox"/> Replace O-Rings <input checked="" type="checkbox"/> Instrument Set Up Verified <input checked="" type="checkbox"/> R-Value <u>2.0</u> <input checked="" type="checkbox"/> Flow Verification (L/s) Flow Column # <u>AP102</u> 32mm <u>152</u> (.139 - .169) 36mm <u>178</u> (.156 - .190) 53mm <u>254</u> (.228 - .278) 103mm <u>508</u> (.447 - .547) <input checked="" type="checkbox"/> Barometric Pressure Check Gauge ID # <u>28662</u> <input checked="" type="checkbox"/> Stability Checks <table border="1" style="width:100%; border-collapse: collapse;"> <thead> <tr> <th>Simulator</th> <th>Serial #</th> <th>Lot #/Exp</th> </tr> </thead> <tbody> <tr> <td>0.05</td> <td><u>DR2035</u></td> <td><u>20603D</u> <u>3/8/18</u></td> </tr> <tr> <td>0.08</td> <td><u>SD1011</u></td> <td><u>201601F</u> <u>1/26/18</u></td> </tr> <tr> <td>0.20</td> <td><u>SD1025</u></td> <td><u>201604C</u> <u>4/5/18</u></td> </tr> <tr> <td>0.08 DGS</td> <td><u>N/A</u></td> <td><u>AG619605</u> <u>7/14/18</u></td> </tr> </tbody> </table>	Simulator	Serial #	Lot #/Exp	0.05	<u>DR2035</u>	<u>20603D</u> <u>3/8/18</u>	0.08	<u>SD1011</u>	<u>201601F</u> <u>1/26/18</u>	0.20	<u>SD1025</u>	<u>201604C</u> <u>4/5/18</u>	0.08 DGS	<u>N/A</u>	<u>AG619605</u> <u>7/14/18</u>	Flow Calibration Performed By _____ <input checked="" type="checkbox"/> Flow Calibration N/A <input type="checkbox"/> Flow Calibration Complete Flow Column # _____ <input type="checkbox"/> 5L/min - 17mm <input type="checkbox"/> 15L/min - 53mm <input type="checkbox"/> 30L/min - 103mm <input type="checkbox"/> R-Value _____ <input type="checkbox"/> Post Calibration Verification (L/s) Program Flow Column # _____ 32mm _____ (.139 - .169) 36mm _____ (.156 - .190) 53mm _____ (.228 - .278) 103mm _____ (.447 - .547)
Simulator	Serial #	Lot #/Exp															
0.05	<u>DR2035</u>	<u>20603D</u> <u>3/8/18</u>															
0.08	<u>SD1011</u>	<u>201601F</u> <u>1/26/18</u>															
0.20	<u>SD1025</u>	<u>201604C</u> <u>4/5/18</u>															
0.08 DGS	<u>N/A</u>	<u>AG619605</u> <u>7/14/18</u>															
Maintenance Performed By _____ <input type="checkbox"/> Battery Replacement <input type="checkbox"/> Dry Gas Regulator Replacement <input type="checkbox"/> Breath Tube Replacement <input type="checkbox"/> Other _____																	
Suggested Service _____ _____																	

RECEIVED
DEC 27 2016
FDLE
Alcohol Testing Program

Optical Bench Calibration Performed By _____ <input checked="" type="checkbox"/> Optical Bench Calibration N/A <input type="checkbox"/> Optical Bench Calibration Complete Barometric Pressure Gauge _____ ID # _____																							
Simulator	Serial Number	Lot Number	Expiration																				
0.000		N/A	N/A																				
0.040																							
0.100																							
0.200																							
0.400																							
0.080 DGS	N/A																						
<input type="checkbox"/> Post Calibration Stability Checks <table border="1" style="width:100%; border-collapse: collapse;"> <thead> <tr> <th>Simulator</th> <th>Serial Number</th> <th>Lot Number</th> <th>Expiration</th> </tr> </thead> <tbody> <tr> <td>0.05</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.08</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.20</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.08 DGS</td> <td>N/A</td> <td></td> <td></td> </tr> </tbody> </table>				Simulator	Serial Number	Lot Number	Expiration	0.05				0.08				0.20				0.08 DGS	N/A		
Simulator	Serial Number	Lot Number	Expiration																				
0.05																							
0.08																							
0.20																							
0.08 DGS	N/A																						

Department Inspection Performed By <u>[Signature]</u> <input checked="" type="checkbox"/> Barometric Pressure <u>1023</u> Gauge ID# <u>28662</u> <u>1022</u> Instrument Mouth Alcohol Solution Lot # <u>2016-A</u> Acetone Stock Solution Lot # <u>2016-B</u>	
Simulator	Serial Number
0.00	<u>SD1019</u>
Interferent	<u>SD1021</u>
0.05	<u>DR2035</u>
0.08	<u>SD1011</u>
0.20	<u>SD1025</u>
Attachments <input checked="" type="checkbox"/> Form 41 <input checked="" type="checkbox"/> Pre-Stability Tests <input type="checkbox"/> Flow Calibration <input type="checkbox"/> Optical Bench Cal <input type="checkbox"/> Post-Stability Tests <input type="checkbox"/> Other _____	

Notes: QC & DGS

<input checked="" type="checkbox"/> Instrument Complies with Chapter 11D-8, FAC <input type="checkbox"/> Instrument Does Not Comply with Chapter 11D-8, FAC <input checked="" type="checkbox"/> Return to/Place into Evidentiary Use <input type="checkbox"/> Remain Out of Evidentiary Use <input checked="" type="checkbox"/> Conduct an Agency Inspection Before Evidentiary Use

Brett Kirkland

Quality Control Review

12/27/16
Date

80-001293
 Stability Checks
 12/21/16

INTOXILYZER 8000
 Instrument Initialization
 10:22 12/21/2016

WILLISTON PD
 Intoxilyzer - Alcohol Analyzer
 Model 8000 SN 80-001293
 12/21/2016
 Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	10:47
Control Test	0.079	10:48
Air Blank	0.000	10:48
Control Test	0.079	10:49
Air Blank	0.000	10:49
Control Test	0.080	10:49
Air Blank	0.000	10:50
Control Test Stats		
Average	0.0793	
Std Dev	0.0006	
Rel Std Dev(%)	0.7277	

CS

[Signature]
 Operator's Signature

WILLISTON PD
 Intoxilyzer - Alcohol Analyzer
 Model 8000 SN 80-001293
 12/21/2016
 Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	10:51
Control Test	0.050	10:51
Air Blank	0.000	10:52
Control Test	0.050	10:53
Air Blank	0.000	10:53
Control Test	0.050	10:54
Air Blank	0.000	10:54
Control Test Stats		
Average	0.0500	
Std Dev	0.0000	
Rel Std Dev(%)	0.0000	

[Signature]
 Operator's Signature

WILLISTON PD
 Intoxilyzer - Alcohol Analyzer
 Model 8000 SN 80-001293
 12/21/2016
 Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	10:56
Control Test	0.081	10:56
Air Blank	0.000	10:57
Control Test	0.081	10:58
Air Blank	0.000	10:58
Control Test	0.081	10:59
Air Blank	0.000	10:59
Control Test Stats		
Average	0.0810	
Std Dev	0.0000	
Rel Std Dev(%)	0.0000	

[Signature]
 Operator's Signature

WILLISTON PD
 Intoxilyzer - Alcohol Analyzer
 Model 8000 SN 80-001293
 12/21/2016
 Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	11:01
Control Test	0.200	11:01
Air Blank	0.000	11:02
Control Test	0.200	11:03
Air Blank	0.000	11:03
Control Test	0.200	11:04
Air Blank	0.000	11:04
Control Test Stats		
Average	0.2000	
Std Dev	0.0000	
Rel Std Dev(%)	0.0000	

[Signature]
BSK

[Signature]
 Operator's Signature