

INSTRUMENT PROCESSING SHEET

Agency Madison County SO S/N 80-001307
 Date In 3/8/16 Date Out 3/30/16 Ship P/U H/D CMI EE

Intake Performed By [Signature]

Registration
 Annual
 Return from CMI
 Return from Enforcement Electronics
 Other _____

Visual Inspection:
OK Case OK Handle
OK Dry Gas Holder OK Feet
OK Keyboard/Plug OK Back/Plugs
OK Screws tight OK Breath Hose

Other Equipment:
 Power cord
 Printer Cable
 Other _____

Notes: _____

Quality Checks Performed By [Signature]

Breath Tube Screen
 Replace O-Rings
 Instrument Set Up Verified
 R-Value 157
 Flow Verification (L/s)
 Flow Column # ATP03
 32mm 148 (.139 - .169)
 36mm 167 (.156 - .190)
 53mm 238 (.228 - .278)
 103mm 507 (.447 - .547)

Barometric Pressure Check
 Gauge ID # 28427

Stability Checks

Simulator	Serial #	Lot #/Exp
0.05	551018	201507A 7/14/17
0.08	551011	201601F 1/26/18
0.20	64444	201505A 5/12/17
0.08 DGS	N/A	AG511701 4/27/17

Flow Calibration Performed By [Signature]

Flow Calibration N/A
 Flow Calibration Complete
 Flow Column # _____
 5L/min - 17mm
 15L/min - 53mm
 30L/min - 103mm
 R-Value _____
 Post Calibration Verification (L/s)
 Flow Column # _____
 32mm _____ (.139 - .169)
 36mm _____ (.156 - .190)
 53mm _____ (.228 - .278)
 103mm _____ (.447 - .547)

Maintenance Performed By _____

Battery Replacement
 Dry Gas Regulator Replacement
 Breath Tube Replacement
 Other _____

Suggested Service

RECEIVED
 APR 13 2016
 FDLE
 Alcohol Testing Program

Optical Bench Calibration Performed By _____

Optical Bench Calibration N/A
 Optical Bench Calibration Complete

Barometric Pressure Gauge ID # _____

Simulator	Serial Number	Lot Number	Expiration
0.000		N/A	N/A
0.040			
0.100			
0.200			
0.400			
0.080 DGS	N/A		

Post Calibration Stability Checks

Simulator	Serial Number	Lot Number	Expiration
0.05			
0.08			
0.20			
0.08 DGS	N/A		

Department Inspection Performed By [Signature]

Barometric Pressure 1020 Gauge
 ID# 28427 1016 Instrument

Mouth Alcohol Solution Lot # 2015-A
 Acetone Stock Solution Lot # 2016-B

Simulator	Serial Number
0.00	551022
Interferent	551021
0.05	551018
0.08	551011
0.20	64444

Attachments

Form 41
 Pre-Stability Tests
 Flow Calibration
 Optical Bench Cal
 Post-Stability Tests
 Other _____

Notes: OK/OC OK [Signature]

Email Approved 4/1/16 [Signature]

Instrument Complies with Chapter 11D-8, FAC
 Instrument Does Not Comply with Chapter 11D-8, FAC

Return to/Place into Evidentiary Use
 Remain Out of Evidentiary Use

Conduct an Agency Inspection Before Evidentiary Use

Brett Wickland
 Quality Control Review

4/13/16
 Date

Stability Tests - Madison CSO #80-001307 3/30/16
BK

MADISON COUNTY SO
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-001307
03/30/2016
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	09:14
Control Test	0.049	09:15
Air Blank	0.000	09:15
Control Test	0.050	09:16
Air Blank	0.000	09:16
Control Test	0.050	09:17
Air Blank	0.000	09:18
Control Test Stats		
Average	0.0497	
Std Dev	0.0006	
Rel Std Dev(%)	1.1625	

BM

PWS

Operator's Signature

MADISON COUNTY SO
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-001307
03/30/2016
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	09:01
Control Test	0.078	09:02
Air Blank	0.000	09:02
Control Test	0.078	09:03
Air Blank	0.000	09:03
Control Test	0.078	09:04
Air Blank	0.000	09:05
Control Test Stats		
Average	0.0780	
Std Dev	0.0000	
Rel Std Dev(%)	0.0000	

PWS

Operator's Signature

MADISON COUNTY SO
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-001307
03/30/2016
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	09:08
Control Test	0.198	09:09
Air Blank	0.000	09:10
Control Test	0.198	09:10
Air Blank	0.000	09:11
Control Test	0.198	09:11
Air Blank	0.000	09:12
Control Test Stats		
Average	0.1980	
Std Dev	0.0000	
Rel Std Dev(%)	0.0000	

PWS

Operator's Signature

MADISON COUNTY SO
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-001307
03/30/2016
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	09:18
Control Test	0.079	09:19
Air Blank	0.000	09:19
Control Test	0.078	09:20
Air Blank	0.000	09:20
Control Test	0.079	09:20
Air Blank	0.000	09:21
Control Test Stats		
Average	0.0787	
Std Dev	0.0006	
Rel Std Dev(%)	0.7339	

PWS

PWS

Operator's Signature