



Alcohol Testing Program

INSTRUMENT PROCESSING SHEET

Agency Bay CSO

S/N 80-001308

Date In 9/8/16

Date Out 9/8/16

Ship P/U H/D CMI EE

Intake Performed By QNS

Registration
 Annual
 Return from CMI
 Return from Enforcement Electronics
 Other _____

Visual Inspection:
OK Case OK Handle
OK Dry Gas Holder OK Feet
OK Keyboard/Plug OK Back/Plugs
OK Screws tight OK Breath Hose

Other Equipment:
 Power cord
 Printer Cable
 Other _____

Notes: _____

Quality Checks Performed By QNS

Breath Tube Screen
 Replace O-Rings
 Instrument Set Up Verified
 R-Value 200
 Flow Verification (L/s)
 Flow Column # ATP102
 32mm 144 (.139 - .169)
 36mm 1104 (.156 - .190)
 53mm 234 (.228 - .278)
 103mm 496 (.447 - .547)

Barometric Pressure Check
 Gauge ID # 28427

Stability Checks

Simulator	Serial #	Lot #/Exp
0.05	SD108	201507A 7/14/17
0.08	SD1011	201601F 1/26/18
0.20	SD1025	201604E 4/5/18
0.08 DGS	N/A	AG602405 5/3/18

Flow Calibration Performed By _____

Flow Calibration N/A
 Flow Calibration Complete
 Flow Column # _____
 5L/min - 17mm
 15L/min - 53mm
 30L/min - 103mm
 R-Value _____
 Post Calibration Verification (L/s)
 Flow Column # _____
 32mm _____ (.139 - .169)
 36mm _____ (.156 - .190)
 53mm _____ (.228 - .278)
 103mm _____ (.447 - .547)

Maintenance Performed By _____

Battery Replacement
 Dry Gas Regulator Replacement
 Breath Tube Replacement
 Other _____

Suggested Service

RECEIVED
SEP 09 2016
FDLE
Alcohol Testing Program

Optical Bench Calibration Performed By _____

Optical Bench Calibration N/A
 Optical Bench Calibration Complete

Barometric Pressure Gauge _____ ID # _____

Simulator	Serial Number	Lot Number	Expiration
0.000		N/A	N/A
0.040			
0.100			
0.200			
0.400			
0.080 DGS	N/A		

Post Calibration Stability Checks

Simulator	Serial Number	Lot Number	Expiration
0.05			
0.08			
0.20			
0.08 DGS	N/A		

Department Inspection Performed By QNS

Barometric Pressure 1019 Gauge
 ID# 28427 1010 Instrument

Mouth Alcohol Solution Lot # 2016-A
 Acetone Stock Solution Lot # 2016-B

Simulator	Serial Number
0.00	SD1019
Interferent	SD1021
0.05	SD1018
0.08	SD1011
0.20	SD1025

Attachments

Form 41
 Pre-Stability Tests
 Flow Calibration
 Optical Bench Cal
 Post-Stability Tests
 Other _____

Notes: QA/QC OK QNS

Instrument Complies with Chapter 11D-8, FAC
 Instrument Does Not Comply with Chapter 11D-8, FAC

Return to/Place into Evidentiary Use
 Remain Out of Evidentiary Use

Conduct an Agency Inspection Before Evidentiary Use

Brett Highland
Quality Control Review

9/9/16
Date

Stability Tests - Bay CSO # 80-001308 9/8/16

BAY COUNTY SO
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-001308
09/08/2016
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	07:32
Control Test	0.049	07:33
Air Blank	0.000	07:33
Control Test	0.049	07:34
Air Blank	0.000	07:35
Control Test	0.050	07:35
Air Blank	0.000	07:36
Control Test Stats		
Average	0.0493	
Std Dev	0.0006	
Rel Std Dev(%)	1.1703	

Boon

RWS

Operator's Signature

BAY COUNTY SO
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-001308
09/08/2016
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	07:37
Control Test	0.079	07:38
Air Blank	0.000	07:38
Control Test	0.080	07:39
Air Blank	0.000	07:40
Control Test	0.080	07:40
Air Blank	0.000	07:41
Control Test Stats		
Average	0.0797	
Std Dev	0.0006	
Rel Std Dev(%)	0.7247	

RWS

Operator's Signature

BAY COUNTY SO
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-001308
09/08/2016
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	07:42
Control Test	0.199	07:43
Air Blank	0.000	07:43
Control Test	0.198	07:44
Air Blank	0.000	07:45
Control Test	0.199	07:45
Air Blank	0.000	07:46
Control Test Stats		
Average	0.1987	
Std Dev	0.0006	
Rel Std Dev(%)	0.2906	

RWS

Operator's Signature

BAY COUNTY SO
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-001308
09/08/2016
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	07:47
Control Test	0.081	07:47
Air Blank	0.000	07:47
Control Test	0.081	07:48
Air Blank	0.000	07:48
Control Test	0.082	07:49
Air Blank	0.000	07:49
Control Test Stats		
Average	0.0813	
Std Dev	0.0006	
Rel Std Dev(%)	0.7099	

SS

ASK

RWS

Operator's Signature