



Alcohol Testing Program

INSTRUMENT PROCESSING SHEET

Agency Bay CSO

S/N 80-001309

Date In 9/29/16

Date Out 9/29/16

Ship P/U H/D CMI EE

<b>Intake</b> Performed By <u>JWS</u> <input type="checkbox"/> Registration <input checked="" type="checkbox"/> Annual <input type="checkbox"/> Return from CMI <input type="checkbox"/> Return from Enforcement Electronics <input type="checkbox"/> Other _____ <b>Visual Inspection:</b> <input checked="" type="checkbox"/> Case <input checked="" type="checkbox"/> Handle <input checked="" type="checkbox"/> Dry Gas Holder <input checked="" type="checkbox"/> Feet <input checked="" type="checkbox"/> Keyboard/Plug <input checked="" type="checkbox"/> Back/Plugs <input checked="" type="checkbox"/> Screws tight <input checked="" type="checkbox"/> Breath Hose <b>Other Equipment:</b> <input type="checkbox"/> Power cord <input type="checkbox"/> Printer Cable <input type="checkbox"/> Other _____ Notes: _____ _____ _____	<b>Quality Checks</b> Performed By <u>JWS</u> <input checked="" type="checkbox"/> Breath Tube Screen <input checked="" type="checkbox"/> Replace O-Rings <input checked="" type="checkbox"/> Instrument Set Up Verified <input checked="" type="checkbox"/> R-Value <u>188</u> <input checked="" type="checkbox"/> Flow Verification (L/s) Flow Column # <u>APP102</u> 32mm <u>144</u> (.139 - .169) 36mm <u>167</u> (.156 - .190) 53mm <u>234</u> (.228 - .278) 103mm <u>503</u> (.447 - .547) <input checked="" type="checkbox"/> Barometric Pressure Check Gauge ID # <u>28427</u> <input checked="" type="checkbox"/> Stability Checks <table border="1"> <thead> <tr> <th>Simulator</th> <th>Serial #</th> <th>Lot #/Exp</th> </tr> </thead> <tbody> <tr> <td>0.05</td> <td>SD1018</td> <td>201507A 7/14/17</td> </tr> <tr> <td>0.08</td> <td>SD104</td> <td>201601F 1/26/18</td> </tr> <tr> <td>0.20</td> <td>SD1025</td> <td>201604C 4/5/18</td> </tr> <tr> <td>0.08 DGS</td> <td>N/A</td> <td>AG612405 5/3/18</td> </tr> </tbody> </table>	Simulator	Serial #	Lot #/Exp	0.05	SD1018	201507A 7/14/17	0.08	SD104	201601F 1/26/18	0.20	SD1025	201604C 4/5/18	0.08 DGS	N/A	AG612405 5/3/18	<b>Flow Calibration</b> Performed By _____ <input checked="" type="checkbox"/> Flow Calibration N/A <input type="checkbox"/> Flow Calibration Complete Flow Column # _____ <input type="checkbox"/> 5L/min - 17mm <input type="checkbox"/> 15L/min - 53mm <input type="checkbox"/> 30L/min - 103mm <input type="checkbox"/> R-Value _____ <input type="checkbox"/> Post Calibration Verification (L/s) Flow Column # _____ 32mm _____ (.139 - .169) 36mm _____ (.156 - .190) 53mm _____ (.228 - .278) 103mm _____ (.447 - .547) <b>Maintenance</b> Performed By _____ <input type="checkbox"/> Battery Replacement <input type="checkbox"/> Dry Gas Regulator Replacement <input type="checkbox"/> Breath Tube Replacement <input type="checkbox"/> Other _____ <b>Suggested Service</b> _____ _____
Simulator	Serial #	Lot #/Exp															
0.05	SD1018	201507A 7/14/17															
0.08	SD104	201601F 1/26/18															
0.20	SD1025	201604C 4/5/18															
0.08 DGS	N/A	AG612405 5/3/18															

RECEIVED  
SEP 29 2016  
FDLE  
Alcohol Testing Program

<b>Optical Bench Calibration</b> Performed By _____ <input checked="" type="checkbox"/> Optical Bench Calibration N/A <input type="checkbox"/> Optical Bench Calibration Complete Barometric Pressure Gauge ID # _____ <table border="1"> <thead> <tr> <th>Simulator</th> <th>Serial Number</th> <th>Lot Number</th> <th>Expiration</th> </tr> </thead> <tbody> <tr> <td>0.000</td> <td></td> <td>N/A</td> <td>N/A</td> </tr> <tr> <td>0.040</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.100</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.200</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.400</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.080 DGS</td> <td>N/A</td> <td></td> <td></td> </tr> </tbody> </table> <input type="checkbox"/> Post Calibration Stability Checks <table border="1"> <thead> <tr> <th>Simulator</th> <th>Serial Number</th> <th>Lot Number</th> <th>Expiration</th> </tr> </thead> <tbody> <tr> <td>0.05</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.08</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.20</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.08 DGS</td> <td>N/A</td> <td></td> <td></td> </tr> </tbody> </table>	Simulator	Serial Number	Lot Number	Expiration	0.000		N/A	N/A	0.040				0.100				0.200				0.400				0.080 DGS	N/A			Simulator	Serial Number	Lot Number	Expiration	0.05				0.08				0.20				0.08 DGS	N/A		
Simulator	Serial Number	Lot Number	Expiration																																													
0.000		N/A	N/A																																													
0.040																																																
0.100																																																
0.200																																																
0.400																																																
0.080 DGS	N/A																																															
Simulator	Serial Number	Lot Number	Expiration																																													
0.05																																																
0.08																																																
0.20																																																
0.08 DGS	N/A																																															

<b>Department Inspection</b> Performed By <u>JWS</u> <input checked="" type="checkbox"/> Barometric Pressure <u>1015</u> Gauge ID# <u>28427</u> <u>1014</u> Instrument Mouth Alcohol Solution Lot # <u>2016-A</u> Acetone Stock Solution Lot # <u>2016-B</u> <table border="1"> <thead> <tr> <th>Simulator</th> <th>Serial Number</th> </tr> </thead> <tbody> <tr> <td>0.00</td> <td>SD1019</td> </tr> <tr> <td>Interferent</td> <td>SD1021</td> </tr> <tr> <td>0.05</td> <td>SD1018</td> </tr> <tr> <td>0.08</td> <td>SD1011</td> </tr> <tr> <td>0.20</td> <td>SD1025</td> </tr> </tbody> </table> <b>Attachments</b> <input checked="" type="checkbox"/> Form 41 <input checked="" type="checkbox"/> Pre-Stability Tests <input type="checkbox"/> Flow Calibration <input type="checkbox"/> Optical Bench Cal <input type="checkbox"/> Post-Stability Tests <input type="checkbox"/> Other _____	Simulator	Serial Number	0.00	SD1019	Interferent	SD1021	0.05	SD1018	0.08	SD1011	0.20	SD1025
Simulator	Serial Number											
0.00	SD1019											
Interferent	SD1021											
0.05	SD1018											
0.08	SD1011											
0.20	SD1025											

Notes: QC: SP  
 \_\_\_\_\_  
 \_\_\_\_\_  
 \_\_\_\_\_

 Instrument Complies with Chapter 11D-8, FAC  
 Instrument Does Not Comply with Chapter 11D-8, FAC  
 Return to/Place into Evidentiary Use  
 Remain Out of Evidentiary Use  
 Conduct an Agency Inspection Before Evidentiary Use

Brett Kirkland

Quality Control Review

9/29/16

Date

Stability Tests - Bay CSO # 80-001309 9/29/16

BAY COUNTY SO  
Intoxilyzer - Alcohol Analyzer  
Model 8000 SN 80-001309  
09/29/2016  
Software: 8100.27

Test	9/21/0L	Time
Air Blank	0.000	07:29
Control Test	0.048	07:30
Air Blank	0.000	07:30
Control Test	0.048	07:31
Air Blank	0.000	07:32
Control Test	0.048	07:32
Air Blank	0.000	07:33
Control Test Stats		
Average	0.0480	
Std Dev	0.0000	
Rel Std Dev(%)	0.0000	

*[Signature]*  
Operator's Signature

BAY COUNTY SO  
Intoxilyzer - Alcohol Analyzer  
Model 8000 SN 80-001309  
09/29/2016  
Software: 8100.27

Test	9/21/0L	Time
Air Blank	0.000	07:34
Control Test	0.078	07:35
Air Blank	0.000	07:36
Control Test	0.078	07:36
Air Blank	0.000	07:37
Control Test	0.079	07:38
Air Blank	0.000	07:38
Control Test Stats		
Average	0.0783	
Std Dev	0.0006	
Rel Std Dev(%)	0.7370	

*[Signature]*  
Operator's Signature

BAY COUNTY SO  
Intoxilyzer - Alcohol Analyzer  
Model 8000 SN 80-001309  
09/29/2016  
Software: 8100.27

Test	9/21/0L	Time
Air Blank	0.000	07:39
Control Test	0.195	07:40
Air Blank	0.000	07:40
Control Test	0.195	07:41
Air Blank	0.000	07:42
Control Test	0.197	07:42
Air Blank	0.000	07:43
Control Test Stats		
Average	0.1957	
Std Dev	0.0012	
Rel Std Dev(%)	0.5901	

*[Signature]*  
Operator's Signature

BAY COUNTY SO  
Intoxilyzer - Alcohol Analyzer  
Model 8000 SN 80-001309  
09/29/2016  
Software: 8100.27

Test	9/21/0L	Time
Air Blank	0.000	07:44
Control Test	0.077	07:44
Air Blank	0.000	07:45
Control Test	0.077	07:45
Air Blank	0.000	07:46
Control Test	0.078	07:46
Air Blank	0.000	07:46
Control Test Stats		
Average	0.0773	
Std Dev	0.0006	
Rel Std Dev(%)	0.7466	

*[Signature]*  
*BSK*  
*SP*

*[Signature]*  
Operator's Signature