



Alcohol Testing Program

INSTRUMENT PROCESSING SHEET

Agency Niceville PD

S/N 80-001310

Date In 10/4/16

Date Out 10/4/16

Ship P/U H/D CMI EE

Intake Performed By <u>PWS</u> <input type="checkbox"/> Registration <input checked="" type="checkbox"/> Annual <input type="checkbox"/> Return from CMI <input type="checkbox"/> Return from Enforcement Electronics <input type="checkbox"/> Other _____ Visual Inspection: <input checked="" type="checkbox"/> Case <input checked="" type="checkbox"/> Handle <input checked="" type="checkbox"/> Dry Gas Holder <input checked="" type="checkbox"/> Feet <input checked="" type="checkbox"/> Keyboard/Plug <input checked="" type="checkbox"/> Back/Plugs <input checked="" type="checkbox"/> Screws tight <input checked="" type="checkbox"/> Breath Hose Other Equipment: <input type="checkbox"/> Power cord <input type="checkbox"/> Printer Cable <input type="checkbox"/> Other _____ Notes: _____ _____ _____	Quality Checks Performed By <u>PWS</u> <input checked="" type="checkbox"/> Breath Tube Screen <input checked="" type="checkbox"/> Replace O-Rings <input checked="" type="checkbox"/> Instrument Set Up Verified <input checked="" type="checkbox"/> R-Value <u>1.85</u> <input checked="" type="checkbox"/> Flow Verification (L/s) Flow Column # <u>ADP102</u> 32mm <u>144</u> (.139 - .169) 36mm <u>164</u> (.156 - .190) 53mm <u>238</u> (.228 - .278) 103mm <u>511</u> (.447 - .547) <input checked="" type="checkbox"/> Barometric Pressure Check Gauge ID # <u>28427</u> $> 1\%$ ¹⁰⁰¹ / ₁₀₁₄ <input checked="" type="checkbox"/> Stability Checks <table border="1"> <thead> <tr> <th>Simulator</th> <th>Serial #</th> <th>Lot #/Exp</th> </tr> </thead> <tbody> <tr> <td>0.05</td> <td>SD1018</td> <td>201507A 7/14/17</td> </tr> <tr> <td>0.08</td> <td>SD1011</td> <td>201601F 1/26/18</td> </tr> <tr> <td>0.20</td> <td>SD1025</td> <td>201604C 4/5/18</td> </tr> <tr> <td>0.08 DGS</td> <td>N/A</td> <td>AG612405 5/3/18</td> </tr> </tbody> </table>	Simulator	Serial #	Lot #/Exp	0.05	SD1018	201507A 7/14/17	0.08	SD1011	201601F 1/26/18	0.20	SD1025	201604C 4/5/18	0.08 DGS	N/A	AG612405 5/3/18	Flow Calibration Performed By _____ <input checked="" type="checkbox"/> Flow Calibration N/A <input type="checkbox"/> Flow Calibration Complete Flow Column # _____ <input type="checkbox"/> 5L/min - 17mm <input type="checkbox"/> 15L/min - 33mm <input type="checkbox"/> 30L/min - 103mm <input type="checkbox"/> R-Value _____ <input type="checkbox"/> Post Calibration Verification (L/s) Flow Column # _____ 32mm _____ (.139 - .169) 36mm _____ (.156 - .190) 53mm _____ (.228 - .278) 103mm _____ (.447 - .547) Maintenance Performed By _____ <input type="checkbox"/> Battery Replacement <input type="checkbox"/> Dry Gas Regulator Replacement <input type="checkbox"/> Breath Tube Replacement <input type="checkbox"/> Other _____ Suggested Service _____ _____
Simulator	Serial #	Lot #/Exp															
0.05	SD1018	201507A 7/14/17															
0.08	SD1011	201601F 1/26/18															
0.20	SD1025	201604C 4/5/18															
0.08 DGS	N/A	AG612405 5/3/18															

RECEIVED
OCT 10 2016
LABORATORY

Optical Bench Calibration Performed By <u>PWS</u> <input type="checkbox"/> Optical Bench Calibration N/A <input checked="" type="checkbox"/> Optical Bench Calibration Complete Barometric Pressure Gauge <u>1012</u> ID # <u>26932</u> <table border="1"> <thead> <tr> <th>Simulator</th> <th>Serial Number</th> <th>Lot Number</th> <th>Expiration</th> </tr> </thead> <tbody> <tr> <td>0.000</td> <td>SD1016</td> <td>N/A</td> <td>N/A</td> </tr> <tr> <td>0.040</td> <td>SD1024</td> <td>1201</td> <td>2/2/18</td> </tr> <tr> <td>0.100</td> <td>SD1013</td> <td>16001</td> <td>5/8/18</td> </tr> <tr> <td>0.200</td> <td>SD1012</td> <td>16103</td> <td>6/14/18</td> </tr> <tr> <td>0.400</td> <td>66621</td> <td>116102</td> <td>3/22/18</td> </tr> <tr> <td>0.080 DGS</td> <td>N/A</td> <td>1561508042</td> <td>7/5/17</td> </tr> </tbody> </table> <input checked="" type="checkbox"/> Post Calibration Stability Checks <table border="1"> <thead> <tr> <th>Simulator</th> <th>Serial Number</th> <th>Lot Number</th> <th>Expiration</th> </tr> </thead> <tbody> <tr> <td>0.05</td> <td>SD1018</td> <td>201507A</td> <td>7/14/17</td> </tr> <tr> <td>0.08</td> <td>SD1011</td> <td>201601F</td> <td>1/26/18</td> </tr> <tr> <td>0.20</td> <td>SD1025</td> <td>201604C</td> <td>4/5/18</td> </tr> <tr> <td>0.08 DGS</td> <td>N/A</td> <td>AG612405</td> <td>5/3/18</td> </tr> </tbody> </table>	Simulator	Serial Number	Lot Number	Expiration	0.000	SD1016	N/A	N/A	0.040	SD1024	1201	2/2/18	0.100	SD1013	16001	5/8/18	0.200	SD1012	16103	6/14/18	0.400	66621	116102	3/22/18	0.080 DGS	N/A	1561508042	7/5/17	Simulator	Serial Number	Lot Number	Expiration	0.05	SD1018	201507A	7/14/17	0.08	SD1011	201601F	1/26/18	0.20	SD1025	201604C	4/5/18	0.08 DGS	N/A	AG612405	5/3/18	Department Inspection Performed By <u>PWS</u> <input checked="" type="checkbox"/> Barometric Pressure <u>1014</u> Gauge ID# <u>28427</u> <u>1011</u> Instrument Mouth Alcohol Solution Lot # <u>2016-A</u> Acetone Stock Solution Lot # <u>2016-B</u> <table border="1"> <thead> <tr> <th>Simulator</th> <th>Serial Number</th> </tr> </thead> <tbody> <tr> <td>0.00</td> <td>SD1019</td> </tr> <tr> <td>Interferent</td> <td>SD1021</td> </tr> <tr> <td>0.05</td> <td>SD1018</td> </tr> <tr> <td>0.08</td> <td>SD1011</td> </tr> <tr> <td>0.20</td> <td>SD1025</td> </tr> </tbody> </table> Attachments <input checked="" type="checkbox"/> Form 41 <input checked="" type="checkbox"/> Pre-Stability Tests <input type="checkbox"/> Flow Calibration <input checked="" type="checkbox"/> Optical Bench Cal <input checked="" type="checkbox"/> Post-Stability Tests <input type="checkbox"/> Other _____	Simulator	Serial Number	0.00	SD1019	Interferent	SD1021	0.05	SD1018	0.08	SD1011	0.20	SD1025
Simulator	Serial Number	Lot Number	Expiration																																																										
0.000	SD1016	N/A	N/A																																																										
0.040	SD1024	1201	2/2/18																																																										
0.100	SD1013	16001	5/8/18																																																										
0.200	SD1012	16103	6/14/18																																																										
0.400	66621	116102	3/22/18																																																										
0.080 DGS	N/A	1561508042	7/5/17																																																										
Simulator	Serial Number	Lot Number	Expiration																																																										
0.05	SD1018	201507A	7/14/17																																																										
0.08	SD1011	201601F	1/26/18																																																										
0.20	SD1025	201604C	4/5/18																																																										
0.08 DGS	N/A	AG612405	5/3/18																																																										
Simulator	Serial Number																																																												
0.00	SD1019																																																												
Interferent	SD1021																																																												
0.05	SD1018																																																												
0.08	SD1011																																																												
0.20	SD1025																																																												

Notes: Calibrated Optical Bench for Barometric Pressure difference greater than 1% (PWS)

QC: SP

 Instrument Complies with Chapter 11D-8, FAC
 Instrument Does Not Comply with Chapter 11D-8, FAC
 Return to/Place into Evidentiary Use
 Remain Out of Evidentiary Use
 Conduct an Agency Inspection Before Evidentiary Use

Brett Hankland
Quality Control Review

10/10/16
Date

Stability Tests
Pre-Calibration

Niceville PD #80-001310

10/4/16

NICEVILLE PD
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-001310
10/04/2016
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	10:15
Control Test	0.049	10:16
Air Blank	0.000	10:17
Control Test	0.049	10:17
Air Blank	0.000	10:18
Control Test	0.049	10:18
Air Blank	0.000	10:19
Control Test Stats		
Average	0.0490	
Std Dev	0.0000	
Rel Std Dev(%)	0.0000	

[Signature]
Operator's Signature

NICEVILLE PD
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-001310
10/04/2016
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	10:20
Control Test	0.080	10:21
Air Blank	0.000	10:21
Control Test	0.080	10:22
Air Blank	0.000	10:22
Control Test	0.080	10:23
Air Blank	0.000	10:24
Control Test Stats		
Average	0.0800	
Std Dev	0.0000	
Rel Std Dev(%)	0.0000	

[Signature]
Operator's Signature

NICEVILLE PD
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-001310
10/04/2016
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	10:24
Control Test	0.197	10:25
Air Blank	0.000	10:26
Control Test	0.197	10:26
Air Blank	0.000	10:27
Control Test	0.196	10:28
Air Blank	0.000	10:28
Control Test Stats		
Average	0.1967	
Std Dev	0.0006	
Rel Std Dev(%)	0.2936	

[Signature]
Operator's Signature

NICEVILLE PD
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-001310
10/04/2016
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	10:30
Control Test	0.078	10:30
Air Blank	0.000	10:31
Control Test	0.078	10:31
Air Blank	0.000	10:31
Control Test	0.078	10:32
Air Blank	0.000	10:32
Control Test Stats		
Average	0.0780	
Std Dev	0.0000	
Rel Std Dev(%)	0.0000	

[Signature]

[Signature]
Operator's Signature

optical bench
calibration
Nireville PD
80-001310
10/4/16

Soi Value = 0.040 g/210L ***
Fit Value = 0.1905 mg/l %%%
Samples Taken = 4, Discarded = 1
Sum Io = 12862, Sum Io = 12917
<<<<< CHANNEL 1 >>>>>
Sample % Abs (% Abs Ref)
Sample #1 = 0.7210 (-0.0220)
Sample #2 = 0.7190 (-0.0120)
Sample #3 = 0.7360 (-0.0210)
Sample #4 = 0.7630 (-0.0170)
Avg % Abs = 0.7393 (-0.0167)
STD DEV = 0.0222 (0.0045)
REL STD DEV = 3.001 (27.055)

<<<<< CHANNEL 2 >>>>>
Sample % Abs (% Abs Ref)
Sample #1 = 1.4510 (-0.0020)
Sample #2 = 1.4640 (-0.0150)
Sample #3 = 1.4740 (-0.0140)
Sample #4 = 1.4800 (-0.0100)
Avg % Abs = 1.4727 (-0.0130)
STD DEV = 0.0081 (0.0026)
REL STD DEV = 0.549 (20.352)

<<<<< CHANNEL 2 >>>>>
Sample % Abs (% Abs Ref)
Sample #1 = 6.9050 (-0.0050)
Sample #2 = 6.9270 (-0.0060)
Sample #3 = 6.9150 (0.0140)
Sample #4 = 6.9210 (0.0130)
Avg % Abs = 6.9210 (0.0070)
STD DEV = 0.0060 (0.0113)
REL STD DEV = 0.087 (160.992)

Soi Value = 0.100 g/210L ***
Fit Value = 0.4762 mg/l %%%
Samples Taken = 4, Discarded = 1
Sum Io = 12871, Sum Io = 12916
<<<<< CHANNEL 1 >>>>>
Sample % Abs (% Abs Ref)
Sample #1 = 1.7670 (-0.0140)
Sample #2 = 1.7770 (-0.0200)
Sample #3 = 1.7900 (-0.0020)
Sample #4 = 1.7860 (-0.0090)
Avg % Abs = 1.7843 (-0.0103)
STD DEV = 0.0067 (0.0091)
REL STD DEV = 0.373 (87.811)

Soi Value = 0.400 g/210L ***
Fit Value = 1.9048 mg/l %%%
Samples Taken = 4, Discarded = 1
Sum Io = 12855, Sum Io = 12914
<<<<< CHANNEL 1 >>>>>
Sample % Abs (% Abs Ref)
Sample #1 = 6.5450 (-0.0110)
Sample #2 = 6.5650 (0.0000)
Sample #3 = 6.5610 (0.0060)
Sample #4 = 6.5920 (0.0040)
Avg % Abs = 6.5727 (0.0033)
STD DEV = 0.0169 (0.0031)
REL STD DEV = 0.257 (91.652)

<<<<< CHANNEL 2 >>>>>
Sample % Abs (% Abs Ref)
Sample #1 = 3.5570 (-0.0040)
Sample #2 = 3.5580 (-0.0030)
Sample #3 = 3.5570 (-0.0040)
Sample #4 = 3.5790 (-0.0080)
Avg % Abs = 3.5647 (-0.0023)
STD DEV = 0.0124 (0.0060)
REL STD DEV = 0.349 (258.331)

<<<<< CHANNEL 2 >>>>>
Sample % Abs (% Abs Ref)
Sample #1 = 12.7840 (0.0000)
Sample #2 = 12.8230 (0.0000)
Sample #3 = 12.8080 (0.0090)
Sample #4 = 12.8490 (0.0050)
Avg % Abs = 12.8267 (0.0047)
STD DEV = 0.0207 (0.0045)
REL STD DEV = 0.162 (96.627)

NICEVILLE PC
Intoxilyzer - Alcohol Analyzer
Model 8000
10/04/2016
SN 80-001310
10:39:20

Auto Calibration
Max Power Res Value = 38
Auto Range Res Value = 26

Soi Value = 0.008 g/210L ***
Fit Value = 0.0000 mg/l %%%
Samples Taken = 4, Discarded = 1
Sum Io = 12871, Sum Io = 12918
<<<<< CHANNEL 1 >>>>>
Sample % Abs (% Abs Ref)
Sample #1 = 0.0560 (-0.0180)
Sample #2 = 0.0520 (0.0020)
Sample #3 = 0.0660 (0.0030)
Sample #4 = 0.0580 (0.0350)
Avg % Abs = 0.0587 (0.0133)
STD DEV = 0.0070 (0.0188)
REL STD DEV = 11.972 (140.779)

<<<<< CHANNEL 2 >>>>>
Sample % Abs (% Abs Ref)
Sample #1 = 0.0730 (-0.0090)
Sample #2 = 0.0890 (-0.0040)
Sample #3 = 0.0810 (-0.0070)
Sample #4 = 0.0860 (-0.0070)
Avg % Abs = 0.0853 (-0.0068)
STD DEV = 0.0040 (0.0017)
REL STD DEV = 4.736 (28.868)

**** AUTO CAL DATA ****
<<<<< CHANNEL 1 >>>>>
Soi Val = 0.0000 mg/l or 0.000 g/210L
% Abs = 0.059
Std Dev = 0.01 Rel Std Dev = 11.97
Soi Val = 0.1905 mg/l or 0.040 g/210L
% Abs = 0.739
Std Dev = 0.02 Rel Std Dev = 3.00
Soi Val = 0.4762 mg/l or 0.100 g/210L
% Abs = 1.784
Std Dev = 0.01 Rel Std Dev = 0.37
Soi Val = 0.9524 mg/l or 0.200 g/210L
% Abs = 3.478
Std Dev = 0.00 Rel Std Dev = 0.08
Soi Val = 1.9048 mg/l or 0.400 g/210L
% Abs = 6.573
Std Dev = 0.02 Rel Std Dev = 0.26
Zero Order Coef = -110.70
First Order Coef = 2635.75
Second Order Coef = 42.25
Standard Deviation = 42.047478

<<<<< CHANNEL 2 >>>>>
Soi Val = 0.0000 mg/l or 0.000 g/210L
% Abs = 0.085
Std Dev = 0.00 Rel Std Dev = 4.74
Soi Val = 0.1905 mg/l or 0.040 g/210L
% Abs = 1.473
Std Dev = 0.01 Rel Std Dev = 0.55
Soi Val = 0.4762 mg/l or 0.100 g/210L
% Abs = 3.565
Std Dev = 0.01 Rel Std Dev = 0.35
Soi Val = 0.9524 mg/l or 0.200 g/210L
% Abs = 6.921
Std Dev = 0.01 Rel Std Dev = 0.09
Soi Val = 1.9048 mg/l or 0.400 g/210L
% Abs = 12.827
Std Dev = 0.02 Rel Std Dev = 0.16
Zero Order Coef = -1280.93
First Order Coef = 16.20
Second Order Coef = 50.657597
Standard Deviation = 50.657597

Sol Value = 0.080 g/210L ***
Fit Value = 0.3810 mg/l %%%
Samples Taken = 4, Discarded = 1
**** CHANNEL 1

Sample #1 = 3340.00
Sample #2 = 3285.00
Sample #3 = 3348.00
Sample #4 = 3330.00
Average Result = 3321.0000
STD DEV = 32.4500
REL STD DEV = 0.977

**** CHANNEL 2

Sample #1 = 3420.00
Sample #2 = 3385.00
Sample #3 = 3418.00
Sample #4 = 3442.00
Average Result = 3415.0000
STD DEV = 28.6182
REL STD DEV = 0.838

Dry Gas H2O Adjust Results *****
Barometric Pressure = 1012
3 um H2O Adjust (mg/l*10,000) = 988
9 um H2O Adjust (mg/l*10,000) = 394
**** AUTO CAL PASS

BK
SP

Stability Tests
Post-Calibration

Niceville PD

80-001310

10/4/16

NICEVILLE PD
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-001310
10/04/2016
Software: 8100.27

Test	9/210L	Time
Air Blank	0.000	11:31
Control Test	0.050	11:32
Air Blank	0.000	11:32
Control Test	0.049	11:33
Air Blank	0.000	11:33
Control Test	0.050	11:34
Air Blank	0.000	11:35
Control Test Stats		
Average	0.0497	
Std Dev	0.0006	
Rel Std Dev(%)	1.1625	

[Signature]

Operator's Signature

NICEVILLE PD
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-001310
10/04/2016
Software: 8100.27

Test	9/210L	Time
Air Blank	0.000	11:26
Control Test	0.079	11:27
Air Blank	0.000	11:28
Control Test	0.080	11:28
Air Blank	0.000	11:29
Control Test	0.079	11:29
Air Blank	0.000	11:30
Control Test Stats		
Average	0.0793	
Std Dev	0.0006	
Rel Std Dev(%)	0.7277	

[Signature]

Operator's Signature

NICEVILLE PD
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-001310
10/04/2016
Software: 8100.27

Test	9/210L	Time
Air Blank	0.000	11:40
Control Test	0.198	11:41
Air Blank	0.000	11:41
Control Test	0.198	11:42
Air Blank	0.000	11:43
Control Test	0.197	11:43
Air Blank	0.000	11:44
Control Test Stats		
Average	0.1977	
Std Dev	0.0006	
Rel Std Dev(%)	0.2921	

[Signature]

Operator's Signature

NICEVILLE PD
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-001310
10/04/2016
Software: 8100.27

Test	9/210L	Time
Air Blank	0.000	11:36
Control Test	0.081	11:36
Air Blank	0.000	11:37
Control Test	0.081	11:37
Air Blank	0.000	11:38
Control Test	0.081	11:38
Air Blank	0.000	11:38
Control Test Stats		
Average	0.0810	
Std Dev	0.0000	
Rel Std Dev(%)	0.0000	

[Signature]

Operator's Signature

SP
BK