

INSTRUMENT PROCESSING SHEET

Agency Desoto County S/N 80-001341

Date In 12/13/16 Date Out 12/14/16 Ship P/U H/D CMI EE

Intake Performed By <u>SP</u> <input type="checkbox"/> Registration <input checked="" type="checkbox"/> Annual <input type="checkbox"/> Return from CMI <input type="checkbox"/> Return from Enforcement Electronics <input type="checkbox"/> Other _____ Visual Inspection: <u>ok</u> Case <u>ok</u> Handle <u>ok</u> Dry Gas Holder <u>ok</u> Feet <u>ok</u> Keyboard/Plug <u>ok</u> Back/Plugs <u>ok</u> Screws tight <u>ok</u> Breath Hose Other Equipment: <input type="checkbox"/> Power cord <input type="checkbox"/> Printer Cable <input checked="" type="checkbox"/> Other <u>Static Bag</u> Notes: _____ _____ _____	Quality Checks Performed By <u>SP</u> <input checked="" type="checkbox"/> Breath Tube Screen <input checked="" type="checkbox"/> Replace O-Rings <input checked="" type="checkbox"/> Instrument Set Up Verified <input checked="" type="checkbox"/> R-Value <u>209</u> <input checked="" type="checkbox"/> Flow Verification (L/s) Flow Column # <u>ATP105</u> 32mm <u>.144</u> (.139 - .169) 36mm <u>.152</u> (.156 - .190) 53mm <u>.222</u> (.228 - .278) 103mm <u>0.500</u> (.447 - .547) <input checked="" type="checkbox"/> Barometric Pressure Check Gauge ID # <u>28421</u> <input checked="" type="checkbox"/> Stability Checks <table border="1" style="width:100%; border-collapse: collapse;"> <thead> <tr> <th>Simulator</th> <th>Serial #</th> <th>Lot #/Exp</th> </tr> </thead> <tbody> <tr> <td>0.05</td> <td><u>SD3962</u></td> <td><u>201603D</u> <u>3-8-18</u></td> </tr> <tr> <td>0.08</td> <td><u>SD3964</u></td> <td><u>201601F</u> <u>1-26-18</u></td> </tr> <tr> <td>0.20</td> <td><u>SD3933</u></td> <td><u>201604C</u> <u>4-5-18</u></td> </tr> <tr> <td>0.08 DGS</td> <td>N/A</td> <td><u>AG1626605</u> <u>9-22-18</u></td> </tr> </tbody> </table>	Simulator	Serial #	Lot #/Exp	0.05	<u>SD3962</u>	<u>201603D</u> <u>3-8-18</u>	0.08	<u>SD3964</u>	<u>201601F</u> <u>1-26-18</u>	0.20	<u>SD3933</u>	<u>201604C</u> <u>4-5-18</u>	0.08 DGS	N/A	<u>AG1626605</u> <u>9-22-18</u>	Flow Calibration Performed By <u>SP</u> <input type="checkbox"/> Flow Calibration N/A <input checked="" type="checkbox"/> Flow Calibration Complete Flow Column # <u>ATP102</u> <input checked="" type="checkbox"/> 5L/min - 17mm <input checked="" type="checkbox"/> 15L/min - 53mm <input checked="" type="checkbox"/> 30L/min - 103mm <input checked="" type="checkbox"/> R-Value <u>205</u> <input checked="" type="checkbox"/> Post Calibration Verification (L/s) Flow Column # <u>ATP105</u> 32mm <u>.148</u> (.139 - .169) 36mm <u>.164</u> (.156 - .190) 53mm <u>.234</u> (.228 - .278) 103mm <u>.503</u> (.447 - .547)
Simulator	Serial #	Lot #/Exp															
0.05	<u>SD3962</u>	<u>201603D</u> <u>3-8-18</u>															
0.08	<u>SD3964</u>	<u>201601F</u> <u>1-26-18</u>															
0.20	<u>SD3933</u>	<u>201604C</u> <u>4-5-18</u>															
0.08 DGS	N/A	<u>AG1626605</u> <u>9-22-18</u>															
		Maintenance Performed By _____ <input type="checkbox"/> Battery Replacement <input type="checkbox"/> Dry Gas Regulator Replacement <input type="checkbox"/> Breath Tube Replacement <input type="checkbox"/> Other _____ Suggested Service _____ _____															

RECEIVED
DEC 16 2016
FDLE
Alcohol Testing Program

Optical Bench Calibration Performed By _____ <input checked="" type="checkbox"/> Optical Bench Calibration N/A <input type="checkbox"/> Optical Bench Calibration Complete Barometric Pressure Gauge _____ ID # _____ <table border="1" style="width:100%; border-collapse: collapse;"> <thead> <tr> <th>Simulator</th> <th>Serial Number</th> <th>Lot Number</th> <th>Expiration</th> </tr> </thead> <tbody> <tr> <td>0.000</td> <td></td> <td>N/A</td> <td>N/A</td> </tr> <tr> <td>0.040</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.100</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.200</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.400</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.080 DGS</td> <td>N/A</td> <td></td> <td></td> </tr> </tbody> </table> <input type="checkbox"/> Post Calibration Stability Checks <table border="1" style="width:100%; border-collapse: collapse;"> <thead> <tr> <th>Simulator</th> <th>Serial Number</th> <th>Lot Number</th> <th>Expiration</th> </tr> </thead> <tbody> <tr> <td>0.05</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.08</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.20</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.08 DGS</td> <td>N/A</td> <td></td> <td></td> </tr> </tbody> </table>	Simulator	Serial Number	Lot Number	Expiration	0.000		N/A	N/A	0.040				0.100				0.200				0.400				0.080 DGS	N/A			Simulator	Serial Number	Lot Number	Expiration	0.05				0.08				0.20				0.08 DGS	N/A			Department Inspection Performed By <u>SP</u> <input checked="" type="checkbox"/> Barometric Pressure <u>1015</u> Gauge ID# <u>28421</u> <u>1014</u> Instrument Mouth Alcohol Solution Lot # <u>2016-A</u> Acetone Stock Solution Lot # <u>2016-B</u> <table border="1" style="width:100%; border-collapse: collapse;"> <thead> <tr> <th>Simulator</th> <th>Serial Number</th> </tr> </thead> <tbody> <tr> <td>0.00</td> <td><u>G2880</u></td> </tr> <tr> <td>Interferent</td> <td><u>G2834</u></td> </tr> <tr> <td>0.05</td> <td><u>SD3962</u></td> </tr> <tr> <td>0.08</td> <td><u>SD3964</u></td> </tr> <tr> <td>0.20</td> <td><u>SD3933</u></td> </tr> </tbody> </table> Attachments <input checked="" type="checkbox"/> Form 41 <input checked="" type="checkbox"/> Pre-Stability Tests <input checked="" type="checkbox"/> Flow Calibration <input type="checkbox"/> Optical Bench Cal <input type="checkbox"/> Post-Stability Tests <input type="checkbox"/> Other _____	Simulator	Serial Number	0.00	<u>G2880</u>	Interferent	<u>G2834</u>	0.05	<u>SD3962</u>	0.08	<u>SD3964</u>	0.20	<u>SD3933</u>
Simulator	Serial Number	Lot Number	Expiration																																																										
0.000		N/A	N/A																																																										
0.040																																																													
0.100																																																													
0.200																																																													
0.400																																																													
0.080 DGS	N/A																																																												
Simulator	Serial Number	Lot Number	Expiration																																																										
0.05																																																													
0.08																																																													
0.20																																																													
0.08 DGS	N/A																																																												
Simulator	Serial Number																																																												
0.00	<u>G2880</u>																																																												
Interferent	<u>G2834</u>																																																												
0.05	<u>SD3962</u>																																																												
0.08	<u>SD3964</u>																																																												
0.20	<u>SD3933</u>																																																												

Notes: QC PAS

<input checked="" type="checkbox"/> Instrument Complies with Chapter 11D-8, FAC <input type="checkbox"/> Instrument Does Not Comply with Chapter 11D-8, FAC <input checked="" type="checkbox"/> Return to/Place into Evidentiary Use <input type="checkbox"/> Remain Out of Evidentiary Use <input checked="" type="checkbox"/> Conduct an Agency Inspection Before Evidentiary Use

Brett H. Huntland

12/16/16

Quality Control Review

Date

STABILITY CHECKS - INSTRUMENT #80-001341 - DESOTO COUNTY - 12/14/16 SP
SD

DESOTO COUNTY SO
Intoxilyzer - Alcohol Analyzer
Model 8000
12/14/2016
Software: 8100.27
SN 80-001341

Test	g/210L	Time
Air Blank	0.000	10:04
Control Test	0.080	10:04
Air Blank	0.000	10:05
Control Test	0.079	10:05
Air Blank	0.000	10:06
Control Test	0.079	10:06
Air Blank	0.000	10:06
Control Test	0.079	10:06
Average	0.0793	
Std Dev	0.0006	
Rel Std Dev(%)	0.7277	

SP
Operator's Signature

BSK

DESOTO COUNTY SO
Intoxilyzer - Alcohol Analyzer
Model 8000
12/14/2016
Software: 8100.27
SN 80-001341

Test	g/210L	Time
Air Blank	0.000	10:14
Control Test	0.199	10:14
Air Blank	0.000	10:15
Control Test	0.199	10:15
Air Blank	0.000	10:16
Control Test	0.199	10:16
Air Blank	0.000	10:17
Control Test	0.199	10:17
Average	0.1990	
Std Dev	0.0000	
Rel Std Dev(%)	0.0000	

SP
Operator's Signature

DESOTO COUNTY SO
Intoxilyzer - Alcohol Analyzer
Model 8000
12/14/2016
Software: 8100.27
SN 80-001341

Test	g/210L	Time
Air Blank	0.000	10:08
Control Test	0.079	10:09
Air Blank	0.000	10:09
Control Test	0.080	10:10
Air Blank	0.000	10:10
Control Test	0.080	10:11
Air Blank	0.000	10:11
Control Test	0.000	10:11
Average	0.0797	
Std Dev	0.0006	
Rel Std Dev(%)	0.7247	

SP
Operator's Signature

DESOTO COUNTY SO
Intoxilyzer - Alcohol Analyzer
Model 8000
12/14/2016
Software: 8100.27
SN 80-001341

Test	g/210L	Time
Air Blank	0.000	10:18
Control Test	0.049	10:19
Air Blank	0.000	10:19
Control Test	0.050	10:20
Air Blank	0.000	10:21
Control Test	0.050	10:21
Air Blank	0.000	10:22
Control Test	0.000	10:22
Average	0.0497	
Std Dev	0.0006	
Rel Std Dev(%)	1.1625	

SP
Operator's Signature

BSK

FLOW CALIBRATION

80-001341

12/14/16

SP

DESOTO COUNTY SO
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-001341
12/14/2016
Software: 8100.27

Flow Rate Calibration*****
1: Rate (Liters/min) = 5
 SQRT(Diff)) = 7.000
2: Rate (Liters/min) = 15
 SQRT(Diff)) = 12.121
3: Rate (Liters/min) = 30
 SQRT(Diff)) = 21.793
Dependent Data Scale Factor = 100000 L/min
Independent Data Scale Factor = 256
Rounded Slope = 653
Rounded Intercept = -613487
Correlation = 0.99815

DB

BK