



Alcohol Testing Program

**INSTRUMENT PROCESSING SHEET**

Agency Savanna County SO S/N 80-001347  
Date In 5/2/16 Date Out 5/2/16  Ship  P/U  H/D  CMI  EE

**Intake** Performed By TP

Registration  
 Annual  
 Return from CMI  
 Return from Enforcement  
 Electronics  
 Other \_\_\_\_\_

Visual Inspection:  
OK Case OK Handle  
OK Dry Gas Holder OK Feet  
OK Keyboard/Plug OK Back/Plugs  
OK Screws tight OK Breath Hose

Other Equipment:  
 Power cord  
 Printer Cable  
 Other Static Bag

Notes: \_\_\_\_\_

**Quality Checks** Performed By TP

Breath Tube Screen  
 Replace O-Rings  
 Instrument Set Up Verified  
 R-Value 220  
 Flow Verification (L/s)  
 Flow Column # AP103  
 32mm 144 (.139 - .169)  
 36mm 164 (.156 - .190)  
 53mm 234 (.228 - .278)  
 103mm 520 (.447 - .547)

Barometric Pressure Check  
 Gauge ID # 28427

Stability Checks

Simulator	Serial #	Lot #/Exp
0.05	G2403	201507A 7/14/17
0.08	SD3964	201601F 1/26/18
0.20	G4444	201505A 5/12/17
0.08 DGS	N/A	AG1600504 1/5/18

**Flow Calibration** Performed By \_\_\_\_\_

Flow Calibration N/A  
 Flow Calibration Complete  
 Flow Column # \_\_\_\_\_  
 5L/min - 17mm  
 15L/min - 46mm  
 30L/min - 103mm

R-Value \_\_\_\_\_  
 Post Calibration Verification (L/s)  
 Flow Column # \_\_\_\_\_  
 32mm \_\_\_\_\_ (.139 - .169)  
 36mm \_\_\_\_\_ (.156 - .190)  
 53mm \_\_\_\_\_ (.228 - .278)  
 103mm \_\_\_\_\_ (.447 - .547)

**Maintenance** Performed By \_\_\_\_\_

Battery Replacement  
 Dry Gas Regulator Replacement  
 Breath Tube Replacement  
 Other \_\_\_\_\_

**Suggested Service**

RECEIVED  
MAY 16 2016  
FDLE  
Alcohol Testing Program

**Optical Bench Calibration** Performed By \_\_\_\_\_

Optical Bench Calibration N/A  
 Optical Bench Calibration Complete

Barometric Pressure Gauge ID # \_\_\_\_\_

Simulator	Serial Number	Lot Number	Expiration
0.000		N/A	N/A
0.040			
0.100			
0.200			
0.400			
0.080 DGS	N/A		

Post Calibration Stability Checks

Simulator	Serial Number	Lot Number	Expiration
0.05			
0.08			
0.20			
0.08 DGS	N/A		

**Department Inspection** Performed By QWS

Barometric Pressure 1018 Gauge  
 ID# 28427 1017 Instrument

Mouth Alcohol Solution Lot # 2015-A  
 Acetone Stock Solution Lot # 2016-B

Simulator	Serial Number
0.00	G2880
Interferent	G2834
0.05	G2403
0.08	SD3964
0.20	G4444

**Attachments**

<input checked="" type="checkbox"/> Form 41	<input type="checkbox"/> Optical Bench Cal
<input checked="" type="checkbox"/> Pre-Stability Tests	<input type="checkbox"/> Post-Stability Tests
<input type="checkbox"/> Flow Calibration	<input type="checkbox"/> Other _____

Notes: DA/nc ok QSPM

Instrument Complies with Chapter 11D-8, FAC  
 Instrument Does Not Comply with Chapter 11D-8, FAC

Return to/Place into Evidentiary Use  
 Remain Out of Evidentiary Use

Conduct an Agency Inspection Before Evidentiary Use

Smith MacLeland  
Quality Control Review

5/16/16  
Date

Stability Tests - Sarasota CSO # 80-001347 5/12/16

SARASOTA COUNTY SO  
Intoxilyzer - Alcohol Analyzer  
Model 8000 SN 80-001347  
05/12/2016  
Software: 8100.27

Test	9/21/0L	Time
Air Blank	0.000	10:23
Control Test	0.050	10:24
Air Blank	0.000	10:25
Control Test	0.050	10:25
Air Blank	0.000	10:26
Control Test	0.050	10:26
Air Blank	0.000	10:27
Control Test Stats		
Average	0.0500	
Std Dev	0.0000	
Rel Std Dev(%)	0.0000	

WON

SARASOTA COUNTY SO  
Intoxilyzer - Alcohol Analyzer  
Model 8000 SN 80-001347  
05/12/2016  
Software: 8100.27

Test	9/21/0L	Time
Air Blank	0.000	10:28
Control Test	0.080	10:29
Air Blank	0.000	10:29
Control Test	0.080	10:30
Air Blank	0.000	10:31
Control Test	0.080	10:31
Air Blank	0.000	10:32
Control Test Stats		
Average	0.0800	
Std Dev	0.0000	
Rel Std Dev(%)	0.0000	

SARASOTA COUNTY SO  
Intoxilyzer - Alcohol Analyzer  
Model 8000 SN 80-001347  
05/12/2016  
Software: 8100.27

Test	9/21/0L	Time
Air Blank	0.000	10:33
Control Test	0.200	10:34
Air Blank	0.000	10:34
Control Test	0.200	10:35
Air Blank	0.000	10:35
Control Test	0.200	10:36
Air Blank	0.000	10:37
Control Test Stats		
Average	0.2000	
Std Dev	0.0000	
Rel Std Dev(%)	0.0000	

SARASOTA COUNTY SO  
Intoxilyzer - Alcohol Analyzer  
Model 8000 SN 80-001347  
05/12/2016  
Software: 8100.27

Test	9/21/0L	Time
Air Blank	0.000	10:54
Control Test	0.060	10:55
Air Blank	0.000	10:55
Control Test	0.079	10:56
Air Blank	0.000	10:56
Control Test	0.079	10:56
Air Blank	0.000	10:57
Control Test Stats		
Average	0.0793	
Std Dev	0.0006	
Rel Std Dev(%)	0.7277	

DAS ASK

WON  
Operator's Signature

WON  
Operator's Signature

WON  
Operator's Signature

WON  
Operator's Signature