

INSTRUMENT PROCESSING SHEET

Agency Sarasota County SO S/N 80-001348
 Date In 3/29/16 Date Out 3/30/16 Ship P/U H/D CMI EE

Intake Performed By JP

Registration
 Annual
 Return from CMI
 Return from Enforcement
 Electronics
 Other _____

Visual Inspection:
 Case Handle
 Dry Gas Holder Feet
 Keyboard/Plug Back/Plugs
 Screws tight Breath Hose

Other Equipment:
 Power cord
 Printer Cable
 Other Static Bag

Notes: _____

Quality Checks Performed By PWS

Breath Tube Screen
 Replace O-Rings
 Instrument Set Up Verified
 R-Value 152
 Flow Verification (L/s)
 Flow Column # ATP103
 32mm 113 (.139 - .169)
 36mm 128 (.156 - .190)
 53mm 207 (.228 - .278)
 103mm 488 (.447 - .547)

Barometric Pressure Check
 Gauge ID # 28427 > 1%

Stability Checks

Simulator	Serial #	Lot #/Exp
0.05	SD1018	201507A 7/14/17
0.08	SD1011	201601F 1/26/18
0.20	<u>SD</u> G4444	201505A 5/12/17
0.08 DGS	N/A	AG 511701 4/27/17

Flow Calibration Performed By PWS

Flow Calibration N/A
 Flow Calibration Complete
 Flow Column # ATP102
 5L/min - 17mm
 15L/min - 53mm
 30L/min - 103mm
 R-Value 151
 Post Calibration Verification (L/s)
 Flow Column # ATP103
 32mm 144 (.139 - .169)
 36mm 164 (.156 - .190)
 53mm 230 (.228 - .278)
 103mm 500 (.447 - .547)

Maintenance Performed By _____

Battery Replacement
 Dry Gas Regulator Replacement
 Breath Tube Replacement
 Other _____

Suggested Service

RECEIVED
 APR 11 2016
 FDLE
 Alcohol Testing Program

Optical Bench Calibration Performed By PWS

Optical Bench Calibration N/A
 Optical Bench Calibration Complete
 Barometric Pressure Gauge 1020 ID # 26932

Simulator	Serial Number	Lot Number	Expiration
0.000	DR1275	N/A	N/A
0.040	SD3962	16101	2/2/18
0.100	G2078	15001	5/20/17
0.200	B 2408	15104	5/27/17
0.400	SD3933	15105	6/10/17
0.080 DGS	N/A	12014080A1	6/1/16

Post Calibration Stability Checks

Simulator	Serial Number	Lot Number	Expiration
0.05	SD1018	201507A	7/14/17
0.08	SD1011	201601F	1/26/18
0.20	B4444	201505A	5/12/17
0.08 DGS	N/A	AG 511701	4/27/17

Department Inspection Performed By PWS

Barometric Pressure
 ID# 28427 Gauge 1020
 Instrument 1019

Mouth Alcohol Solution Lot # 2015-A
 Acetone Stock Solution Lot # 2016-B

Simulator	Serial Number
0.00	SD1022
Interferent	SD1021
0.05	SD1018
0.08	SD1011
0.20	B4444

Attachments

Form 41
 Pre-Stability Tests
 Flow Calibration
 Optical Bench Cal
 Post-Stability Tests
 Other _____

Notes: Calibrated Optical Bench due to Barometric pressure being off by more than 1%. All measurements before were within range. PWS
QA/QC O.K. BSM

Instrument Complies with Chapter 11D-8, FAC
 Instrument Does Not Comply with Chapter 11D-8, FAC

Return to/Place into Evidentiary Use
 Remain Out of Evidentiary Use

Conduct an Agency Inspection Before Evidentiary Use

Brett Kirkland
 Quality Control Review

4/11/16
 Date

Stability Tests - Sarasota CSO # 80-001348 3/30/16
 Pre Calibration + Flow Calibration BK

SARASOTA COUNTY SO
 Intoxilyzer - Alcohol Analyzer
 Model 8000 SN 80-001348
 03/30/2016
 Software: 8100.27

Test	9/210L	Time
Air Blank	0.000	09:16
Control Test	0.050	09:19
Air Blank	0.000	09:20
Control Test	0.051	09:20
Air Blank	0.000	09:21
Control Test	0.050	09:21
Air Blank	0.000	09:22
Control Test Stats		
Average	0.0503	
Std Dev	0.0006	
Rel Std Dev(%)	1.1471	

BM

PWS

Operator's Signature

SARASOTA COUNTY SO
 Intoxilyzer - Alcohol Analyzer
 Model 8000 SN 80-001348
 03/30/2016
 Software: 8100.27

Test	9/210L	Time
Air Blank	0.000	09:10
Control Test	0.078	09:11
Air Blank	0.000	09:11
Control Test	0.079	09:12
Air Blank	0.000	09:13
Control Test	0.078	09:13
Air Blank	0.000	09:14
Control Test Stats		
Average	0.0783	
Std Dev	0.0006	
Rel Std Dev(%)	0.7370	

PWS

Operator's Signature

SARASOTA COUNTY SO
 Intoxilyzer - Alcohol Analyzer
 Model 8000 SN 80-001348
 03/30/2016
 Software: 8100.27

Test	9/210L	Time
Air Blank	0.000	09:28
Control Test	0.200	09:29
Air Blank	0.000	09:29
Control Test	0.201	09:30
Air Blank	0.000	09:30
Control Test	0.199	09:31
Air Blank	0.000	09:32
Control Test Stats		
Average	0.2000	
Std Dev	0.0010	
Rel Std Dev(%)	0.5000	

PWS

Operator's Signature

SARASOTA COUNTY SO
 Intoxilyzer - Alcohol Analyzer
 Model 8000 SN 80-001348
 03/30/2016
 Software: 8100.27

Test	9/210L	Time
Air Blank	0.000	09:24
Control Test	0.080	09:24
Air Blank	0.000	09:25
Control Test	0.081	09:25
Air Blank	0.000	09:25
Control Test	0.081	09:26
Air Blank	0.000	09:26
Control Test Stats		
Average	0.0807	
Std Dev	0.0006	
Rel Std Dev(%)	0.7157	

PWS

Operator's Signature

SARASOTA COUNTY SO
 Intoxilyzer - Alcohol Analyzer
 Model 8000 SN 80-001348
 Software: 8100.27

Flow Rate Calibration*****
 1: Rate (Liters/min) = 5
 SORT(Off) = 6.480
 2: Rate (Liters/min) = 15
 SORT(Off) = 11.531
 3: Rate (Liters/min) = 30
 SORT(Off) = 20.758
 Dependent Data Scale Factor = 10000 L/min
 Independent Data Scale Factor = 256
 Rounded Slope = 678
 Rounded Intercept = -576379
 Correlation = 0.99863

Optical Bench
Calibration
Sara Satter CSO
80-001348

3/30/16 ASK

QSM

SARASOTA COUNTY SO
Intoxilyzer - Alcohol Analyzer
Model 8000
SN 80-001348
10:28:45

Auto Calibration
Max Power Res Value = 32
Auto Range Res Value = 21

Sol Value = 0.040 g/210L
Fit Value = 0.1905 mg/l
Samples Taken = 4, Discard = 1
Sum Io = 12879, Sum Io = 14203
<<<<< CHANNEL 1 >>>>>
Sample % Abs (% Abs Ref)
Sample #1 = 0.8310 (-0.0240)
Sample #2 = 0.8340 (-0.0110)
Sample #3 = 0.8310 (-0.0080)
Sample #4 = 0.8450 (-0.0070)
Avg % Abs = 0.8357 (-0.0087)
STD DEV = 0.0074 (0.0021)
REL STD DEV = 0.881 (24.019)

<<<<< CHANNEL 2 >>>>>
Sample % Abs (% Abs Ref)
Sample #1 = 1.5240 (-0.0040)
Sample #2 = 1.5270 (0.0040)
Sample #3 = 1.5200 (-0.0070)
Sample #4 = 1.5180 (0.0090)
Avg % Abs = 1.5217 (0.0020)
STD DEV = 0.0047 (0.0082)
REL STD DEV = 0.311 (409.268)

Sol Value = 0.100 g/210L ***
Fit Value = 0.4762 mg/l %>>>
Samples Taken = 4, Discard = 1
Sum Io = 12879, Sum Io = 14199
<<<<< CHANNEL 1 >>>>>
Sample % Abs (% Abs Ref)
Sample #1 = 1.9560 (-0.0080)
Sample #2 = 1.9610 (0.0030)
Sample #3 = 1.9700 (0.0020)
Sample #4 = 1.9660 (0.0060)
Avg % Abs = 1.9657 (0.0037)
STD DEV = 0.0045 (0.0021)
REL STD DEV = 0.229 (56.773)

<<<<< CHANNEL 2 >>>>>
Sample % Abs (% Abs Ref)
Sample #1 = 3.6670 (-0.0190)
Sample #2 = 3.6580 (-0.0100)
Sample #3 = 3.6710 (-0.0140)
Sample #4 = 3.6630 (-0.0150)
Avg % Abs = 3.6648 (-0.0138)
STD DEV = 0.0081 (0.0081)
REL STD DEV = 7.959 (349.1)

210L ***
mg/l %>>>
Sum Io = 12879, Sum Io = 14200
<<<<< CHANNEL 1 >>>>>
Sample % Abs (% Abs Ref)
Sample #1 = 3.7810 (-0.0080)
Sample #2 = 3.7990 (-0.0080)
Sample #3 = 3.8020 (0.0070)
Sample #4 = 3.7870 (0.0090)
Avg % Abs = 3.7960 (0.0027)
STD DEV = 0.0079 (0.0093)
REL STD DEV = 0.203 (24.019)

Sol Value = 0.080 g/210L ***
Fit Value = 0.3810 mg/l %>>>
Samples Taken = 4, Discard = 1
Sum Io = 12879, Sum Io = 14200
<<<<< CHANNEL 1 >>>>>
Sample % Abs (% Abs Ref)
Sample #1 = 7.0400 (-0.0030)
Sample #2 = 7.0790 (-0.0070)
Sample #3 = 7.0720 (-0.0020)
Sample #4 = 7.0890 (-0.0050)
Avg % Abs = 7.0800 (-0.0047)
STD DEV = 0.0085 (0.0025)
REL STD DEV = 0.121 (53.927)

Sol Value = 0.400 g/210L ***
Fit Value = 1.9048 mg/l %>>>
Samples Taken = 4, Discard = 1
Sum Io = 12877, Sum Io = 14201
<<<<< CHANNEL 1 >>>>>
Sample % Abs (% Abs Ref)
Sample #1 = 7.2050 (-0.0120)
Sample #2 = 7.1610 (0.0080)
Sample #3 = 7.1460 (0.0140)
Sample #4 = 7.1730 (0.0180)
Avg % Abs = 7.1600 (0.0133)
STD DEV = 0.0135 (0.0050)
REL STD DEV = 0.189 (37.749)

<<<<< CHANNEL 2 >>>>>
Sample % Abs (% Abs Ref)
Sample #1 = 13.2220 (-0.0010)
Sample #2 = 13.1540 (0.0140)
Sample #3 = 13.1610 (0.0070)
Sample #4 = 13.1590 (0.0230)
Avg % Abs = 13.1588 (0.0147)
STD DEV = 0.0147 (0.0147)
REL STD DEV = 0.112 (37.749)

**** AUTO CAL DATA
<<<<< CHANNEL 1 >>>>>
Sol Val = 0.0000 mg/l or 0.000 g/210L
% Abs = 0.091
Std Dev = 0.00 Rel Std Dev = 4.15
Sol Val = 0.1905 mg/l or 0.040 g/210L
% Abs = 0.837
Std Dev = 0.01 Rel Std Dev = 0.88
Sol Val = 0.4762 mg/l or 0.100 g/210L
% Abs = 1.966
Std Dev = 0.00 Rel Std Dev = 0.23
Sol Val = 0.9524 mg/l or 0.210 g/210L
% Abs = 3.796
Std Dev = 0.01 Rel Std Dev = 0.21
Sol Val = 1.9048 mg/l or 0.400 g/210L
% Abs = 7.160
Std Dev = 0.01 Rel Std Dev = 0.19
Zero Order Coef = -189.16
First Order Coef = 2435.09
Second Order Coef = 35.00
Standard Deviation = 32.725254

<<<<< CHANNEL 2 >>>>>
Sol Val = 0.0000 mg/l or 0.000 g/210L
% Abs = 0.102
Std Dev = 0.01 Rel Std Dev = 7.96
Sol Val = 0.1905 mg/l or 0.040 g/210L
% Abs = 1.522
Std Dev = 0.00 Rel Std Dev = 0.31
Sol Val = 0.4762 mg/l or 0.100 g/210L
% Abs = 3.664
Std Dev = 0.01 Rel Std Dev = 0.18
Sol Val = 0.9524 mg/l or 0.200 g/210L
% Abs = 7.080
Std Dev = 0.01 Rel Std Dev = 0.12
Sol Val = 1.9048 mg/l or 0.400 g/210L
% Abs = 13.158
Std Dev = 0.00 Rel Std Dev = 0.03
Zero Order Coef = -86.85
First Order Coef = 1259.35
Second Order Coef = 14.75
Standard Deviation = 41.603371

**** CHANNEL 1
Sample #1 = 3118.00
Sample #2 = 3147.00
Sample #3 = 3117.00
Sample #4 = 3112.00
Average Result = 3125.3333
STD DEV = 18.9297
REL STD DEV = 0.606
**** CHANNEL 2
Sample #1 = 3409.00
Sample #2 = 3414.00
Sample #3 = 3409.00
Sample #4 = 3404.00
Average Result = 3409.0000
STD DEV = 5.6000
REL STD DEV = 0.147
**** CHANNEL 1
Dry Gas H2O Adjust Results *****
Barometric Pressure = 1019
Temperature = 10.000 = 684
Humidity = 10.000 = 400

Dratic Fit Chan 1

g/210L	g/210L	Residual
0.000	0.001	-0.0007
0.040	0.039	0.0007
0.100	0.099	0.0006
0.200	0.201	-0.0007
0.400	0.400	0.0002

Solution Stats Quadratic Fit Chan 2

Act	Fit	Residual
g/210L	g/210L	g/210L
0.000	0.001	-0.0009
0.040	0.039	0.0009
0.100	0.099	0.0008
0.200	0.201	-0.0009
0.400	0.400	0.0002

Stability Tests - Sarasota
Post Calibration

#80-001348

3/30/16

BK

SARASOTA COUNTY SO
Intoxilyzer - Alcohol Analyzer
Model: 8000 SN 80-001348
03/30/2016
Software: 8100.27

Test	9/210L	Time
Air Blank	0.000	12:55
Control Test	0.048	12:56
Air Blank	0.000	12:56
Control Test	0.049	12:57
Air Blank	0.000	12:57
Control Test	0.049	12:58
Air Blank	0.000	12:59
Control Test Stats		
Average	0.0487	
Std Dev	0.0006	
Rel Std Dev(%)	1.1663	

3041

CMW

Operator's Signature

SARASOTA COUNTY SO
Intoxilyzer - Alcohol Analyzer
Model: 8000 SN 80-001348
03/30/2016
Software: 8100.27

Test	9/210L	Time
Air Blank	0.000	13:00
Control Test	0.078	13:00
Air Blank	0.000	13:01
Control Test	0.079	13:01
Air Blank	0.000	13:02
Control Test	0.079	13:03
Air Blank	0.000	13:03
Control Test Stats		
Average	0.0787	
Std Dev	0.0006	
Rel Std Dev(%)	0.7339	

PMW

Operator's Signature

SARASOTA COUNTY SO
Intoxilyzer - Alcohol Analyzer
Model: 8000 SN 80-001348
03/30/2016
Software: 8100.27

Test	9/210L	Time
Air Blank	0.000	13:04
Control Test	0.196	13:05
Air Blank	0.000	13:06
Control Test	0.198	13:06
Air Blank	0.000	13:07
Control Test	0.199	13:08
Air Blank	0.000	13:08
Control Test Stats		
Average	0.1977	
Std Dev	0.0015	
Rel Std Dev(%)	0.7728	

PMW

Operator's Signature

SARASOTA COUNTY SO
Intoxilyzer - Alcohol Analyzer
Model: 8000 SN 80-001348
03/30/2016
Software: 8100.27

Test	9/210L	Time
Air Blank	0.000	13:10
Control Test	0.079	13:10
Air Blank	0.000	13:11
Control Test	0.080	13:11
Air Blank	0.000	13:12
Control Test	0.079	13:12
Air Blank	0.000	13:12
Control Test Stats		
Average	0.0793	
Std Dev	0.0006	
Rel Std Dev(%)	0.7277	

DAS

PMW

Operator's Signature