



Alcohol Testing Program

INSTRUMENT PROCESSING SHEET

Agency Fruitland Park PD S/N 80-001350  
Date In 2/8/16 Date Out 2/23/16  Ship  P/U  H/D  CMI  EE

**Intake** Performed By TP

Registration  
 Annual  
 Return from CMI  
 Return from Enforcement Electronics  
 Other \_\_\_\_\_

Visual Inspection:  
OK Case OK Handle  
OK Dry Gas Holder OK Feet  
OK Keyboard/Plug OK Back/Plugs  
OK Screws tight OK Breath Hose

Other Equipment:  
 Power cord  
 Printer Cable  
 Other Static Bag

Notes: \_\_\_\_\_

**Quality Checks** Performed By DMS

Breath Tube Screen  
 Replace O-Rings  
 Instrument Set Up Verified  
 R-Value 1165  
 Flow Verification (L/s)  
 Flow Column # ATP103  
 32mm 0.144 (.139 - .169)  
 36mm 0.160 (.156 - .190)  
 53mm 0.234 (.228 - .278)  
 103mm 0.496 (.447 - .547)

Barometric Pressure Check  
 Gauge ID # 26932

Stability Checks

Simulator	Serial #	Lot #/Exp
0.05	G11739	201507A 7/14/17
0.08	G8149	201502G 2/24/17
0.20	G11621	201505A 5/12/17
0.08 DGS	N/A	AG511701 4/27/17

**Flow Calibration** Performed By \_\_\_\_\_

Flow Calibration N/A  
 Flow Calibration Complete  
 Flow Column # \_\_\_\_\_  
 5L/min - 7mm  
 15L/min - 53mm  
 30L/min - 103mm

R-Value \_\_\_\_\_  
 Post Calibration Verification (L/min)  
 Flow Column # \_\_\_\_\_  
 32mm \_\_\_\_\_ (.139 - .169)  
 36mm \_\_\_\_\_ (.156 - .190)  
 53mm \_\_\_\_\_ (.228 - .278)  
 103mm \_\_\_\_\_ (.447 - .547)

**Maintenance** Performed By \_\_\_\_\_

Battery Replacement  
 Dry Gas Regulator Replacement  
 Breath Tube Replacement  
 Other \_\_\_\_\_

**Suggested Service**

RECEIVED  
FEB 23 2016  
FDLE  
Alcohol Testing Program

**Optical Bench Calibration** Performed By \_\_\_\_\_

Optical Bench Calibration N/A  
 Optical Bench Calibration Complete  
 Barometric Pressure Gauge ID # \_\_\_\_\_

Simulator	Serial Number	Lot Number	Expiration
0.000		N/A	N/A
0.040			
0.100			
0.200			
0.400			
0.080 DGS	N/A		

Post Calibration Stability Checks

Simulator	Serial Number	Lot Number	Expiration
0.05			
0.08			
0.20			
0.08 DGS	N/A		

**Department Inspection** Performed By DMS

Barometric Pressure 1012 Gauge  
 ID# 26932 1010 Instrument

Mouth Alcohol Solution Lot # 2015-A  
 Acetone Stock Solution Lot # 2015-B

Simulator	Serial Number
0.00	G2879
Interferent	G8149
0.05	G11739
0.08	G8149
0.20	G11621

**Attachments**

Form 41  Optical Bench Cal  
 Pre-Stability Tests  Post-Stability Tests  
 Flow Calibration  Other \_\_\_\_\_

Notes: QC-98

Instrument Complies with Chapter 11D-8, FAC  
 Instrument Does Not Comply with Chapter 11D-8, FAC

Return to/Place into Evidentiary Use  
 Remain Out of Evidentiary Use

Conduct an Agency Inspection Before Evidentiary Use

Patrick Murphy  
Quality Control Review

2/23/16  
Date

Stability Checks

80-001350

Fruitland Park P.D.

2/22/16 DMS

*DMS*

*DMS*

FRUITLAND PARK P.D.  
 Intoxilyzer - Alcohol Analyzer  
 Model 8000 SN 80-001350  
 02/22/2016  
 Software: 8100.27

Test	9/21/0L	Time
Air Blank	0.000	13:55
Control Test	0.049	13:56
Air Blank	0.000	13:56
Control Test	0.049	13:57
Air Blank	0.000	13:58
Control Test	0.049	13:58
Air Blank	0.000	13:59
Control Test Stats		
Average	0.0490	
Std Dev	0.0000	
Rel Std Dev(%)	0.0000	

*DMS*

Operator's Signature

FRUITLAND PARK P.D.  
 Intoxilyzer - Alcohol Analyzer  
 Model 8000 SN 80-001350  
 02/22/2016  
 Software: 8100.27

Test	9/21.0L	Time
Air Blank	0.000	14:01
Control Test	0.079	14:02
Air Blank	0.000	14:02
Control Test	0.079	14:03
Air Blank	0.000	14:04
Control Test	0.079	14:04
Air Blank	0.000	14:05
Control Test Stats		
Average	0.0790	
Std Dev	0.0000	
Rel Std Dev(%)	0.0000	

*DMS*

Operator's Signature

FRUITLAND PARK P.D.  
 Intoxilyzer - Alcohol Analyzer  
 Model 8000 SN 80-001350  
 02/22/2016  
 Software: 8100.27

Test	9/21.0L	Time
Air Blank	0.000	14:16
Control Test	0.202	14:17
Air Blank	0.000	14:18
Control Test	0.201	14:18
Air Blank	0.000	14:19
Control Test	0.201	14:20
Air Blank	0.000	14:20
Control Test Stats		
Average	0.2013	
Std Dev	0.0006	
Rel Std Dev(%)	0.2868	

*DMS*

Operator's Signature

FRUITLAND PARK P.D.  
 Intoxilyzer - Alcohol Analyzer  
 Model 8000 SN 80-001350  
 02/22/2016  
 Software: 8100.27

Test	9/21.0L	Time
Air Blank	0.000	13:51
Control Test	0.080	13:51
Air Blank	0.000	13:52
Control Test	0.080	13:52
Air Blank	0.000	13:53
Control Test	0.080	13:53
Air Blank	0.000	13:53
Control Test Stats		
Average	0.0800	
Std Dev	0.0000	
Rel Std Dev(%)	0.0000	

*988*

*DMS*

Operator's Signature