

INSTRUMENT PROCESSING SHEET

Agency Gainesville PD S/N 80-001364
 Date In 6/1/16 Date Out 6/27/16 Ship P/U H/D CMI EE

Intake Performed By (TP)

Registration
 Annual
 Return from CMI
 Return from Enforcement Electronics
 Other _____

Visual Inspection:
 Case Handle
 Dry Gas Holder Feet
 Keyboard/Plug Back/Plugs
 Screws tight Breath Hose

Other Equipment:
 Power cord
 Printer Cable
 Other Static Bag

Notes: _____

Quality Checks Performed By (JPM)

Breath Tube Screen
 Replace O-Rings
 Instrument Set Up Verified
 R-Value 204
 Flow Verification (L/s)
 Flow Column # ATP102
 32mm 1164 (.139 - .169)
 36mm 179 (.156 - .190)
 53mm 253 (.228 - .278)
 103mm 509 (.447 - .547)

Barometric Pressure Check
 Gauge ID # 28427

Stability Checks

Simulator	Serial #	Lot #/Exp
0.05	5D3962	201507A 7/14/17
0.08	503964	201601F 1/26/18
0.20	64444	201563A 5/12/17
0.08 DGS	N/A	AG605301 2/22/18

Flow Calibration Performed By _____

Flow Calibration N/A
 Flow Calibration Complete
 Flow Column # _____
 5L/min - 17mm
 15L/min - 53mm
 30L/min - 103mm

R-Value _____
 Post Calibration Verification (L/s)
 Flow Column # _____
 32mm _____ (.139 - .169)
 36mm _____ (.156 - .190)
 53mm _____ (.228 - .278)
 103mm _____ (.447 - .547)

Maintenance Performed By _____

Battery Replacement
 Dry Gas Regulator Replacement
 Breath Tube Replacement
 Other _____

Suggested Service

RECEIVED
 JUN 28 2016
 FDLE
 Alcohol Testing Program

Optical Bench Calibration Performed By _____

Optical Bench Calibration N/A
 Optical Bench Calibration Complete

Barometric Pressure Gauge ID # _____

Simulator	Serial Number	Lot Number	Expiration
0.000		N/A	N/A
0.040			
0.100			
0.200			
0.400			
0.080 DGS	N/A		

Post Calibration Stability Checks

Simulator	Serial Number	Lot Number	Expiration
0.05			
0.08			
0.20			
0.08 DGS	N/A		

Department Inspection Performed By (JPM)

Barometric Pressure 1020 Gauge
 ID# 28427 1817 Instrument

Mouth Alcohol Solution Lot # 2016-A
 Acetone Stock Solution Lot # 2016-B

Simulator	Serial Number
0.00	62880
Interferent	62834
0.05	5D3962
0.08	503964
0.20	64444

Attachments

Form 41 Optical Bench Cal
 Pre-Stability Tests Post-Stability Tests
 Flow Calibration Other _____

Notes: QC (TP) 6/27/16

Brett Wickland
 Quality Control Review

Instrument Complies with Chapter 11D-8, FAC
 Instrument Does Not Comply with Chapter 11D-8, FAC


Return to/Place into Evidentiary Use
 Remain Out of Evidentiary Use

Conduct an Agency Inspection Before Evidentiary Use

6/28/16
 Date

GAINESVILLE PD
Intoxilyzer - Alcohol Analyzer
Mode: 8000 SN 80-001364
06/24/2016
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	10:52
Control Test	0.047	10:53
Air Blank	0.000	10:53
Control Test	0.047	10:54
Air Blank	0.000	10:55
Control Test	0.048	10:55
Air Blank	0.000	10:56
Control Test Stats		
Average	0.0473	
Std Dev	0.0006	
Rel Std Dev(%)	1.2198	


Operator's Signature


GAINESVILLE PD
Intoxilyzer - Alcohol Analyzer
Mode: 8000 SN 80-001364
06/24/2016
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	10:46
Control Test	0.078	10:47
Air Blank	0.000	10:47
Control Test	0.078	10:48
Air Blank	0.000	10:49
Control Test	0.077	10:49
Air Blank	0.000	10:50
Control Test Stats		
Average	0.0777	
Std Dev	0.0006	
Rel Std Dev(%)	0.7434	


Operator's Signature

GAINESVILLE PD
Intoxilyzer - Alcohol Analyzer
Mode: 8000 SN 80-001364
06/24/2016
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	11:03
Control Test	0.199	11:04
Air Blank	0.000	11:04
Control Test	0.199	11:05
Air Blank	0.000	11:05
Control Test	0.199	11:06
Air Blank	0.000	11:07
Control Test Stats		
Average	0.1990	
Std Dev	0.0000	
Rel Std Dev(%)	0.0000	



Operator's Signature

GAINESVILLE PD
Intoxilyzer - Alcohol Analyzer
Mode: 8000 SN 80-001364
06/24/2016
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	11:14
Control Test	0.080	11:15
Air Blank	0.000	11:15
Control Test	0.080	11:16
Air Blank	0.000	11:16
Control Test	0.080	11:16
Air Blank	0.000	11:17
Control Test Stats		
Average	0.0800	
Std Dev	0.0000	
Rel Std Dev(%)	0.0000	



DGS


Operator's Signature