

INSTRUMENT PROCESSING SHEET

Agency Pinellas County SO S/N 80-001415
 Date In 1/7/16 Date Out 1/25/16 Ship P/U H/D CMI EE

Intake <input type="checkbox"/> Registration <input checked="" type="checkbox"/> Annual <input type="checkbox"/> Return from CMI <input type="checkbox"/> Return from Enforcement Electronics <input type="checkbox"/> Other _____ Visual Inspection: OK Case OK Handle OK Dry Gas Holder OK Feet OK Keyboard/Plug OK Back/Plugs OK Screws tight OK Breath Hose Other Equipment: <input type="checkbox"/> Power cord <input type="checkbox"/> Printer Cable <input checked="" type="checkbox"/> Other <u>Static Bag</u> Notes: _____ _____ _____	Quality Checks Performed By <u>DMS</u> <input checked="" type="checkbox"/> Breath Tube Screen <input checked="" type="checkbox"/> Replace O-Rings <input checked="" type="checkbox"/> Instrument Set Up Verified <input checked="" type="checkbox"/> R-Value <u>207</u> <input checked="" type="checkbox"/> Flow Verification (L/s) Flow Column # <u>ATP103</u> 32mm <u>0.148</u> (.139 - .169) 36mm <u>0.167</u> (.156 - .190) 53mm <u>0.234</u> (.228 - .278) 103mm <u>0.507</u> (.447 - .547) <input checked="" type="checkbox"/> Barometric Pressure Check Gauge ID # <u>28427</u> <input checked="" type="checkbox"/> Stability Checks <table border="1"> <thead> <tr> <th>Simulator</th> <th>Serial #</th> <th>Lot #/Exp</th> </tr> </thead> <tbody> <tr> <td>0.05</td> <td>SD1018</td> <td>201507A 7/14/17</td> </tr> <tr> <td>0.08</td> <td>SD1011</td> <td>201502G 2/24/17</td> </tr> <tr> <td>0.20</td> <td>SD1025</td> <td>201505A 5/12/17</td> </tr> <tr> <td>0.08 DGS</td> <td>N/A</td> <td>AG511701 4/27/17</td> </tr> </tbody> </table>	Simulator	Serial #	Lot #/Exp	0.05	SD1018	201507A 7/14/17	0.08	SD1011	201502G 2/24/17	0.20	SD1025	201505A 5/12/17	0.08 DGS	N/A	AG511701 4/27/17	Flow Calibration Performed By <u>DMS</u> <input checked="" type="checkbox"/> Flow Calibration N/A <input type="checkbox"/> Flow Calibration Complete Flow Column # _____ <input type="checkbox"/> 5L/min - 17mm <input type="checkbox"/> 15L/min - 53mm <input type="checkbox"/> 30L/min - 103mm <input type="checkbox"/> R-Value _____ <input type="checkbox"/> Post Calibration Verification (L/s) Flow Column # _____ 32mm _____ (.139 - .169) 36mm _____ (.156 - .190) 53mm _____ (.228 - .278) 103mm _____ (.447 - .547)
Simulator	Serial #	Lot #/Exp															
0.05	SD1018	201507A 7/14/17															
0.08	SD1011	201502G 2/24/17															
0.20	SD1025	201505A 5/12/17															
0.08 DGS	N/A	AG511701 4/27/17															
		Maintenance Performed By _____ <input type="checkbox"/> Battery Replacement <input type="checkbox"/> Dry Gas Regulator Replacement <input type="checkbox"/> Breath Tube Replacement <input type="checkbox"/> Other _____ Suggested Service _____ _____															

RECEIVED
 JAN 25 2016
 FDLE
 Alcohol Testing Program

Optical Bench Calibration Performed By _____ <input checked="" type="checkbox"/> Optical Bench Calibration N/A <input type="checkbox"/> Optical Bench Calibration Complete Barometric Pressure Gauge ID # _____ <table border="1"> <thead> <tr> <th>Simulator</th> <th>Serial Number</th> <th>Lot Number</th> <th>Expiration</th> </tr> </thead> <tbody> <tr> <td>0.000</td> <td></td> <td>N/A</td> <td>N/A</td> </tr> <tr> <td>0.040</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.100</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.200</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.400</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.080 DGS</td> <td>N/A</td> <td></td> <td></td> </tr> </tbody> </table> <input type="checkbox"/> Post Calibration Stability Checks <table border="1"> <thead> <tr> <th>Simulator</th> <th>Serial Number</th> <th>Lot Number</th> <th>Expiration</th> </tr> </thead> <tbody> <tr> <td>0.05</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.08</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.20</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.08 DGS</td> <td>N/A</td> <td></td> <td></td> </tr> </tbody> </table> Notes: <u>QC-7BK</u> _____ _____ _____	Simulator	Serial Number	Lot Number	Expiration	0.000		N/A	N/A	0.040				0.100				0.200				0.400				0.080 DGS	N/A			Simulator	Serial Number	Lot Number	Expiration	0.05				0.08				0.20				0.08 DGS	N/A			Department Inspection Performed By <u>DMS</u> <input checked="" type="checkbox"/> Barometric Pressure Gauge ID# <u>28427</u> <u>1025</u> <u>1024</u> Mouth Alcohol Solution Lot # <u>2015-A</u> Acetone Stock Solution Lot # <u>2015-B</u> <table border="1"> <thead> <tr> <th>Simulator</th> <th>Serial Number</th> </tr> </thead> <tbody> <tr> <td>0.00</td> <td>SD1020</td> </tr> <tr> <td>Interferent</td> <td>SD1021</td> </tr> <tr> <td>0.05</td> <td>SD1018</td> </tr> <tr> <td>0.08</td> <td>SD1011</td> </tr> <tr> <td>0.20</td> <td>SD1025</td> </tr> </tbody> </table> Attachments <input checked="" type="checkbox"/> Form 41 <input type="checkbox"/> Optical Bench Cal <input checked="" type="checkbox"/> Pre-Stability Tests <input type="checkbox"/> Post-Stability Tests <input type="checkbox"/> Flow Calibration <input checked="" type="checkbox"/> Other <u>Form 40</u> <input checked="" type="checkbox"/> Instrument Complies with Chapter 11D-8, FAC <input type="checkbox"/> Instrument Does Not Comply with Chapter 11D-8, FAC <input checked="" type="checkbox"/> Return to/Place into Evidentiary Use <input type="checkbox"/> Remain Out of Evidentiary Use <input checked="" type="checkbox"/> Conduct an Agency Inspection Before Evidentiary Use Date <u>1/25/16</u>	Simulator	Serial Number	0.00	SD1020	Interferent	SD1021	0.05	SD1018	0.08	SD1011	0.20	SD1025
Simulator	Serial Number	Lot Number	Expiration																																																										
0.000		N/A	N/A																																																										
0.040																																																													
0.100																																																													
0.200																																																													
0.400																																																													
0.080 DGS	N/A																																																												
Simulator	Serial Number	Lot Number	Expiration																																																										
0.05																																																													
0.08																																																													
0.20																																																													
0.08 DGS	N/A																																																												
Simulator	Serial Number																																																												
0.00	SD1020																																																												
Interferent	SD1021																																																												
0.05	SD1018																																																												
0.08	SD1011																																																												
0.20	SD1025																																																												

Arthur Murphy
 Quality Control Review

Stability Checks
 80-001415
 Pinellas County SO
 1/25/16
 RMB

PINELLAS COUNTY SO
 Intoxilyzer - Alcohol Analyzer
 Model 8000 SN 80-001415
 01/25/2016
 Software: 8100.27

RPM

Test	g/210L	Time
Air Blank	0.000	08:30
Control Test	0.048	08:30
Air Blank	0.000	08:31
Control Test	0.049	08:32
Air Blank	0.000	08:32
Control Test	0.050	08:33
Air Blank	0.000	08:33
Control Test Stats		
Average	0.0490	
Std Dev	0.0010	
Rel Std Dev(%)	2.0408	

RMB

Operator's Signature

PINELLAS COUNTY SO
 Intoxilyzer - Alcohol Analyzer
 Model 8000 SN 80-001415
 01/25/2016
 Software: 8100.27

PINELLAS COUNTY SO
 Intoxilyzer - Alcohol Analyzer
 Model 8000 SN 80-001415
 01/25/2016
 Software: 8100.27

DGS

PINELLAS COUNTY SO
 Intoxilyzer - Alcohol Analyzer
 Model 8000 SN 80-001415
 01/25/2016
 Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	08:34
Control Test	0.077	08:35
Air Blank	0.000	08:35
Control Test	0.078	08:36
Air Blank	0.000	08:37
Control Test	0.079	08:37
Air Blank	0.000	08:38
Control Test Stats		
Average	0.0780	
Std Dev	0.0010	
Rel Std Dev(%)	1.2821	

Test	g/210L	Time
Air Blank	0.000	08:39
Control Test	0.200	08:40
Air Blank	0.000	08:40
Control Test	0.202	08:41
Air Blank	0.000	08:41
Control Test	0.203	08:42
Air Blank	0.000	08:43
Control Test Stats		
Average	0.2017	
Std Dev	0.0015	
Rel Std Dev(%)	0.7574	

BSK

Test	g/210L	Time
Air Blank	0.000	08:44
Control Test	0.078	08:44
Air Blank	0.000	08:44
Control Test	0.080	08:45
Air Blank	0.000	08:45
Control Test	0.079	08:46
Air Blank	0.000	08:46
Control Test Stats		
Average	0.0790	
Std Dev	0.0010	
Rel Std Dev(%)	1.2658	

RMB

Operator's Signature

RMB

Operator's Signature

RMB

Operator's Signature