



Alcohol Testing Program

INSTRUMENT PROCESSING SHEET

Agency Orange County SO S/N 80-001418

Date In 5/16/16 Date Out 5/23/16 Ship P/U H/D CMI EE

Intake Performed By TV

Registration
 Annual
 Return from CMI
 Return from Enforcement Electronics
 Other _____

Visual Inspection:
OK Case OK Handle
OK Dry Gas Holder OK Feet
OK Keyboard/Plug OK Back/Plugs
OK Screws tight OK Breath Hose

Other Equipment:
 Power cord
 Printer Cable
 Other Static Bag

Notes: _____

Quality Checks Performed By Q

Breath Tube Screen
 Replace O-Rings
 Instrument Set Up Verified
 R-Value 208
 Flow Verification (L/s)
 Flow Column # ATP105
 32mm 151 (.139 - .169)
 36mm 182 (.156 - .190)
 53mm 257 (.228 - .278)
 103mm 519 (.447 - .547)

Barometric Pressure Check
 Gauge ID # 28427

Stability Checks

Simulator	Serial #	Lot #/Exp
0.05	G2403	201507A 7/14/17
0.08	SD3964	201601F 1/26/18
0.20	G4444	201505A 5/12/17
0.08 DGS	N/A	AG605301 2/22/18

Flow Calibration Performed By Q

Flow Calibration N/A
 Flow Calibration Complete
 Flow Column # _____
 5L/min - 17mm
 15L/min - 40mm
 30L/min - 103mm

R-Value _____
 Post Calibration Verification (L/s)
 Flow Column # _____
 32mm _____ (.139 - .169)
 36mm _____ (.156 - .190)
 53mm _____ (.228 - .278)
 103mm _____ (.447 - .547)

Maintenance Performed By _____

Battery Replacement
 Dry Gas Regulator Replacement
 Breath Tube Replacement
 Other _____

Suggested Service

RECEIVED
MAY 24 2016
FDLE
Alcohol Testing Program

Optical Bench Calibration Performed By _____

Optical Bench Calibration N/A
 Optical Bench Calibration Complete

Barometric Pressure Gauge ID # _____

Simulator	Serial Number	Lot Number	Expiration
0.000		N/A	N/A
0.040			
0.100			
0.200			
0.400			
0.080 DGS	N/A		

Post Calibration Stability Checks

Simulator	Serial Number	Lot Number	Expiration
0.05			
0.08			
0.20			
0.08 DGS	N/A		

Department Inspection Performed By Q

Barometric Pressure 1013 Gauge
 ID# 28427 1010 Instrument

Mouth Alcohol Solution Lot # 2016-A
 Acetone Stock Solution Lot # 2016-B

Simulator	Serial Number
0.00	G2880
Interferent	G2834
0.05	G2403
0.08	SD3964
0.20	G4444

Attachments

Form 41
 Pre-Stability Tests
 Flow Calibration

Optical Bench Cal
 Post-Stability Tests
 Other _____

Notes: QA/OK OK (SPM)

Instrument Complies with Chapter 11D-8, FAC
 Instrument Does Not Comply with Chapter 11D-8, FAC

Return to/Place into Evidentiary Use
 Remain Out of Evidentiary Use

Conduct an Agency Inspection Before Evidentiary Use

Ruth Whitland
Quality Control Review


5/24/16
Date

80-001418
 Stability checks
 5/23/16

ORANGE COUNTY S.O.
 Intoxilyzer - Alcohol Analyzer
 Model 8000 SN 80-001418
 05/23/2016
 Software: 8100.27


Test	g/210L	Time
Air Blank	0.000	09:25
Control Test	0.081	09:26
Air Blank	0.000	09:26
Control Test	0.081	09:26
Air Blank	0.000	09:27
Control Test	0.082	09:27
Air Blank	0.000	09:28
Control Test Stats		
Average	0.0813	
Std Dev	0.0006	
Rel Std Dev(%)	0.7099	

OGS


 Operator's Signature

ORANGE COUNTY S.O.
 Intoxilyzer - Alcohol Analyzer
 Model 8000 SN 80-001418
 05/23/2016
 Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	09:29
Control Test	0.199	09:30
Air Blank	0.000	09:30
Control Test	0.198	09:31
Air Blank	0.000	09:32
Control Test	0.200	09:32
Air Blank	0.000	09:33
Control Test Stats		
Average	0.1990	
Std Dev	0.0010	
Rel Std Dev(%)	0.5025	


 Operator's Signature

ORANGE COUNTY S.O.
 Intoxilyzer - Alcohol Analyzer
 Model 8000 SN 80-001418
 05/23/2016
 Software: 8100.27


Test	g/210L	Time
Air Blank	0.000	09:37
Control Test	0.050	09:38
Air Blank	0.000	09:38
Control Test	0.050	09:39
Air Blank	0.000	09:39
Control Test	0.050	09:40
Air Blank	0.000	09:41
Control Test Stats		
Average	0.0500	
Std Dev	0.0000	
Rel Std Dev(%)	0.0000	


 Operator's Signature

ORANGE COUNTY S.O.
 Intoxilyzer - Alcohol Analyzer
 Model 8000 SN 80-001418
 05/23/2016
 Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	09:41
Control Test	0.080	09:42
Air Blank	0.000	09:43
Control Test	0.080	09:43
Air Blank	0.000	09:44
Control Test	0.079	09:45
Air Blank	0.000	09:45
Control Test Stats		
Average	0.0797	
Std Dev	0.0006	
Rel Std Dev(%)	0.7247	

BSK


 Operator's Signature