



Alcohol Testing Program

INSTRUMENT PROCESSING SHEET

Agency Kennedy Space Center S/N 80-00140
 Date In 2/24/16 Date Out 3/9/16 Ship P/U H/D CMI EE

Intake Performed By FD

Registration
 Annual
 Return from CMI
 Return from Enforcement Electronics
 Other _____

Visual Inspection:
OK Case OK Handle
OK Dry Gas Holder OK Feet
OK Keyboard/Plug OK Back/Plugs
OK Screws tight OK Breath Hose

Other Equipment:
 Power cord
 Printer Cable
 Other _____

Notes: _____

Quality Checks Performed By Q

Breath Tube Screen
 Replace O-Rings
 Instrument Set Up Verified
 R-Value 220
 Flow Verification (L/s)
 Flow Column # A7P102
 32mm 152 (.139 - .169)
 36mm 177 (.156 - .190)
 53mm 253 (.228 - .278)
 103mm 505 (.447 - .547)

Barometric Pressure Check
 Gauge ID # 28427

Stability Checks

Simulator	Serial #	Lot #/Exp
0.05	SD1018	201507A 7/14/17
0.08	SD1011	201502G 2/24/17
0.20	G4444	201505A 5/12/17
0.08 DGS	N/A	AG6005M 1/5/18

Flow Calibration Performed By _____

Flow Calibration N/A
 Flow Calibration Complete
 Flow Column # _____
 5L/min - 17mm
 15L/min - 53mm
 30L/min - 103mm

R-Value _____
 Post Calibration Verification (L/s)
 Flow Column # _____
 32mm _____ (.139 - .169)
 36mm _____ (.156 - .190)
 53mm _____ (.228 - .278)
 103mm _____ (.447 - .547)

Maintenance Performed By _____

Battery Replacement
 Dry Gas Regulator Replacement
 Breath Tube Replacement
 Other _____

Suggested Service _____

RECEIVED
 MAR 09 2016
 FDLE
 Alcohol Testing Program

Optical Bench Calibration Performed By _____

Optical Bench Calibration N/A
 Optical Bench Calibration Complete

Barometric Pressure Gauge ID # _____

Simulator	Serial Number	Lot Number	Expiration
0.000		N/A	N/A
0.040			
0.100			
0.200			
0.400			
0.080 DGS	N/A		

Post Calibration Stability Checks

Simulator	Serial Number	Lot Number	Expiration
0.05			
0.08			
0.20			
0.08 DGS	N/A		

Department Inspection Performed By Q

Barometric Pressure Gauge ID# 28427 1023 Gauge
1023 Instrument

Mouth Alcohol Solution Lot # 2016-A
 Acetone Stock Solution Lot # 2016-B

Simulator	Serial Number
0.00	SD1022
Interferent	SD1021
0.05	SD1018
0.08	SD1011
0.20	G4444

Attachments

Form 41
 Pre-Stability Tests
 Flow Calibration

Optical Bench Cal
 Post-Stability Tests
 Other Form 41

Notes: QC - BK

Instrument Complies with Chapter 11D-8, FAC
 Instrument Does Not Comply with Chapter 11D-8, FAC

Return to/Place into Evidentiary Use
 Remain Out of Evidentiary Use

Conduct an Agency Inspection Before Evidentiary Use

Patrick Murphy
 Quality Control Review

3/9/16
 Date

80-001440
 Standard Checks
 3/7/16

INTOXILYZER 8000
 Instrument Initialization
 10:36 03/07/2016

PPM

KENNEDY SPACE CENTER
 Intoxilyzer - Alcohol Analyzer
 Model 8000 SN 80-001440
 03/07/2016
 Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	10:59
Control Test	0.078	11:00
Air Blank	0.000	11:00
Control Test	0.078	11:00
Air Blank	0.000	11:01
Control Test	0.078	11:01
Air Blank	0.000	11:02
Control Test Stats		
Average	0.0780	
Std Dev	0.0000	
Rel Std Dev(%)	0.0000	

100

Operator Signature

KENNEDY SPACE CENTER
 Intoxilyzer - Alcohol Analyzer
 Model 8000 SN 80-001440
 03/07/2016
 Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	11:03
Control Test	0.052	11:04
Air Blank	0.000	11:04
Control Test	0.052	11:05
Air Blank	0.000	11:05
Control Test	0.052	11:06
Air Blank	0.000	11:06
Control Test Stats		
Average	0.0520	
Std Dev	0.0000	
Rel Std Dev(%)	0.0000	

Operator Signature

KENNEDY SPACE CENTER
 Intoxilyzer - Alcohol Analyzer
 Model 8000 SN 80-001440
 03/07/2016
 Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	11:13
Control Test	0.081	11:13
Air Blank	0.000	11:14
Control Test	0.079	11:14
Air Blank	0.000	11:15
Control Test	0.079	11:15
Air Blank	0.000	11:16
Control Test Stats		
Average	0.0797	
Std Dev	0.0012	
Rel Std Dev(%)	1.4494	

Operator Signature

KENNEDY SPACE CENTER
 Intoxilyzer - Alcohol Analyzer
 Model 8000 SN 80-001440
 03/07/2016
 Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	11:16
Control Test	0.203	11:17
Air Blank	0.000	11:17
Control Test	0.203	11:18
Air Blank	0.000	11:19
Control Test	0.204	11:19
Air Blank	0.000	11:20
Control Test Stats		
Average	0.2133	
Std Dev	0.0006	
Rel Std Dev(%)	0.2839	

AK

Operator Signature