



INSTRUMENT PROCESSING SHEET

Agency Kennedy Space Center S/N 80-00441
Date In 5/13/16 Date Out 5/23/16 Ship P/U H/D CMI EE

Intake Performed By
Registration
Annual
Return from CMI
Return from Enforcement Electronics
Other
Visual Inspection:
Case Handle
Dry Gas Holder Feet
Keyboard/Plug Back/Plugs
Screws tight Breath Hose
Other Equipment:
Power cord
Printer Cable
Other Static Bag

Quality Checks Performed By
Breath Tube Screen
Replace O-Rings
Instrument Set Up Verified
R-Value 175
Flow Verification (L/s)
Flow Column # ATP105
32mm 153 (.139 - .169)
36mm 178 (.156 - .190)
53mm 259 (.228 - .278)
103mm 503 (.447 - .547)
Barometric Pressure Check
Gauge ID # 28427
Stability Checks
Table with Simulator, Serial #, Lot #/Exp

Flow Calibration Performed By
Flow Calibration N/A
Flow Calibration Complete
Flow Column #
5L/min 17mm
15L/min 50mm
30L/min - 103mm
R-Value
Post Calibration Verification (L/s)
Flow Column #
32mm (.139 - .169)
36mm (.156 - .190)
53mm (.228 - .278)
103mm (.447 - .547)
Maintenance Performed By
Battery Replacement
Dry Gas Regulator Replacement
Breath Tube Replacement
Other
Suggested Service

RECEIVED
MAY 24 2016
FDLE
Alcohol Testing Program

Optical Bench Calibration Performed By
Optical Bench Calibration N/A
Optical Bench Calibration Complete
Barometric Pressure Gauge ID #
Table with Simulator, Serial Number, Lot Number, Expiration
Post Calibration Stability Checks
Table with Simulator, Serial Number, Lot Number, Expiration

Department Inspection Performed By
Barometric Pressure 1013 Gauge
ID# 28427 1015 Instrument
Mouth Alcohol Solution Lot # 2016-A
Acetone Stock Solution Lot # 2016-B
Table with Simulator, Serial Number
Attachments
Form 41
Pre-Stability Tests
Flow Calibration
Optical Bench Cal
Post-Stability Tests
Other Form 40

Notes: QA/OK/OK/OK

Instrument Complies with Chapter 11D-8, FAC
Instrument Does Not Comply with Chapter 11D-8, FAC
Return to/Place into Evidentiary Use
Remain Out of Evidentiary Use
Conduct an Agency Inspection Before Evidentiary Use

Britt Kirkland
Quality Control Review

5/27/16
Date

80-001441
 Steb. liny clears
 5/23/16

INTOXILYZER 8000
 Instrument Initialization
 08:24 05/23/2016

KENNEDY SPACE CENTER
 Intoxilyzer - Alcohol Analyzer
 Model 8000 SN 80-001441
 05/23/2016
 Software: 8100.27

Test	9/210L	Time
Air Blank	0.000	09:25
Control Test	0.050	09:26
Air Blank	0.000	09:26
Control Test	0.051	09:27
Air Blank	0.000	09:28
Control Test	0.051	09:28
Air Blank	0.000	09:29
Control Test Stats	0.0507	
Average	0.0006	
Std Dev	1.1395	
Rel Std Dev(%)		

Operator's Signature

KENNEDY SPACE CENTER
 Intoxilyzer - Alcohol Analyzer
 Model 8000 SN 80-001441
 05/23/2016
 Software: 8100.27

Test	9/210L	Time
Air Blank	0.000	09:35
Control Test	0.077	09:35
Air Blank	0.000	09:36
Control Test	0.077	09:36
Air Blank	0.000	09:37
Control Test	0.078	09:37
Air Blank	0.000	09:38
Control Test Stats	0.0773	
Average	0.0016	
Std Dev	0.7466	
Rel Std Dev(%)		

Operator's Signature

KENNEDY SPACE CENTER
 Intoxilyzer - Alcohol Analyzer
 Model 8000 SN 80-001441
 05/23/2016
 Software: 8100.27

Test	9/210L	Time
Air Blank	0.000	09:39
Control Test	0.203	09:40
Air Blank	0.000	09:40
Control Test	0.202	09:41
Air Blank	0.000	09:42
Control Test	0.203	09:42
Air Blank	0.000	09:43
Control Test Stats	0.2027	
Average	0.0006	
Std Dev	0.2849	
Rel Std Dev(%)		

BK

Operator's Signature

KENNEDY SPACE CENTER
 Intoxilyzer - Alcohol Analyzer
 Model 8000 SN 80-001441
 05/23/2016
 Software: 8100.27

Test	9/210L	Time
Air Blank	0.000	09:21
Control Test	0.079	09:22
Air Blank	0.000	09:22
Control Test	0.077	09:22
Air Blank	0.000	09:23
Control Test	0.077	09:23
Air Blank	0.000	09:24
Control Test Stats	0.0777	
Average	0.0012	
Std Dev	1.4867	
Rel Std Dev(%)		

OS

QPM

Operator's Signature