

INSTRUMENT PROCESSING SHEET

Agency Clay County SO S/N 80-00443
 Date In 5/15/16 Date Out 5/23/16 Ship P/U H/D CMI EE

Intake Performed By [Signature]

Registration
 Annual
 Return from CMI
 Return from Enforcement
 Electronics
 Other _____

Visual Inspection:
OK Case OK Handle
OK Dry Gas Holder OK Feet
OK Keyboard/Plug OK Back/Plugs
OK Screws tight OK Breath Hose

Other Equipment:
 Power cord
 Printer Cable
 Other Static Bag

Notes: _____

Quality Checks Performed By [Signature]

Breath Tube Screen
 Replace O-Rings
 Instrument Set Up Verified
 R-Value 241
 Flow Verification (L/s)
 Flow Column # ATP103
 32mm 160 (.139 - .169)
 36mm 175 (.156 - .190)
 53mm 242 (.228 - .278)
 103mm 503 (.447 - .547)

Barometric Pressure Check
 Gauge ID # 28427

Stability Checks

Simulator	Serial #	Lot #/Exp
0.05	G2403	201507A 7/14/17
0.08	SD3964	201601F 1/26/18
0.20	64444	201505A 5/12/17
0.08 DGS	N/A	R6600504 1/5/18

Flow Calibration Performed By _____

Flow Calibration N/A
 Flow Calibration Complete
 Flow Column # _____
 5L/min - 17mm
 15L/min - 153mm
 30L/min - 103mm

R-Value _____
 Post Calibration Verification (L/s)
 Flow Column # _____
 32mm _____ (.139 - .169)
 36mm _____ (.156 - .190)
 53mm _____ (.228 - .278)
 103mm _____ (.447 - .547)

Maintenance Performed By _____

Battery Replacement
 Dry Gas Regulator Replacement
 Breath Tube Replacement
 Other _____

Suggested Service

RECEIVED
 MAY 24 2016
 FDLE
 Alcohol Testing Program

Optical Bench Calibration Performed By _____

Optical Bench Calibration N/A
 Optical Bench Calibration Complete
 Barometric Pressure Gauge ID # _____

Simulator	Serial Number	Lot Number	Expiration
0.000		N/A	N/A
0.040			
0.100			
0.200			
0.400			
0.080 DGS	N/A		

Post Calibration Stability Checks

Simulator	Serial Number	Lot Number	Expiration
0.05			
0.08			
0.20			
0.08 DGS	N/A		

Department Inspection Performed By [Signature]

Barometric Pressure 1016 Gauge
 ID# 28427 1014 Instrument

Mouth Alcohol Solution Lot # 2016-A
 Acetone Stock Solution Lot # 2016-B

Simulator	Serial Number
0.00	G2880
Interferent	G2834
0.05	G2403
0.08	SD3964
0.20	64444

Attachments

Form 41 Optical Bench Cal
 Pre-Stability Tests Post-Stability Tests
 Flow Calibration Other _____

Notes: OK [Signature] 5/23/16

Instrument Complies with Chapter 11D-8, FAC
 Instrument Does Not Comply with Chapter 11D-8, FAC

Return to/Place into Evidentiary Use
 Remain Out of Evidentiary Use

Conduct an Agency Inspection Before Evidentiary Use

[Signature]
 Quality Control Review

5/24/16
 Date

CLAY COUNTY SO
 Intoxilyzer - Alcohol Analyzer
 Model 8000 SN 80-001443
 05/21/2016
 Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	11:28
Control Test	0.050	11:29
Air Blank	0.000	11:30
Control Test	0.050	11:30
Air Blank	0.000	11:31
Control Test	0.049	11:31
Air Blank	0.000	11:32
Control Test Stats		
Average	0.0497	
Std Dev	0.0006	
Rel Std Dev(%)	1.1625	

P Murphy

 Operator's Signature

CLAY COUNTY SO
 Intoxilyzer - Alcohol Analyzer
 Model 8000 SN 80-001443
 05/21/2016
 Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	11:34
Control Test	0.078	11:35
Air Blank	0.000	11:35
Control Test	0.079	11:36
Air Blank	0.000	11:36
Control Test	0.079	11:37
Air Blank	0.000	11:38
Control Test Stats		
Average	0.0787	
Std Dev	0.0006	
Rel Std Dev(%)	0.7339	

P Murphy

 Operator's Signature

CLAY COUNTY SO
 Intoxilyzer - Alcohol Analyzer
 Model 8000 SN 80-001443
 05/21/2016
 Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	11:39
Control Test	0.199	11:40
Air Blank	0.000	11:40
Control Test	0.199	11:41
Air Blank	0.000	11:41
Control Test	0.200	11:42
Air Blank	0.000	11:43
Control Test Stats		
Average	0.1993	
Std Dev	0.0006	
Rel Std Dev(%)	0.2896	

P Murphy

 Operator's Signature

BK

CLAY COUNTY SO
 Intoxilyzer - Alcohol Analyzer
 Model 8000 SN 80-001443
 05/21/2016
 Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	11:44
Control Test	0.080	11:45
Air Blank	0.000	11:45
Control Test	0.079	11:45
Air Blank	0.000	11:46
Control Test	0.079	11:46
Air Blank	0.000	11:47
Control Test Stats		
Average	0.0793	
Std Dev	0.0006	
Rel Std Dev(%)	0.7277	

[Handwritten mark]

D6S

P Murphy

 Operator's Signature