



Alcohol Testing Program

INSTRUMENT PROCESSING SHEET

Agency St. Petersburg P.D.

S/N 8D-DD11653

Date In 10/11/16

Date Out 10/27/16

Ship  P/U  H/D  CMI  EE

<b>Intake</b> <input type="checkbox"/> Registration <input checked="" type="checkbox"/> Annual <input type="checkbox"/> Return from CMI <input checked="" type="checkbox"/> Return from Enforcement Electronics <input type="checkbox"/> Other _____ Visual Inspection: <u>OK</u> Case <u>OK</u> Handle <u>OK</u> Dry Gas Holder <u>OK</u> Feet <u>OK</u> Keyboard/Plug <u>OK</u> Back/Plugs <u>OK</u> Screws tight <u>OK</u> Breath Hose Other Equipment: <input type="checkbox"/> Power cord <input type="checkbox"/> Printer Cable <input checked="" type="checkbox"/> Other <u>Static Bag</u> Notes: _____ _____ _____	<b>Quality Checks</b> Performed By <u>KMS</u> <input checked="" type="checkbox"/> Breath Tube Screen <input checked="" type="checkbox"/> Replace O-Rings <input checked="" type="checkbox"/> Instrument Set Up Verified <input checked="" type="checkbox"/> R-Value <u>138</u> <input checked="" type="checkbox"/> Flow Verification (L/s) Flow Column # <u>ATP 102</u> 32mm <u>D.113</u> (.139 - .169) 36mm <u>D.140</u> (.156 - .190) 53mm <u>D.210</u> (.228 - .278) 103mm <u>D.492</u> (.447 - .547) <input checked="" type="checkbox"/> Barometric Pressure Check Gauge ID # <u>28427</u> <input checked="" type="checkbox"/> Stability Checks <table border="1"> <thead> <tr> <th>Simulator</th> <th>Serial #</th> <th>Lot #/Exp</th> </tr> </thead> <tbody> <tr> <td>0.05</td> <td>SD1018</td> <td>201507A 7/14/17</td> </tr> <tr> <td>0.08</td> <td>SD1011</td> <td>201601F 1/26/18</td> </tr> <tr> <td>0.20</td> <td>SD1025</td> <td>201604C 4/5/18</td> </tr> <tr> <td>0.08 DGS</td> <td>N/A</td> <td>AG1019605 7/14/18</td> </tr> </tbody> </table>	Simulator	Serial #	Lot #/Exp	0.05	SD1018	201507A 7/14/17	0.08	SD1011	201601F 1/26/18	0.20	SD1025	201604C 4/5/18	0.08 DGS	N/A	AG1019605 7/14/18	<b>Flow Calibration</b> Performed By <u>KMS</u> <input type="checkbox"/> Flow Calibration N/A <input checked="" type="checkbox"/> Flow Calibration Complete Flow Column # <u>ATP 105</u> <input checked="" type="checkbox"/> 5L/min - 17mm <input checked="" type="checkbox"/> 15L/min - 53mm <input type="checkbox"/> 30L/min - 103mm <input checked="" type="checkbox"/> R-Value <u>138</u> <input checked="" type="checkbox"/> Post Calibration Verification (Testing Program) Flow Column # <u>ATP 102</u> 32mm <u>D.136</u> (.139 - .169) 36mm <u>D.152</u> (.156 - .190) 53mm <u>D.214</u> (.228 - .278) 103mm <u>D.476</u> (.447 - .547) See Flow Calibrations #2 & #3 for values. <b>Maintenance</b> Performed By _____ <input type="checkbox"/> Battery Replacement <input type="checkbox"/> Dry Gas Regulator Replacement <input type="checkbox"/> Breath Tube Replacement <input type="checkbox"/> Other _____ <b>Suggested Service</b> _____ _____
Simulator	Serial #	Lot #/Exp															
0.05	SD1018	201507A 7/14/17															
0.08	SD1011	201601F 1/26/18															
0.20	SD1025	201604C 4/5/18															
0.08 DGS	N/A	AG1019605 7/14/18															

RECEIVED  
OCT 27 2016  
FDLE  
Alcohol Testing Program

<b>Optical Bench Calibration</b> Performed By _____ <input checked="" type="checkbox"/> Optical Bench Calibration N/A <input type="checkbox"/> Optical Bench Calibration Complete Barometric Pressure Gauge ID # _____ <table border="1"> <thead> <tr> <th>Simulator</th> <th>Serial Number</th> <th>Lot Number</th> <th>Expiration</th> </tr> </thead> <tbody> <tr> <td>0.000</td> <td></td> <td>N/A</td> <td>N/A</td> </tr> <tr> <td>0.040</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.100</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.200</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.400</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.080 DGS</td> <td>N/A</td> <td></td> <td></td> </tr> </tbody> </table> <input type="checkbox"/> Post Calibration Stability Checks <table border="1"> <thead> <tr> <th>Simulator</th> <th>Serial Number</th> <th>Lot Number</th> <th>Expiration</th> </tr> </thead> <tbody> <tr> <td>0.05</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.08</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.20</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.08 DGS</td> <td>N/A</td> <td></td> <td></td> </tr> </tbody> </table>	Simulator	Serial Number	Lot Number	Expiration	0.000		N/A	N/A	0.040				0.100				0.200				0.400				0.080 DGS	N/A			Simulator	Serial Number	Lot Number	Expiration	0.05				0.08				0.20				0.08 DGS	N/A		
Simulator	Serial Number	Lot Number	Expiration																																													
0.000		N/A	N/A																																													
0.040																																																
0.100																																																
0.200																																																
0.400																																																
0.080 DGS	N/A																																															
Simulator	Serial Number	Lot Number	Expiration																																													
0.05																																																
0.08																																																
0.20																																																
0.08 DGS	N/A																																															

<b>Department Inspection</b> Performed By <u>KMS</u> <input checked="" type="checkbox"/> Barometric Pressure <u>1025</u> Gauge ID# <u>28427</u> <u>1023</u> Instrument Mouth Alcohol Solution Lot # <u>2015-A</u> Acetone Stock Solution Lot # <u>2016-B</u> <table border="1"> <thead> <tr> <th>Simulator</th> <th>Serial Number</th> </tr> </thead> <tbody> <tr> <td>0.00</td> <td>SD1019</td> </tr> <tr> <td>Interferent</td> <td>SD1021</td> </tr> <tr> <td>0.05</td> <td>SD1018</td> </tr> <tr> <td>0.08</td> <td>SD1011</td> </tr> <tr> <td>0.20</td> <td>SD1025</td> </tr> </tbody> </table> <b>Attachments</b> <input checked="" type="checkbox"/> Form 41 <input checked="" type="checkbox"/> Pre-Stability Tests <input checked="" type="checkbox"/> Flow Calibration x 3 <input type="checkbox"/> Optical Bench Cal <input type="checkbox"/> Post-Stability Tests <input type="checkbox"/> Other _____	Simulator	Serial Number	0.00	SD1019	Interferent	SD1021	0.05	SD1018	0.08	SD1011	0.20	SD1025
Simulator	Serial Number											
0.00	SD1019											
Interferent	SD1021											
0.05	SD1018											
0.08	SD1011											
0.20	SD1025											

Notes: Flow calibrations performed to bring values to nominal. KMS  
\_\_\_\_\_  
QC: SP

<input checked="" type="checkbox"/> Instrument Complies with Chapter 11D-8, FAC <input type="checkbox"/> Instrument Does Not Comply with Chapter 11D-8, FAC <input checked="" type="checkbox"/> Return to/Place into Evidentiary Use <input type="checkbox"/> Remain Out of Evidentiary Use <input checked="" type="checkbox"/> Conduct an Agency Inspection Before Evidentiary Use
-------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------

Scott Kirkland  
Quality Control Review

10/27/16  
Date

# Stability Checks # 80-001653 St. Petersburg P.D. 10/27/16 EWS

ES

ST PETERSBURG PD  
Intoxilyzer - Alcohol Analyzer  
Model 8000 SN 80-001653  
10/27/2016  
Software: 8100.27

Test	9/21/0L	Time
Air Blank	0.000	10:48
Control Test	0.051	10:49
Air Blank	0.000	10:49
Control Test	0.051	10:50
Air Blank	0.000	10:51
Control Test	0.050	10:51
Air Blank	0.000	10:52
Control Test Stats		
Average	0.0507	
Std Dev	0.0006	
Rel Std Dev(%)	1.1395	

*[Signature]*

Operator's Signature

ST PETERSBURG PD  
Intoxilyzer - Alcohol Analyzer  
Model 8000 SN 80-001653  
10/27/2016  
Software: 8100.27

Test	9/21/0L	Time
Air Blank	0.000	10:54
Control Test	0.081	10:54
Air Blank	0.000	10:55
Control Test	0.081	10:56
Air Blank	0.000	10:56
Control Test	0.081	10:57
Air Blank	0.000	10:57
Control Test Stats		
Average	0.0810	
Std Dev	0.0000	
Rel Std Dev(%)	0.0000	

*[Signature]*

Operator's Signature

ST PETERSBURG PD  
Intoxilyzer - Alcohol Analyzer  
Model 8000 SN 80-001653  
10/27/2016  
Software: 8100.27

Test	9/21/0L	Time
Air Blank	0.000	10:59
Control Test	0.197	11:00
Air Blank	0.000	11:00
Control Test	0.198	11:01
Air Blank	0.000	11:01
Control Test	0.197	11:02
Air Blank	0.000	11:03
Control Test Stats		
Average	0.1973	
Std Dev	0.0006	
Rel Std Dev(%)	0.2926	

*[Signature]*

Operator's Signature

ST PETERSBURG PD  
Intoxilyzer - Alcohol Analyzer  
Model 8000 SN 80-001653  
10/27/2016  
Software: 8100.27

Test	9/21/0L	Time
Air Blank	0.000	11:05
Control Test	0.079	11:05
Air Blank	0.000	11:05
Control Test	0.080	11:06
Air Blank	0.000	11:06
Control Test	0.079	11:07
Air Blank	0.000	11:07
Control Test Stats		
Average	0.0793	
Std Dev	0.0006	
Rel Std Dev(%)	0.7277	

8P

13K

*[Signature]*

Operator's Signature

# Flow Calibrations # 80-001653 St. Petersburg P.D. 10/27/16 <sup>QWS</sup>

#1  
 ST. PETERSBURG PD  
 Intoxilyzer - Alconol Analyzer  
 Model 8000 SN 80-001653  
 10/27/2016  
 Software: 8100.27

Flow Rate Calibration\*\*\*\*\*  
 1: Rate (Liters/min) = 5  
 SORT(Diff) ) = 6.781  
 2: Rate (Liters/min) = 15  
 SORT(Diff) ) = 11.746  
 3: Rate (Liters/min) = 30  
 SORT(Diff) ) = 21.398  
 Dependent Data Scale Factor = 100000 L/min  
 Independent Data Scale Factor = 256  
 Rounded Slope = 660  
 Rounded Intercept = -581131  
 Correlation = 0.99768

*Flow rate is not correct. Previous attempts.*

Cal Column ATP105 <sup>QWS</sup>

Post  
 R-value 138  
 ATP102  
 32mm 0.136  
 36mm 0.152  
 53mm 0.214  
 103mm 0.476

Did not check

#2  
 ST. PETERSBURG PD  
 Intoxilyzer - Alconol Analyzer  
 Model 8000 SN 80-001653  
 10/27/2016  
 Software: 8100.27

Flow Rate Calibration\*\*\*\*\*  
 1: Rate (Liters/min) = 5  
 SORT(Diff) ) = 7.070  
 2: Rate (Liters/min) = 15  
 SORT(Diff) ) = 11.223  
 3: Rate (Liters/min) = 30  
 SORT(Diff) ) = 19.922  
 Dependent Data Scale Factor = 100000 L/min  
 Independent Data Scale Factor = 256  
 Rounded Slope = 747  
 Rounded Intercept = -768151  
 Correlation = 0.99626

*Flow rate is not correct. Low air tank. required pressure.*

Cal Column ATP105 <sup>QWS</sup>

Post  
 R-value 137  
 ATP102  
 32mm 0.132  
 36mm  
 53mm  
 103mm

#3  
 ST. PETERSBURG PD  
 Intoxilyzer - Alconol Analyzer  
 Model 8000 SN 80-001653  
 10/27/2016  
 Software: 8100.27

Flow Rate Calibration\*\*\*\*\*  
 1: Rate (Liters/min) = 5  
 SORT(Diff) ) = 6.633  
 2: Rate (Liters/min) = 15  
 SORT(Diff) ) = 11.832  
 3: Rate (Liters/min) = 30  
 SORT(Diff) ) = 20.902  
 Dependent Data Scale Factor = 100000 L/min  
 Independent Data Scale Factor = 256  
 Rounded Slope = 680  
 Rounded Intercept = -617891  
 Correlation = 0.99918

Cal Column ATP105 <sup>QWS</sup>

Post  
 R-value 137  
 ATP102  
 32mm 0.144  
 36mm 0.160  
 53mm 0.230  
 103mm 0.484

*QWS*



Alcohol Testing Program

INSTRUMENT PROCESSING SHEET

Agency St. Petersburg PD

S/N 80-001653

Date In 7/20/16 Date Out 7/26/16

Ship  P/U  H/D  CMI  XEE

QMS

RECEIVED  
JUL 27 2016  
FDLE  
Alcohol Testing Program

**Intake** Performed By [Signature]

Registration  
 Annual  
 Return from CMI  
 Return from Enforcement  
 Electronics  
 Other \_\_\_\_\_

Visual Inspection:  
OK Case OK Handle  
OK Dry Gas Holder OK Feet  
OK Keyboard/Plug OK Back/Plugs  
OK Screws tight OK Breath Hose

Other Equipment:  
 Power cord  
 Printer Cable  
 Other Static Bag

Notes: \_\_\_\_\_

**Quality Checks** Performed By [Signature]

Breath Tube Screen  
 Replace O-Rings  
 Instrument Set Up Verified  
 R-Value 1104  
 Flow Verification (L/s)  
 Flow Column # APP102  
 32mm 2.167 (.139 - .169)  
 36mm 0.183 (.156 - .190)  
 53mm 0.257 (.228 - .278)  
 103mm 0.523 (.447 - .547)

Barometric Pressure Check  
 Gauge ID # 28427

Stability Checks

Simulator	Serial #	Lot #/Exp
0.05	SD1018	201507A 7/14/17
0.08	SD1011	201601F 1/26/18
0.20	SD1025	201505A 5/12/17
0.08 DGS	N/A	AG102405 5/31/18

**Flow Calibration** Performed By \_\_\_\_\_

Flow Calibration N/A  
 Flow Calibration Complete  
 Flow Column # \_\_\_\_\_  
 5L/min - 17mm  
 15L/min - 53mm  
 30L/min - 103mm

R-Value \_\_\_\_\_  
 Post Calibration Verification (L/s)  
 Flow Column # \_\_\_\_\_  
 32mm \_\_\_\_\_ (.139 - .169)  
 36mm \_\_\_\_\_ (.156 - .190)  
 53mm \_\_\_\_\_ (.228 - .278)  
 103mm \_\_\_\_\_ (.447 - .547)

**Maintenance** Performed By \_\_\_\_\_

Battery Replacement  
 Dry Gas Regulator Replacement  
 Breath Tube Replacement  
 Other \_\_\_\_\_

**Suggested Service**  
Dry Gas Digital Pressure  
Sensor Reading 0-  
Send for servicing QMS

**Optical Bench Calibration** Performed By \_\_\_\_\_

Optical Bench Calibration N/A  
 Optical Bench Calibration Complete

Barometric Pressure Gauge ID # \_\_\_\_\_

Simulator	Serial Number	Lot Number	Expiration
0.000		N/A	N/A
0.040			
0.100			
0.200			
0.400			
0.080 DGS	N/A		

Post Calibration Stability Checks

Simulator	Serial Number	Lot Number	Expiration
0.05			
0.08			
0.20			
0.08 DGS	N/A		

**Department Inspection** Performed By [Signature]

Barometric Pressure ID# 28427 Gauge 1016 Instrument 1016

Mouth Alcohol Solution Lot # 2015-A  
 Acetone Stock Solution Lot # 2016-B

Simulator	Serial Number
0.00	SD1019
Interferent	SD1021
0.05	SD1018
0.08	SD1011
0.20	SD1025

**Attachments**

Form 41  
 Pre-Stability Tests  
 Flow Calibration  
 Optical Bench Cal  
 Post-Stability Tests  
 Other Form 40  
Diagnostics

Notes: [Signature] 7/27/16

Instrument Complies with Chapter 11D-8, FAC  
 Instrument Does Not Comply with Chapter 11D-8, FAC  
 Return to/Place into Evidentiary Use  
 Remain Out of Evidentiary Use  
 Conduct an Agency Inspection Before Evidentiary Use

[Signature]  
Quality Control Review

7/27/16  
Date

Stability Checks #80-001653

St. Petersburg P.D. 7/26/16 RWS

AK

ST PETERSBURG PD  
Intoxilyzer - Alcohol Analyzer  
Model 8000 SN 80-001653  
07/26/2016  
Software: 8100.27

ST PETERSBURG PD  
Intoxilyzer - Alcohol Analyzer  
Model 8000 SN 80-001653  
07/26/2016  
Software: 8100.27

ST PETERSBURG PD  
Intoxilyzer - Alcohol Analyzer  
Model 8000 SN 80-001653  
07/26/2016  
Software: 8100.27

ST PETERSBURG PD  
Intoxilyzer - Alcohol Analyzer  
Model 8000 SN 80-001653  
07/26/2016  
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	17:24
Control Test	0.080	17:25
Air Blank	0.000	17:25
Control Test	0.081	17:25
Air Blank	0.000	17:26
Control Test	0.080	17:26
Air Blank	0.000	17:27
Control Test Stats		
Average	0.0803	
Std Dev	0.0016	
Rel Std Dev(%)	0.7187	

Test	g/210L	Time
Air Blank	0.000	17:08
Control Test	0.200	17:08
Air Blank	0.000	17:09
Control Test	0.200	17:10
Air Blank	0.000	17:10
Control Test	0.197	17:11
Air Blank	0.000	17:11
Control Test Stats		
Average	0.1990	
Std Dev	0.0017	
Rel Std Dev(%)	0.8704	

Test	g/210L	Time
Air Blank	0.000	17:19
Control Test	0.080	17:20
Air Blank	0.000	17:20
Control Test	0.079	17:21
Air Blank	0.000	17:21
Control Test	0.080	17:22
Air Blank	0.000	17:23
Control Test Stats		
Average	0.0797	
Std Dev	0.0006	
Rel Std Dev(%)	0.7247	

Test	g/210L	Time
Air Blank	0.000	17:12
Control Test	0.048	17:13
Air Blank	0.000	17:13
Control Test	0.049	17:14
Air Blank	0.000	17:15
Control Test	0.049	17:15
Air Blank	0.000	17:16
Control Test Stats		
Average	0.0487	
Std Dev	0.0006	
Rel Std Dev(%)	1.1863	

Operator's Signature

Operator's Signature

Operator's Signature

Operator's Signature

Inadvertent Key Stroke  
resulting in a Diagnostics Check

# 80-001653

St. Petersburg P.D.

7/25/16

AMS

ASK

ST PETERSBURG PD  
Intoxilyzer - Alcohol Analyzer  
Model 8000 SN 80-001653  
07/25/2016  
Software: 8100.27

-----  
: DIAGNOSTICS :  
-----

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

10



Alcohol Testing Program

INSTRUMENT PROCESSING SHEET

Agency St. Petersburg PD

S/N 80-001653

Date In 2/29/16

Date Out 3/8/16

Ship  P/U  H/D  CMI  EE

Intake Performed By TP

- Registration
- Annual
- Return from CMI
- Return from Enforcement Electronics
- Other \_\_\_\_\_

Visual Inspection:

OK Case OK Handle OK Feet

OK Dry Gas Holder OK Keyboard/Plug OK Back/Plugs

OK Screws tight OK Breath Hose

- Other Equipment:
- Power cord
  - Printer Cable
  - Other \_\_\_\_\_

Notes: \_\_\_\_\_

Quality Checks Performed By RMB

- Breath Tube Screen
- Replace O-Rings
- Instrument Set Up Verified
- R-Value 150
- Flow Verification (L/s)

Flow Column # ATP 103

32mm 0.113 (.139 - .169)

36mm 0.128 (.156 - .190)

53mm 0.207 (.228 - .278)

103mm 0.484 (.447 - .547)

Simulator	Serial #	Lot #/Exp
0.05	SD1018	201507A 7/14/17
0.08	SD1011	201502G 2/24/17
0.20	<sup>RMB</sup> SG4444	201505A 5/12/17
0.08 DGS	N/A	AG1005301 2/22/18

- Barometric Pressure Check Gauge ID # 28427
- Stability Checks

Flow Calibration Performed By RMB

- Flow Calibration N/A
- Flow Calibration Complete
- Flow Column # ATP102
- 5L/min - 17mm
- 15L/min - ~~50mm~~
- 30L/min - 103mm

R-Value 147

Post Calibration Verification (L/s)

Flow Column # ATP103

32mm 0.144 (.139 - .169)

36mm 0.160 (.156 - .190)

53mm 0.230 (.228 - .278)

103mm 0.511 (.447 - .547)

Maintenance Performed By \_\_\_\_\_

- Battery Replacement
- Dry Gas Regulator Replacement
- Breath Tube Replacement
- Other \_\_\_\_\_

Suggested Service

Digital Dry Gas Sensor needs service.

Optical Bench Calibration Performed By \_\_\_\_\_

- Optical Bench Calibration N/A
- Optical Bench Calibration Complete

Barometric Pressure Gauge ID # \_\_\_\_\_

Simulator	Serial Number	Lot Number	Expiration
0.000		N/A	N/A
0.040			
0.100			
0.200			
0.400			
0.080 DGS	N/A		

Post Calibration Stability Checks

Simulator	Serial Number	Lot Number	Expiration
0.05			
0.08			
0.20			
0.08 DGS	N/A		

Notes: QC - 13K

\_\_\_\_\_

Patrick Murphy

Quality Control Review

Department Inspection Performed By RMB

- Barometric Pressure Gauge ID# 28427 1022 Gauge 1020 Instrument

Mouth Alcohol Solution Lot # 2015-A

Acetone Stock Solution Lot # 2016-B

Simulator	Serial Number
0.00	SD1022
Interferent	SD1021
0.05	SD1018
0.08	SD1011
0.20	G4444

Attachments

- Form 41
- Pre-Stability Tests
- Flow Calibration
- Optical Bench Cal
- Post-Stability Tests
- Other \_\_\_\_\_

Instrument Complies with Chapter 11D-8, FAC

Instrument Does Not Comply with Chapter 11D-8, FAC

Return to/Place into Evidentiary Use

Remain Out of Evidentiary Use

Conduct an Agency Inspection Before Evidentiary Use

3/9/16

Date

Stability Checks 80-001653 St. Petersburg P.D.

3/8/16 *RMB*

*PGM*

*80*

ST PETERSBURG PD  
Intoxilyzer - Alcohol Analyzer  
Model 8000 SN 80-001653  
03/08/2016  
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	12:40
Control Test	0.052	12:40
Air Blank	0.000	12:41
Control Test	0.051	12:41
Air Blank	0.000	12:42
Control Test	0.052	12:43
Air Blank	0.000	12:43
Control Test Stats		
Average	0.0517	
Std Dev	0.0006	
Rel Std Dev(%)	1.1175	

*RMB*

Operator's Signature

ST PETERSBURG PD  
Intoxilyzer - Alcohol Analyzer  
Model 8000 SN 80-001653  
03/08/2016  
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	12:45
Control Test	0.081	12:46
Air Blank	0.000	12:47
Control Test	0.080	12:47
Air Blank	0.000	12:48
Control Test	0.081	12:49
Air Blank	0.000	12:49
Control Test Stats		
Average	0.0807	
Std Dev	0.0006	
Rel Std Dev(%)	0.7157	

*RMB*

Operator's Signature

ST PETERSBURG PD  
Intoxilyzer - Alcohol Analyzer  
Model 8000 SN 80-001653  
03/08/2016  
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	12:55
Control Test	0.201	12:55
Air Blank	0.000	12:56
Control Test	0.200	12:56
Air Blank	0.000	12:57
Control Test	0.201	12:58
Air Blank	0.000	12:58
Control Test Stats		
Average	0.2007	
Std Dev	0.0006	
Rel Std Dev(%)	0.2877	

*RMB*

Operator's Signature

ST PETERSBURG PD  
Intoxilyzer - Alcohol Analyzer  
Model 8000 SN 80-001653  
03/08/2016  
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	12:35
Control Test	0.081	12:36
Air Blank	0.000	12:36
Control Test	0.082	12:36
Air Blank	0.000	12:37
Control Test	0.082	12:37
Air Blank	0.000	12:38
Control Test Stats		
Average	0.0817	
Std Dev	0.0006	
Rel Std Dev(%)	0.7070	

*ASK*

*RMB*

Operator's Signature



Flow Calibration

80-001653

St. Petersburg P.D.

3/8/16

DUB

ST PETERSBURG PD  
Intoxilyzer - Alcohol Analyzer  
Model 8000 SN 80-001653  
03/08/2016  
Software: 8100.27

Flow Rate Calibration\*\*\*\*\*

1: Rate (Liters/min) = 5

  SQR(Diff) ) = 7.070

2: Rate (Liters/min) = 15

  SQR(Diff) ) = 11.355

3: Rate (Liters/min) = 30

  SQR(Diff) ) = 20.195

Dependent Data Scale Factor = 100000 L/min

Independent Data Scale Factor = 256

Rounded Slope = 732

Rounded Intercept = -745372

Correlation = 0.99658

pgm

DUB

SK