



Alcohol Testing Program

INSTRUMENT PROCESSING SHEET

Agency University of FL. PD. S/N 80-001714

Date In 9/30/16 Date Out 10/14/16 Ship P/U H/D CMI EE

Intake Performed By <u>[Signature]</u> <input type="checkbox"/> Registration <input checked="" type="checkbox"/> Annual <input type="checkbox"/> Return from CMI <input type="checkbox"/> Return from Enforcement Electronics <input type="checkbox"/> Other _____ Visual Inspection: <u>ok</u> Case <u>ok</u> Handle <u>ok</u> Dry Gas Holder <u>ok</u> Feet <u>ok</u> Keyboard/Plug <u>ok</u> Back/Plugs <u>ok</u> Screws tight <u>ok</u> Breath Hose Other Equipment: <input type="checkbox"/> Power cord <input type="checkbox"/> Printer Cable <input checked="" type="checkbox"/> Other <u>Static Bag x 2</u> Notes: _____ _____ _____	Quality Checks Performed By <u>[Signature]</u> <input checked="" type="checkbox"/> Breath Tube Screen <input checked="" type="checkbox"/> Replace O-Rings <input checked="" type="checkbox"/> Instrument Set Up Verified <input checked="" type="checkbox"/> R-Value <u>179</u> <input checked="" type="checkbox"/> Flow Verification (L/s) Flow Column # <u>ATP102</u> 32mm <u>.142</u> (.139 - .169) 36mm <u>.156</u> (.156 - .190) 53mm <u>.234</u> (.228 - .278) 103mm <u>.492</u> (.447 - .547) <input checked="" type="checkbox"/> Barometric Pressure Check Gauge ID # <u>28427</u> <input checked="" type="checkbox"/> Stability Checks <table border="1"> <thead> <tr> <th>Simulator</th> <th>Serial #</th> <th>Lot #/Exp</th> </tr> </thead> <tbody> <tr> <td>0.05</td> <td>SD1018</td> <td>501507A 7/14/17</td> </tr> <tr> <td>0.08</td> <td>SD1011</td> <td>201601F 1/26/18</td> </tr> <tr> <td>0.20</td> <td>SD1025</td> <td>201604C 4/5/18</td> </tr> <tr> <td>0.08 DGS</td> <td>N/A</td> <td>AG619605 7/14/18</td> </tr> </tbody> </table>	Simulator	Serial #	Lot #/Exp	0.05	SD1018	501507A 7/14/17	0.08	SD1011	201601F 1/26/18	0.20	SD1025	201604C 4/5/18	0.08 DGS	N/A	AG619605 7/14/18	Flow Calibration Performed By _____ <input checked="" type="checkbox"/> Flow Calibration N/A <input type="checkbox"/> Flow Calibration Complete Flow Column # _____ <input type="checkbox"/> 5L/min - 17mm <input type="checkbox"/> 15L/min - 53mm <input type="checkbox"/> 30L/min - 103mm <input type="checkbox"/> R-Value _____ <input type="checkbox"/> Post Calibration Verification (L/s) Flow Column # _____ 32mm _____ (.139 - .169) 36mm _____ (.156 - .190) 53mm _____ (.228 - .278) 103mm _____ (.447 - .547) Maintenance Performed By _____ <input type="checkbox"/> Battery Replacement <input type="checkbox"/> Dry Gas Regulator Replacement <input type="checkbox"/> Breath Tube Replacement <input type="checkbox"/> Other _____ Suggested Service _____ _____
Simulator	Serial #	Lot #/Exp															
0.05	SD1018	501507A 7/14/17															
0.08	SD1011	201601F 1/26/18															
0.20	SD1025	201604C 4/5/18															
0.08 DGS	N/A	AG619605 7/14/18															

RECEIVED
OCT 14 2016
FDLE
Alcohol Testing Program

Optical Bench Calibration Performed By _____ <input checked="" type="checkbox"/> Optical Bench Calibration N/A <input type="checkbox"/> Optical Bench Calibration Complete Barometric Pressure Gauge ID # _____ <table border="1"> <thead> <tr> <th>Simulator</th> <th>Serial Number</th> <th>Lot Number</th> <th>Expiration</th> </tr> </thead> <tbody> <tr> <td>0.000</td> <td></td> <td>N/A</td> <td>N/A</td> </tr> <tr> <td>0.040</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.100</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.200</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.400</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.080 DGS</td> <td>N/A</td> <td></td> <td></td> </tr> </tbody> </table> <input type="checkbox"/> Post Calibration Stability Checks <table border="1"> <thead> <tr> <th>Simulator</th> <th>Serial Number</th> <th>Lot Number</th> <th>Expiration</th> </tr> </thead> <tbody> <tr> <td>0.05</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.08</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.20</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.08 DGS</td> <td>N/A</td> <td></td> <td></td> </tr> </tbody> </table>	Simulator	Serial Number	Lot Number	Expiration	0.000		N/A	N/A	0.040				0.100				0.200				0.400				0.080 DGS	N/A			Simulator	Serial Number	Lot Number	Expiration	0.05				0.08				0.20				0.08 DGS	N/A			Department Inspection Performed By <u>[Signature]</u> <input checked="" type="checkbox"/> Barometric Pressure <u>1022</u> Gauge ID# <u>28427</u> <u>1020</u> Instrument Mouth Alcohol Solution Lot # <u>2016-A</u> Acetone Stock Solution Lot # <u>2016-B</u> <table border="1"> <thead> <tr> <th>Simulator</th> <th>Serial Number</th> </tr> </thead> <tbody> <tr> <td>0.00</td> <td>SD1019</td> </tr> <tr> <td>Interferent</td> <td>SD1021</td> </tr> <tr> <td>0.05</td> <td>SD1018</td> </tr> <tr> <td>0.08</td> <td>SD1011</td> </tr> <tr> <td>0.20</td> <td>SD1025</td> </tr> </tbody> </table> Attachments <input checked="" type="checkbox"/> Form 41 <input type="checkbox"/> Optical Bench Cal <input checked="" type="checkbox"/> Pre-Stability Tests <input type="checkbox"/> Post-Stability Tests <input type="checkbox"/> Flow Calibration <input type="checkbox"/> Other _____ _____	Simulator	Serial Number	0.00	SD1019	Interferent	SD1021	0.05	SD1018	0.08	SD1011	0.20	SD1025
Simulator	Serial Number	Lot Number	Expiration																																																										
0.000		N/A	N/A																																																										
0.040																																																													
0.100																																																													
0.200																																																													
0.400																																																													
0.080 DGS	N/A																																																												
Simulator	Serial Number	Lot Number	Expiration																																																										
0.05																																																													
0.08																																																													
0.20																																																													
0.08 DGS	N/A																																																												
Simulator	Serial Number																																																												
0.00	SD1019																																																												
Interferent	SD1021																																																												
0.05	SD1018																																																												
0.08	SD1011																																																												
0.20	SD1025																																																												

Notes: QC:SP

 Instrument Complies with Chapter 11D-8, FAC
 Instrument Does Not Comply with Chapter 11D-8, FAC
 Return to/Place into Evidentiary Use
 Remain Out of Evidentiary Use
 Conduct an Agency Inspection Before Evidentiary Use

Brett Kirkland

Quality Control Review

10/14/16
Date

UNIVERSITY OF FL PD
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-001714
10/14/2016
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	10:24
Control Test	0.050	10:25
Air Blank	0.000	10:25
Control Test	0.050	10:26
Air Blank	0.000	10:26
Control Test	0.050	10:27
Air Blank	0.000	10:28
Control Test Stats		
Average	0.0500	
Std Dev	0.0000	
Rel Std Dev(%)	0.0000	

UNIVERSITY OF FL PD
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-001714
10/14/2016
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	10:18
Control Test	0.081	10:19
Air Blank	0.000	10:19
Control Test	0.080	10:20
Air Blank	0.000	10:20
Control Test	0.081	10:21
Air Blank	0.000	10:21
Control Test Stats		
Average	0.0807	
Std Dev	0.0006	
Rel Std Dev(%)	0.7157	

UNIVERSITY OF FL PD
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-001714
10/14/2016
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	10:30
Control Test	0.196	10:31
Air Blank	0.000	10:31
Control Test	0.197	10:32
Air Blank	0.000	10:32
Control Test	0.197	10:33
Air Blank	0.000	10:34
Control Test Stats		
Average	0.1967	
Std Dev	0.0006	
Rel Std Dev(%)	0.2936	

BK

P Murphy
Operator's Signature

P Murphy
Operator's Signature

P Murphy
Operator's Signature

UNIVERSITY OF FL PD
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-001714
10/14/2016
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	10:40
Control Test	0.079	10:40
Air Blank	0.000	10:41
Control Test	0.078	10:41
Air Blank	0.000	10:42
Control Test	0.079	10:42
Air Blank	0.000	10:43
Control Test Stats		
Average	0.0787	
Std Dev	0.0006	
Rel Std Dev(%)	0.7339	

SP

DGS

P Murphy
Operator's Signature