



INSTRUMENT PROCESSING SHEET

Agency Clewiston PD S/N 80-001740
Date In 1/14/16 Date Out 1/25/16 Ship P/U H/D CMI EE

Intake Performed By TP

Registration
 Annual
 Return from CMI
 Return from Enforcement Electronics
 Other _____

Visual Inspection:
OK Case OK Handle
OK Dry Gas Holder OK Feet
OK Keyboard/Plug OK Back/Plugs
OK Screws tight OK Breath Hose

Other Equipment:
 Power cord
 Printer Cable
 Other _____

Notes: _____

Quality Checks Performed By BUS

Breath Tube Screen
 Replace O-Rings
 Instrument Set Up Verified
 R-Value 169
 Flow Verification (L/s)
 Flow Column # ATP.103
 32mm D.144 (.139 - .169)
 36mm D.164 (.156 - .190)
 53mm D.234 (.228 - .278)
 103mm D.503 (.447 - .547)

Barometric Pressure Check
 Gauge ID # 28427

Stability Checks

Simulator	Serial #	Lot #/Exp
0.05	SD1018	201507A 7/14/17
0.08	SD1011	201502G 2/24/17
0.20	SD1025	201505A 5/12/17
0.08 DGS	N/A	A8511701 4/27/17

Flow Calibration Performed By BUS

Flow Calibration N/A
 Flow Calibration Complete
 Flow Column # _____
 5L/min - 17mm
 15L/min - 53mm
 30L/min - 103mm

R-Value _____
 Post Calibration Verification (L/s)
 Flow Column # _____
 32mm _____ (.139 - .169)
 36mm _____ (.156 - .190)
 53mm _____ (.228 - .278)
 103mm _____ (.447 - .547)

Maintenance Performed By _____

Battery Replacement
 Dry Gas Regulator Replacement
 Breath Tube Replacement
 Other _____

Suggested Service

RECEIVED
JAN 26 2016
FDLE
Alcohol Testing Program

Optical Bench Calibration Performed By _____

Optical Bench Calibration N/A
 Optical Bench Calibration Complete

Barometric Pressure Gauge ID # _____

Simulator	Serial Number	Lot Number	Expiration
0.000		N/A	N/A
0.040			
0.100			
0.200			
0.400			
0.080 DGS	N/A		

Post Calibration Stability Checks

Simulator	Serial Number	Lot Number	Expiration
0.05			
0.08			
0.20			
0.08 DGS	N/A		

Department Inspection Performed By BUS

Barometric Pressure 1024 Gauge
 ID# 28427 1022 Instrument

Mouth Alcohol Solution Lot # 2015-A
 Acetone Stock Solution Lot # 2015-B

Simulator	Serial Number
0.00	SD1022
Interferent	SD1021
0.05	SD1018
0.08	SD1011
0.20	SD1025

Attachments

Form 41
 Pre-Stability Tests
 Flow Calibration

Optical Bench Cal
 Post-Stability Tests
 Other _____

Notes: QC-13K

Instrument Complies with Chapter 11D-8, FAC
 Instrument Does Not Comply with Chapter 11D-8, FAC

Return to/Place into Evidentiary Use
 Remain Out of Evidentiary Use

Conduct an Agency Inspection Before Evidentiary Use

Patrick Murphy
Quality Control Review

1/26/16
Date

Stability Checks 80-001740 Clearwater P.D. 1/25/16 *DMS*

Q9M

DMS

CLEARWATER PD
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-001740
01/25/2016
Software: 8100.27

CLEARWATER PD
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-001740
01/25/2016
Software: 8100.27

CLEARWATER PD
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-001740
01/25/2016
Software: 8100.27

CLEARWATER PD
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-001740
01/25/2016
Software: 8100.27

Test	9/21/0L	Time	Test	9/21/0L	Time	Test	9/21/0L	Time	Test	9/21/0L	Time
Air Blank	0.000	08:33	Air Blank	0.000	08:38	Air Blank	0.000	08:43	Air Blank	0.000	08:55
Control Test	0.051	08:34	Control Test	0.081	08:39	Control Test	0.203	08:44	Control Test	0.081	08:55
Air Blank	0.000	08:35	Air Blank	0.000	08:39	Air Blank	0.000	08:44	Air Blank	0.000	08:56
Control Test	0.052	08:35	Control Test	0.081	08:40	Control Test	0.203	08:45	Control Test	0.082	08:56
Air Blank	0.000	08:36	Air Blank	0.000	08:41	Air Blank	0.000	08:45	Air Blank	0.000	08:56
Control Test	0.051	08:36	Control Test	0.080	08:41	Control Test	0.204	08:46	Control Test	0.081	08:57
Air Blank	0.000	08:37	Air Blank	0.000	08:42	Air Blank	0.000	08:47	Air Blank	0.000	08:57
Control Test Stats			Control Test Stats			Control Test Stats			Control Test Stats		
Average	0.0513		Average	0.0807		Average	0.2033		Average	0.0813	
Std Dev	0.0006		Std Dev	0.0006		Std Dev	0.0006		Std Dev	0.0006	
Rel Std Dev(%)	1.1247		Rel Std Dev(%)	0.7157		Rel Std Dev(%)	0.2839		Rel Std Dev(%)	0.7199	

AK

DMS
Operator's Signature

DMS
Operator's Signature

DMS
Operator's Signature

DMS
Operator's Signature