



Alcohol Testing Program

INSTRUMENT PROCESSING SHEET

RECEIVED

Agency Manatee County SO S/N 80-001742

Date In 7/1/16 Date Out 7/5/16

Ship P/U H/D CMI EEE

JUL 06 2016

FDLE

Intake Performed By DR

- Registration
- Annual
- Return from CMI
- Return from Enforcement Electronics
- Other _____

Visual Inspection:

Case OK Handle OK

Dry Gas Holder OK Feet OK

Keyboard/Plug OK Back/Plugs OK

Screws tight OK Breath Hose OK

Other Equipment:

- Power cord
- Printer Cable
- Other Static Bag

Notes: _____

Quality Checks Performed By DWS

- Breath Tube Screen
- Replace O-Rings
- Instrument Set Up Verified
- R-Value 206
- Flow Verification (L/s)

Flow Column #	<u>ATP102</u>
32mm	<u>152</u> (.139 - .169)
36mm	<u>171</u> (.156 - .190)
53mm	<u>238</u> (.228 - .278)
103mm	<u>503</u> (.447 - .547)

- Barometric Pressure Check Gauge ID # 28427
- Stability Checks

Simulator	Serial #	Lot #/Exp
0.05	SD1018	201507A 7/14/17
0.08	SD1011	201601F 1/26/18
0.20	SD1025	201505A 5/12/17
0.08 DGS	N/A	AG600SD4 1/5/18

Flow Calibration Performed By DWS

- Flow Calibration N/A
- Flow Calibration Complete
- Flow Column # _____
- 5L/min - 17mm
- 15L/min - 53mm
- 30L/min - 103mm
- R-Value _____
- Post Calibration Verification (L/s)
- Flow Column # _____
- 32mm _____ (.139 - .169)
- 36mm _____ (.156 - .190)
- 53mm _____ (.228 - .278)
- 103mm _____ (.447 - .547)

Maintenance Performed By _____

- Battery Replacement
- Dry Gas Regulator Replacement
- Breath Tube Replacement
- Other _____

Suggested Service _____

Optical Bench Calibration Performed By _____

- Optical Bench Calibration N/A
- Optical Bench Calibration Complete

Barometric Pressure Gauge ID # _____

Simulator	Serial Number	Lot Number	Expiration
0.000		N/A	N/A
0.040			
0.100			
0.200			
0.400			
0.080 DGS	N/A		

Post Calibration Stability Checks

Simulator	Serial Number	Lot Number	Expiration
0.05			
0.08			
0.20			
0.08 DGS	N/A		

Notes: OC-DWS

Department Inspection Performed By DWS

- Barometric Pressure ID# 28427 Gauge 1020 Instrument 1016

Mouth Alcohol Solution Lot # 2015-A

Acetone Stock Solution Lot # 2016-B

Simulator	Serial Number
0.00	SD1019
Interferent	SD1021
0.05	SD1018
0.08	SD1011
0.20	SD1025

Attachments

- Form 41
- Pre-Stability Tests
- Flow Calibration
- Optical Bench Cal
- Post-Stability Tests
- Other _____

- Instrument Complies with Chapter 11D-8, FAC
- Instrument Does Not Comply with Chapter 11D-8, FAC
- Return to/Place Into Evidentiary Use
- Remain Out of Evidentiary Use
- Conduct an Agency Inspection Before Evidentiary Use

Brett Kuitland

Quality Control Review

7/5/16

Date

Stability Tests - Manatee CSD # 80-001742 7/5/16

MANATEE COUNTY SO
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-001742
07/05/2016
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	10:32
Control Test	0.049	10:33
Air Blank	0.000	10:33
Control Test	0.049	10:34
Air Blank	0.000	10:35
Control Test	0.049	10:35
Air Blank	0.000	10:36
Control Test Stats		
Average	0.0490	
Std Dev	0.0000	
Rel Std Dev(%)	0.0000	

gws
Operator's Signature

MANATEE COUNTY SO
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-001742
07/05/2016
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	10:38
Control Test	0.079	10:39
Air Blank	0.000	10:39
Control Test	0.080	10:40
Air Blank	0.000	10:40
Control Test	0.080	10:41
Air Blank	0.000	10:42
Control Test Stats		
Average	0.0797	
Std Dev	0.0006	
Rel Std Dev(%)	0.7247	

gws
Operator's Signature

MANATEE COUNTY SO
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-001742
07/05/2016
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	10:43
Control Test	0.203	10:44
Air Blank	0.000	10:45
Control Test	0.203	10:45
Air Blank	0.000	10:46
Control Test	0.204	10:46
Air Blank	0.000	10:47
Control Test Stats		
Average	0.2033	
Std Dev	0.0006	
Rel Std Dev(%)	0.2839	

gws
Operator's Signature

MANATEE COUNTY SO
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-001742
07/05/2016
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	10:48
Control Test	0.080	10:49
Air Blank	0.000	10:49
Control Test	0.080	10:50
Air Blank	0.000	10:50
Control Test	0.079	10:50
Air Blank	0.000	10:51
Control Test Stats		
Average	0.0797	
Std Dev	0.0006	
Rel Std Dev(%)	0.7247	

gws

gws
Operator's Signature

BSK

DGS