

INSTRUMENT PROCESSING SHEET

Agency Jacksonville SO S/N 80-001743
 Date In 11/14/16 Date Out 12/5/16 Ship P/U H/D CMI EE

Intake Performed By [Signature]

Registration
 Annual
 Return from CMI
 Return from Enforcement
 Electronics
 Other _____

Visual Inspection:
ok Case ok Handle
ok Dry Gas Holder ok Feet
ok Keyboard/Plug ok Back/Plugs
ok Screws tight ok Breath Hose

Other Equipment:
 Power cord
 Printer Cable
 Other Static Bag

Notes: _____

Quality Checks Performed By [Signature]

Breath Tube Screen
 Replace O-Rings
 Instrument Set Up Verified
 R-Value 219
 Flow Verification (L/s)
 Flow Column # A7P102
 32mm 160 (.139 - .169)
 36mm 182 (.156 - .190)
 53mm 255 (.228 - .278)
 103mm 516 (.447 - .547)

Barometric Pressure Check
 Gauge ID # 28427

Stability Checks

Simulator	Serial #	Lot #/Exp
0.05	DR2035	201507A 7/14/17
0.08	SD1011	201601F 1/26/18
0.20	SD1025	201604C 4/5/18
0.08 DGS	N/A	AG609605 7/14/18

Flow Calibration Performed By _____

Flow Calibration N/A
 Flow Calibration Complete
 Flow Column # _____
 5L/min - 17mm
 15L/min - 453mm
 30L/min - 103mm

R-Value _____
 Post Calibration Verification (L/s) Program
 Flow Column # _____
 32mm _____ (.139 - .169)
 36mm _____ (.156 - .190)
 53mm _____ (.228 - .278)
 103mm _____ (.447 - .547)

Maintenance Performed By _____

Battery Replacement
 Dry Gas Regulator Replacement
 Breath Tube Replacement
 Other _____

Suggested Service

RECEIVED
 DEC 05 2016
 FDLE
 Alcohol Testing Program

Optical Bench Calibration Performed By _____

Optical Bench Calibration N/A
 Optical Bench Calibration Complete
 Barometric Pressure Gauge _____ ID # _____

Simulator	Serial Number	Lot Number	Expiration
0.000		N/A	N/A
0.040			
0.100			
0.200			
0.400			
0.080 DGS	N/A		

Post Calibration Stability Checks

Simulator	Serial Number	Lot Number	Expiration
0.05			
0.08			
0.20			
0.08 DGS	N/A		

Department Inspection Performed By [Signature]

Barometric Pressure 1014 Gauge
 ID# 28427 1012 Instrument

Mouth Alcohol Solution Lot # 2016-A
 Acetone Stock Solution Lot # 2016-B

Simulator	Serial Number
0.00	SD1019
Interferent	SD1021
0.05	DR2035
0.08	SD1011
0.20	SD1025

Attachments

Form 41
 Pre-Stability Tests
 Flow Calibration
 Optical Bench Cal
 Post-Stability Tests
 Other _____

Notes: QC & DB

Instrument Complies with Chapter 11D-8, FAC
 Instrument Does Not Comply with Chapter 11D-8, FAC
 Return to/Place into Evidentiary Use
 Remain Out of Evidentiary Use
 Conduct an Agency Inspection Before Evidentiary Use

[Signature]
 Quality Control Review

[Signature] 12/5/16
 Date

JACKSONVILLE SO
 Intoxilyzer - Alcohol Analyzer
 Model 8000 SN 80-001743
 12/05/2016
 Software: 8100.27

Test	9/210L	Time
Air Blank	0.000	09:06
Control Test	0.078	09:06
Air Blank	0.000	09:07
Control Test	0.079	09:08
Air Blank	0.000	09:08
Control Test	0.078	09:09
Air Blank	0.000	09:09
Control Test Stats	0.0783	
Average	0.0006	
Std Dev	0.7370	
Rel Std Dev(%)		

Operator's Signature

JACKSONVILLE SO
 Intoxilyzer - Alcohol Analyzer
 Model 8000 SN 80-001743
 12/05/2016
 Software: 8100.27

Test	9/210L	Time
Air Blank	0.000	09:16
Control Test	0.048	09:16
Air Blank	0.000	09:17
Control Test	0.048	09:17
Air Blank	0.000	09:18
Control Test	0.048	09:19
Air Blank	0.000	09:19
Control Test Stats		
Average	0.0480	
Std Dev	0.0000	
Rel Std Dev(%)	0.0000	

Operator's Signature

JACKSONVILLE SO
 Intoxilyzer - Alcohol Analyzer
 Model 8000 SN 80-001743
 12/05/2016
 Software: 8100.27

Test	9/210L	Time
Air Blank	0.000	09:21
Control Test	0.199	09:21
Air Blank	0.000	09:22
Control Test	0.198	09:22
Air Blank	0.000	09:23
Control Test	0.199	09:24
Air Blank	0.000	09:24
Control Test Stats		
Average	0.1987	
Std Dev	0.0006	
Rel Std Dev(%)	0.2906	

TSK

Operator's Signature

JACKSONVILLE SO
 Intoxilyzer - Alcohol Analyzer
 Model 8000 SN 80-001743
 12/05/2016
 Software: 8100.27

Test	9/210L	Time
Air Blank	0.000	09:02
Control Test	0.081	09:02
Air Blank	0.000	09:03
Control Test	0.079	09:03
Air Blank	0.000	09:04
Control Test	0.080	09:04
Air Blank	0.000	09:04
Control Test Stats		
Average	0.0800	
Std Dev	0.0010	
Rel Std Dev(%)	1.2500	

DS

Operator's Signature

1005