



Alcohol Testing Program

INSTRUMENT PROCESSING SHEET

Agency Miami Beach Police Department S/N 80-002232

Date In 9/7/2016 Date Out 9/23/2016 Ship P/U H/D CMI EE

Intake Performed By DELL

Registration
 Annual
 Return from CMI
 Return from Enforcement
 Electronics
 Other _____

Visual Inspection:
OK Case OK Handle
OK Dry Gas Holder OK Feet
OK Keyboard/Plug OK Back/Plugs
OK Screws tight OK Breath Hose

Other Equipment:
 Power cord
 Printer Cable
 Other ANTI STATIC BAG

Notes: _____

Quality Checks Performed By DELL

Breath Tube Screen
 Replace O-Rings
 Instrument Set Up Verified
 R-Value 232
 Flow Verification (L/s)
 Flow Column # ATP 101
 32mm 148 (.139 - .169)
 36mm 164 (.156 - .190)
 53mm 234 (.228 - .278)
 103mm 496 (.447 - .547)

Barometric Pressure Check
 Gauge ID # 28663
 Stability Checks

Simulator	Serial #	Lot #/Exp
0.05	SD3967	201507A 07/14/2017
0.08	SD3968	201601F 01/26/2018
0.20	SD3969	201505A 05/12/2017
0.08 DGS	N/A	A6600504 01/05/2018

Flow Calibration Performed By DELL

Flow Calibration N/A
 Flow Calibration Complete
 Flow Column # _____
 5L/min - 17mm
 15L/min - 53mm
 30L/min - 103mm
 R-Value _____
 Post Calibration Verification (L/s)
 Flow Column # _____
 32mm _____ (.139 - .169)
 36mm _____ (.156 - .190)
 53mm _____ (.228 - .278)
 103mm _____ (.447 - .547)

Maintenance Performed By _____

Battery Replacement
 Dry Gas Regulator Replacement
 Breath Tube Replacement
 Other _____

Suggested Service

RECEIVED
SEP 23 2016
FDLE
Alcohol Testing Program

Optical Bench Calibration Performed By DELL

Optical Bench Calibration N/A
 Optical Bench Calibration Complete
 Barometric Pressure Gauge 1016 ID # 28663

Simulator	Serial Number	Lot Number	Expiration
0.000	2235	N/A	N/A
0.040	2236	16101	02-02-2018
0.100	2237	15001	05-20-2017
0.200	2238	15104	05-27-2017
0.400	2239	16102	03-22-2018
0.080 DGS	N/A	0341508DA1	03-05-2017

Post Calibration Stability Checks

Simulator	Serial Number	Lot Number	Expiration
0.05	SD3967	201507A	07/14/2017
0.08	SD3968	201601F	01/26/2018
0.20	SD3963	201505A	05/12/2017
0.08 DGS	N/A	A6600504	01/05/2018

Department Inspection Performed By DELL

Barometric Pressure 1016-1016 Gauge
 ID# 28199-28199 1011-1016 Instrument

Mouth Alcohol Solution Lot # 2016-A
 Acetone Stock Solution Lot # 2016-B

Simulator	Serial Number
0.00	SD 3965 - SD 3965
Interferent	SD 3966 - SD 3966
0.05	SD 3967 - SD 3967
0.08	SD 3968 - SD 3968
0.20	SD 3969 - SD 3969

Attachments

Form 41
 Pre-Stability Tests
 Flow Calibration
 Optical Bench Cal
 Post-Stability Tests
 Other _____

Notes: E-MAILED APPROVED
9/23/2016

Recalibrated to bring values closer to nominal
AA/PC OK PSPM 9/20/16

Brett Kirkland
 Quality Control Review

Instrument Complies with Chapter 11D-8, FAC
 Instrument Does Not Comply with Chapter 11D-8, FAC

Return to/Place into Evidentiary Use
 Remain Out of Evidentiary Use
 Conduct an Agency Inspection Before Evidentiary Use

9/23/16
 Date

TYPE OF TEST	SERIAL NUMBER	AGENCY	DATE	PERFORMED BY
Post Stabilities	80-002232	Miami Beach Police Department	9/14/2016	<i>[Signature]</i>

0.05g/210L <input checked="" type="checkbox"/>	0.08g/210L <input checked="" type="checkbox"/>	0.20g/210L <input checked="" type="checkbox"/>	DGS 0.08g/210L <input checked="" type="checkbox"/>
0.047 to 0.053 <input checked="" type="checkbox"/>	0.077 to 0.083 <input checked="" type="checkbox"/>	0.194 to 0.206 <input checked="" type="checkbox"/>	0.077 to 0.083 <input checked="" type="checkbox"/>

MIAMI BEACH PD
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-002232
09/14/2016
Software: 8100.27

Test g/210L Time

RfR Blank 0.000 08:55
Control Test 0.051 08:55
RfR Blank 0.000 08:56
Control Test 0.051 08:56
RfR Blank 0.000 08:57
Control Test 0.051 08:58
RfR Blank 0.000 08:58
Control Test Status
Average 0.0510
Std Dev 0.0000
Rel Std Dev(%) 0.0000

[Signature]
Operator's Signature

MIAMI BEACH PD
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-002232
09/14/2016
Software: 8100.27

Test g/210L Time

RfR Blank 0.000 08:59
Control Test 0.080 09:00
RfR Blank 0.000 09:00
Control Test 0.080 09:01
RfR Blank 0.000 09:02
Control Test 0.080 09:02
RfR Blank 0.000 09:03
Control Test Status
Average 0.0800
Std Dev 0.0000
Rel Std Dev(%) 0.0000

[Signature]
Operator's Signature

MIAMI BEACH PD
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-002232
09/14/2016
Software: 8100.27

Test g/210L Time

RfR Blank 0.000 09:04
Control Test 0.202 09:05
RfR Blank 0.000 09:05
Control Test 0.202 09:06
RfR Blank 0.000 09:06
Control Test 0.202 09:07
RfR Blank 0.000 09:07
Control Test Status
Average 0.2020
Std Dev 0.0000
Rel Std Dev(%) 0.0000

[Signature]
Operator's Signature

MIAMI BEACH PD
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-002232
09/14/2016
Software: 8100.27

Test g/210L Time

RfR Blank 0.000 09:10
Control Test 0.080 09:10
RfR Blank 0.000 09:11
Control Test 0.080 09:11
RfR Blank 0.000 09:11
Control Test 0.080 09:12
RfR Blank 0.000 09:12
Control Test Status
Average 0.0800
Std Dev 0.0000
Rel Std Dev(%) 0.0000

[Signature]
Operator's Signature

[Handwritten initials]

TYPE OF TEST	SERIAL NUMBER	AGENCY	DATE	PERFORMED BY
Stabilities	80-002232	Miami Beach Police Department	9/8/2016	WZL

0.05g/210L	0.08g/210L	0.20g/210L	DGS 0.08g/210L
0.047 to 0.053 <input checked="" type="checkbox"/>	0.077 to 0.083 <input checked="" type="checkbox"/>	0.194 to 0.206 <input checked="" type="checkbox"/>	0.077 to 0.083 <input checked="" type="checkbox"/>

MIAMI BEACH PD Intoxilizer - Alcohol Analyzer Model 8000 SN 80-002232 09/08/2016 Software: 8100.27	MIAMI BEACH PD Intoxilizer - Alcohol Analyzer Model 8000 SN 80-002232 09/08/2016 Software: 8100.27	MIAMI BEACH PD Intoxilizer - Alcohol Analyzer Model 8000 SN 80-002232 09/08/2016 Software: 8100.27	MIAMI BEACH PD Intoxilizer - Alcohol Analyzer Model 8000 SN 80-002232 09/08/2016 Software: 8100.27
--	--	--	--

Test	g/210L	Time	Test	g/210L	Time	Test	g/210L	Time	Test	g/210L	Time
Air Blank	0.000	08:11	Air Blank	0.000	08:15	Air Blank	0.000	08:21	Air Blank	0.000	08:27
Control Test	0.051	08:11	Control Test	0.082	08:16	Control Test	0.204	08:22	Control Test	0.081	08:27
Air Blank	0.000	08:12	Air Blank	0.000	08:17	Air Blank	0.000	08:22	Air Blank	0.000	08:28
Control Test	0.051	08:13	Control Test	0.082	08:17	Control Test	0.205	08:23	Control Test	0.084	08:28
Air Blank	0.000	08:13	Air Blank	0.000	08:18	Air Blank	0.000	08:23	Air Blank	0.000	08:29
Control Test	0.051	08:14	Control Test	0.082	08:19	Control Test	0.206	08:24	Control Test	0.081	08:29
Air Blank	0.000	08:14	Air Blank	0.000	08:19	Air Blank	0.000	08:25	Air Blank	0.000	08:30
Control Test Stats			Control Test Stats			Control Test Stats			Control Test Stats		
Average	0.0510		Average	0.0820		Average	0.2050		Average	0.0820	
Std Dev	0.0000		Std Dev	0.0000		Std Dev	0.0010		Std Dev	0.0017	
Rel Std Dev(%)	0.0000		Rel Std Dev(%)	0.0000		Rel Std Dev(%)	0.4878		Rel Std Dev(%)	2.1123	

Operator's Signature *WZL*

Operator's Signature *WZL*

Operator's Signature *WZL*

Operator's Signature *WZL*

WZL

MIAMI BEACH PD
 Intercooler - Alcotest Analyzer
 Model 8000 SN 80-002232
 09/14/2016 08:08:46

Auto Calibration
 Max Power Res Value = 32
 Auto Range Res Value = 15

Sol Value = 0.000 g/210L ***
 Fit value = 0.000 mg/l ****
 Samples Taken = 4, Discarded = 1
 Sum Io = 12794, Sum Io = 13596

Channel 1
 Sample #1 = 0.0261 (% Abs Ref)
 Sample #2 = 0.0440 (% Abs Ref)
 Sample #3 = 0.0210 (% Abs Ref)
 Sample #4 = 0.0260 (% Abs Ref)
 Avg % Abs = 0.0333 (0.0303)
 STD DEV = 0.0121 (0.0285)
 REL STD DEV = 39.880 (93.961)

Channel 2
 Sample #1 = -0.0090 (% Abs Ref)
 Sample #2 = 0.0200 (% Abs Ref)
 Sample #3 = 0.0100 (% Abs Ref)
 Sample #4 = 0.0210 (% Abs Ref)
 Avg % Abs = 0.0170 (-0.0267)
 STD DEV = 0.0061 (0.0166)
 REL STD DEV = 35.781 (62.337)

Sol Value = 0.040 g/210L ***
 Fit value = 0.1905 mg/l ****
 Samples Taken = 4, Discarded = 1
 Sum Io = 12780, Sum Io = 13588
 Channel 1
 Sample #1 = 0.7740 (% Abs Ref)
 Sample #2 = 0.7710 (% Abs Ref)
 Sample #3 = 0.7690 (% Abs Ref)
 Sample #4 = 0.7750 (% Abs Ref)
 Avg % Abs = 0.7717 (0.0263)
 STD DEV = 0.0031 (0.0202)
 REL STD DEV = 0.796 (75.736)

Channel 2
 Sample #1 = 1.4940 (% Abs Ref)
 Sample #2 = 1.5200 (% Abs Ref)
 Sample #3 = 1.5170 (% Abs Ref)
 Sample #4 = 1.5190 (% Abs Ref)
 Avg % Abs = 1.5187 (-0.0133)
 STD DEV = 0.0015 (0.0040)
 REL STD DEV = 0.101 (30.311)

Sol Value = 0.100 g/210L ***
 Fit value = 0.4762 mg/l ****
 Samples Taken = 4, Discarded = 1
 Sum Io = 12771, Sum Io = 13588
 Channel 1
 Sample #1 = 1.8730 (% Abs Ref)
 Sample #2 = 1.8610 (% Abs Ref)
 Sample #3 = 1.8470 (% Abs Ref)
 Sample #4 = 1.8550 (% Abs Ref)
 Avg % Abs = 1.8543 (0.0260)
 STD DEV = 0.0070 (0.0251)
 REL STD DEV = 0.379 (96.384)

Channel 2
 Sample #1 = 3.6620 (% Abs Ref)
 Sample #2 = 3.6890 (% Abs Ref)
 Sample #3 = 3.6690 (% Abs Ref)
 Sample #4 = 3.6640 (% Abs Ref)
 Avg % Abs = 3.6740 (0.0217)
 STD DEV = 0.0132 (0.0189)
 REL STD DEV = 0.360 (91.339)

Sol Value = 0.200 g/210L ***
 Fit value = 0.9524 mg/l ****
 Samples Taken = 4, Discarded = 1
 Sum Io = 12763, Sum Io = 13582
 Channel 1
 Sample #1 = 3.5300 (% Abs Ref)
 Sample #2 = 3.5060 (% Abs Ref)
 Sample #3 = 3.5190 (% Abs Ref)
 Sample #4 = 3.5080 (% Abs Ref)
 Avg % Abs = 3.5110 (0.0183)
 STD DEV = 0.0070 (0.0131)
 REL STD DEV = 0.199 (71.188)

Channel 2
 Sample #1 = 12.9160 (% Abs Ref)
 Sample #2 = 12.9110 (% Abs Ref)
 Sample #3 = 12.9340 (% Abs Ref)
 Sample #4 = 12.9360 (% Abs Ref)
 Avg % Abs = 12.9270 (0.0233)
 STD DEV = 0.0139 (0.0015)
 REL STD DEV = 0.107 (6.547)

Channel 2
 Sample #1 = 6.9130 (% Abs Ref)
 Sample #2 = 6.9300 (% Abs Ref)
 Sample #3 = 6.9430 (% Abs Ref)
 Sample #4 = 6.9250 (% Abs Ref)
 Avg % Abs = 6.9327 (0.0040)
 STD DEV = 0.0093 (0.0144)
 REL STD DEV = 0.134 (360.555)

Sol Value = 0.400 g/210L ***
 Fit value = 1.9048 mg/l ****
 Samples Taken = 4, Discarded = 1
 Sum Io = 12759, Sum Io = 13581
 Channel 1
 Sample #1 = 6.6830 (% Abs Ref)
 Sample #2 = 6.6720 (% Abs Ref)
 Sample #3 = 6.6790 (% Abs Ref)
 Sample #4 = 6.6960 (% Abs Ref)
 Avg % Abs = 6.6823 (0.0363)
 STD DEV = 0.0123 (0.0075)
 REL STD DEV = 0.185 (20.657)

Channel 2
 Sample #1 = 12.9160 (% Abs Ref)
 Sample #2 = 12.9110 (% Abs Ref)
 Sample #3 = 12.9340 (% Abs Ref)
 Sample #4 = 12.9360 (% Abs Ref)
 Avg % Abs = 12.9270 (0.0233)
 STD DEV = 0.0139 (0.0015)
 REL STD DEV = 0.107 (6.547)

Sol Value = 0.01 Rel Std Dev = 0.11
 Zero Order Coef = -37.79
 First Order Coef = 1254.95
 Second Order Coef = 17.17
 Standard Deviation = 29.421921

Channel 2
 Sample #1 = 3.5700 (% Abs Ref)
 Sample #2 = 3.4400 (% Abs Ref)
 Sample #3 = 3.4600 (% Abs Ref)
 Sample #4 = 3.4690 (% Abs Ref)
 Avg % Abs = 3.4978 (0.0000)
 STD DEV = 15.9478
 REL STD DEV = 0.461

Channel 1
 Sample #1 = 3.5500 (% Abs Ref)
 Sample #2 = 3.2550 (% Abs Ref)
 Sample #3 = 3.2650 (% Abs Ref)
 Sample #4 = 3.3330 (% Abs Ref)
 Avg % Abs = 3.2777 (0.0000)
 STD DEV = 50.2129
 REL STD DEV = 1.532

Sol Value = 0.080 g/210L ***
 Fit value = 0.3810 mg/l ****
 Samples Taken = 4, Discarded = 1
 Channel 1
 Sample #1 = 3350.00
 Sample #2 = 3235.00
 Sample #3 = 3265.00
 Sample #4 = 3333.00
 Average Result = 3277.6667
 STD DEV = 50.2129
 REL STD DEV = 1.532

Channel 2
 Sample #1 = 3.5700 (% Abs Ref)
 Sample #2 = 3.4400 (% Abs Ref)
 Sample #3 = 3.4600 (% Abs Ref)
 Sample #4 = 3.4690 (% Abs Ref)
 Avg % Abs = 3.4978 (0.0000)
 STD DEV = 15.9478
 REL STD DEV = 0.461

Sol Value = 0.01 Rel Std Dev = 0.11
 Zero Order Coef = -37.79
 First Order Coef = 1254.95
 Second Order Coef = 17.17
 Standard Deviation = 29.421921

Channel 2
 Sample #1 = 12.9160 (% Abs Ref)
 Sample #2 = 12.9110 (% Abs Ref)
 Sample #3 = 12.9340 (% Abs Ref)
 Sample #4 = 12.9360 (% Abs Ref)
 Avg % Abs = 12.9270 (0.0233)
 STD DEV = 0.0139 (0.0015)
 REL STD DEV = 0.107 (6.547)

Solution Stats Quadratic Fit Chan 2
 Act Fit Residual
 g/210L g/210L g/210L
 0.000 -0.000 0.0003
 0.040 0.040 -0.0001
 0.100 0.101 -0.0009
 0.200 0.199 0.0008
 0.400 0.400 -0.0001

Sol Value = 0.080 g/210L ***
 Fit value = 0.3810 mg/l ****
 Samples Taken = 4, Discarded = 1
 Channel 1
 Sample #1 = 3350.00
 Sample #2 = 3235.00
 Sample #3 = 3265.00
 Sample #4 = 3333.00
 Average Result = 3277.6667
 STD DEV = 50.2129
 REL STD DEV = 1.532

Channel 2
 Sample #1 = 3.5700 (% Abs Ref)
 Sample #2 = 3.4400 (% Abs Ref)
 Sample #3 = 3.4600 (% Abs Ref)
 Sample #4 = 3.4690 (% Abs Ref)
 Avg % Abs = 3.4978 (0.0000)
 STD DEV = 15.9478
 REL STD DEV = 0.461

Sol Value = 0.01 Rel Std Dev = 0.11
 Zero Order Coef = -37.79
 First Order Coef = 1254.95
 Second Order Coef = 17.17
 Standard Deviation = 29.421921

Channel 2
 Sample #1 = 12.9160 (% Abs Ref)
 Sample #2 = 12.9110 (% Abs Ref)
 Sample #3 = 12.9340 (% Abs Ref)
 Sample #4 = 12.9360 (% Abs Ref)
 Avg % Abs = 12.9270 (0.0233)
 STD DEV = 0.0139 (0.0015)
 REL STD DEV = 0.107 (6.547)

Optical Calibration
 SN: 80-002232
 Agency: Miami Beach PD
 Date: 09/14/2016
 Quadratic Fit: +/-0.002g/210L
 By: *WLL*

Act	Fit	Residual
g/210L	g/210L	g/210L
0.000	-0.001	0.0005
0.040	0.040	-0.0001
0.100	0.101	-0.0012
0.200	0.199	0.0010
0.400	0.400	-0.0002

Dry Gas H2O Adjust Results *****
 Barometric Pressure = 1016
 3 um H2O Adjust (mg/l*(10,000) = 532
 9 um H2O Adjust (mg/l*(10,000) = 351

 gpm
 R