

## INSTRUMENT PROCESSING SHEET

Agency Highland Police Department S/N 80-002462  
 Date In 4/22/2016 Date Out 4/25/2016  Ship  P/U  H/D  CMI  EE

<b>Intake</b> Performed By <u>DELR</u> <input type="checkbox"/> Registration <input checked="" type="checkbox"/> Annual <input type="checkbox"/> Return from CMI <input type="checkbox"/> Return from Enforcement Electronics <input type="checkbox"/> Other _____ Visual Inspection: OK Case <u>OK</u> Handle <u>OK</u> OK Dry Gas Holder <u>OK</u> Feet <u>OK</u> OK Keyboard/Plug <u>OK</u> Back/Plugs <u>OK</u> OK Screws tight <u>OK</u> Breath Hose <u>OK</u> Other Equipment: <input checked="" type="checkbox"/> Power cord <input type="checkbox"/> Printer Cable <input checked="" type="checkbox"/> Other <u>ADD STATIC BAG</u> Notes: _____ _____ _____	<b>Quality Checks</b> Performed By <u>DELR</u> <input checked="" type="checkbox"/> Breath Tube Screen <input checked="" type="checkbox"/> Replace O-Rings <input checked="" type="checkbox"/> Instrument Set Up Verified <input checked="" type="checkbox"/> R-Value <u>203</u> <input checked="" type="checkbox"/> Flow Verification (L/s) Flow Column # <u>ATP101</u> 32mm <u>156</u> (.139 - .169) 36mm <u>171</u> (.156 - .190) 53mm <u>246</u> (.228 - .278) 103mm <u>511</u> (.447 - .547) <input checked="" type="checkbox"/> Barometric Pressure Check Gauge ID # <u>28663</u> <input checked="" type="checkbox"/> Stability Checks <table border="1"> <thead> <tr> <th>Simulator</th> <th>Serial #</th> <th>Lot #/Exp</th> </tr> </thead> <tbody> <tr> <td>0.05</td> <td>SD1014</td> <td>201507A 07/14/2017</td> </tr> <tr> <td>0.08</td> <td>SD1015</td> <td>2015026 02/24/2017</td> </tr> <tr> <td>0.20</td> <td>SD1017</td> <td>201505A 05/12/2017</td> </tr> <tr> <td>0.08 DGS</td> <td>N/A</td> <td>A6519701 07/16/2017</td> </tr> </tbody> </table>	Simulator	Serial #	Lot #/Exp	0.05	SD1014	201507A 07/14/2017	0.08	SD1015	2015026 02/24/2017	0.20	SD1017	201505A 05/12/2017	0.08 DGS	N/A	A6519701 07/16/2017	<b>Flow Calibration</b> Performed By _____ <input checked="" type="checkbox"/> Flow Calibration N/A <input type="checkbox"/> Flow Calibration Complete Flow Column # _____ <input type="checkbox"/> 5L/min - 1mm <input type="checkbox"/> 15L/min - 53mm <input type="checkbox"/> 30L/min - 103mm <input type="checkbox"/> R-Value _____ <input type="checkbox"/> Post Calibration Verification (L/s) Flow Column # _____ 32mm _____ (.139 - .169) 36mm _____ (.156 - .190) 53mm _____ (.228 - .278) 103mm _____ (.447 - .547)
Simulator	Serial #	Lot #/Exp															
0.05	SD1014	201507A 07/14/2017															
0.08	SD1015	2015026 02/24/2017															
0.20	SD1017	201505A 05/12/2017															
0.08 DGS	N/A	A6519701 07/16/2017															
		<b>Maintenance</b> Performed By _____ <input type="checkbox"/> Battery Replacement <input type="checkbox"/> Dry Gas Regulator Replacement <input type="checkbox"/> Breath Tube Replacement <input type="checkbox"/> Other _____ <b>Suggested Service</b> _____ _____ _____															

RECEIVED  
 MAY 06 2016  
 FDLE  
 Alcohol Testing Program

<b>Optical Bench Calibration</b> Performed By <u>DELR</u> <input type="checkbox"/> Optical Bench Calibration N/A <input checked="" type="checkbox"/> Optical Bench Calibration Complete Barometric Pressure Gauge <u>1016</u> ID# <u>28199</u>																							
Simulator	Serial Number	Lot Number	Expiration																				
0.000	2235	N/A	N/A																				
0.040	2236	15108	08/18/2017																				
0.100	2237	15001	05/20/2017																				
0.200	2238	15104	05/27/2017																				
0.400	2239	15105	06/10/2017																				
0.080 DGS	N/A	03415080A1	03/05/2017																				
<input checked="" type="checkbox"/> Post Calibration Stability Checks <table border="1"> <thead> <tr> <th>Simulator</th> <th>Serial Number</th> <th>Lot Number</th> <th>Expiration</th> </tr> </thead> <tbody> <tr> <td>0.05</td> <td>SD1014</td> <td>201507A</td> <td>07/14/2017</td> </tr> <tr> <td>0.08</td> <td>SD1015</td> <td>2015026</td> <td>02/24/2017</td> </tr> <tr> <td>0.20</td> <td>SD1017</td> <td>201505A</td> <td>05/12/2017</td> </tr> <tr> <td>0.08 DGS</td> <td>N/A</td> <td>A6519701</td> <td>07/16/2017</td> </tr> </tbody> </table>				Simulator	Serial Number	Lot Number	Expiration	0.05	SD1014	201507A	07/14/2017	0.08	SD1015	2015026	02/24/2017	0.20	SD1017	201505A	05/12/2017	0.08 DGS	N/A	A6519701	07/16/2017
Simulator	Serial Number	Lot Number	Expiration																				
0.05	SD1014	201507A	07/14/2017																				
0.08	SD1015	2015026	02/24/2017																				
0.20	SD1017	201505A	05/12/2017																				
0.08 DGS	N/A	A6519701	07/16/2017																				

<b>Department Inspection</b> Performed By <u>DELR</u> <input checked="" type="checkbox"/> Barometric Pressure <u>1016</u> Gauge ID# <u>28663</u> <u>1015</u> Instrument	
Mouth Alcohol Solution Lot # <u>2015-A</u> Acetone Stock Solution Lot # <u>2015-B</u>	
Simulator	Serial Number
0.00	G8147
Interferent	G12000 DELR G12100
0.05	SD1014
0.08	SD1015
0.20	SD1017

<b>Attachments</b> <input checked="" type="checkbox"/> Form 41 <input checked="" type="checkbox"/> Pre-Stability Tests <input type="checkbox"/> Flow Calibration <input checked="" type="checkbox"/> Optical Bench Cal <input checked="" type="checkbox"/> Post-Stability Tests <input checked="" type="checkbox"/> Other <u>one additional 0.20 pre stability</u>	
--	--

Notes: **E-MAILED**  **APPROVED**  
04/25/2016  
Replaced printer paper  
PA/OC OK 04/25/16

<input checked="" type="checkbox"/> Instrument Complies with Chapter 11D-8, FAC <input type="checkbox"/> Instrument Does Not Comply with Chapter 11D-8, FAC <input checked="" type="checkbox"/> Return to/Place into Evidentiary Use <input type="checkbox"/> Remain Out of Evidentiary Use <input checked="" type="checkbox"/> Conduct an Agency Inspection Before Evidentiary Use
---

Smith Kirkland  
 Quality Control Review

5/6/16  
 Date

<b>TYPE OF TEST</b>	<b>SERIAL NUMBER</b>	<b>AGENCY</b>	<b>DATE</b>	<b>PERFORMED BY</b>
Post Stabilities	80-002462	Hialeah Police Department	04/25/2016	<i>[Signature]</i> KPK

0.05g/210L  0.077 to 0.083  0.20g/210L  DGS 0.08g/210L  0.077 to 0.083

MQ-594 PD  
Intoxilyzer - Alcotest Analyzer  
Model: 8000 SN: 80-002462  
04/25/2016  
Software: 8100.27

Test 9/210L Time

Air Blank 0.000 11:04  
Control Test 0.050 11:05  
Air Blank 0.000 11:05  
Control Test 0.051 11:06  
Air Blank 0.000 11:06  
Control Test 0.051 11:07  
Air Blank 0.000 11:07  
Control Test Status  
Average 0.0507  
Std Dev 0.0006  
Rel. Std Dev(%) 1.1395

*[Signature]*  
Operator's Signature

MQ-594 PD  
Intoxilyzer - Alcotest Analyzer  
Model: 8000 SN: 80-002462  
04/25/2016  
Software: 8100.27

Test 9/210L Time

Air Blank 0.000 11:11  
Control Test 0.079 11:11  
Air Blank 0.000 11:12  
Control Test 0.080 11:13  
Air Blank 0.000 11:13  
Control Test 0.080 11:14  
Air Blank 0.000 11:14  
Control Test Status  
Average 0.0797  
Std Dev 0.0006  
Rel. Std Dev(%) 0.7247

*[Signature]*  
Operator's Signature

MQ-594 PD  
Intoxilyzer - Alcotest Analyzer  
Model: 8000 SN: 80-002462  
04/25/2016  
Software: 8100.27

Test 9/210L Time

Air Blank 0.000 11:15  
Control Test 0.202 11:15  
Air Blank 0.000 11:17  
Control Test 0.204 11:17  
Air Blank 0.000 11:18  
Control Test 0.204 11:19  
Air Blank 0.000 11:19  
Control Test Status  
Average 0.2033  
Std Dev 0.0012  
Rel. Std Dev(%) 0.5679

*[Signature]*  
Operator's Signature

MQ-594 PD  
Intoxilyzer - Alcotest Analyzer  
Model: 8000 SN: 80-002462  
04/25/2016  
Software: 8100.27

Test 9/210L Time

Air Blank 0.000 11:21  
Control Test 0.075 11:21  
Air Blank 0.000 11:21  
Control Test 0.080 11:22  
Air Blank 0.000 11:22  
Control Test 0.080 11:23  
Air Blank 0.000 11:23  
Control Test Status  
Average 0.0797  
Std Dev 0.0006  
Rel. Std Dev(%) 0.7247

*[Signature]*  
Operator's Signature

*[Handwritten]* gpm

HTR-REP-PS  
 Nitro Analyzer - Alconal Analyzer  
 Model 8000 SN 80-002462  
 04/25/2016 10:33:28

Auto Calibration  
 Max Power Res Value = 93  
 Run Range Res Value = 65

Fit Value = 0.005 g/210L \*\*\*  
 Samples Taken = 4, Discarded = 1  
 Sum Io = 12673, Sum Io = 13070

<<<< CHANNEL 1 >>>>  
 Sample % R05 (% R05 Ref)  
 Sample #1 = 0.0680 (-0.0180)  
 Sample #2 = 0.0980 (-0.0090)  
 Sample #3 = 0.0690 (-0.0600)  
 Sample #4 = 0.0680 (-0.0610)  
 Avg % R05 = 0.0917 (0.0433)  
 STD DEV = 0.0055 (0.0297)  
 REL STD DEV = 6.008 (66.626)

<<<< CHANNEL 2 >>>>  
 Sample % R05 (% R05 Ref)  
 Sample #1 = 0.0970 (-0.0110)  
 Sample #2 = 0.0630 (-0.0070)  
 Sample #3 = 0.0600 (-0.0110)  
 Sample #4 = 0.0660 (-0.0000)  
 Avg % R05 = 0.0830 (0.0013)  
 STD DEV = 0.0030 (0.0091)  
 REL STD DEV = 3.614 (680.533)

Sol Value = 0.040 g/210L \*\*\*  
 Fit Value = 0.1905 mg/l %%%  
 Samples Taken = 4, Discarded = 1  
 Sum Io = 12681, Sum Io = 13073  
 <<<< CHANNEL 1 >>>>  
 Sample % R05 (% R05 Ref)  
 Sample #1 = 0.8010 (-0.0130)  
 Sample #2 = 0.7730 (-0.0150)  
 Sample #3 = 0.8020 (-0.0330)  
 Sample #4 = 0.7900 (-0.0320)  
 Avg % R05 = 0.7883 (0.0267)  
 STD DEV = 0.0146 (0.0101)

<<<< CHANNEL 2 >>>>  
 Sample % R05 (% R05 Ref)  
 Sample #1 = 1.4180 (-0.0030)  
 Sample #2 = 1.4170 (-0.0100)  
 Sample #3 = 1.4480 (-0.0060)  
 Sample #4 = 1.4240 (-0.0000)  
 Avg % R05 = 1.4297 (-0.0053)  
 STD DEV = 0.0163 (0.0050)  
 REL STD DEV = 1.137 (94.373)

Sol Value = 0.100 g/210L \*\*\*  
 Fit Value = 0.4762 mg/l %%%  
 Samples Taken = 4, Discarded = 1  
 Sum Io = 12673, Sum Io = 13070

<<<< CHANNEL 1 >>>>  
 Sample % R05 (% R05 Ref)  
 Sample #1 = 1.6570 (-0.0170)  
 Sample #2 = 1.9090 (-0.0080)  
 Sample #3 = 1.6790 (-0.0170)  
 Sample #4 = 1.9020 (-0.0000)  
 Avg % R05 = 1.8970 (-0.0083)  
 STD DEV = 0.1019 (0.0085)  
 REL STD DEV = 0.837 (102.059)

<<<< CHANNEL 2 >>>>  
 Sample % R05 (% R05 Ref)  
 Sample #1 = 3.4760 (-0.0260)  
 Sample #2 = 3.5000 (-0.0150)  
 Sample #3 = 3.4830 (-0.0280)  
 Sample #4 = 3.4830 (-0.0100)  
 Avg % R05 = 3.4887 (-0.0177)  
 STD DEV = 0.0099 (0.0093)  
 REL STD DEV = 0.281 (52.594)

Sol Value = 0.210 g/210L \*\*\*  
 Fit Value = 1.9524 mg/l %%%  
 Samples Taken = 4, Discarded = 1  
 Sum Io = 12669, Sum Io = 13069  
 <<<< CHANNEL 1 >>>>  
 Sample % R05 (% R05 Ref)  
 Sample #1 = 3.5710 (-0.0230)  
 Sample #2 = 3.6070 (-0.0300)  
 Sample #3 = 3.5980 (-0.0180)  
 Sample #4 = 3.6080 (-0.0070)  
 Avg % R05 = 3.6043 (-0.0137)  
 STD DEV = 0.0055 (0.0189)  
 REL STD DEV = 0.153 (138.123)

<<<< CHANNEL 2 >>>>  
 Sample % R05 (% R05 Ref)  
 Sample #1 = 6.6180 (-0.0150)  
 Sample #2 = 6.6090 (-0.0210)  
 Sample #3 = 6.6220 (-0.0170)  
 Sample #4 = 6.6340 (-0.0000)  
 Avg % R05 = 6.6217 (-0.0127)  
 STD DEV = 0.0125 (0.0112)  
 REL STD DEV = 0.189 (88.030)

Sol Value = 0.400 g/210L \*\*\*  
 Fit Value = 1.9548 mg/l %%%  
 Samples Taken = 4, Discarded = 1  
 Sum Io = 12667, Sum Io = 13069

<<<< CHANNEL 1 >>>>  
 Sample % R05 (% R05 Ref)  
 Sample #1 = 6.8870 (-0.0300)  
 Sample #2 = 6.9260 (-0.0110)  
 Sample #3 = 6.9390 (-0.0080)  
 Sample #4 = 6.9340 (-0.0210)  
 Avg % R05 = 6.9300 (0.0007)  
 STD DEV = 0.0066 (0.0177)  
 REL STD DEV = 0.095 (2650.944)

<<<< CHANNEL 2 >>>>  
 Sample % R05 (% R05 Ref)  
 Sample #1 = 12.4500 (0.0020)  
 Sample #2 = 12.5580 (-0.0090)  
 Sample #3 = 12.5760 (-0.0030)  
 Sample #4 = 12.6190 (-0.0130)  
 Avg % R05 = 12.5847 (0.0003)  
 STD DEV = 0.0309 (0.0114)  
 REL STD DEV = 0.246 (3411.744)

Sol Value = 0.0000 mg/l or 0.000 g/210L  
 Fit Value = 0.9524 mg/l or 0.200 g/210L  
 % R05 = 3.534  
 Std Dev = 0.01 Rel Std Dev = 0.15  
 Sol Val = 1.9048 mg/l or 0.400 g/210L  
 % R05 = 6.933  
 Std Dev = 0.01 Rel Std Dev = 0.09  
 Zero Order Coef = -220.16  
 First Order Coef = 2607.90  
 Second Order Coef = 24.75  
 Standard Deviation = 40.927250

\*\*\*\*\* AUTO CAL DATA \*\*\*\*\*  
 <<<< CHANNEL 1 >>>>  
 Sol Val = 0.0000 mg/l or 0.000 g/210L  
 % R05 = 0.092  
 Std Dev = 0.01 Rel Std Dev = 6.01  
 Sol Val = 0.1905 mg/l or 0.040 g/210L  
 % R05 = 0.788  
 Std Dev = 0.01 Rel Std Dev = 1.85  
 Sol Val = 0.4762 mg/l or 0.100 g/210L  
 % R05 = 1.997  
 Std Dev = 0.02 Rel Std Dev = 0.84  
 Sol Val = 0.9524 mg/l or 0.200 g/210L  
 % R05 = 3.534  
 Std Dev = 0.01 Rel Std Dev = 0.15  
 Sol Val = 1.9048 mg/l or 0.400 g/210L  
 % R05 = 6.933  
 Std Dev = 0.01 Rel Std Dev = 0.09  
 Zero Order Coef = -220.16  
 First Order Coef = 2607.90  
 Second Order Coef = 24.75  
 Standard Deviation = 40.927250

<<<< CHANNEL 2 >>>>  
 Sol Val = 0.0000 mg/l or 0.000 g/210L  
 % R05 = 0.093  
 Std Dev = 0.00 Rel Std Dev = 3.61  
 Sol Val = 0.1905 mg/l or 0.040 g/210L  
 % R05 = 1.433  
 Std Dev = 0.02 Rel Std Dev = 1.14  
 Sol Val = 0.4762 mg/l or 0.100 g/210L  
 % R05 = 3.485  
 Std Dev = 0.01 Rel Std Dev = 0.28  
 Sol Val = 0.9524 mg/l or 0.200 g/210L  
 % R05 = 6.522  
 Std Dev = 0.01 Rel Std Dev = 0.19  
 Sol Val = 1.9048 mg/l or 0.400 g/210L  
 % R05 = 12.585  
 Std Dev = 0.03 Rel Std Dev = 0.25  
 Zero Order Coef = -110.22  
 First Order Coef = 1369.77  
 Second Order Coef = 12.15  
 Standard Deviation = 35.275009

<<<< CHANNEL 1 >>>>  
 Sol Val = 0.0000 mg/l or 0.000 g/210L  
 % R05 = 0.093  
 Std Dev = 0.00 Rel Std Dev = 3.61  
 Sol Val = 0.1905 mg/l or 0.040 g/210L  
 % R05 = 1.433  
 Std Dev = 0.02 Rel Std Dev = 1.14  
 Sol Val = 0.4762 mg/l or 0.100 g/210L  
 % R05 = 3.485  
 Std Dev = 0.01 Rel Std Dev = 0.28  
 Sol Val = 0.9524 mg/l or 0.200 g/210L  
 % R05 = 6.522  
 Std Dev = 0.01 Rel Std Dev = 0.19  
 Sol Val = 1.9048 mg/l or 0.400 g/210L  
 % R05 = 12.585  
 Std Dev = 0.03 Rel Std Dev = 0.25  
 Zero Order Coef = -110.22  
 First Order Coef = 1369.77  
 Second Order Coef = 12.15  
 Standard Deviation = 35.275009

<<<< CHANNEL 2 >>>>  
 Sol Val = 0.0000 mg/l or 0.000 g/210L  
 % R05 = 0.093  
 Std Dev = 0.00 Rel Std Dev = 3.61  
 Sol Val = 0.1905 mg/l or 0.040 g/210L  
 % R05 = 1.433  
 Std Dev = 0.02 Rel Std Dev = 1.14  
 Sol Val = 0.4762 mg/l or 0.100 g/210L  
 % R05 = 3.485  
 Std Dev = 0.01 Rel Std Dev = 0.28  
 Sol Val = 0.9524 mg/l or 0.200 g/210L  
 % R05 = 6.522  
 Std Dev = 0.01 Rel Std Dev = 0.19  
 Sol Val = 1.9048 mg/l or 0.400 g/210L  
 % R05 = 12.585  
 Std Dev = 0.03 Rel Std Dev = 0.25  
 Zero Order Coef = -110.22  
 First Order Coef = 1369.77  
 Second Order Coef = 12.15  
 Standard Deviation = 35.275009

Solution Stats Quadratic Fit Chan 1

Act	Fit	Residual
g/210L	g/210L	g/210L
0.000	0.000	-0.000
0.046	0.039	0.007
0.100	0.101	-0.001
0.200	0.199	0.002
0.400	0.400	-0.000

Sol Value = 0.000 g/210L \*\*\*  
 Fit Value = 0.3810 mg/l %%%  
 Samples Taken = 4, Discarded = 1  
 <<<< CHANNEL 1 >>>>  
 Sample #1 = 3200.00  
 Sample #2 = 3189.00  
 Sample #3 = 3194.00  
 Sample #4 = 3284.00  
 Average Result = 3222.3333  
 STD DEV = 53.4634  
 REL STD DEV = 1.659

<<<< CHANNEL 2 >>>>  
 Sample #1 = 3437.00  
 Sample #2 = 3452.00  
 Sample #3 = 3461.00  
 Sample #4 = 3473.00  
 Average Result = 3462.0000  
 STD DEV = 11.5357  
 REL STD DEV = 0.304

Dry Gas H2O Adjust Results \*\*\*\*\*  
 Barometric Pressure = 1015  
 3 um H2O Adjust (mg/l\*10,000) = 587  
 9 um H2O Adjust (mg/l\*10,000) = 347  
 \*\*\*\*\* AUTO CAL PRESS \*\*\*\*\*

Solution Stats Quadratic Fit Chan 2

Act	Fit	Residual
g/210L	g/210L	g/210L
0.000	0.000	-0.000
0.046	0.039	0.007
0.100	0.101	-0.001
0.200	0.199	0.002
0.400	0.400	-0.000

**Optical Calibration**  
**SN: 80-002462**  
**Agency: Hialeah Police Dept**  
**Date: 04/25/2016**  
**Quadratic Fit: +/-0.002g/210L**

Bpk  
 Bpm

TYPE OF TEST	SERIAL NUMBER	AGENCY	DATE	PERFORMED BY
Stabilities	80-002462	Haleah Police Department	04/25/2016	<i>WJL</i>

*AK*

0.05g/210L	0.08g/210L	0.20g/210L	DGS 0.08g/210L
<p>0.047 to 0.053 <input checked="" type="checkbox"/></p> <p>Haleah PD Intoxilizer - Alcohol Analyzer Model: 8100 04/25/2016 Software: 8100.27</p> <p>Test: 9/21/0L Time</p> <p>Air Blank 0.000 08:16 Control Test 0.052 08:17 Air Blank 0.003 08:17 Control Test 0.051 08:18 Air Blank 0.003 08:18 Control Test 0.052 08:19 Air Blank 0.100 08:20 Control Test Status Average 0.0517 Std Dev 0.0035 Rel. Std Dev(%) 1.1175</p> <p>Operator's Signature <i>WJL</i></p>	<p>0.077 to 0.083 <input checked="" type="checkbox"/></p> <p>Haleah PD Intoxilizer - Alcohol Analyzer Model: 8100 04/25/2016 Software: 8100.27</p> <p>Test: 9/21/0L Time</p> <p>Air Blank 0.006 08:21 Control Test 0.081 08:21 Air Blank 0.000 08:22 Control Test 0.081 08:23 Air Blank 0.000 08:23 Control Test 0.081 08:24 Air Blank 0.000 08:24 Control Test Status Average 0.0810 Std Dev 0.0000 Rel. Std Dev(%) 0.0000</p> <p>Operator's Signature <i>WJL</i></p>	<p>0.194 to 0.206 <input checked="" type="checkbox"/></p> <p>Haleah PD Intoxilizer - Alcohol Analyzer Model: 8100 04/25/2016 Software: 8100.27</p> <p>Test: 9/21/0L Time</p> <p>Air Blank 0.058 08:26 Control Test 0.206 08:26 Air Blank 0.008 08:27 Control Test 0.207 08:28 Air Blank 0.008 08:28 Control Test 0.207 08:29 Air Blank 0.000 08:29 Control Test Status Average 0.2067 Std Dev 0.0005 Rel. Std Dev(%) 0.2794</p> <p>Operator's Signature <i>WJL</i></p>	<p>0.077 to 0.083 <input checked="" type="checkbox"/></p> <p>Haleah PD Intoxilizer - Alcohol Analyzer Model: 8100 04/25/2016 Software: 8100.27</p> <p>Test: 9/21/0L Time</p> <p>Air Blank 0.080 08:31 Control Test 0.080 08:32 Air Blank 0.000 08:32 Control Test 0.080 08:33 Air Blank 0.000 08:33 Control Test 0.080 08:34 Control Test Status Average 0.0800 Std Dev 0.0000 Rel. Std Dev(%) 0.0000</p> <p>Operator's Signature <i>WJL</i></p>

*WJL*

TYPE OF TEST	SERIAL NUMBER	AGENCY	DATE	PERFORMED BY
Stabilities	80-002462	Hialeah Police Department	04/25/2016	<i>MLL</i>

151

0.05g/210L 0.047 to 0.053 <input type="checkbox"/>	0.08g/210L 0.077 to 0.083 <input type="checkbox"/>	0.20g/210L 0.194 to 0.206 <input checked="" type="checkbox"/>	DGS 0.08g/210L 0.077 to 0.083 <input type="checkbox"/>																																				
<p>           HIQ-EQU-2J            Intoxilyzer - Alcohol Analyzer            Model 8000            04/25/2016            Software: 8100.27            SN: 80-102462         </p> <table border="1"> <thead> <tr> <th>Test</th> <th>g/210L</th> <th>Time</th> </tr> </thead> <tbody> <tr> <td>RIr Blank</td> <td>0.000</td> <td>08:36</td> </tr> <tr> <td>Control Test</td> <td>0.207</td> <td>08:37</td> </tr> <tr> <td>RIr Blank</td> <td>0.000</td> <td>08:38</td> </tr> <tr> <td>Control Test</td> <td>0.206</td> <td>08:38</td> </tr> <tr> <td>RIr Blank</td> <td>0.000</td> <td>08:39</td> </tr> <tr> <td>Control Test</td> <td>0.208</td> <td>08:39</td> </tr> <tr> <td>RIr Blank</td> <td>0.000</td> <td>08:40</td> </tr> <tr> <td colspan="3">Control Test Stats</td> </tr> <tr> <td>Average</td> <td>0.2070</td> <td></td> </tr> <tr> <td>Std Dev</td> <td>0.0010</td> <td></td> </tr> <tr> <td>Rel Std Dev(%)</td> <td>0.4831</td> <td></td> </tr> </tbody> </table>				Test	g/210L	Time	RIr Blank	0.000	08:36	Control Test	0.207	08:37	RIr Blank	0.000	08:38	Control Test	0.206	08:38	RIr Blank	0.000	08:39	Control Test	0.208	08:39	RIr Blank	0.000	08:40	Control Test Stats			Average	0.2070		Std Dev	0.0010		Rel Std Dev(%)	0.4831	
Test	g/210L	Time																																					
RIr Blank	0.000	08:36																																					
Control Test	0.207	08:37																																					
RIr Blank	0.000	08:38																																					
Control Test	0.206	08:38																																					
RIr Blank	0.000	08:39																																					
Control Test	0.208	08:39																																					
RIr Blank	0.000	08:40																																					
Control Test Stats																																							
Average	0.2070																																						
Std Dev	0.0010																																						
Rel Std Dev(%)	0.4831																																						
<p>Operator's Signature: <i>MLL</i></p>																																							

ggm

HIALEAH PD  
Intoxilyzer - Alcohol Analyzer  
Model 8000 SN 80-002462  
04/25/2016  
Software: 8100.27

-----  
DIAGNOSTICS

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

*OGM*

*BK*