



Alcohol Testing Program

**INSTRUMENT PROCESSING SHEET**

Agency Hialeah Police Department S/N 80-002463  
Date In 4/22/2016 Date Out 04/25/2016  Ship  P/U  H/D  CMI  EE

**Intake** Performed By DEER

Registration  
 Annual  
 Return from CMI  
 Return from Enforcement Electronics  
 Other \_\_\_\_\_

Visual Inspection:  
OK Case OK Handle  
OK Dry Gas Holder OK Feet  
OK Keyboard/Plug OK Back/Plugs  
OK Screws tight OK Breath Hose

Other Equipment:  
 Power cord  
 Printer Cable  
 Other ANTI-STATIC BAG

Notes: \_\_\_\_\_

**Quality Checks** Performed By DEER

Breath Tube Screen  
 Replace O-Rings  
 Instrument Set Up Verified  
 R-Value 156  
 Flow Verification (L/s)  
 Flow Column # ATP101  
 32mm 152 (.139 - .169)  
 36mm 171 (.156 - .190)  
 53mm 234 (.228 - .278)  
 103mm 492 (.447 - .547)

Barometric Pressure Check  
 Gauge ID # 28663

Stability Checks

Simulator	Serial #	Lot #/Exp
0.05	SD1014	201507A 07/14/2017
0.08	SD1015	2015026 02/24/2017
0.20	SD1017	201505A 05/12/2017
0.08 DGS	N/A	R659701 07/16/2017

**Flow Calibration** Performed By \_\_\_\_\_

Flow Calibration N/A  
 Flow Calibration Complete  
 Flow Column # \_\_\_\_\_  
 5L/min - 17mm  
 15L/min - 33mm  
 30L/min - 103mm

R-Value \_\_\_\_\_  
 Post Calibration Verification (L/s)  
 Flow Column # \_\_\_\_\_  
 32mm \_\_\_\_\_ (.139 - .169)  
 36mm \_\_\_\_\_ (.156 - .190)  
 53mm \_\_\_\_\_ (.228 - .278)  
 103mm \_\_\_\_\_ (.447 - .547)

**Maintenance** Performed By \_\_\_\_\_

Battery Replacement  
 Dry Gas Regulator Replacement  
 Breath Tube Replacement  
 Other \_\_\_\_\_

**Suggested Service**

**Optical Bench Calibration** Performed By \_\_\_\_\_

Optical Bench Calibration N/A  
 Optical Bench Calibration Complete

Barometric Pressure Gauge ID # \_\_\_\_\_

Simulator	Serial Number	Lot Number	Expiration
0.000		N/A	N/A
0.040			
0.100			
0.200			
0.400			
0.080 DGS	N/A		

Post Calibration Stability Checks

Simulator	Serial Number	Lot Number	Expiration
0.05			
0.08			
0.20			
0.08 DGS	N/A		

**Department Inspection** Performed By DEER

Barometric Pressure 1015 Gauge  
 ID# 28199 1011 Instrument

Mouth Alcohol Solution Lot # 2015A  
 Acetone Stock Solution Lot # 2015B

Simulator	Serial Number
0.00	G8147
Interferent	G12100
0.05	SD1014
0.08	SD1015
0.20	SD1017

**Attachments**

Form 41  Optical Bench Cal  
 Pre-Stability Tests  Post-Stability Tests  
 Flow Calibration  Other \_\_\_\_\_

Notes:  **E-MAILED**  **APPROVED**  
04/25/2016  
RA/OC OK PPM

Instrument Complies with Chapter 11D-8, FAC  
 Instrument Does Not Comply with Chapter 11D-8, FAC  
 Return to/Place into Evidentiary Use  
 Remain Out of Evidentiary Use  
 Conduct an Agency Inspection Before Evidentiary Use

Brett Hurlbut  
 Quality Control Review

5/6/16  
 Date

TYPE OF TEST	SERIAL NUMBER	AGENCY	DATE	PERFORMED BY
Stabilities	80-002463	Hialeah Police Department	04/25/2016	<i>WLL</i>

0.05g/210L <input checked="" type="checkbox"/>	0.08g/210L <input checked="" type="checkbox"/>	0.20g/210L <input checked="" type="checkbox"/>	DGS 0.08g/210L <input checked="" type="checkbox"/>
0.047 to 0.053 <input checked="" type="checkbox"/>	0.077 to 0.083 <input checked="" type="checkbox"/>	0.194 to 0.206 <input checked="" type="checkbox"/>	0.077 to 0.083 <input checked="" type="checkbox"/>

HIALEAH PD  
 Intoxilyzer - 9i Coronol Analyzer  
 Model: 8009 SN: 80-002463  
 04/25/2016  
 Software: 8:00:27

Test g/210L Time

Air Blank 0.000 08:21  
 Control Test 0.051 08:22  
 Air Blank 0.000 08:23  
 Control Test 0.051 08:23  
 Air Blank 0.000 08:24  
 Control Test 0.051 08:25  
 Air Blank 0.000 08:25  
 Control Test Stats  
 Average 0.0510  
 Std Dev 0.0008  
 Rel Std Dev(%) 0.0006

*WLL*  
 Operator's Signature

HIALEAH PD  
 Intoxilyzer - Alcoron Analyzer  
 Model: 8100 SN: 80-002463  
 04/25/2016  
 Software: 8:00:27

Test g/210L Time

Air Blank 0.000 08:26  
 Control Test 0.080 08:27  
 Air Blank 0.000 08:28  
 Control Test 0.080 08:28  
 Air Blank 0.000 08:29  
 Control Test 0.080 08:29  
 Air Blank 0.000 08:30  
 Control Test Stats  
 Average 0.0800  
 Std Dev 0.0006  
 Rel Std Dev(%) 0.0006

*WLL*  
 Operator's Signature

HIALEAH PD  
 Intoxilyzer - Alcoron Analyzer  
 Model: 8100 SN: 80-002463  
 04/25/2016  
 Software: 8:00:27

Test g/210L Time

Air Blank 0.000 08:32  
 Control Test 0.201 08:32  
 Air Blank 0.000 08:33  
 Control Test 0.200 08:34  
 Air Blank 0.000 08:34  
 Control Test 0.200 08:35  
 Air Blank 0.000 08:35  
 Control Test Stats  
 Average 0.2003  
 Std Dev 0.0006  
 Rel Std Dev(%) 0.2882

*WLL*  
 Operator's Signature

HIALEAH PD  
 Intoxilyzer - Alcoron Analyzer  
 Model: 8009 SN: 80-002463  
 04/25/2016  
 Software: 8:00:27

Test g/210L Time

Air Blank 0.000 08:37  
 Control Test 0.081 08:37  
 Air Blank 0.000 08:38  
 Control Test 0.081 08:38  
 Air Blank 0.000 08:39  
 Control Test 0.081 08:39  
 Air Blank 0.000 08:40  
 Control Test Stats  
 Average 0.0810  
 Std Dev 0.0008  
 Rel Std Dev(%) 0.0008

*WLL*  
 Operator's Signature

*WLL*