



Alcohol Testing Program

INSTRUMENT PROCESSING SHEET

Agency Miami Gardens Police Department S/N 80-002987

RECEIVED

AUG 15 2016

Date In 8/8/2016

Date Out 8/8/2016

Ship P/U H/D CMIFDLE

Alcohol Testing Program

Intake Performed By DERR

Registration
 Annual
 Return from CMI
 Return from Enforcement Electronics
 Other _____

Visual Inspection:
OK Case OK Handle
OK Dry Gas Holder OK Feet
OK Keyboard/Plug OK Back/Plugs
OK Screws tight OK Breath Hose

Other Equipment:
 Power cord
 Printer Cable
 Other Anti Static Bag

Notes: _____

Quality Checks Performed By DERR

Breath Tube Screen
 Replace O-Rings
 Instrument Set Up Verified
 R-Value 231
 Flow Verification (L/s)
 Flow Column # ATP.101
 32mm 156 (.139 - .169)
 36mm 171 (.156 - .190)
 53mm 242 (.228 - .278)
 103mm 500 (.447 - .547)

Barometric Pressure Check
 Gauge ID # 28663

Stability Checks

Simulator	Serial #	Lot #/Exp
0.05	SD 3963	201507A 07/14/2017
0.08	SD 3968	201601F 01/26/2018
0.20	SD 3969	201505A 05/12/2017
0.08 DGS	N/A	196600504 01/05/2018

Flow Calibration Performed By _____

Flow Calibration N/A
 Flow Calibration Complete
 Flow Column # _____
 5L/min - 17mm
 15L/min - 53mm
 30L/min - 103mm

R-Value _____
 Post Calibration Verification (L/s)
 Flow Column # _____
 32mm _____ (.139 - .169)
 36mm _____ (.156 - .190)
 53mm _____ (.228 - .278)
 103mm _____ (.447 - .547)

Maintenance Performed By _____

Battery Replacement
 Dry Gas Regulator Replacement
 Breath Tube Replacement
 Other _____

Suggested Service

Optical Bench Calibration Performed By _____

Optical Bench Calibration N/A
 Optical Bench Calibration Complete

Barometric Pressure Gauge ID # _____

Simulator	Serial Number	Lot Number	Expiration
0.000		N/A	N/A
0.040			
0.100			
0.200			
0.400			
0.080 DGS	N/A		

Post Calibration Stability Checks

Simulator	Serial Number	Lot Number	Expiration
0.05			
0.08			
0.20			
0.08 DGS	N/A		

Department Inspection Performed By DERR

Barometric Pressure 1014 Gauge
 ID# 28199 1014 Instrument

Mouth Alcohol Solution Lot # 2016-A
 Acetone Stock Solution Lot # 2016-B

Simulator	Serial Number
0.00	SD 3963
Interferent	SD 3966
0.05	SD 3967
0.08	SD 3968
0.20	SD 3969

Attachments

Form 41
 Pre-Stability Tests
 Flow Calibration

Optical Bench Cal
 Post-Stability Tests
 Other _____

Notes: **E-MAILED** **APPROVED**
8/8/2016

QC/QA OK upon 8/12/16

Brett Wickland

Instrument Complies with Chapter 11D-8, FAC
 Instrument Does Not Comply with Chapter 11D-8, FAC

Return to/Place into Evidentiary Use
 Remain Out of Evidentiary Use

Conduct an Agency Inspection Before Evidentiary Use

8/15/16

Quality Control Review

Date

TYPE OF TEST	SERIAL NUMBER	AGENCY	DATE	PERFORMED BY
Stabilities	80-002987	Miami Gardens Police Department	08/08/2016	<i>ML</i>

0.05g/210L	0.08g/210L	0.20g/210L	DGS 0.08g/210L
0.047 to 0.053 <input checked="" type="checkbox"/>	0.077 to 0.083 <input checked="" type="checkbox"/>	0.194 to 0.206 <input checked="" type="checkbox"/>	0.077 to 0.083 <input checked="" type="checkbox"/>

Test	Time	Test	Time	Test	Time	Test	Time
MIAMI GARDENS PD Intoxilyzer - Alcotest Analyzer Model 8000 08/08/2016 Software: 8100.27	08:06	MIAMI GARDENS PD Intoxilyzer - Alcotest Analyzer Model 8000 08/08/2016 Software: 8100.27	08:11	MIAMI GARDENS PD Intoxilyzer - Alcotest Analyzer Model 8000 08/08/2016 Software: 8100.27	08:16	MIAMI GARDENS PD Intoxilyzer - Alcotest Analyzer Model 8000 08/08/2016 Software: 8100.27	08:21
Air Blank 0.008	08:06	Air Blank 0.003	08:11	Air Blank 0.006	08:16	Air Blank 0.002	08:21
Control Test 0.051	08:07	Control Test 0.080	08:12	Control Test 0.203	08:17	Control Test 0.079	08:22
Air Blank 0.009	08:08	Air Blank 0.000	08:12	Air Blank 0.000	08:17	Air Blank 0.000	08:22
Control Test 0.056	08:08	Control Test 0.091	08:13	Control Test 0.201	08:18	Control Test 0.079	08:23
Air Blank 0.003	08:09	Air Blank 0.000	08:14	Air Blank 0.000	08:18	Air Blank 0.000	08:23
Control Test 0.051	08:10	Control Test 0.086	08:14	Control Test 0.201	08:19	Control Test 0.080	08:23
Air Blank 0.000	08:10	Air Blank 0.000	08:15	Air Blank 0.000	08:20	Air Blank 0.000	08:24
Control Test Status		Control Test Status		Control Test Status		Control Test Status	
Average 0.0507		Average 0.0903		Average 0.2007		Average 0.0793	
Std Dev 0.0006		Std Dev 0.0036		Std Dev 0.0006		Std Dev 0.0006	
Rel Std Dev(%) 1.1395		Rel Std Dev(%) 0.7187		Rel Std Dev(%) 0.2877		Rel Std Dev(%) 0.7277	
<i>ML</i> Operator's Signature		<i>ML</i> Operator's Signature		<i>ML</i> Operator's Signature		<i>ML</i> Operator's Signature	

ASK