



INSTRUMENT PROCESSING SHEET

Agency FWC S/N 80-003195  
Date In 4/20/16 Date Out 4/25/16  Ship  P/U  H/D  CMI  EE

**Intake** Performed By (Signature)

Registration  
 Annual  
 Return from CMI  
 Return from Enforcement  
 Electronics  
 Other \_\_\_\_\_

Visual Inspection:  
 Case OK  Handle OK  
 Dry Gas Holder OK  Feet OK  
 Keyboard/Plug OK  Back/Plugs OK  
 Screws tight OK  Breath Hose OK

Other Equipment:  
 Power cord  
 Printer Cable  
 Other Cox Check

Notes: \_\_\_\_\_  
 \_\_\_\_\_  
 \_\_\_\_\_

**Quality Checks** Performed By (Signature)

Breath Tube Screen  
 Replace O-Rings  
 Instrument Set Up Verified  
 R-Value 127  
 Flow Verification (L/s)  
 Flow Column # ATP103  
 32mm .148 (.139 - .169)  
 36mm .160 (.156 - .190)  
 53mm .234 (.228 - .278)  
 103mm .507 (.447 - .547)

Barometric Pressure Check  
 Gauge ID # 28427

Stability Checks

Simulator	Serial #	Lot #/Exp
0.05	SD1018	201507A 7/14/17
0.08	SD1011	201601E 1/26/18
0.20	G4444	201505A 5/12/17
0.08 DGS	N/A	AG600504 1/5/18

**Flow Calibration** Performed By \_\_\_\_\_

Flow Calibration N/A **RECEIVED**  
 Flow Calibration Complete  
 Flow Column # \_\_\_\_\_  
 5L/min - 17mm APR 25 2016  
 15L/min - 53mm **FDLE**  
 30L/min - 103mm **Alcohol Testing Program**

R-Value \_\_\_\_\_  
 Post Calibration Verification (L/s)  
 Flow Column # \_\_\_\_\_  
 32mm \_\_\_\_\_ (.139 - .169)  
 36mm \_\_\_\_\_ (.156 - .190)  
 53mm \_\_\_\_\_ (.228 - .278)  
 103mm \_\_\_\_\_ (.447 - .547)

**Maintenance** Performed By \_\_\_\_\_

Battery Replacement  
 Dry Gas Regulator Replacement  
 Breath Tube Replacement  
 Other \_\_\_\_\_

**Suggested Service**  
 \_\_\_\_\_  
 \_\_\_\_\_

**Optical Bench Calibration** Performed By \_\_\_\_\_

Optical Bench Calibration N/A  
 Optical Bench Calibration Complete  
 Barometric Pressure Gauge ID # \_\_\_\_\_

Simulator	Serial Number	Lot Number	Expiration
0.000		N/A	N/A
0.040			
0.100			
0.200			
0.400			
0.080 DGS	N/A		

Post Calibration Stability Checks

Simulator	Serial Number	Lot Number	Expiration
0.05			
0.08			
0.20			
0.08 DGS	N/A		

**Department Inspection** Performed By (Signature)

Barometric Pressure 1014 Gauge  
 ID# 28427 1013 Instrument

Mouth Alcohol Solution Lot # 2016-A  
 Acetone Stock Solution Lot # 2016-B

Simulator	Serial Number
0.00	SD1022
Interferent	SD1021
0.05	SD1018
0.08	SD1011
0.20	G4444

**Attachments**

Form 41  
 Pre-Stability Tests  
 Flow Calibration

Optical Bench Cal  
 Post-Stability Tests  
 Other \_\_\_\_\_

Notes: OK - PWS  
 \_\_\_\_\_  
 \_\_\_\_\_

Instrument Complies with Chapter 11D-8, FAC  
 Instrument Does Not Comply with Chapter 11D-8, FAC

Return to/Place into Evidentiary Use  
 Remain Out of Evidentiary Use

Conduct an Agency Inspection Before Evidentiary Use

Brett Hinkland  
Quality Control Review

4/25/16  
Date

BK

FWC  
Intoxilyzer - Alcohol Analyzer  
Model 8000 SN 80-003195  
04/22/2016  
Software: 8100.27

FWC  
Intoxilyzer - Alcohol Analyzer  
Model 8000 SN 80-003195  
04/22/2016  
Software: 8100.27

FWC  
Intoxilyzer - Alcohol Analyzer  
Model 8000 SN 80-003195  
04/22/2016  
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	11:23
Control Test	0.049	11:24
Air Blank	0.000	11:25
Control Test	0.049	11:25
Air Blank	0.000	11:26
Control Test	0.050	11:27
Air Blank	0.000	11:27
Control Test Stats		
Average	0.0493	
Std Dev	0.0006	
Rel Std Dev(%)	1.1703	

Test	g/210L	Time
Air Blank	0.000	11:18
Control Test	0.078	11:18
Air Blank	0.000	11:19
Control Test	0.078	11:19
Air Blank	0.000	11:20
Control Test	0.079	11:21
Air Blank	0.000	11:21
Control Test Stats		
Average	0.0783	
Std Dev	0.0006	
Rel Std Dev(%)	0.7370	

Test	g/210L	Time
Air Blank	0.000	11:30
Control Test	0.199	11:30
Air Blank	0.000	11:31
Control Test	0.199	11:31
Air Blank	0.000	11:32
Control Test	0.200	11:33
Air Blank	0.000	11:33
Control Test Stats		
Average	0.1993	
Std Dev	0.0006	
Rel Std Dev(%)	0.2896	

*P Murphy*  
Operator's Signature

*P Murphy*  
Operator's Signature

*P Murphy*  
Operator's Signature

PWS

FWC  
Intoxilyzer - Alcohol Analyzer  
Model 8000 SN 80-003195  
04/22/2016  
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	11:42
Control Test	0.080	11:42
Air Blank	0.000	11:42
Control Test	0.080	11:43
Air Blank	0.000	11:43
Control Test	0.079	11:43
Air Blank	0.000	11:44
Control Test Stats		
Average	0.0797	
Std Dev	0.0006	
Rel Std Dev(%)	0.7247	

DGS

*P Murphy*  
Operator's Signature