



Alcohol Testing Program

**INSTRUMENT PROCESSING SHEET**

Agency FWC

S/N 80-003196

Date In 4/20/16

Date Out 4/25/16

Ship  P/U  H/D  CMI  EE

**RECEIVED**  
APR 25 2016  
FDLE  
Alcohol Testing Program

**Intake** Performed By JD

Registration  
 Annual  
 Return from CMI  
 Return from Enforcement  
 Electronics  
 Other \_\_\_\_\_

Visual Inspection:  
 Case  Handle  
 Dry Gas Holder  Feet  
 Keyboard/Plug  Back/Plugs  
 Screws tight  Breath Hose

Other Equipment:  
 Power cord  
 Printer Cable  
 Other Car Charger

Notes: \_\_\_\_\_

**Quality Checks** Performed By WSON

Breath Tube Screen  
 Replace O-Rings  
 Instrument Set Up Verified  
 R-Value 144  
 Flow Verification (L/s)  
 Flow Column # ATP103  
 32mm .140 (.139 - .169)  
 36mm .156 (.156 - .190)  
 53mm .230 (.228 - .278)  
 103mm .492 (.447 - .547)

Barometric Pressure Check  
 Gauge ID # 28427

Stability Checks

Simulator	Serial #	Lot #/Exp
0.05	SD1018	201507A 07/14/17
0.08	SD1011	201601F 1/26/18
0.20	64444	201565A 5/12/17
0.08 DGS	N/A	AG600504 1/5/18

**Flow Calibration** Performed By \_\_\_\_\_

Flow Calibration N/A  
 Flow Calibration Complete  
 Flow Column # \_\_\_\_\_  
 5L/min - 17mm  
 15L/min - 53mm  
 30L/min - 103mm

R-Value \_\_\_\_\_  
 Post Calibration Verification (L/s)  
 Flow Column # \_\_\_\_\_  
 32mm \_\_\_\_\_ (.139 - .169)  
 36mm \_\_\_\_\_ (.156 - .190)  
 53mm \_\_\_\_\_ (.228 - .278)  
 103mm \_\_\_\_\_ (.447 - .547)

**Maintenance** Performed By \_\_\_\_\_

Battery Replacement  
 Dry Gas Regulator Replacement  
 Breath Tube Replacement  
 Other \_\_\_\_\_

**Suggested Service**

**Optical Bench Calibration** Performed By \_\_\_\_\_

Optical Bench Calibration N/A  
 Optical Bench Calibration Complete

Barometric Pressure Gauge ID # \_\_\_\_\_

Simulator	Serial Number	Lot Number	Expiration
0.000		N/A	N/A
0.040			
0.100			
0.200			
0.400			
0.080 DGS	N/A		

Post Calibration Stability Checks

Simulator	Serial Number	Lot Number	Expiration
0.05			
0.08			
0.20			
0.08 DGS	N/A		

**Department Inspection** Performed By WSON

Barometric Pressure  
 ID# 28427 Gauge 1014  
 Instrument 1013

Mouth Alcohol Solution Lot # 2016-A  
 Acetone Stock Solution Lot # 2016-B

Simulator	Serial Number
0.00	SD1022
Interferent	SD1021
0.05	SD1018
0.08	SD1011
0.20	64444

**Attachments**

Form 41  
 Pre-Stability Tests  
 Flow Calibration

Optical Bench Cal  
 Post-Stability Tests  
 Other \_\_\_\_\_

Notes: OK - PWS

Instrument Complies with Chapter 11D-8, FAC  
 Instrument Does Not Comply with Chapter 11D-8, FAC

Return to/Place into Evidentiary Use  
 Remain Out of Evidentiary Use

Conduct an Agency Inspection Before Evidentiary Use

Brett Hinkland

Quality Control Review

4/25/16

Date

FWC  
 Intoxilyzer - Alcohol Analyzer  
 Model 8000 SN 80-003196  
 04/22/2016  
 Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	11:17
Control Test	0.048	11:18
Air Blank	0.000	11:19
Control Test	0.049	11:19
Air Blank	0.000	11:20
Control Test	0.049	11:20
Air Blank	0.000	11:21
Control Test Stats		
Average	0.0487	
Std Dev	0.0006	
Rel Std Dev(%)	1.1863	

*P Murphy*  
 Operator's Signature

FWC  
 Intoxilyzer - Alcohol Analyzer  
 Model 8000 SN 80-003196  
 04/22/2016  
 Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	11:30
Control Test	0.078	11:30
Air Blank	0.000	11:31
Control Test	0.079	11:32
Air Blank	0.000	11:32
Control Test	0.079	11:33
Air Blank	0.000	11:33
Control Test Stats		
Average	0.0787	
Std Dev	0.0006	
Rel Std Dev(%)	0.7339	

*P Murphy*  
 Operator's Signature

FWC  
 Intoxilyzer - Alcohol Analyzer  
 Model 8000 SN 80-003196  
 04/22/2016  
 Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	11:23
Control Test	0.198	11:24
Air Blank	0.000	11:24
Control Test	0.198	11:25
Air Blank	0.000	11:26
Control Test	0.198	11:26
Air Blank	0.000	11:27
Control Test Stats		
Average	0.1980	
Std Dev	0.0000	
Rel Std Dev(%)	0.0000	

*P Murphy*  
 Operator's Signature

BSK

FWC  
 Intoxilyzer - Alcohol Analyzer  
 Model 8000 SN 80-003196  
 04/22/2016  
 Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	11:38
Control Test	0.079	11:38
Air Blank	0.000	11:38
Control Test	0.078	11:39
Air Blank	0.000	11:39
Control Test	0.078	11:40
Air Blank	0.000	11:40
Control Test Stats		
Average	0.0793	
Std Dev	0.0006	
Rel Std Dev(%)	0.7370	

DGS

*P Murphy*  
 Operator's Signature

PWS