



Alcohol Testing Program

INSTRUMENT PROCESSING SHEET

Agency Sweetwater Police Department S/N 80-003348

Date In 9/20/2016 Date Out 9/20/2016  Ship  P/U  H/D  CMI  EE

<b>Intake</b> Performed By <u>DEER</u> <input checked="" type="checkbox"/> Registration <input checked="" type="checkbox"/> Annual <input type="checkbox"/> Return from CMI <input type="checkbox"/> Return from Enforcement <input type="checkbox"/> Electronics <input checked="" type="checkbox"/> Other <u>DE REGISTER</u> Visual Inspection: <u>OK</u> Case <u>OK</u> Handle <u>OK</u> Dry Gas Holder <u>OK</u> Feet <u>OK</u> Keyboard/Plug <u>OK</u> Back/Plugs <u>OK</u> Screws tight <u>OK</u> Breath Hose Other Equipment: <input type="checkbox"/> Power cord <input type="checkbox"/> Printer Cable <input checked="" type="checkbox"/> Other <u>ANTI STATIC BAG</u> Notes: _____ _____ _____		<b>Quality Checks</b> Performed By <u>DEER</u> <input checked="" type="checkbox"/> Breath Tube Screen <input checked="" type="checkbox"/> Replace O-Rings <input checked="" type="checkbox"/> Instrument Set Up Verified <input checked="" type="checkbox"/> R-Value <u>207</u> <input checked="" type="checkbox"/> Flow Verification (L/s) Flow Column # <u>ATP 101</u> 32mm <u>156</u> (.139 - .169) 36mm <u>171</u> (.156 - .190) 53mm <u>242</u> (.228 - .278) 103mm <u>503</u> (.447 - .547) <input checked="" type="checkbox"/> Barometric Pressure Check Gauge ID # <u>28663</u> <input checked="" type="checkbox"/> Stability Checks		<b>Flow Calibration</b> Performed By _____ <input checked="" type="checkbox"/> Flow Calibration N/A <input type="checkbox"/> Flow Calibration Complete Flow Column # _____ <input type="checkbox"/> 5L/min - 17mm <input type="checkbox"/> 15L/min - 53mm <input type="checkbox"/> 30L/min - 103mm <input type="checkbox"/> R-Value _____ <input type="checkbox"/> Post Calibration Verification (L/s) Flow Column # _____ 32mm _____ (.139 - .169) 36mm _____ (.156 - .190) 53mm _____ (.228 - .278) 103mm _____ (.447 - .547)																
		<table border="1"> <thead> <tr> <th>Simulator</th> <th>Serial #</th> <th>Lot #/Exp</th> </tr> </thead> <tbody> <tr> <td>0.05</td> <td>SD 3967</td> <td>201507A 07/14/2017</td> </tr> <tr> <td>0.08</td> <td>SD 3968</td> <td>201601F 01/26/2018</td> </tr> <tr> <td>0.20</td> <td>SD 3963</td> <td>201505A 05/12/2017</td> </tr> <tr> <td>0.08 DGS</td> <td>N/A</td> <td>A6600SD4 01/05/2018</td> </tr> </tbody> </table>		Simulator	Serial #	Lot #/Exp	0.05	SD 3967	201507A 07/14/2017	0.08	SD 3968	201601F 01/26/2018	0.20	SD 3963	201505A 05/12/2017	0.08 DGS	N/A	A6600SD4 01/05/2018	<b>Maintenance</b> Performed By _____ <input type="checkbox"/> Battery Replacement <input type="checkbox"/> Dry Gas Regulator Replacement <input type="checkbox"/> Breath Tube Replacement <input type="checkbox"/> Other _____ <b>Suggested Service</b> _____ _____	
Simulator	Serial #	Lot #/Exp																		
0.05	SD 3967	201507A 07/14/2017																		
0.08	SD 3968	201601F 01/26/2018																		
0.20	SD 3963	201505A 05/12/2017																		
0.08 DGS	N/A	A6600SD4 01/05/2018																		

RECEIVED  
OCT 18 2016  
FDLE  
Alcohol Testing Program

<b>Optical Bench Calibration</b> Performed By _____ <input checked="" type="checkbox"/> Optical Bench Calibration N/A <input type="checkbox"/> Optical Bench Calibration Complete Barometric Pressure Gauge ID # _____			
Simulator	Serial Number	Lot Number	Expiration
0.000		N/A	N/A
0.040			
0.100			
0.200			
0.400			
0.080 DGS	N/A		
<input type="checkbox"/> Post Calibration Stability Checks			
Simulator	Serial Number	Lot Number	Expiration
0.05			
0.08			
0.20			
0.08 DGS	N/A		

<b>Department Inspection</b> Performed By <u>DEER</u> <input checked="" type="checkbox"/> Barometric Pressure <u>1015</u> Gauge ID# <u>28199</u> <u>1015</u> Instrument	
Mouth Alcohol Solution Lot # <u>2016-A</u> Acetone Stock Solution Lot # <u>2016-B</u>	
Simulator	Serial Number
0.00	SD 3965
Interferent	SD 3966
0.05	SD 3967
0.08	SD 3968
0.20	SD 3969
<b>Attachments</b> <input checked="" type="checkbox"/> Form 41 <input checked="" type="checkbox"/> Pre-Stability Tests <input type="checkbox"/> Flow Calibration <input type="checkbox"/> Optical Bench Cal <input type="checkbox"/> Post-Stability Tests <input checked="" type="checkbox"/> Other <u>Registration</u>	

Notes: **E-MAILED**  
09/20/2016  
Deregister from Sweetwater; register to FHP  
DR/OC OK upon 10/18/2016  
Brett Kirkland

<input checked="" type="checkbox"/> Instrument Complies with Chapter 11D-8, FAC <input type="checkbox"/> Instrument Does Not Comply with Chapter 11D-8, FAC <input checked="" type="checkbox"/> Return to/Place into Evidentiary Use <input type="checkbox"/> Remain Out of Evidentiary Use <input checked="" type="checkbox"/> Conduct an Agency Inspection Before Evidentiary Use	<u>10/18/16</u> Date
---	-------------------------

Quality Control Review

Date

TYPE OF TEST	SERIAL NUMBER	AGENCY	DATE	PERFORMED BY
Stabilities	80-003348	Sweetwater Police Department	9/20/2016	<i>Bill</i>

0.05g/210L

0.047 to 0.053

0.08g/210L

0.077 to 0.083

0.20g/210L

0.194 to 0.206

DGS 0.08g/210L

0.077 to 0.083

FPP TROOP E MIAMI  
Intoxilyzer - Alcohol Analyzer  
Model 8000 SN 80-003348  
09/20/2016  
Software: 8100.27

Test 9/21/16 Time

Air-Blank 0.800 10:13  
Control Test 0.149 10:14  
Air-Blank 0.000 10:14  
Control Test 0.050 10:15  
Air-Blank 0.000 10:15  
Control Test 0.051 10:15  
Air-Blank 0.000 10:17  
Control Test Stats  
Average 0.0500  
Std Dev 0.0010  
Rel Std Dev(%) 2.0000

FPP TROOP E MIAMI  
Intoxilyzer - Alcohol Analyzer  
Model 8000 SN 80-003348  
09/20/2016  
Software: 8100.27

Test 9/21/16 Time

Air-Blank 0.000 10:18  
Control Test 0.080 10:19  
Air-Blank 0.000 10:19  
Control Test 0.081 10:20  
Air-Blank 0.000 10:20  
Control Test 0.079 10:21  
Air-Blank 0.000 10:21  
Control Test Stats  
Average 0.0800  
Std Dev 0.0010  
Rel Std Dev(%) 1.2500

FPP TROOP E MIAMI  
Intoxilyzer - Alcohol Analyzer  
Model 8000 SN 80-003348  
09/20/2016  
Software: 8100.27

Test 9/21/16 Time

Air-Blank 0.000 10:23  
Control Test 0.202 10:24  
Air-Blank 0.000 10:24  
Control Test 0.201 10:25  
Air-Blank 0.000 10:25  
Control Test 0.201 10:25  
Air-Blank 0.000 10:26  
Control Test Stats  
Average 0.2013  
Std Dev 0.0006  
Rel Std Dev(%) 0.2868

FPP TROOP E MIAMI  
Intoxilyzer - Alcohol Analyzer  
Model 8000 SN 80-003348  
09/20/2016  
Software: 8100.27

Test 9/21/16 Time

Air-Blank 0.000 10:28  
Control Test 0.080 10:28  
Air-Blank 0.000 10:28  
Control Test 0.080 10:29  
Air-Blank 0.000 10:29  
Control Test 0.080 10:29  
Air-Blank 0.000 10:30  
Control Test Stats  
Average 0.0800  
Std Dev 0.0000  
Rel Std Dev(%) 0.0000

*Bill*  
Operator's Signature

*Bill*  
Operator's Signature

*Bill*  
Operator's Signature

*Bill*  
Operator's Signature

*ggm*