



INSTRUMENT PROCESSING SHEET

Agency Hillsborough County SO S/N 80-003388

Date In 11/18/15 Date Out 1/20/16 Ship P/U H/D CMI EE

Intake Performed By [Signature]
Registration
Annual
Return from CMI
Return from Enforcement Electronics
Other
Visual Inspection:
OK Case OK Handle
OK Dry Gas Holder OK Feet
OK Keyboard/Plug OK Back/Plugs
OK Screws tight OK Breath Hose
Other Equipment:
Power cord
Printer Cable
Other Static bag
Notes:

Quality Checks Performed By [Signature]
Breath Tube Screen
Replace O-Rings
Instrument Set Up Verified
R-Value 190
Flow Verification (L/s)
Flow Column # ATP102
32mm D.152 (.139-.169)
36mm D.171 (.156-.190)
53mm D.238 (.228-.278)
103mm D.507 (.447-.547)
Barometric Pressure Check
Gauge ID # 26932
Stability Checks
Simulator Serial # Lot #/Exp
0.05 SD1024 201507A 7/14/17
0.08 DR2035 201502G 2/24/17
0.20 DR3856 201505A 5/12/17
0.08 DGS N/A AG511701 4/27/17

Flow Calibration Performed By [Signature]
Flow Calibration N/A
Flow Calibration Complete
Flow Column #
5L/min - 17mm
15L/min - 53mm
30L/min - 103mm
R-Value
Post Calibration Verification (L/s)
Flow Column #
32mm (.139-.169)
36mm (.156-.190)
53mm (.228-.278)
103mm (.447-.547)
Maintenance Performed By
Battery Replacement
Dry Gas Regulator Replacement
Breath Tube Replacement
Other
Suggested Service

RECEIVED
JAN 20 2016
FDLE
Alcohol Testing Program

Optical Bench Calibration Performed By [Signature]
Optical Bench Calibration N/A
Optical Bench Calibration Complete
Barometric Pressure Gauge ID #
Simulator Serial Number Lot Number Expiration
0.000 N/A N/A
0.040
0.100
0.200
0.400
0.080 DGS N/A
Post Calibration Stability Checks
Simulator Serial Number Lot Number Expiration
0.05
0.08
0.20
0.08 DGS N/A

Department Inspection Performed By [Signature]
Barometric Pressure 1024 Gauge
ID# 26932 1023 Instrument
Mouth Alcohol Solution Lot # 2015-A
Acetone Stock Solution Lot # 2015-B
Simulator Serial Number
0.00 SD1016
Interferent G2840
0.05 SD1024
0.08 DR DR2035
0.20 DR3856
Attachments
Form 41
Pre-Stability Tests
Flow Calibration
Optical Bench Cal
Post-Stability Tests
Other FORM 40

Notes: [Signature] 1/20/16

Instrument Complies with Chapter 11D-8, FAC
Instrument Does Not Comply with Chapter 11D-8, FAC
Return to/Place into Evidentiary Use
Remain Out of Evidentiary Use
Conduct an Agency Inspection Before Evidentiary Use

Brett Huikland

Quality Control Review

1/20/16

Date

Stability Checks 80-003388 Hillsborough County SO. 1/20/16 *DMS*

800

HILLSBOROUGH CO SO
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-003388
01/20/2016
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	10:54
Control Test	0.051	10:55
Air Blank	0.000	10:55
Control Test	0.052	10:56
Air Blank	0.000	10:57
Control Test	0.052	10:57
Air Blank	0.000	10:58
Control Test Stats		
Average	0.0517	
Std Dev	0.0006	
Rel Std Dev(%)	1.1175	

DMS
Operator's Signature

HILLSBOROUGH CO SO
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-003388
01/20/2016
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	10:43
Control Test	0.204	10:43
Air Blank	0.000	10:44
Control Test	0.203	10:44
Air Blank	0.000	10:45
Control Test	0.204	10:46
Air Blank	0.000	10:46
Control Test Stats		
Average	0.2037	
Std Dev	0.0006	
Rel Std Dev(%)	0.2835	

DMS
Operator's Signature

HILLSBOROUGH CO SO
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-003388
01/20/2016
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	10:50
Control Test	0.082	10:50
Air Blank	0.000	10:51
Control Test	0.081	10:51
Air Blank	0.000	10:52
Control Test	0.081	10:52
Air Blank	0.000	10:53
Control Test Stats		
Average	0.0813	
Std Dev	0.0006	
Rel Std Dev(%)	0.7099	

BSK

DMS
Operator's Signature

INSTRUMENT PROCESSING SHEET

Agency Hillsborough Co SD S/N 80-00 3388
 Date In 8/12/15 Date Out 8/13/15 Ship P/U H/D CMI EE

Intake Performed By [Signature]

Registration
 Annual
 Return from CMI
 Return from Enforcement Electronics
 Other _____

Visual Inspection:
OK Case OK Handle
OK Dry Gas Holder OK Feet
OK Keyboard/Plug OK Back/Plugs
OK Screws tight OK Breath Hose

Other Equipment:
 Power cord
 Printer Cable
 Other Static Bag

Notes: _____

Quality Checks Performed By _____

Breath Tube Screen
 Replace O-Rings
 Instrument Set Up Verified
 R-Value _____
 Flow Verification (L/s)
 Flow Column # _____
 32mm _____ (.139 - .169)
 36mm _____ (.156 - .190)
 53mm _____ (.228 - .278)
 103mm _____ (.447 - .547)

Barometric Pressure Check
 Gauge ID # _____

Stability Checks

Simulator	Serial #	Lot #/Exp
0.05		
0.08		
0.20		
0.08 DGS	N/A	

Flow Calibration Performed By _____

Flow Calibration N/A
 Flow Calibration Complete
 Flow Column # _____
 5L/min - 17mm
 15L/min - 53mm
 30L/min - 103mm

R-Value _____
 Post Calibration Verification (L/s)
 Flow Column # _____
 32mm _____ (.139 - .169)
 36mm _____ (.156 - .190)
 53mm _____ (.228 - .278)
 103mm _____ (.447 - .547)

Maintenance Performed By _____

Battery Replacement
 Dry Gas Regulator Replacement
 Breath Tube Replacement
 Other _____

Suggested Service
Modem repair

RECEIVED
 JAN 20 2016
 FDLE
 Alcohol Testing Program

Optical Bench Calibration Performed By _____

Optical Bench Calibration N/A
 Optical Bench Calibration Complete

Barometric Pressure Gauge ID # _____

Simulator	Serial Number	Lot Number	Expiration
0.000		N/A	N/A
0.040			
0.100			
0.200			
0.400			
0.080 DGS	N/A		

Post Calibration Stability Checks

Simulator	Serial Number	Lot Number	Expiration
0.05			
0.08			
0.20			
0.08 DGS	N/A		

Department Inspection Performed By _____

Barometric Pressure _____ Gauge
 ID# _____ Instrument

Mouth Alcohol Solution Lot # _____
 Acetone Stock Solution Lot # _____

Simulator	Serial Number
0.00	
Interferent	
0.05	
0.08	
0.20	

Attachments

<input type="checkbox"/> Form 41	<input type="checkbox"/> Optical Bench Cal
<input type="checkbox"/> Pre-Stability Tests	<input type="checkbox"/> Post-Stability Tests
<input type="checkbox"/> Flow Calibration	<input type="checkbox"/> Other _____

Notes: Instrument uploaded via direct connect. No Dial Tone confirmed. Compliance not determined.

Instrument Complies with Chapter 11D-8, FAC
 Instrument Does Not Comply with Chapter 11D-8, FAC

Return to/Place into Evidentiary Use
 Remain Out of Evidentiary Use
 Conduct an Agency Inspection Before Evidentiary Use