

INSTRUMENT PROCESSING SHEET

Agency FHP S/N 80-003406
 Date In 5/24/16 Date Out 5/26/16 Ship P/U H/D CMI EE

Intake Performed By PP

Registration
 Annual
 Return from CMI
 Return from Enforcement
 Electronics
 Other _____

Visual Inspection:
 Case Handle
 Dry Gas Holder Feet
 Keyboard/Plug Back/Plugs
 Screws tight Breath Hose

Other Equipment:
 Power cord
 Printer Cable
 Other _____

Notes: * Missing two Feet

Quality Checks Performed By PWS

Breath Tube Screen
 Replace O-Rings
 Instrument Set Up Verified
 R-Value 196
 Flow Verification (L/s)
 Flow Column # APDS
 32mm 144 (.139 - .169)
 36mm 164 (.156 - .190)
 53mm 234 (.228 - .278)
 103mm 523 (.447 - .547)

Barometric Pressure Check
 Gauge ID # 28427

Stability Checks

Simulator	Serial #	Lot #/Exp
0.05	G 2403	201507A 7/14/17
0.08	SD3964	201601F 1/20/18
0.20	G4444	201505A 5/2/17
0.08 DGS	N/A	AG 600504 1/5/18

Flow Calibration Performed By _____

Flow Calibration N/A
 Flow Calibration Complete
 Flow Column # _____
 5L/min - 17mm
 15L/min - 53mm
 30L/min - 103mm
 R-Value _____
 Post Calibration Verification (L/s)
 Flow Column # _____
 32mm _____ (.139 - .169)
 36mm _____ (.156 - .190)
 53mm _____ (.228 - .278)
 103mm _____ (.447 - .547)

Maintenance Performed By _____

Battery Replacement
 Dry Gas Regulator Replacement
 Breath Tube Replacement
 Other _____

Suggested Service

RECEIVED
 MAY 31 2016
 FDLE
 Alcohol Testing Program

Optical Bench Calibration Performed By _____

Optical Bench Calibration N/A
 Optical Bench Calibration Complete

Barometric Pressure Gauge ID # _____

Simulator	Serial Number	Lot Number	Expiration
0.000		N/A	N/A
0.040			
0.100			
0.200			
0.400			
0.080 DGS	N/A		

Post Calibration Stability Checks

Simulator	Serial Number	Lot Number	Expiration
0.05			
0.08			
0.20			
0.08 DGS	N/A		

Department Inspection Performed By PWS

Barometric Pressure ID# 28427 Gauge 1019
 Instrument 1017

Mouth Alcohol Solution Lot # 2015-A
 Acetone Stock Solution Lot # 2016-B

Simulator	Serial Number
0.00	G 2880
Interferent	G 2834
0.05	G 2403
0.08	SD 3964
0.20	G 4444

Attachments

Form 41 Optical Bench Cal
 Pre-Stability Tests Post-Stability Tests
 Flow Calibration Other _____

Notes: QA/PC OK PPM

Instrument Complies with Chapter 11D-8, FAC
 Instrument Does Not Comply with Chapter 11D-8, FAC

Return to/Place into Evidentiary Use
 Remain Out of Evidentiary Use

Conduct an Agency Inspection Before Evidentiary Use

Brett Huckelband
 Quality Control Review

5/31/16
 Date

Stability Tests - FHP #80-003406

5/26/16

FL HIGHWAY PATROL
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-003406
05/26/2016
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	12:16
Control Test	0.049	12:17
Air Blank	0.000	12:17
Control Test	0.048	12:18
Air Blank	0.000	12:18
Control Test	0.050	12:19
Air Blank	0.000	12:20
Control Test Stats		
Average	0.0490	
Std Dev	0.0010	
Rel Std Dev(%)	2.0408	

Handwritten initials

Handwritten signature

Operator's Signature

FL HIGHWAY PATROL
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-003406
05/26/2016
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	12:06
Control Test	0.079	12:07
Air Blank	0.000	12:07
Control Test	0.079	12:08
Air Blank	0.000	12:08
Control Test	0.080	12:09
Air Blank	0.000	12:09
Control Test Stats		
Average	0.0793	
Std Dev	0.0006	
Rel Std Dev(%)	0.7277	

Handwritten signature

Operator's Signature

FL HIGHWAY PATROL
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-003406
05/26/2016
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	12:11
Control Test	0.198	12:11
Air Blank	0.000	12:12
Control Test	0.199	12:13
Air Blank	0.000	12:13
Control Test	0.199	12:14
Air Blank	0.000	12:14
Control Test Stats		
Average	0.1987	
Std Dev	0.0006	
Rel Std Dev(%)	0.2906	

Handwritten signature

Operator's Signature

FL HIGHWAY PATROL
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-003406
05/26/2016
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	12:25
Control Test	0.091	12:25
Air Blank	0.000	12:26
Control Test	0.080	12:26
Air Blank	0.000	12:26
Control Test	0.081	12:27
Air Blank	0.000	12:27
Control Test Stats		
Average	0.0807	
Std Dev	0.0006	
Rel Std Dev(%)	0.7157	

Handwritten initials

Handwritten initials

Handwritten signature

Operator's Signature