

INSTRUMENT PROCESSING SHEET

Agency Florida Highway Patrol

S/N 80-003408

Date In 5/10/2016

Date Out 5/11/2016

Ship P/U H/D CMI EE

Intake Performed By <u>DERE</u> <input type="checkbox"/> Registration <input checked="" type="checkbox"/> Annual <input type="checkbox"/> Return from CMI <input type="checkbox"/> Return from Enforcement Electronics <input type="checkbox"/> Other _____ Visual Inspection: <u>OK</u> Case <u>OK</u> Handle <u>OK</u> Dry Gas Holder <u>OK</u> Feet <u>OK</u> Keyboard/Plug <u>OK</u> Back/Plugs <u>OK</u> Screws tight <u>OK</u> Breath Hose Other Equipment: <input type="checkbox"/> Power cord <input type="checkbox"/> Printer Cable <input checked="" type="checkbox"/> Other <u>ADD STATIC BAG</u> Notes: _____ _____ _____	Quality Checks Performed By <u>DERE</u> <input checked="" type="checkbox"/> Breath Tube Screen <input checked="" type="checkbox"/> Replace O-Rings <input checked="" type="checkbox"/> Instrument Set Up Verified <input checked="" type="checkbox"/> R-Value <u>163</u> <input checked="" type="checkbox"/> Flow Verification (L/s) Flow Column # <u>ATP101</u> 32mm <u>152</u> (.139 - .169) 36mm <u>171</u> (.156 - .190) 53mm <u>242</u> (.228 - .278) 103mm <u>507</u> (.447 - .547) <input checked="" type="checkbox"/> Barometric Pressure Check Gauge ID # <u>28663</u> <input checked="" type="checkbox"/> Stability Checks <table border="1" style="width:100%; border-collapse: collapse;"> <thead> <tr> <th>Simulator</th> <th>Serial #</th> <th>Lot #/Exp</th> </tr> </thead> <tbody> <tr> <td>0.05</td> <td>SD1014</td> <td>201507A 07/14/2017</td> </tr> <tr> <td>0.08</td> <td>SD1015</td> <td>2015026 02/24/2017</td> </tr> <tr> <td>0.20</td> <td>SD1017</td> <td>201505A 05/12/2017</td> </tr> <tr> <td>0.08 DGS</td> <td>N/A</td> <td>AG600504 01/05/2018</td> </tr> </tbody> </table>	Simulator	Serial #	Lot #/Exp	0.05	SD1014	201507A 07/14/2017	0.08	SD1015	2015026 02/24/2017	0.20	SD1017	201505A 05/12/2017	0.08 DGS	N/A	AG600504 01/05/2018	Flow Calibration Performed By _____ <input checked="" type="checkbox"/> Flow Calibration N/A <input type="checkbox"/> Flow Calibration Complete Flow Column # _____ <input type="checkbox"/> 5L/min - 17mm <input type="checkbox"/> 15L/min - 30mm <input type="checkbox"/> 30L/min - 103mm <input type="checkbox"/> R-Value _____ <input type="checkbox"/> Post Calibration Verification (L/s) Flow Column # _____ 32mm _____ (.139 - .169) 36mm _____ (.156 - .190) 53mm _____ (.228 - .278) 103mm _____ (.447 - .547) Maintenance Performed By _____ <input type="checkbox"/> Battery Replacement <input type="checkbox"/> Dry Gas Regulator Replacement <input type="checkbox"/> Breath Tube Replacement <input type="checkbox"/> Other _____ Suggested Service _____ _____
Simulator	Serial #	Lot #/Exp															
0.05	SD1014	201507A 07/14/2017															
0.08	SD1015	2015026 02/24/2017															
0.20	SD1017	201505A 05/12/2017															
0.08 DGS	N/A	AG600504 01/05/2018															

RECEIVED
 MAY 20 2016
 FDLE
 Alcohol Testing Program

Optical Bench Calibration Performed By _____ <input checked="" type="checkbox"/> Optical Bench Calibration N/A <input type="checkbox"/> Optical Bench Calibration Complete Barometric Pressure Gauge ID # _____																							
Simulator	Serial Number	Lot Number	Expiration																				
0.000		N/A	N/A																				
0.040																							
0.100																							
0.200																							
0.400																							
0.080 DGS	N/A																						
<input type="checkbox"/> Post Calibration Stability Checks <table border="1" style="width:100%; border-collapse: collapse;"> <thead> <tr> <th>Simulator</th> <th>Serial Number</th> <th>Lot Number</th> <th>Expiration</th> </tr> </thead> <tbody> <tr> <td>0.05</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.08</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.20</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.08 DGS</td> <td>N/A</td> <td></td> <td></td> </tr> </tbody> </table>				Simulator	Serial Number	Lot Number	Expiration	0.05				0.08				0.20				0.08 DGS	N/A		
Simulator	Serial Number	Lot Number	Expiration																				
0.05																							
0.08																							
0.20																							
0.08 DGS	N/A																						

Department Inspection Performed By <u>DERE</u> <input checked="" type="checkbox"/> Barometric Pressure <u>1019</u> Gauge ID# <u>28199</u> <u>1018</u> Instrument Mouth Alcohol Solution Lot # <u>2015-A</u> Acetone Stock Solution Lot # <u>2015-B</u>	
Simulator	Serial Number
0.00	G8147
Interferent	G12100
0.05	SD1014
0.08	SD1015
0.20	SD1017

Attachments <input checked="" type="checkbox"/> Form 41 <input checked="" type="checkbox"/> Pre-Stability Tests <input type="checkbox"/> Flow Calibration <input type="checkbox"/> Optical Bench Cal <input type="checkbox"/> Post-Stability Tests <input type="checkbox"/> Other _____	
--	--

Notes: **E-MAILED** **APPROVED**
5/11/2016
QA/QC OK 09PM

Instrument Complies with Chapter 11D-8, FAC
 Instrument Does Not Comply with Chapter 11D-8, FAC
 Return to/Place into Evidentiary Use
 Remain Out of Evidentiary Use
 Conduct an Agency Inspection Before Evidentiary Use

Beth Hartland
Quality Control Review

5/20/16
Date

TYPE OF TEST	SERIAL NUMBER	AGENCY	DATE	PERFORMED BY
Stabilities	80-003408	Florida Highway Patrol	05/11/2016	<i>[Signature]</i>

AK

0.05g/210L	0.08g/210L	0.20g/210L	DGS 0.08g/210L
0.047 to 0.053 <input checked="" type="checkbox"/>	0.077 to 0.083 <input checked="" type="checkbox"/>	0.194 to 0.206 <input checked="" type="checkbox"/>	0.077 to 0.083 <input checked="" type="checkbox"/>

FHP TROOP E MIAMI
Intoxilyzer - Alcotest Analyzer
Model 8000 SN 80-003408
05/11/2016
Software: 8100.27

FHP TROOP E MIAMI
Intoxilyzer - Alcotest Analyzer
Model 8000 SN 80-003408
05/11/2016
Software: 8100.27

FHP TROOP E MIAMI
Intoxilyzer - Alcotest Analyzer
Model 8000 SN 80-003408
05/11/2016
Software: 8100.27

FHP TROOP E MIAMI
Intoxilyzer - Alcotest Analyzer
Model 8000 SN 80-003408
05/11/2016
Software: 8100.27

[Signature]

Test g/210L Time
RfR Blank 0.000 09:09
Control Test 0.051 09:00
RfR Blank 0.000 09:01
Control Test 0.051 09:01
RfR Blank 0.000 09:02
Control Test 0.051 09:03
RfR Blank 0.000 09:03
Control Test Stats
Average 0.0510
Std Dev 0.0000
Rel Std Dev(%) 0.0000

Test g/210L Time
RfR Blank 0.000 09:05
Control Test 0.081 09:05
RfR Blank 0.000 09:06
Control Test 0.081 09:07
RfR Blank 0.000 09:07
Control Test 0.081 09:08
RfR Blank 0.000 09:08
Control Test Stats
Average 0.0810
Std Dev 0.0000
Rel Std Dev(%) 0.0000

Test g/210L Time
RfR Blank 0.000 09:10
Control Test 0.202 09:10
RfR Blank 0.000 09:11
Control Test 0.201 09:12
RfR Blank 0.000 09:12
Control Test 0.202 09:13
RfR Blank 0.000 09:13
Control Test Stats
Average 0.2017
Std Dev 0.0006
Rel Std Dev(%) 0.2863

Test g/210L Time
RfR Blank 0.000 09:15
Control Test 0.081 09:15
RfR Blank 0.000 09:16
Control Test 0.080 09:16
RfR Blank 0.000 09:17
Control Test 0.080 09:17
RfR Blank 0.000 09:17
Control Test Stats
Average 0.0803
Std Dev 0.0006
Rel Std Dev(%) 0.7187

[Signature]
Operator's Signature

[Signature]
Operator's Signature

[Signature]
Operator's Signature

[Signature]
Operator's Signature