

INSTRUMENT PROCESSING SHEET

Agency FHP Troop D S/N 80-003410
 Date In 8/29/16 Date Out 9/12/16 Ship P/U H/D CMI EE

Intake	Performed By <u>SP</u>	Quality Checks	Performed By <u>@</u>	Flow Calibration	Performed By															
<input type="checkbox"/> Registration <input checked="" type="checkbox"/> Annual <input checked="" type="checkbox"/> Return from CMI <input type="checkbox"/> Return from Enforcement Electronics <input type="checkbox"/> Other _____ Visual Inspection: <input checked="" type="checkbox"/> Case <input checked="" type="checkbox"/> Handle <input checked="" type="checkbox"/> Dry Gas Holder <input checked="" type="checkbox"/> Feet <input checked="" type="checkbox"/> Keyboard/Plug <input checked="" type="checkbox"/> Back/Plugs <input checked="" type="checkbox"/> Screws tight <input checked="" type="checkbox"/> Breath Hose Other Equipment: <input type="checkbox"/> Power cord <input checked="" type="checkbox"/> Printer Cable <input checked="" type="checkbox"/> Other <u>Static Bag</u>		<input checked="" type="checkbox"/> Breath Tube Screen <input checked="" type="checkbox"/> Replace O-Rings <input checked="" type="checkbox"/> Instrument Set Up Verified <input checked="" type="checkbox"/> R-Value <u>241</u> <input checked="" type="checkbox"/> Flow Verification (L/s) Flow Column # <u>079102</u> 32mm <u>154</u> (.139 - .169) 36mm <u>182</u> (.156 - .190) 53mm <u>266</u> (.228 - .278) 103mm <u>521</u> (.447 - .547) <input checked="" type="checkbox"/> Barometric Pressure Check Gauge ID # <u>28427</u> <input checked="" type="checkbox"/> Stability Checks		<input checked="" type="checkbox"/> Flow Calibration N/A <input type="checkbox"/> Flow Calibration Complete Flow Column # _____ <input type="checkbox"/> 5L/min - 7mm <input type="checkbox"/> 15L/min - 53mm <input type="checkbox"/> 30L/min - 103mm <input type="checkbox"/> R-Value _____ <input type="checkbox"/> Post Calibration Verification (L/s) Flow Column # _____ 32mm _____ (.139 - .169) 36mm _____ (.156 - .190) 53mm _____ (.228 - .278) 103mm _____ (.447 - .547)																
Notes: _____		<table border="1" style="width:100%; border-collapse: collapse;"> <thead> <tr> <th>Simulator</th> <th>Serial #</th> <th>Lot #/Exp</th> </tr> </thead> <tbody> <tr> <td>0.05</td> <td>SD1018</td> <td>201507A 7/14/17</td> </tr> <tr> <td>0.08</td> <td>SD1011</td> <td>201601F 1/26/18</td> </tr> <tr> <td>0.20</td> <td>SD1025</td> <td>201604C 4/5/18</td> </tr> <tr> <td>0.08 DGS</td> <td>N/A</td> <td>AG612405 5/3/18</td> </tr> </tbody> </table>		Simulator	Serial #	Lot #/Exp	0.05	SD1018	201507A 7/14/17	0.08	SD1011	201601F 1/26/18	0.20	SD1025	201604C 4/5/18	0.08 DGS	N/A	AG612405 5/3/18	Maintenance Performed By _____ <input type="checkbox"/> Battery Replacement <input type="checkbox"/> Dry Gas Regulator Replacement <input type="checkbox"/> Breath Tube Replacement <input type="checkbox"/> Other _____ Suggested Service _____ _____ _____	
Simulator	Serial #	Lot #/Exp																		
0.05	SD1018	201507A 7/14/17																		
0.08	SD1011	201601F 1/26/18																		
0.20	SD1025	201604C 4/5/18																		
0.08 DGS	N/A	AG612405 5/3/18																		

RECEIVED
 ROAD TO ZERO
 SEP 13 2016
 FDLE
 ALCOHOL TESTING PROGRAM

Optical Bench Calibration	Department Inspection																																																												
<input checked="" type="checkbox"/> Optical Bench Calibration N/A <input type="checkbox"/> Optical Bench Calibration Complete Barometric Pressure Gauge _____ ID # _____ <table border="1" style="width:100%; border-collapse: collapse;"> <thead> <tr> <th>Simulator</th> <th>Serial Number</th> <th>Lot Number</th> <th>Expiration</th> </tr> </thead> <tbody> <tr> <td>0.000</td> <td></td> <td>N/A</td> <td>N/A</td> </tr> <tr> <td>0.040</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.100</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.200</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.400</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.080 DGS</td> <td>N/A</td> <td></td> <td></td> </tr> </tbody> </table> <input type="checkbox"/> Post Calibration Stability Checks <table border="1" style="width:100%; border-collapse: collapse;"> <thead> <tr> <th>Simulator</th> <th>Serial Number</th> <th>Lot Number</th> <th>Expiration</th> </tr> </thead> <tbody> <tr> <td>0.05</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.08</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.20</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.08 DGS</td> <td>N/A</td> <td></td> <td></td> </tr> </tbody> </table>	Simulator	Serial Number	Lot Number	Expiration	0.000		N/A	N/A	0.040				0.100				0.200				0.400				0.080 DGS	N/A			Simulator	Serial Number	Lot Number	Expiration	0.05				0.08				0.20				0.08 DGS	N/A			<input checked="" type="checkbox"/> Barometric Pressure <u>1017</u> Gauge ID# <u>28427</u> <u>1014</u> Instrument Mouth Alcohol Solution Lot # <u>2016-A</u> Acetone Stock Solution Lot # <u>2016-B</u> <table border="1" style="width:100%; border-collapse: collapse;"> <thead> <tr> <th>Simulator</th> <th>Serial Number</th> </tr> </thead> <tbody> <tr> <td>0.00</td> <td>SD1019</td> </tr> <tr> <td>Interferent</td> <td>SD1021</td> </tr> <tr> <td>0.05</td> <td>SD1018</td> </tr> <tr> <td>0.08</td> <td>SD1011</td> </tr> <tr> <td>0.20</td> <td>SD1025</td> </tr> </tbody> </table>	Simulator	Serial Number	0.00	SD1019	Interferent	SD1021	0.05	SD1018	0.08	SD1011	0.20	SD1025
Simulator	Serial Number	Lot Number	Expiration																																																										
0.000		N/A	N/A																																																										
0.040																																																													
0.100																																																													
0.200																																																													
0.400																																																													
0.080 DGS	N/A																																																												
Simulator	Serial Number	Lot Number	Expiration																																																										
0.05																																																													
0.08																																																													
0.20																																																													
0.08 DGS	N/A																																																												
Simulator	Serial Number																																																												
0.00	SD1019																																																												
Interferent	SD1021																																																												
0.05	SD1018																																																												
0.08	SD1011																																																												
0.20	SD1025																																																												
Attachments <input checked="" type="checkbox"/> Form 41 <input type="checkbox"/> Optical Bench Cal <input checked="" type="checkbox"/> Pre-Stability Tests <input type="checkbox"/> Post-Stability Tests <input type="checkbox"/> Flow Calibration <input type="checkbox"/> Other _____																																																													

Notes: Time was incorrect. Changed, corrected, inspected. In Compliance.
QA/OC OK GDM 9/12/16

Instrument Complies with Chapter 11D-8, FAC
 Instrument Does Not Comply with Chapter 11D-8, FAC
 Return to/Place into Evidentiary Use
 Remain Out of Evidentiary Use
 Conduct an Agency Inspection Before Evidentiary Use

Brett Kiviland
 Quality Control Review

9/13/16
 Date

80-003410
 Stability Checks
 9/12/16
 R

INTOXILYZER 8000
 Instrument Initialization
 07:12 09/12/2016

FL HIGHWAY PATROL
 Intoxilyzer - Alcohol Analyzer
 Model 8000 SN 80-003410
 09/12/2016
 Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	08:03
Control Test	0.078	08:04
Air Blank	0.000	08:04
Control Test	0.079	08:05
Air Blank	0.000	08:05
Control Test	0.079	08:05
Air Blank	0.000	08:07
Control Test Stats		
Average	0.0787	
Std Dev	0.0006	
Rel Std Dev(%)	0.7339	

Operator's Signature

FL HIGHWAY PATROL
 Intoxilyzer - Alcohol Analyzer
 Model 8000 SN 80-003410
 09/12/2016
 Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	08:08
Control Test	0.049	08:08
Air Blank	0.000	08:09
Control Test	0.049	08:09
Air Blank	0.000	08:10
Control Test	0.049	08:11
Air Blank	0.000	08:11
Control Test Stats		
Average	0.0490	
Std Dev	0.0000	
Rel Std Dev(%)	0.0000	

Operator's Signature

FL HIGHWAY PATROL
 Intoxilyzer - Alcohol Analyzer
 Model 8000 SN 80-003410
 09/12/2016
 Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	08:12
Control Test	0.197	08:13
Air Blank	0.000	08:14
Control Test	0.197	08:14
Air Blank	0.000	08:15
Control Test	0.198	08:16
Air Blank	0.000	08:16
Control Test Stats		
Average	0.1973	
Std Dev	0.0006	
Rel Std Dev(%)	0.2926	

BK

Operator's Signature

FL HIGHWAY PATROL
 Intoxilyzer - Alcohol Analyzer
 Model 8000 SN 80-003410
 09/12/2016
 Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	07:53
Control Test	0.080	07:53
Air Blank	0.000	07:54
Control Test	0.079	07:54
Air Blank	0.000	07:55
Control Test	0.080	07:55
Air Blank	0.000	07:56
Control Test Stats		
Average	0.0797	
Std Dev	0.0006	
Rel Std Dev(%)	0.7247	

065

099M

Operator's Signature